

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name:	<u>Alliance</u>	Contract:	<u>DAYE01</u>
Lab Code:	<u>ACE</u>	SDG NO.:	<u>Q2816</u>
Lab File ID:	<u>VN087502.D</u>	Date Analyzed:	<u>08/12/2025</u>
Instrument ID:	<u>MSVOA_N</u>	Time Analyzed:	<u>10:24</u>
GC Column:	<u>RXI-624</u>	ID:	<u>0.25</u> (mm)
		Heated Purge:	(Y/N) <u>N</u>

	IS4 AREA #	RT #				
12 HOUR STD	291692	13.77				
UPPER LIMIT	583384	14.27				
LOWER LIMIT	145846	13.27				
EPA SAMPLE NO.						
1055-MW-01 (23)	246455	13.77				
1056-MW-02 (23.8)	220538	13.77				
1056-MW-02 (23.8)MS	243472	13.77				
1056-MW-02 (23.8)MSD	251240	13.77				
1057-MW-03A (17)	233578	13.77				
1058-MW-11 (15)	234425	13.77				
1059-MW-17A (15.5)	232946	13.77				
1061-TB080725	215113	13.77				
VN0812WBL01	242445	13.77				
VN0812WBS01	256652	13.77				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.