

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: DAYE01
 Lab Code: ACE SDG NO.: Q2816
 Lab File ID: VN087525.D Date Analyzed: 08/13/2025
 Instrument ID: MSVOA_N Time Analyzed: 10:57
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	261423	13.77				
UPPER LIMIT	522846	14.27				
LOWER LIMIT	130712	13.27				
EPA SAMPLE NO.						
1055-MW-01 (23) DL	235933	13.77				
1057-MW-03A (17) DL	224016	13.77				
1060-FB080725	210279	13.77				
VN0813WBL01	217582	13.77				
VN0813WBS01	246357	13.77				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.