

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Homeowner

SDG No.: Q2817

Contract: HOME01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Lead	3770	4000	94	90 - 110	P	08/13/2025	12:21	LB136801

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Homeowner

SDG No.: Q2817

Contract: HOME01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Lead	11.2	12.0	93	80 - 120	P	08/13/2025	12:28	LB136801

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Homeowner

SDG No.: Q2817

Contract: HOME01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Lead	4860	5000	97	90 - 110	P	08/13/2025	12:58	LB136801
CCV02	Lead	4850	5000	97	90 - 110	P	08/13/2025	13:42	LB136801
CCV03	Lead	4860	5000	97	90 - 110	P	08/13/2025	14:33	LB136801
CCV04	Lead	4860	5000	97	90 - 110	P	08/13/2025	15:23	LB136801
CCV05	Lead	4810	5000	96	90 - 110	P	08/13/2025	16:14	LB136801
CCV06	Lead	4850	5000	97	90 - 110	P	08/13/2025	17:05	LB136801
CCV07	Lead	4880	5000	98	90 - 110	P	08/13/2025	17:46	LB136801