



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Homeowner

SDG No.: Q2817

Contract: HOME01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Lead	2.30	+/-6	U	12.0	P	08/13/2025	12:33	LB136801

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Homeowner

SDG No.: Q2817

Contract: HOME01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Lead	2.30	+/-6	U	12.0	P	08/13/2025	13:03	LB136801
CCB02	Lead	2.30	+/-6	U	12.0	P	08/13/2025	13:46	LB136801
CCB03	Lead	2.30	+/-6	U	12.0	P	08/13/2025	14:37	LB136801
CCB04	Lead	2.30	+/-6	U	12.0	P	08/13/2025	15:27	LB136801
CCB05	Lead	2.30	+/-6	U	12.0	P	08/13/2025	16:19	LB136801
CCB06	Lead	2.30	+/-6	U	12.0	P	08/13/2025	17:09	LB136801
CCB07	Lead	2.30	+/-6	U	12.0	P	08/13/2025	17:50	LB136801