

Metals

- 4 -

INTERFERENCE CHECK SAMPLE

Client: Homeowner

SDG No.: Q2817

Contract: HOME01

Lab Code: ACE

ICS Source: EPA

Instrument ID: P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Lead	-6.38			-12	12	08/13/2025	12:41	LB136801
ICSAB01	Lead	42.2	49.0	86	37	61	08/13/2025	12:45	LB136801