

metals
- 5a -
MATRIX SPIKE SUMMARY

client: Homeowner **level:** low **sdg no.:** Q2817
contract: HOME01 **lab code:** ACE
matrix: Solid **sample id:** Q2828-01 **client id:** POWDERMS
Percent Solids for Sample: 99.4 **Spiked ID:** Q2828-01MS **Percent Solids for Spike Sample:** 99.4

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Lead	mg/Kg	81 - 112	110		71.4		41.1	95		P

metals
- 5a -
MATRIX SPIKE DUPLICATE SUMMARY

client: Homeowner **level:** low **sdg no.:** Q2817
contract: HOME01 **lab code:** ACE
matrix: Solid **sample id:** Q2828-01 **client id:** POWDERMSD
Percent Solids for Sample: 99.4 **Spiked ID:** Q2828-01MSD **Percent Solids for Spike Sample:** 99.4

Analyte	Units	Acceptance Limit %R	MSD Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Lead	mg/Kg	81 - 112	116		71.4		43.7	102		P