

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Homeowner

SDG No.: Q2817

Contract: HOME01

Lab Code: ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB169211BS Lead	mg/Kg	50.0	47.4		95	81 - 112	P