

Initial and Continuing Calibration Verification

Client: Earth Engineering Inc.

SDG No.: Q2818

Project: Reserve Turgyan Farms

RunNo.: LB136777

Analyte		Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: Cyanide	ICV1	mg/L	0.096	0.099	97	90-110	08/11/2025
Sample ID: Cyanide	CCV1	mg/L	0.25	0.25	100	90-110	08/11/2025
Sample ID: Cyanide	CCV2	mg/L	0.25	0.25	100	90-110	08/11/2025

Initial and Continuing Calibration Verification

Client: Earth Engineering Inc.

SDG No.: Q2818

Project: Reserve Turgyan Farms

RunNo.: LB136872

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV Hexavalent Chromium	mg/L	0.499	0.5	100	90-110	08/19/2025
Sample ID: CCV1 Hexavalent Chromium	mg/L	0.496	0.5	99	90-110	08/19/2025
Sample ID: CCV2 Hexavalent Chromium	mg/L	0.487	0.5	97	90-110	08/19/2025
Sample ID: CCV3 Hexavalent Chromium	mg/L	0.493	0.5	99	90-110	08/19/2025