

Initial and Continuing Calibration Blank Summary

Client: Earth Engineering Inc.

SDG No.: Q2818

Project: Reserve Turgyan Farms

RunNo.: LB136777

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: ICB1 Cyanide	mg/L	0.0014	0.0025	J	0.00096	0.005	08/11/2025
Sample ID: CCB1 Cyanide	mg/L	0.0016	0.0025	J	0.00096	0.005	08/11/2025
Sample ID: CCB2 Cyanide	mg/L	0.0014	0.0025	J	0.00096	0.005	08/11/2025

Initial and Continuing Calibration Blank Summary

Client: Earth Engineering Inc.
Project: Reserve Turgyan Farms

SDG No.: Q2818
RunNo.: LB136872

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: ICB Hexavalent Chromium	mg/L	< 0.0050	0.0050	U	0.0029	0.01	08/19/2025
Sample ID: CCB1 Hexavalent Chromium	mg/L	< 0.0050	0.0050	U	0.0029	0.01	08/19/2025
Sample ID: CCB2 Hexavalent Chromium	mg/L	< 0.0050	0.0050	U	0.0029	0.01	08/19/2025
Sample ID: CCB3 Hexavalent Chromium	mg/L	< 0.0050	0.0050	U	0.0029	0.01	08/19/2025