


LB136879

7470A

INSTRUMENT ID: CV1



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	g/n
0.0	0.000	0.012	0.012	1220	0.000	1220	0				-
0.2	0.200	0.212	0.012	2127	0.0 %	2127					6
2.5	2.500	2.508	0.008	12528	0.0 %	12528	0				0
5.0	5.000	4.780	-0.220	22817	0.0 %	22817					-4
7.5	7.500	7.940	0.440	37134	0.0 %	37134					6
10.0	10.000	9.747	-0.253	45316	0.0 %	45316					-3

LB136879 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 19 Aug 2025 13:43:23

Type	Sample ID	Extended ID	Conc	µ Abs	Units	Type	Std Conc	Method	Date
S	0.0 - 1	50	-	1220	PPB	Std	0.0000	7470A	19 Aug 2025 13:52:47
S	0.2 - 1	50.2	-	2127	PPB	Std	0.2000	7470A	19 Aug 2025 13:55:05
S	2.5 - 1	52.5	-	12528	PPB	Std	2.5000	7470A	19 Aug 2025 13:57:22
S	5.0 - 1	55	-	22817	PPB	Std	5.0000	7470A	19 Aug 2025 13:59:39
S	7.5 - 1	57.5	-	37134	PPB	Std	7.5000	7470A	19 Aug 2025 14:01:56
S	10.0 - 1	57.5	-	45316	PPB	Std	10.0000	7470A	19 Aug 2025 14:04:13
U	ICV23 - 1	ICV23	4.1057	19763	PPB	SMPL	-7470A	-7470A	19 Aug 2025 14:07:25
U	ICB23 - 1	ICB23	-0.0238	1057	PPB	SMPL	-7470A	-7470A	19 Aug 2025 14:09:40
U	CCV71 - 1	CCV71	4.7092	22497	PPB	SMPL	-7470A	-7470A	19 Aug 2025 14:11:56
U	CCB71 - 1	CCB71	-0.0238	1057	PPB	SMPL	-7470A	-7470A	19 Aug 2025 14:14:12
U	CRA - 1	CRA	0.2053	2095	PPB	SMPL	-7470A	-7470A	19 Aug 2025 14:16:28
U	HighStd - 1	HighStd	9.7491	45327	PPB	SMPL	-7470A	-7470A	19 Aug 2025 14:18:44
U	ChkStd - 1	ChkStd	6.5837	30988	PPB	SMPL	-7470A	-7470A	19 Aug 2025 14:20:59
U	PB169293BL - 1	PBW	-0.0055	1140	PPB	SMPL	-7470A	-7470A	19 Aug 2025 14:25:54
U	PB169293BS - 1	LCSW	4.1500	19964	PPB	SMPL	-7470A	-7470A	19 Aug 2025 14:31:42
U	Q2818-04 - 1	B-2-5-1	0.5250	3543	PPB	SMPL	-7470A	-7470A	19 Aug 2025 14:33:57
U	Q2818-05 - 1	B-3-5-2	0.7603	4609	PPB	SMPL	-7470A	-7470A	19 Aug 2025 14:36:14
U	Q2879-02 - 1	OU4-TS-GRILLO-TSCP03-081425	-0.0878	767	PPB	SMPL	-7470A	-7470A	19 Aug 2025 14:38:36
U	Q2879-04 - 1	OU4-TS-GRILLO-TSCP04-081425	-0.1187	627	PPB	SMPL	-7470A	-7470A	19 Aug 2025 14:40:55
U	Q2879-06 - 1	OU4-TS-GRILLO-TSCP05-081425	-0.1245	601	PPB	SMPL	-7470A	-7470A	19 Aug 2025 14:43:12
U	CCV72 - 1	CCV72	4.7033	22470	PPB	SMPL	-7470A	-7470A	19 Aug 2025 14:45:30
U	CCB72 - 1	CCB72	-0.0333	1014	PPB	SMPL	-7470A	-7470A	19 Aug 2025 14:47:47
U	Q2879-08 - 1	OU4-TS-GRILLO-TSCP06-081425	0.1994	2068	PPB	SMPL	-7470A	-7470A	19 Aug 2025 14:50:12
U	Q2879-10 - 1	OU4-TS-GRILLO-TSCP07-081425	0.2016	2078	PPB	SMPL	-7470A	-7470A	19 Aug 2025 14:52:27
U	Q2879-12 - 1	OU4-TS-GRILLO-TSCP08-081425	0.4071	3009	PPB	SMPL	-7470A	-7470A	19 Aug 2025 14:54:44
U	Q2879-14 - 1	OU4-TS-GRILLO-TSCP09-081425	0.2636	2358	PPB	SMPL	-7470A	-7470A	19 Aug 2025 14:57:00
U	Q2879-16 - 1	OU4-TS-GRILLO-TSCP10-081425	0.2115	2123	PPB	SMPL	-7470A	-7470A	19 Aug 2025 14:59:18
U	Q2879-18 - 1	OU4-TS-GRILLO-TSCP11-081425	-0.1018	704	PPB	SMPL	-7470A	-7470A	19 Aug 2025 15:01:36
U	Q2879-18DUP - 1	OU4-TS-GRILLO-TSCP11-081425DUP	-0.0759	821	PPB	SMPL	-7470A	-7470A	19 Aug 2025 15:03:54
U	Q2879-18MS - 1	OU4-TS-GRILLO-TSCP11-081425MS	4.0253	19399	PPB	SMPL	-7470A	-7470A	19 Aug 2025 15:09:51
U	Q2879-18MSD - 1	OU4-TS-GRILLO-TSCP11-081425MSD	3.9995	19282	PPB	SMPL	-7470A	-7470A	19 Aug 2025 15:12:08
U	CCV73 - 1	CCV73	4.7501	22682	PPB	SMPL	-7470A	-7470A	19 Aug 2025 15:14:24
U	CCB73 - 1	CCB73	-0.0280	1038	PPB	SMPL	-7470A	-7470A	19 Aug 2025 15:16:40
U	PB169293TB - 1	PB169276TB	-0.1132	652	PPB	SMPL	-7470A	-7470A	19 Aug 2025 15:22:24
U	PB169293TB - 1	PB169294TB	-0.1123	656	PPB	SMPL	-7470A	-7470A	19 Aug 2025 15:26:46
U	Q2879-18LX5 - 1		-0.2154	189	PPB	SMPL	-7470A	-7470A	19 Aug 2025 15:29:01
U	Q2879-18A - 1		4.0260	19402	PPB	SMPL	-7470A	-7470A	19 Aug 2025 15:31:17
U	CCV74 - 1	CCV74	5.2370	24888	PPB	SMPL	-7470A	-7470A	19 Aug 2025 15:37:18
U	CCB73 - 1	CCB73	-0.0269	1043	PPB	SMPL	-7470A	-7470A	19 Aug 2025 15:39:33