



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: Earth Engineering Inc.

SDG No.: Q2818

Contract: EARTH03

Lab Code: ACE

Initial Calibration Source: _____

Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRA	Mercury	0.21	0.2	103	70 - 130	CV	08/19/2025	14:16	LB136879
CRI01	Beryllium	6.30	6.0	105	65 - 135	P	08/20/2025	13:26	LB136898
	Silver	9.95	10.0	100	65 - 135	P	08/20/2025	13:26	LB136898