

Metals

- 4 -

INTERFERENCE CHECK SAMPLE

Client: Earth Engineering Inc.

SDG No.: Q2818

Contract: EARTH03

Lab Code: ACE

ICS Source: EPA

Instrument ID: P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Beryllium	1.50			-6	6	08/20/2025	13:30	LB136898
	Silver	2.09			-10	10	08/20/2025	13:30	LB136898
ICSAB01	Beryllium	482	495	97	420	570	08/20/2025	13:34	LB136898
	Silver	212	201	106	170	232	08/20/2025	13:34	LB136898