

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: EARTH03

Lab Code: ACE

SDG NO.: Q2818

Continuing Calib Date: 08/12/2025

Initial Calibration Date(s): 08/01/2025

08/01/2025

Continuing Calib Time: 10:27

Initial Calibration Time(s): 12:05

20:28

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1254-1	(1)	6.66	6.66	6.56	6.76	0.00
Aroclor-1254-2	(2)	6.87	6.87	6.77	6.97	0.00
Aroclor-1254-3	(3)	7.23	7.24	7.14	7.34	0.01
Aroclor-1254-4	(4)	7.52	7.52	7.42	7.62	0.01
Aroclor-1254-5	(5)	7.93	7.93	7.83	8.03	0.00
Tetrachloro-m-xylene		4.66	4.66	4.56	4.76	0.00
Decachlorobiphenyl		10.43	10.43	10.33	10.53	0.00

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: EARTH03

Lab Code: ACE

SDG NO.: Q2818

Continuing Calib Date: 08/12/2025

Initial Calibration Date(s): 08/01/2025

08/01/2025

Continuing Calib Time: 10:27

Initial Calibration Time(s): 12:05

20:28

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1254-1 (1)	5.68	5.69	5.59	5.79	0.01
Aroclor-1254-2 (2)	5.83	5.83	5.73	5.93	0.00
Aroclor-1254-3 (3)	6.23	6.24	6.14	6.34	0.01
Aroclor-1254-4 (4)	6.46	6.46	6.36	6.56	0.00
Aroclor-1254-5 (5)	6.88	6.88	6.78	6.98	0.00
Tetrachloro-m-xylene	3.80	3.80	3.70	3.90	0.00
Decachlorobiphenyl	8.82	8.82	8.72	8.92	0.00