

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** EARTH03
Lab Code: ACE **SDG NO.:** Q2818
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/23/2025 07/23/2025

Client Sample No.: CCAL02 **Date Analyzed:** 08/12/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO112865.D **Time Analyzed:** 15:50

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.751	4.655	4.855	524.640	500.000	4.9
Aroclor-1016-2	4.769	4.674	4.874	553.430	500.000	10.7
Aroclor-1016-3	4.827	4.731	4.931	551.220	500.000	10.2
Aroclor-1016-4	4.947	4.851	5.051	550.720	500.000	10.1
Aroclor-1016-5	5.203	5.108	5.308	584.150	500.000	16.8
Aroclor-1260-1	6.240	6.145	6.345	550.350	500.000	10.1
Aroclor-1260-2	6.430	6.333	6.533	543.410	500.000	8.7
Aroclor-1260-3	6.795	6.700	6.900	570.950	500.000	14.2
Aroclor-1260-4	7.055	6.959	7.159	542.260	500.000	8.5
Aroclor-1260-5	7.297	7.203	7.403	529.970	500.000	6.0
Decachlorobiphenyl	8.687	8.593	8.793	48.970	50.000	-2.1
Tetrachloro-m-xylene	3.667	3.569	3.769	57.760	50.000	15.5

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** EARTH03
Lab Code: ACE **SDG NO.:** Q2818
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/23/2025 07/23/2025

Client Sample No.: CCAL02 **Date Analyzed:** 08/12/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO112865.D **Time Analyzed:** 15:50

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.734	4.638	4.838	522.080	500.000	4.4
Aroclor-1016-2	4.751	4.656	4.856	517.090	500.000	3.4
Aroclor-1016-3	4.926	4.831	5.031	516.260	500.000	3.3
Aroclor-1016-4	4.969	4.873	5.073	515.840	500.000	3.2
Aroclor-1016-5	5.181	5.085	5.285	527.110	500.000	5.4
Aroclor-1260-1	6.208	6.113	6.313	523.960	500.000	4.8
Aroclor-1260-2	6.396	6.302	6.502	531.850	500.000	6.4
Aroclor-1260-3	6.547	6.453	6.653	550.200	500.000	10.0
Aroclor-1260-4	7.017	6.923	7.123	543.780	500.000	8.8
Aroclor-1260-5	7.260	7.164	7.364	560.820	500.000	12.2
Decachlorobiphenyl	8.633	8.540	8.740	53.180	50.000	6.4
Tetrachloro-m-xylene	3.659	3.562	3.762	55.270	50.000	10.5