

Cover Page

Order ID : Q2818

Project ID : Reserve Turgyan Farms

Client : Earth Engineering Inc.

Lab Sample Number

Q2818-01
Q2818-02

Client Sample Number

B-2-5-1
B-3-5-2

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. Release of the data contained in this hard copy data package has been authorized by the laboratory manager or his designee, as verified by the following signature.

Signature : _____

Date: 8/14/2025

NYDOH CERTIFICATION NO - 11376

NJDEP CERTIFICATION NO - 20012

CASE NARRATIVE

Earth Engineering Inc.

Project Name: Reserve Turgyan Farms

Project # N/A

Order ID # Q2818

Test Name: PCB

A. Number of Samples and Date of Receipt:

B. Parameters

According to the Chain of Custody document, the following analyses were requested: PCB. This data package contains results for PCB.

C. Analytical Techniques:

The analyses were performed on instrument GCECD_P. The front column is ZB-MR1 which is 30 meters, 0.32 mm ID, 0.5 um df, Catalogue # 7HM-G016-17. The rear column is ZB-MR2 which is 30 meters, 0.32 mm ID, 0.25 µm; Catalogue # 7HM-G017-11. The analyses were performed on instrument GCECD_O. The front column is ZB-MR1 which is 30 meters, 0.32 mm ID, 0.5 um df, Catalogue # 7HM-G016-17. The rear column is ZB-MR2 which is 30 meters, 0.32 mm ID, 0.25 µm; Catalogue # 7HM-G017-11. The analysis of PCBs was based on method 8082A and extraction was done based on method 3541.

D. QA/ QC Samples:

The Holding Times were met for all analysis.

The Surrogate recoveries were met for all analysis.

The Retention Times were met for all analysis.

The MS recoveries met the requirements for all compounds.

The MSD recoveries met the requirements for all compounds.

The RPD were met for all analysis.

The Blank Spike met requirements for all compounds.

The Blank analysis did not indicate the presence of lab contamination.

The Initial Calibration met the requirements.

The Continuous Calibration File ID PO112850.D met the requirements except for Aroclor-1016(Peak-01),Tetrachloro-m-xylene is failing in 1st column however 2nd column passing for that therefore no corrective action taken.

The Continuous Calibration File ID PP074302.D met the requirements except for Aroclor-1016(Peak-02),Tetrachloro-m-xylene is failing in 2nd column however 1st column passing for that therefore no corrective action taken.



284 Sheffield Street, Mountainside, NJ 07092
Phone: 908 789 8900 Fax: 908 789 8922

E. Additional Comments:

The temperature of the samples at the time of receipt was 6.1°C.

The soil samples results are based on a dry weight basis.

F. Manual Integration Comments:

Please refer to the Manual integration Report included with the Run Logs for information on the manual integrations performed.

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature_____

DATA REPORTING QUALIFIERS- ORGANIC

For reporting results, the following “ Results Qualifiers” are used:

Value	If the result is a value greater than or equal to the detection limit, report the value
U	Indicates the compound was analyzed for but was not detected. Report the minimum detection limit for the sample with the U, i.e. “10 U”. This is not necessarily the instrument detection limit attainable for this particular sample based on any concentration or dilution that may have been required.
ND	Indicates the analyte was analyzed for, but not detected
J	Indicates an estimated value. This flag is used: (1) When estimating a concentration for a tentatively identified compound (library search hits, where a 1:1 response is assumed.) (2) When the mass spectral data indicated the identification, however the result was less than the specified detection limit greater than zero. If the detection limit was 10ug/L and a concentration of 3 ug/L was calculated report as 3 J. This flag is used when similar situation arise on any organic parameter i.e. Pest, PCB and others.
B	Indicates the analyte was found in the blank as well as the sample report as “12 B”.
E	Indicates the analyte ‘s concentration exceeds the calibrated range of the instrument for that specific analysis.
D	This flag identifies all compounds identified in an analysis at a secondary dilution factor.
P	This flag is used for Pesticide/PCB target analyte when there is >25% difference for detected concentrations between the two GC columns. The lower of the two values is reported on Form 1 and flagged with a “P”.
N	This flag indicates presumptive evidence of a compound. This is only used for tentatively identified compounds (TICs), where the identification is based on a mass spectral library search. It applies to all TIC results. For generic characterization of a TIC, such as chlorinated hydrocarbon, the flag is not used.
A	This flag indicates that a Tentatively Identified Compound is a suspected aldol-condensation product.
Q	Indicates the LCS did not meet the control limits requirements

APPENDIX A

QA REVIEW GENERAL DOCUMENTATION

Project #: Q2818

Completed

For thorough review, the report must have the following:

GENERAL:

Are all original paperwork present (chain of custody, record of communication,airbill, sample management lab chronicle, login page)

✓

Check chain-of-custody for proper relinquish/return of samples

✓

Is the chain of custody signed and complete

✓

Check internal chain-of-custody for proper relinquish/return of samples /sample extracts

✓

Collect information for each project id from server. Were all requirements followed

✓

COVER PAGE:

Do numbers of samples correspond to the number of samples in the Chain of Custody on login page

✓

Do lab numbers and client Ids on cover page agree with the Chain of Custody

✓

CHAIN OF CUSTODY:

Do requested analyses on Chain of Custody agree with form I results

✓

Do requested analyses on Chain of Custody agree with the log-in page

✓

Were the correct method log-in for analysis according to the Analytical Request and Chain of Custody

✓

Were the samples received within hold time

✓

Were any problems found with the samples at arrival recorded in the Sample Management Laboratory Chronicle

✓

ANALYTICAL:

Was method requirement followed?

✓

Was client requirement followed?

✓

Does the case narrative summarize all QC failure?

✓

All runlogs and manual integration are reviewed for requirements

✓

All manual calculations and /or hand notations verified

✓

QA Review Signature: MAHESH PATEL

Date: 08/14/2025

LAB CHRONICLE

OrderID:	Q2818	OrderDate:	8/11/2025 11:57:00 AM
Client:	Earth Engineering Inc.	Project:	Reserve Turgyan Farms
Contact:	Frank Dougherty, LSRP	Location:	J21

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q2818-01	B-2-5-1	SOIL	PCB	8082A	08/11/25	08/12/25	08/12/25	08/11/25
Q2818-02	B-3-5-2	SOIL	PCB	8082A	08/11/25	08/12/25	08/12/25	08/11/25



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Hit Summary Sheet
SW-846

SDG No.: Q2818

Order ID: Q2818

Client: Earth Engineering Inc.

Project ID: Reserve Turgyan Farms

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
-----------	-----------	--------	-----------	---------------	---	-----	-----	-------

Client ID :

Total Concentration: 0.000



QC SUMMARY

Surrogate Summary

SDG No.: **Q2818**

Client: **Earth Engineering Inc.**

Analytical Method: **8082A**

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Recovery(%)	Qual	Limits(%)	
								Low	High
I.BLK-PO112413.D	PIBLK-PO112413.D	Tetrachloro-m-xyl	1	20	19.1	95		60	140
		Decachlorobiphen	1	20	20.0	100		60	140
		Tetrachloro-m-xyl	2	20	19.5	98		60	140
		Decachlorobiphen	2	20	20.4	102		60	140
I.BLK-PO112854.D	PIBLK-PO112854.D	Tetrachloro-m-xyl	1	20	20.9	105		60	140
		Decachlorobiphen	1	20	17.2	86		60	140
		Tetrachloro-m-xyl	2	20	20.5	102		60	140
		Decachlorobiphen	2	20	20.4	102		60	140
Q2818-01	B-2-5-1	Tetrachloro-m-xyl	1	20	23.7	118		32	144
		Decachlorobiphen	1	20	19.9	100		32	175
		Tetrachloro-m-xyl	2	20	23.4	117		32	144
		Decachlorobiphen	2	20	23.0	115		32	175
Q2818-02	B-3-5-2	Tetrachloro-m-xyl	1	20	21.1	106		32	144
		Decachlorobiphen	1	20	19.6	98		32	175
		Tetrachloro-m-xyl	2	20	21.2	106		32	144
		Decachlorobiphen	2	20	21.5	107		32	175
I.BLK-PO112869.D	PIBLK-PO112869.D	Tetrachloro-m-xyl	1	20	20.7	103		60	140
		Decachlorobiphen	1	20	19.2	96		60	140
		Tetrachloro-m-xyl	2	20	20.3	102		60	140
		Decachlorobiphen	2	20	20.5	103		60	140
I.BLK-PP074167.D	PIBLK-PP074167.D	Tetrachloro-m-xyl	1	20	19.1	95		60	140
		Decachlorobiphen	1	20	19.7	98		60	140
		Tetrachloro-m-xyl	2	20	17.3	86		60	140
		Decachlorobiphen	2	20	18.7	93		60	140
I.BLK-PP074305.D	PIBLK-PP074305.D	Tetrachloro-m-xyl	1	20	22.0	110		60	140
		Decachlorobiphen	1	20	21.1	105		60	140
		Tetrachloro-m-xyl	2	20	19.7	98		60	140
		Decachlorobiphen	2	20	18.5	93		60	140
PB169205BL	PB169205BL	Tetrachloro-m-xyl	1	20	23.1	116		32	144
		Decachlorobiphen	1	20	22.2	111		32	175
		Tetrachloro-m-xyl	2	20	20.1	101		32	144
		Decachlorobiphen	2	20	20.1	100		32	175
PB169205BS	PB169205BS	Tetrachloro-m-xyl	1	20	24.3	121		32	144
		Decachlorobiphen	1	20	22.5	112		32	175
		Tetrachloro-m-xyl	2	20	20.8	104		32	144
		Decachlorobiphen	2	20	20.0	100		32	175
Q2830-01MS	BIN0009-DRIVEWAY-TP-SOUT	Tetrachloro-m-xyl	1	20	23.5	117		32	144
		Decachlorobiphen	1	20	19.9	100		32	175
		Tetrachloro-m-xyl	2	20	22.0	110		32	144
		Decachlorobiphen	2	20	17.5	87		32	175
Q2830-01MSD	BIN0009-DRIVEWAY-TP-SOUT	Tetrachloro-m-xyl	1	20	23.3	116		32	144

Surrogate Summary

SDG No.: Q2818

Client: Earth Engineering Inc.

Analytical Method: 8082A

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Recovery(%)	Qual	Limits(%)	
								Low	High
Q2830-01MSD	BIN0009-DRIVEWAY-TP-SOUT	Decachlorobiphen	1	20	19.9	99		32	175
		Tetrachloro-m-xyl	2	20	22.1	111		32	144
		Decachlorobiphen	2	20	17.9	89		32	175
I.BLK-PP074319.D	PIBLK-PP074319.D	Tetrachloro-m-xyl	1	20	22.9	115		60	140
		Decachlorobiphen	1	20	21.7	109		60	140
		Tetrachloro-m-xyl	2	20	19.8	99		60	140
		Decachlorobiphen	2	20	19.8	99		60	140

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.:	<u>Q2818</u>	Analytical Method:	<u>8082A</u>
Client:	<u>Earth Engineering Inc.</u>	DataFile :	<u>PP074311.D</u>

		Sample				Rec		RPD		Limits	
	Parameter	Spike	Result	Result	Units	Rec	Qual	RPD	Qual	Low	High
Lab Sample ID:	Q2830-01MS	Client Sample ID:		BIN0009-DRIVEWAY-TP-							
	(Column 1)										
	AR1016	191.7	0	201	ug/kg	105				55	146
	AR1260	191.7	0	173	ug/kg	90				54	119
Lab Sample ID:	Q2830-01MS	Client Sample ID:		BIN0009-DRIVEWAY-TP-							
	(Column 2)										
	AR1016	191.7	0	213	ug/kg	111				55	146
	AR1260	191.7	0	181	ug/kg	94				54	119

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.:	<u>Q2818</u>	Analytical Method:	<u>8082A</u>
Client:	<u>Earth Engineering Inc.</u>	DataFile :	<u>PP074312.D</u>

		Sample				Rec		RPD		Limits		
	Parameter	Spike	Result	Result	Units	Rec	Qual	RPD	Qual	Low	High	RPD
Lab Sample ID:	Q2830-01MSD	Client Sample ID:		BIN0009-DRIVEWAY-TP-								
	(Column 1)											
	AR1016	191.6	0	213	ug/kg	111		6		55	146	15
	AR1260	191.6	0	177	ug/kg	92		2		54	119	15
Lab Sample ID:	Q2830-01MSD	Client Sample ID:		BIN0009-DRIVEWAY-TP-								
	(Column 2)											
	AR1016	191.6	0	219	ug/kg	114		3		55	146	15
	AR1260	191.6	0	188	ug/kg	98		4		54	119	15



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q2818</u>	Analytical Method:	<u>8082A</u>
Client:	<u>Earth Engineering Inc.</u>	Datafile :	<u>PP074309.D</u>

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD	Low	Limits	RPD
								Qual		High	
PB169205BS (Column 1)	AR1016	166.6	185	ug/kg	111				71	120	
	AR1260	166.6	165	ug/kg	99				65	130	
PB169205BS (Column 2)	AR1016	166.6	183	ug/kg	110				71	120	
	AR1260	166.6	180	ug/kg	108				65	130	

4C

PESTICIDE METHOD BLANK SUMMARY

Client ID

PB169205BL

Lab Name: Alliance

Contract: EARTH03

Lab Code: ACE

SDG NO.: Q2818

Lab Sample ID: PB169205BL

Lab File ID: PP074308.D

Matrix: (soil/water) Solid

Extraction: (Type) SOXH

Sulfur Cleanup: (Y/N) N

Date Extracted: 08/12/2025

Date Analyzed (1): 08/12/2025

Date Analyzed (2): 08/12/2025

Time Analyzed (1): 12:48

Time Analyzed (2): 12:48

Instrument ID (1): ECD_P

Instrument ID (2): ECD_P

GC Column (1): ZB-MR1 ID: 0.32 (mm)

GC Column (2): ZB-MR2 ID: 0.32 (mm)

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED 1	DATE ANALYZED 2
B-2-5-1	Q2818-01	PO112858.D	08/12/2025	08/12/2025
B-3-5-2	Q2818-02	PO112859.D	08/12/2025	08/12/2025
PB169205BS	PB169205BS	PP074309.D	08/12/2025	08/12/2025
BIN0009-DRIVEWAY-TP-SOUTH-EASTMS	Q2830-01MS	PP074311.D	08/12/2025	08/12/2025
BIN0009-DRIVEWAY-TP-SOUTH-EASTMSD	Q2830-01MSD	PP074312.D	08/12/2025	08/12/2025

COMMENTS: _____



SAMPLE DATA

Report of Analysis

Client:	Earth Engineering Inc.	Date Collected:	08/11/25
Project:	Reserve Turgyan Farms	Date Received:	08/11/25
Client Sample ID:	B-2-5-1	SDG No.:	Q2818
Lab Sample ID:	Q2818-01	Matrix:	SOIL
Analytical Method:	8082A	% Solid:	87.5
Sample Wt/Vol:	30.07	Units:	g
Soil Aliquot Vol:			uL
Extraction Type:		Test:	PCB
GPC Factor :	1.0	PH :	
Prep Method :	SW3541B	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO112858.D	1	08/12/25 08:15	08/12/25 12:50	PB169205

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	4.50	U	4.50	19.4	ug/kg
11104-28-2	Aroclor-1221	4.60	U	4.60	19.4	ug/kg
11141-16-5	Aroclor-1232	4.20	U	4.20	19.4	ug/kg
53469-21-9	Aroclor-1242	4.60	U	4.60	19.4	ug/kg
12672-29-6	Aroclor-1248	6.70	U	6.70	19.4	ug/kg
11097-69-1	Aroclor-1254	3.70	U	3.70	19.4	ug/kg
37324-23-5	Aroclor-1262	5.70	U	5.70	19.4	ug/kg
11100-14-4	Aroclor-1268	4.10	U	4.10	19.4	ug/kg
11096-82-5	Aroclor-1260	3.70	U	3.70	19.4	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	23.7		32 - 144	118%	SPK: 20
2051-24-3	Decachlorobiphenyl	23.0		32 - 175	115%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0081225\
 Data File : PO112858.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 12 Aug 2025 12:50
 Operator : YP/AJ
 Sample : Q2818-01
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 B-2-5-1

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/13/2025
 Supervised By :mohammad ahmed 08/14/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 12 13:27:39 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0072325.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Jul 24 04:54:06 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.666	3.656	192.6E6	116.5E6	23.691	23.451
2) SA Decachlor...	8.686	8.630	145.7E6	40707830	19.932m	23.036

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO081225\
Data File : PO112858.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Aug 2025 12:50
Operator : YP/AJ
Sample : Q2818-01
Misc :
ALS Vial : 10 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
B-2-5-1

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/13/2025
Supervised By :mohammad ahmed 08/14/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 12 13:27:39 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO072325.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Jul 24 04:54:06 2025
Response via : Initial Calibration
Integrator: ChemStation


Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Response_

Signal: PO112858.D\ECD1A.ch

#1 Tetrachloro-m-xylene



R.T.: 3.666 min
 Delta R.T.: -0.003 min
 Response: 192644050
 Conc: 23.69 ng/ml

Instrument :
 ECD_O
 ClientSampleId :
 B-2-5-1

Manual Integrations
 APPROVED


Reviewed By :Yogesh Patel 08/13/2025
 Supervised By :mohammad ahmed 08/14/2025

Time

Response_

Signal: PO112858.D\ECD2B.ch

#1 Tetrachloro-m-xylene




R.T.: 3.656 min
 Delta R.T.: -0.007 min
 Response: 116531826
 Conc: 23.45 ng/ml

Time

Response_

Signal: PO112858.D\ECD1A.ch

#2 Decachlorobiphenyl




R.T.: 8.686 min
 Delta R.T.: -0.007 min
 Response: 145735745
 Conc: 19.93 ng/ml m

Time

Response_

Signal: PO112858.D\ECD2B.ch

#2 Decachlorobiphenyl



R.T.: 8.630 min
 Delta R.T.: -0.010 min
 Response: 40707830
 Conc: 23.04 ng/ml

Time

Report of Analysis

Client:	Earth Engineering Inc.	Date Collected:	08/11/25
Project:	Reserve Turgyan Farms	Date Received:	08/11/25
Client Sample ID:	B-3-5-2	SDG No.:	Q2818
Lab Sample ID:	Q2818-02	Matrix:	SOIL
Analytical Method:	8082A	% Solid:	89.1
Sample Wt/Vol:	30.03	Units:	g
Soil Aliquot Vol:			uL
Extraction Type:		Test:	PCB
GPC Factor :	1.0	PH :	
Prep Method :	SW3541B	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO112859.D	1	08/12/25 08:15	08/12/25 13:08	PB169205

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	4.40	U	4.40	19.1	ug/kg
11104-28-2	Aroclor-1221	4.50	U	4.50	19.1	ug/kg
11141-16-5	Aroclor-1232	4.20	U	4.20	19.1	ug/kg
53469-21-9	Aroclor-1242	4.50	U	4.50	19.1	ug/kg
12672-29-6	Aroclor-1248	6.60	U	6.60	19.1	ug/kg
11097-69-1	Aroclor-1254	3.60	U	3.60	19.1	ug/kg
37324-23-5	Aroclor-1262	5.60	U	5.60	19.1	ug/kg
11100-14-4	Aroclor-1268	4.00	U	4.00	19.1	ug/kg
11096-82-5	Aroclor-1260	3.60	U	3.60	19.1	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	21.2		32 - 144	106%	SPK: 20
2051-24-3	Decachlorobiphenyl	21.5		32 - 175	107%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0081225\
Data File : P0112859.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Aug 2025 13:08
Operator : YP/AJ
Sample : Q2818-02
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
B-3-5-2

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 12 13:28:15 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0072325.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Jul 24 04:54:06 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.666	3.657	171.9E6	105.4E6	21.134	21.211
2) SA Decachlor...	8.687	8.632	143.3E6	37976153	19.601	21.491

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO081225\
Data File : PO112859.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Aug 2025 13:08
Operator : YP/AJ
Sample : Q2818-02
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
B-3-5-2

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 12 13:28:15 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO072325.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Jul 24 04:54:06 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.666 min
Delta R.T.: -0.003 min
Response: 171854572
Conc: 21.13 ng/ml

Instrument :
ECD_O
ClientSampleId :
B-3-5-2




#1 Tetrachloro-m-xylene

R.T.: 3.657 min
Delta R.T.: -0.005 min
Response: 105400843
Conc: 21.21 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.687 min
Delta R.T.: -0.006 min
Response: 143315526
Conc: 19.60 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.632 min
Delta R.T.: -0.009 min
Response: 37976153
Conc: 21.49 ng/ml



CALIBRATION SUMMARY

Calibration Times: 11:32 19:47

LAB FILE ID: RT 1000 = PO112414.D RT 750 = PO112415.D
RT 500 = PO112416.D RT 250 = PO112417.D RT 050 = PO112418.D

[illegible]

[illegible]

Contract: EARTH03

SDG NO.: Q2818

Calibration Date(s): 07/23/2025 07/23/2025

Calibration Times: 11:32 19:47

GC Column: ZB-MR2 ID: 0.32 (mm)

LAB FILE ID:		RT 1000 =	<u>PO112414.D</u>	RT 750 =	<u>PO112415.D</u>
RT 500 =	PO112416.D	RT 250 =	PO112417.D	RT 050 =	PO112418.D

[illegible]

[illegible]



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION FACTOR OF INITIAL CALIBRATION

Lab Name: Alliance
Lab Code: ACE
Instrument ID: ECD_O

Contract: EARTH03
SDG NO.: Q2818

Calibration Date(s): 07/23/2025 07/23/2025
Calibration Times: 11:32 19:47

GC Column: ZB-MR1 ID: 0.32 (mm)

LAB FILE ID:		CF 1000 = <u>PO112414.D</u>		CF 750 = <u>PO112415.D</u>			
CF 500 = <u>PO112416.D</u>		CF 250 = <u>PO112417.D</u>		CF 050 = <u>PO112418.D</u>			
COMPOUND		CF 1000	CF 750	CF 500	CF 250	CF 050	% RSD
Aroclor-1016-1	(1)	281481703	285148303	290849518	279501268	204848240	268365806
Aroclor-1016-2	(2)	418001550	431852928	433849136	420704900	326178800	406117463
Aroclor-1016-3	(3)	263638281	273794351	273610504	262887848	234438940	261673985
Aroclor-1016-4	(4)	216165184	219268720	223332708	213740540	186700200	211841470
Aroclor-1016-5	(5)	221113370	228268088	233867428	235369996	178734160	219470608
Aroclor-1260-1	(1)	431773857	450001749	476240286	447612572	391026880	439331069
Aroclor-1260-2	(2)	673380136	698254531	722944530	686506064	658383800	687893812
Aroclor-1260-3	(3)	611274773	627246767	610311884	614713140	476551180	588019549
Aroclor-1260-4	(4)	447285263	461047633	473248904	505121600	368319820	451004644
Aroclor-1260-5	(5)	1262689005	1283961461	1337121492	1347466800	964454900	1239138732
Decachlorobiphenyl		7456684020	7676252293	7829495380	7672255680	5924316000	7311800675
Tetrachloro-m-xylene		8554863660	8750848067	8754072420	8314620000	6283438800	8131568589
Aroclor-1242-1	(1)	256299409	259568919	260660454	265097176	196334180	247592028
Aroclor-1242-2	(2)	381137122	387088803	391556862	393536916	295648160	369793573
Aroclor-1242-3	(3)	241769617	246476011	248315054	252195680	194532760	236657824
Aroclor-1242-4	(4)	195717021	198363989	201564902	203398396	172348620	194278586
Aroclor-1242-5	(5)	215656431	220174447	229182990	229543836	175020200	213915581
Decachlorobiphenyl		8000645330	8037989413	8087967020	8086832920	6389983800	7720683697
Tetrachloro-m-xylene		9125652230	9195232120	9219491440	9207437080	6341808000	8617924174
Aroclor-1248-1	(1)	189314500	193406379	198364040	203445536	151513120	187208715
Aroclor-1248-2	(2)	257058197	266019899	270572928	278752472	203158020	255112303
Aroclor-1248-3	(3)	329824745	342687100	350417966	364320620	331255900	343701266
Aroclor-1248-4	(4)	503150480	523568551	540549606	563745892	409374460	508077798
Aroclor-1248-5	(5)	346508966	363261911	374337920	392493440	271353380	349591123
Decachlorobiphenyl		7900018490	8086308400	8195171040	8262080520	6143274400	7717370570
Tetrachloro-m-xylene		8946590960	9189061507	9273186240	9357342160	6567561400	8666748453
Aroclor-1254-1	(1)	525205609	551869540	568048078	599161796	500508140	548958633
Aroclor-1254-2	(2)	467389736	490253857	501431616	531267632	439250900	485918748
Aroclor-1254-3	(3)	736681841	765368536	778116616	788008044	653257120	744286431
Aroclor-1254-4	(4)	537295054	560529123	571244926	572391952	507057560	549703723
Aroclor-1254-5	(5)	713157717	729355651	738947588	743850736	613809420	707824222
Decachlorobiphenyl		7810644110	7988158653	8118440840	8069654680	6321650600	7661709777
Tetrachloro-m-xylene		8477893750	8928798520	9019842540	9077508800	7004909800	8501790682
Aroclor-1268-1	(1)	1805209287	1836925397	1850534848	1852730580	1410311920	1751142406



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION FACTOR OF INITIAL CALIBRATION

Aroclor-1268-2	(2)	1523811042	1548194379	1559255960	1566806172	1167050180	1473023547	11
Aroclor-1268-3	(3)	1298638460	1316156257	1333295814	1329215988	983852920	1252231888	11
Aroclor-1268-4	(4)	502325303	515160189	514467028	515975792	361392800	481864222	13
Aroclor-1268-5	(5)	3537850645	3568750916	3561527908	3536594816	2581136540	3357172165	12
Decachlorobiphenyl		14538757830	14683927053	14777305700	14801547680	10962984400	13952904533	12
Tetrachloro-m-xylene		8916461140	9229914440	9271431040	9355148600	6516020600	8657795164	14



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION FACTOR OF INITIAL CALIBRATION

Lab Name: Alliance
Lab Code: ACE
Instrument ID: ECD_O

Contract: EARTH03
SDG NO.: Q2818

Calibration Date(s): 07/23/2025 07/23/2025
Calibration Times: 11:32 19:47

GC Column: ZB-MR2 ID: 0.32 (mm)

LAB FILE ID:		CF 1000 = <u>PO112414.D</u>		CF 750 = <u>PO112415.D</u>			
CF 500 = <u>PO112416.D</u>		CF 250 = <u>PO112417.D</u>		CF 050 = <u>PO112418.D</u>			
COMPOUND		CF 1000	CF 750	CF 500	CF 250	CF 050	% RSD
Aroclor-1016-1	(1)	171399488	177928739	180256326	179838608	165612060	175007044
Aroclor-1016-2	(2)	254283756	266059885	273532534	272267680	236331280	260495027
Aroclor-1016-3	(3)	134555314	139369485	142347070	141007248	122553240	135966471
Aroclor-1016-4	(4)	106915956	111683577	115661150	118698688	102464540	111084782
Aroclor-1016-5	(5)	139169322	144750759	147937312	152336452	132321220	143303013
Aroclor-1260-1	(1)	240286237	254043527	261134446	258041412	229050000	248511124
Aroclor-1260-2	(2)	317223235	329626419	335041102	340620804	324956000	329493512
Aroclor-1260-3	(3)	256241072	262701627	269297872	275910760	220454080	256921082
Aroclor-1260-4	(4)	185740526	193469844	197255514	200936168	179151380	191310686
Aroclor-1260-5	(5)	419199446	421594747	433075366	434815624	352283180	412193673
Decachlorobiphenyl		1753295660	1820833173	1877801640	1892185680	1491444000	1767112031
Tetrachloro-m-xylene		5169666700	5314856493	5330554580	5111206560	3919594600	4969175787
Aroclor-1242-1	(1)	153645977	158282413	162201976	169125204	148714620	158394038
Aroclor-1242-2	(2)	230553136	236233123	241816698	253231308	206497880	233666429
Aroclor-1242-3	(3)	120997871	124593004	126633224	130347852	103373540	121189098
Aroclor-1242-4	(4)	116921071	120755713	124229780	129097232	122693120	122739383
Aroclor-1242-5	(5)	154482444	160907067	167116390	163960092	158390400	160971279
Decachlorobiphenyl		1831604040	1871840933	1924104720	1967119920	1579850400	1834904003
Tetrachloro-m-xylene		5483238770	5543459253	5573552360	5630865680	4061456200	5258514453
Aroclor-1248-1	(1)	116797575	121746903	126835386	132022196	110144520	121509316
Aroclor-1248-2	(2)	161755092	168905528	175846314	185366264	154525280	169279696
Aroclor-1248-3	(3)	170672394	178220732	184519824	194073912	160748960	177647164
Aroclor-1248-4	(4)	202462025	210645355	219142380	229572292	179950700	208354550
Aroclor-1248-5	(5)	207118457	216119221	227433754	242834452	180319280	214765033
Decachlorobiphenyl		1836524820	1878481893	1941164520	1982063240	1517022000	1831051295
Tetrachloro-m-xylene		5362990180	5502770333	5558036420	5629888520	4177350800	5246207251
Aroclor-1254-1	(1)	308086628	324598995	339171148	360746064	310502200	328621007
Aroclor-1254-2	(2)	269579969	283423153	295499466	318260532	280690420	289490708
Aroclor-1254-3	(3)	411643535	427646563	444183030	462345912	391487940	427461396
Aroclor-1254-4	(4)	256982627	269810957	279068642	286892132	272060980	272963068
Aroclor-1254-5	(5)	324913084	333876627	346327314	357344936	307636300	334019652
Decachlorobiphenyl		1794065640	1872151533	1910898600	1948005080	1573949600	1819814091
Tetrachloro-m-xylene		5135321880	5340323933	5419556800	5524764520	4404208000	5164835027
Aroclor-1268-1	(1)	514567345	516695037	532735372	551455976	473857920	517862330



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION FACTOR OF INITIAL CALIBRATION

Aroclor-1268-2	(2)	421574479	425971267	436618852	454509592	386645480	425063934	6
Aroclor-1268-3	(3)	318758782	324505707	331241876	345977680	285851460	321267101	7
Aroclor-1268-4	(4)	111516853	115705837	114796592	121616720	92841800	111295560	9
Aroclor-1268-5	(5)	742715083	756736297	758808540	780432416	619995020	731737471	8
Decachlorobiphenyl		3224484190	3300929360	3351395200	3481054400	2726789400	3216930510	9
Tetrachloro-m-xylene		5326223920	5511724707	5541913900	5672741280	4190979000	5248716561	12



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Lab Name: Alliance **Contract:** EARTH03
Lab Code: ACE **SDG NO.:** Q2818
Instrument ID: ECD_O **Date(s) Analyzed:** 07/23/2025 07/23/2025
GC Column: ZB-MR1 **ID:** 0.32 (mm)

COMPOUND	AMOUNT (ng)	PEAK	RT	RT WINDOW		CALIBRATION FACTOR
				FROM	TO	
Aroclor-1221	500	1	3.88	3.78	3.98	96659600
		2	3.97	3.87	4.07	71862000
		3	4.04	3.94	4.14	237712000
		4	0.00			0
		5	0.00			0
Aroclor-1232	500	1	4.04	3.94	4.14	186979000
		2	4.54	4.44	4.64	107625000
		3	4.78	4.68	4.88	206800000
		4	4.95	4.85	5.05	104740000
		5	4.99	4.89	5.09	67880800
Aroclor-1262	500	1	6.80	6.70	6.90	929208000
		2	7.30	7.20	7.40	1487020000
		3	7.59	7.49	7.69	590252000
		4	7.65	7.55	7.75	994166000
		5	8.15	8.05	8.25	434062000



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Lab Name: Alliance **Contract:** EARTH03
Lab Code: ACE **SDG NO.:** Q2818
Instrument ID: ECD_O **Date(s) Analyzed:** 07/23/2025 07/23/2025
GC Column: ZB-MR2 **ID:** 0.32 (mm)

COMPOUND	AMOUNT (ng)	PEAK	RT	RT WINDOW		CALIBRATION FACTOR
				FROM	TO	
Aroclor-1221	500	1	3.87	3.77	3.97	67265800
		2	3.96	3.86	4.06	50728000
		3	4.03	3.93	4.13	158529000
		4	0.00			0
		5	0.00			0
Aroclor-1232	500	1	4.03	3.93	4.13	126550000
		2	4.76	4.66	4.86	131155000
		3	4.93	4.83	5.03	67196800
		4	5.02	4.92	5.12	59829000
		5	5.19	5.09	5.29	65203000
Aroclor-1262	500	1	6.77	6.67	6.87	414828000
		2	7.27	7.17	7.37	515546000
		3	7.55	7.45	7.65	186525000
		4	7.61	7.51	7.71	288104000
		5	8.10	8.00	8.20	99503400

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0072325\
 Data File : P0112414.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 23 Jul 2025 11:32
 Operator : YP/AJ
 Sample : AR1660ICC1000
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1660ICC1000

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 23 12:31:02 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0072325.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Jul 23 12:24:33 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.670	3.663	855.5E6	517.0E6	98.849	98.468
2) SA Decachlor...	8.695	8.641	745.7E6	175.3E6	97.561	96.571
Target Compounds						
3) L1 AR-1016-1	4.756	4.738	281.5E6	171.4E6	983.632	974.814
4) L1 AR-1016-2	4.775	4.756	418.0E6	254.3E6	981.396	963.531
5) L1 AR-1016-3	4.832	4.931	263.6E6	134.6E6	981.438	971.861
6) L1 AR-1016-4	4.951	4.974	216.2E6	106.9E6	983.692	960.709
7) L1 AR-1016-5	5.208	5.186	221.1E6	139.2E6	971.968	969.461
31) L7 AR-1260-1	6.245	6.214	431.8E6	240.3E6	951.029	958.422
32) L7 AR-1260-2	6.435	6.402	673.4E6	317.2E6	964.504	972.683
33) L7 AR-1260-3	6.801	6.554	611.3E6	256.2E6	1000.788	975.155
34) L7 AR-1260-4	7.061	7.023	447.3E6	185.7E6	971.795	969.934
35) L7 AR-1260-5	7.304	7.265	1262.7E6	419.2E6	971.370	983.719



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO072325\
Data File : PO112414.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Jul 2025 11:32
Operator : YP/AJ
Sample : AR1660ICC1000
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1660ICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 23 12:31:02 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO072325.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jul 23 12:24:33 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0072325\
 Data File : P0112415.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 23 Jul 2025 11:50
 Operator : YP/AJ
 Sample : AR1660ICC750
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1660ICC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 23 12:34:32 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0072325.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Jul 23 12:24:33 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.668	3.662	656.3E6	398.6E6	75.555	75.614
2)	SA Decachlor...	8.692	8.640	575.7E6	136.6E6	75.217	75.145
Target Compounds							
3)	L1 AR-1016-1	4.754	4.738	213.9E6	133.4E6	748.220	755.950
4)	L1 AR-1016-2	4.772	4.756	323.9E6	199.5E6	756.926	754.066
5)	L1 AR-1016-3	4.829	4.931	205.3E6	104.5E6	759.562	753.309
6)	L1 AR-1016-4	4.949	4.973	164.5E6	83762683	748.907	751.773
7)	L1 AR-1016-5	5.206	5.185	171.2E6	108.6E6	751.707	754.159
31)	L7 AR-1260-1	6.243	6.213	337.5E6	190.5E6	745.576	756.618
32)	L7 AR-1260-2	6.433	6.401	523.7E6	247.2E6	750.066	755.338
33)	L7 AR-1260-3	6.798	6.553	470.4E6	197.0E6	763.349	749.871
34)	L7 AR-1260-4	7.058	7.022	345.8E6	145.1E6	750.847	755.131
35)	L7 AR-1260-5	7.301	7.264	963.0E6	316.2E6	743.842	744.651



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO072325\
Data File : PO112415.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Jul 2025 11:50
Operator : YP/AJ
Sample : AR1660ICC750
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1660ICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 23 12:34:32 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO072325.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jul 23 12:24:33 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0072325\
 Data File : P0112416.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 23 Jul 2025 12:08
 Operator : YP/AJ
 Sample : AR1660ICC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1660ICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 23 12:27:27 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0072325.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Jul 23 12:24:33 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.669	3.662	437.7E6	266.5E6	50.000	50.000
2) SA Decachlor...	8.693	8.640	391.5E6	93890082	50.000	50.000
Target Compounds						
3) L1 AR-1016-1	4.755	4.738	145.4E6	90128163	500.000	500.000
4) L1 AR-1016-2	4.774	4.756	216.9E6	136.8E6	500.000	500.000
5) L1 AR-1016-3	4.831	4.931	136.8E6	71173535	500.000	500.000
6) L1 AR-1016-4	4.951	4.973	111.7E6	57830575	500.000	500.000
7) L1 AR-1016-5	5.208	5.185	116.9E6	73968656	500.000	500.000
31) L7 AR-1260-1	6.245	6.213	238.1E6	130.6E6	500.000	500.000
32) L7 AR-1260-2	6.433	6.402	361.5E6	167.5E6	500.000	500.000
33) L7 AR-1260-3	6.800	6.553	305.2E6	134.6E6	500.000	500.000
34) L7 AR-1260-4	7.059	7.023	236.6E6	98627757	500.000	500.000
35) L7 AR-1260-5	7.303	7.264	668.6E6	216.5E6	500.000	500.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO072325\
Data File : PO112416.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Jul 2025 12:08
Operator : YP/AJ
Sample : AR1660ICC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1660ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 23 12:27:27 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO072325.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jul 23 12:24:33 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_0\Data\P0072325\
 Data File : P0112417.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 23 Jul 2025 12:27
 Operator : YP/AJ
 Sample : AR1660ICC250
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_0
 ClientSampleId :
 AR1660ICC250

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 07/24/2025
 Supervised By :mohammad ahmed 07/25/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 23 12:41:31 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\P0072325.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Jul 23 12:41:19 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.670	3.662	207.9E6	127.8E6	24.188	24.425
2) SA Decachlor...	8.693	8.639	191.8E6	47304642	25.044	25.765
Target Compounds						
3) L1 AR-1016-1	4.756	4.737	69875317	44959652	245.828	253.500
4) L1 AR-1016-2	4.774	4.756	105.2E6	68066920	246.833	255.376
5) L1 AR-1016-3	4.831	4.931	65721962	35251812	244.790	253.028
6) L1 AR-1016-4	4.951	4.973	53435135	29674672	244.973	262.052
7) L1 AR-1016-5	5.207	5.185	58842499	38084113	256.222	260.764
31) L7 AR-1260-1	6.243	6.213	111.9E6	64510353	248.096m	249.850m
32) L7 AR-1260-2	6.433	6.401	171.6E6	85155201	247.077m	257.556
33) L7 AR-1260-3	6.800	6.552	153.7E6	68977690	249.524	262.028m
34) L7 AR-1260-4	7.059	7.022	126.3E6	50234042	267.727	258.471
35) L7 AR-1260-5	7.302	7.264	336.9E6	108.7E6	257.581	254.474

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO072325\
Data File : PO112417.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Jul 2025 12:27
Operator : YP/AJ
Sample : AR1660ICC250
Misc :
ALS Vial : 6 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
AR1660ICC250

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 07/24/2025
Supervised By :mohammad ahmed 07/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 23 12:41:31 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO072325.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jul 23 12:41:19 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0072325\
 Data File : P0112418.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 23 Jul 2025 12:45
 Operator : YP/AJ
 Sample : AR1660ICC050
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1660ICC050

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 07/24/2025
 Supervised By :mohammad ahmed 07/25/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 23 13:01:11 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0072325.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Jul 23 13:01:00 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.669	3.662	31417194	19597973	3.883	3.906
2) SA Decachlor...	8.693	8.640	29621580	7457220	4.051	4.220
Target Compounds						
3) L1 AR-1016-1	4.756	4.737	10242412	8280603	38.166	47.316
4) L1 AR-1016-2	4.774	4.756	16308940	11816564	40.158	45.362
5) L1 AR-1016-3	4.832	4.931	11721947	6127662	44.796	45.067
6) L1 AR-1016-4	4.951	4.973	9335010	5123227	44.066	46.120
7) L1 AR-1016-5	5.207	5.185	8936708	6616061	41.260m	46.488m
31) L7 AR-1260-1	6.245	6.213	19551344	11452500	44.503	44.442m
32) L7 AR-1260-2	6.434	6.402	32919190	16247800	47.855	49.311
33) L7 AR-1260-3	6.801	6.553	23827559	11022704	40.522	43.782m
34) L7 AR-1260-4	7.060	7.022	18415991	8957569	40.833	46.822
35) L7 AR-1260-5	7.303	7.265	48222745	17614159	38.916	42.733

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO072325\
Data File : PO112418.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Jul 2025 12:45
Operator : YP/AJ
Sample : AR1660ICC050
Misc :
ALS Vial : 7 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
AR1660ICC050

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 07/24/2025
Supervised By :mohammad ahmed 07/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 23 13:01:11 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO072325.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jul 23 13:01:00 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0072325\
 Data File : P0112419.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 23 Jul 2025 13:03
 Operator : YP/AJ
 Sample : AR1221ICC500
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1221ICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 23 13:39:35 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0072325.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Jul 23 13:39:21 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.670	3.663	423.2E6	251.3E6	50.000	50.000
2) SA Decachlor...	8.694	8.641	374.2E6	89450396	50.000	50.000
Target Compounds						
8) L2 AR-1221-1	3.882	3.872	48329823	33632938	500.000	500.000
9) L2 AR-1221-2	3.967	3.958	35930990	25363994	500.000	500.000
10) L2 AR-1221-3	4.044	4.032	118.9E6	79264688	500.000	500.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO072325\
Data File : PO112419.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Jul 2025 13:03
Operator : YP/AJ
Sample : AR1221ICC500
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1221ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 23 13:39:35 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO072325.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jul 23 13:39:21 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0072325\
 Data File : P0112420.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 23 Jul 2025 13:22
 Operator : YP/AJ
 Sample : AR1232ICC500
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1232ICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 23 14:21:11 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0072325.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Jul 23 14:21:01 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.670	3.662	456.9E6	277.1E6	50.000	50.000
2) SA Decachlor...	8.694	8.641	397.2E6	95040741	50.000	50.000
Target Compounds						
11) L3 AR-1232-1	4.043	4.031	93489328	63274969	500.000	500.000
12) L3 AR-1232-2	4.535	4.756	53812362	65577392	500.000	500.000
13) L3 AR-1232-3	4.775	4.931	103.4E6	33598402	500.000	500.000
14) L3 AR-1232-4	4.951	5.015	52370149	29914546	500.000	500.000
15) L3 AR-1232-5	4.994	5.185	33940444	32601518	500.000	500.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO072325\
Data File : PO112420.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Jul 2025 13:22
Operator : YP/AJ
Sample : AR1232ICC500
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1232ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 23 14:21:11 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO072325.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jul 23 14:21:01 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0072325\
 Data File : P0112421.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 23 Jul 2025 13:40
 Operator : YP/AJ
 Sample : AR1242ICC1000
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1242ICC1000

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 23 14:43:34 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0072325.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Jul 23 14:40:09 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.670	3.663	912.6E6	548.3E6	99.488	99.183
2)	SA Decachlor...	8.694	8.641	800.1E6	183.2E6	99.457	97.537
Target Compounds							
16)	L4 AR-1242-1	4.755	4.738	256.3E6	153.6E6	991.564	972.911
17)	L4 AR-1242-2	4.775	4.756	381.1E6	230.6E6	986.515	976.155
18)	L4 AR-1242-3	4.832	4.931	241.8E6	121.0E6	986.644	977.243
19)	L4 AR-1242-4	4.951	5.015	195.7E6	116.9E6	985.280	969.692
20)	L4 AR-1242-5	5.602	5.534	215.7E6	154.5E6	969.592	960.715



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO072325\
Data File : PO112421.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Jul 2025 13:40
Operator : YP/AJ
Sample : AR1242ICC1000
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1242ICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 23 14:43:34 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO072325.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jul 23 14:40:09 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0072325\
 Data File : P0112422.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 23 Jul 2025 13:59
 Operator : YP/AJ
 Sample : AR1242ICC750
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1242ICC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 23 14:46:26 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0072325.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Jul 23 14:40:09 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.670	3.662	689.6E6	415.8E6	75.123	75.136
2) SA Decachlor...	8.694	8.641	602.8E6	140.4E6	74.961	74.840
Target Compounds						
16) L4 AR-1242-1	4.756	4.738	194.7E6	118.7E6	752.104	751.134
17) L4 AR-1242-2	4.774	4.756	290.3E6	177.2E6	750.959	750.102
18) L4 AR-1242-3	4.831	4.931	184.9E6	93444753	752.920	753.133
19) L4 AR-1242-4	4.951	5.015	148.8E6	90566785	749.303	750.747
20) L4 AR-1242-5	5.602	5.534	165.1E6	120.7E6	744.936	750.335



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO072325\
Data File : PO112422.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Jul 2025 13:59
Operator : YP/AJ
Sample : AR1242ICC750
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1242ICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 23 14:46:26 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO072325.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jul 23 14:40:09 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0072325\
 Data File : P0112423.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 23 Jul 2025 14:17
 Operator : YP/AJ
 Sample : AR1242ICC500
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1242ICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 23 14:40:35 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0072325.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Jul 23 14:40:09 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.669	3.662	461.0E6	278.7E6	50.000	50.000
2) SA Decachlor...	8.693	8.641	404.4E6	96205236	50.000	50.000
Target Compounds						
16) L4 AR-1242-1	4.755	4.737	130.3E6	81100988	500.000	500.000
17) L4 AR-1242-2	4.774	4.755	195.8E6	120.9E6	500.000	500.000
18) L4 AR-1242-3	4.831	4.931	124.2E6	63316612	500.000	500.000
19) L4 AR-1242-4	4.951	5.015	100.8E6	62114890	500.000	500.000
20) L4 AR-1242-5	5.602	5.534	114.6E6	83558195	500.000	500.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO072325\
Data File : PO112423.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Jul 2025 14:17
Operator : YP/AJ
Sample : AR1242ICC500
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1242ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 23 14:40:35 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO072325.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jul 23 14:40:09 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0072325\
 Data File : P0112424.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 23 Jul 2025 14:36
 Operator : YP/AJ
 Sample : AR1242ICC250
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1242ICC250

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 07/24/2025
 Supervised By :mohammad ahmed 07/25/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 23 14:50:42 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0072325.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Jul 23 14:50:26 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.669	3.662	230.2E6	140.8E6	25.056	25.329
2) SA Decachlor...	8.692	8.640	202.2E6	49177998	25.104	25.901
Target Compounds						
16) L4 AR-1242-1	4.756	4.737	66274294	42281301	254.503	262.921
17) L4 AR-1242-2	4.774	4.756	98384229	63307827	253.352	263.280
18) L4 AR-1242-3	4.830	4.930	63048920	32586963	255.064	259.362
19) L4 AR-1242-4	4.951	5.015	50849599	32274308	254.552	262.925
20) L4 AR-1242-5	5.601	5.534	57385959	40990023	262.483m	253.808m

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0072325\
Data File : P0112424.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Jul 2025 14:36
Operator : YP/AJ
Sample : AR1242ICC250
Misc :
ALS Vial : 13 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
AR1242ICC250

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 07/24/2025
Supervised By :mohammad ahmed 07/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 23 14:50:42 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0072325.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jul 23 14:50:26 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0072325\
 Data File : P0112425.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 23 Jul 2025 14:54
 Operator : YP/AJ
 Sample : AR1242ICC050
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1242ICC050

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 07/24/2025
 Supervised By :mohammad ahmed 07/25/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 23 15:07:13 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0072325.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Jul 23 15:07:00 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.670	3.662	31709040	20307281	3.679	3.861m
2)	SA Decachlor...	8.695	8.640	31949919	7899252	4.138	4.305
Target Compounds							
16)	L4 AR-1242-1	4.757	4.738	9816709	7435731	39.649	46.945
17)	L4 AR-1242-2	4.776	4.756	14782408	10324894	39.975	44.186
18)	L4 AR-1242-3	4.833	4.932	9726638	5168677	41.100	42.650
19)	L4 AR-1242-4	4.952	5.016	8617431	6134656	44.356	49.981
20)	L4 AR-1242-5	5.603	5.535	8751010	7919520	41.522m	48.545m

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO072325\
Data File : PO112425.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Jul 2025 14:54
Operator : YP/AJ
Sample : AR1242ICC050
Misc :
ALS Vial : 14 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
AR1242ICC050

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 07/24/2025
Supervised By :mohammad ahmed 07/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 23 15:07:13 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO072325.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jul 23 15:07:00 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_0\Data\P0072325\
 Data File : P0112426.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 23 Jul 2025 15:13
 Operator : YP/AJ
 Sample : AR1248ICC1000
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 ECD_0
 ClientSampleId :
 AR1248ICC1000

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 23 17:42:45 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\P0072325.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Jul 23 17:39:09 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.670	3.663	894.7E6	536.3E6	98.207	98.214
2) SA Decachlor...	8.694	8.642	790.0E6	183.7E6	98.166	97.230
Target Compounds						
21) L5 AR-1248-1	4.756	4.738	189.3E6	116.8E6	976.657	958.799
22) L5 AR-1248-2	4.994	4.974	257.1E6	161.8E6	974.386	958.261
23) L5 AR-1248-3	5.208	5.015	329.8E6	170.7E6	969.727	961.014
24) L5 AR-1248-4	5.561	5.185	503.2E6	202.5E6	964.167	960.436
25) L5 AR-1248-5	5.602	5.576	346.5E6	207.1E6	961.394	953.250



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO072325\
Data File : PO112426.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Jul 2025 15:13
Operator : YP/AJ
Sample : AR1248ICC1000
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1248ICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 23 17:42:45 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO072325.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jul 23 17:39:09 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_0\Data\P0072325\
 Data File : P0112427.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 23 Jul 2025 15:31
 Operator : YP/AJ
 Sample : AR1248ICC750
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 ECD_0
 ClientSampleId :
 AR1248ICC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 23 17:46:03 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\P0072325.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Jul 23 17:39:09 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.669	3.662	689.2E6	412.7E6	75.433	75.386
2) SA Decachlor...	8.694	8.640	606.5E6	140.9E6	75.240	74.725
Target Compounds						
21) L5 AR-1248-1	4.755	4.737	145.1E6	91310177	748.883	749.714
22) L5 AR-1248-2	4.993	4.973	199.5E6	126.7E6	754.166	750.310
23) L5 AR-1248-3	5.207	5.015	257.0E6	133.7E6	753.762	751.756
24) L5 AR-1248-4	5.560	5.185	392.7E6	158.0E6	751.645	749.628
25) L5 AR-1248-5	5.601	5.575	272.4E6	162.1E6	753.927	747.333



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO072325\
Data File : PO112427.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Jul 2025 15:31
Operator : YP/AJ
Sample : AR1248ICC750
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1248ICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 23 17:46:03 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO072325.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jul 23 17:39:09 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0072325\
 Data File : P0112428.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 23 Jul 2025 15:48
 Operator : YP/AJ
 Sample : AR1248ICC500
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1248ICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 23 17:39:29 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0072325.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Jul 23 17:39:09 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.668	3.662	463.7E6	277.9E6	50.000	50.000
2) SA Decachlor...	8.692	8.640	409.8E6	97058226	50.000	50.000
Target Compounds						
21) L5 AR-1248-1	4.754	4.737	99182020	63417693	500.000	500.000
22) L5 AR-1248-2	4.992	4.973	135.3E6	87923157	500.000	500.000
23) L5 AR-1248-3	5.205	5.015	175.2E6	92259912	500.000	500.000
24) L5 AR-1248-4	5.559	5.185	270.3E6	109.6E6	500.000	500.000
25) L5 AR-1248-5	5.599	5.575	187.2E6	113.7E6	500.000	500.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO072325\
Data File : PO112428.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Jul 2025 15:48
Operator : YP/AJ
Sample : AR1248ICC500
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1248ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 23 17:39:29 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO072325.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jul 23 17:39:09 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0072325\
 Data File : P0112429.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 23 Jul 2025 16:07
 Operator : YP/AJ
 Sample : AR1248ICC250
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1248ICC250

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 23 17:49:12 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0072325.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Jul 23 17:39:09 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.670	3.663	233.9E6	140.7E6	25.451	25.528
2) SA Decachlor...	8.695	8.641	206.6E6	49551581	25.466	25.949
Target Compounds						
21) L5 AR-1248-1	4.756	4.738	50861384	33005549	259.321	265.423
22) L5 AR-1248-2	4.994	4.974	69688118	46341566	259.932	267.919
23) L5 AR-1248-3	5.208	5.016	91080155	48518478	262.621	266.773
24) L5 AR-1248-4	5.561	5.186	140.9E6	57393073	264.543	266.380
25) L5 AR-1248-5	5.602	5.576	98123360	60708613	265.809	271.777



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO072325\
Data File : PO112429.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Jul 2025 16:07
Operator : YP/AJ
Sample : AR1248ICC250
Misc :
ALS Vial : 18 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1248ICC250

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 23 17:49:12 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO072325.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jul 23 17:39:09 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0072325\
 Data File : P0112430.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 23 Jul 2025 16:25
 Operator : YP/AJ
 Sample : AR1248ICC050
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1248ICC050

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 07/24/2025
 Supervised By :mohammad ahmed 07/25/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 23 17:53:17 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0072325.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Jul 23 17:39:09 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.670	3.662	32837807	20886754	3.789	3.981
2) SA Decachlor...	8.694	8.641	30716372	7585110	3.980	4.142
Target Compounds						
21) L5 AR-1248-1	4.757	4.738	7575656	5507226	40.466	45.323
22) L5 AR-1248-2	4.995	4.973	10157901	7726264	39.817	45.642
23) L5 AR-1248-3	5.208	5.015	16562795	8037448	48.190	45.244
24) L5 AR-1248-4	5.561	5.185	20468723	8997535	40.302m	43.063m
25) L5 AR-1248-5	5.602	5.575	13567669	9015964	38.853m	41.910m

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO072325\
Data File : PO112430.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Jul 2025 16:25
Operator : YP/AJ
Sample : AR1248ICC050
Misc :
ALS Vial : 19 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
AR1248ICC050

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 07/24/2025
Supervised By :mohammad ahmed 07/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 23 17:53:17 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO072325.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jul 23 17:39:09 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0072325\
 Data File : P0112431.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 23 Jul 2025 16:44
 Operator : YP/AJ
 Sample : AR1254ICC1000
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1254ICC1000

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 23 18:03:16 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0072325.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Jul 23 17:59:47 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.670	3.663	847.8E6	513.5E6	96.903	97.307
2) SA Decachlor...	8.694	8.640	781.1E6	179.4E6	98.068	96.847
Target Compounds						
26) L6 AR-1254-1	5.561	5.535	525.2E6	308.1E6	960.812	951.975
27) L6 AR-1254-2	5.710	5.682	467.4E6	269.6E6	964.863	954.131
28) L6 AR-1254-3	6.114	6.083	736.7E6	411.6E6	972.647	961.979
29) L6 AR-1254-4	6.343	6.311	537.3E6	257.0E6	969.374	958.799
30) L6 AR-1254-5	6.762	6.727	713.2E6	324.9E6	982.240	968.098



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO072325\
Data File : PO112431.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Jul 2025 16:44
Operator : YP/AJ
Sample : AR1254ICC1000
Misc :
ALS Vial : 20 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1254ICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 23 18:03:16 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO072325.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jul 23 17:59:47 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0072325\
 Data File : P0112432.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 23 Jul 2025 17:02
 Operator : YP/AJ
 Sample : AR1254ICC750
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1254ICC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 23 18:06:16 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0072325.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Jul 23 17:59:47 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.670	3.662	669.7E6	400.5E6	76.021	75.593
2)	SA Decachlor...	8.694	8.640	599.1E6	140.4E6	75.148	75.529
Target Compounds							
26)	L6 AR-1254-1	5.561	5.535	413.9E6	243.4E6	754.780	751.497
27)	L6 AR-1254-2	5.710	5.683	367.7E6	212.6E6	756.007	751.562
28)	L6 AR-1254-3	6.113	6.084	574.0E6	320.7E6	755.243	749.688
29)	L6 AR-1254-4	6.343	6.311	420.4E6	202.4E6	755.625	753.323
30)	L6 AR-1254-5	6.762	6.728	547.0E6	250.4E6	752.271	747.398



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO072325\
Data File : PO112432.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Jul 2025 17:02
Operator : YP/AJ
Sample : AR1254ICC750
Misc :
ALS Vial : 21 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1254ICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 23 18:06:16 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO072325.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jul 23 17:59:47 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0072325\
 Data File : P0112433.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 23 Jul 2025 17:21
 Operator : YP/AJ
 Sample : AR1254ICC500
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1254ICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 23 18:00:06 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0072325.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Jul 23 17:59:47 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.670	3.663	451.0E6	271.0E6	50.000	50.000
2) SA Decachlor...	8.695	8.641	405.9E6	95544930	50.000	50.000
Target Compounds						
26) L6 AR-1254-1	5.561	5.535	284.0E6	169.6E6	500.000	500.000
27) L6 AR-1254-2	5.711	5.683	250.7E6	147.7E6	500.000	500.000
28) L6 AR-1254-3	6.114	6.084	389.1E6	222.1E6	500.000	500.000
29) L6 AR-1254-4	6.344	6.311	285.6E6	139.5E6	500.000	500.000
30) L6 AR-1254-5	6.762	6.727	369.5E6	173.2E6	500.000	500.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO072325\
Data File : PO112433.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Jul 2025 17:21
Operator : YP/AJ
Sample : AR1254ICC500
Misc :
ALS Vial : 22 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1254ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 23 18:00:06 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO072325.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jul 23 17:59:47 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0072325\
 Data File : P0112434.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 23 Jul 2025 17:39
 Operator : YP/AJ
 Sample : AR1254ICC250
 Misc :
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1254ICC250

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 23 18:10:33 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0072325.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Jul 23 17:59:47 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.669	3.662	226.9E6	138.1E6	25.568	25.793
2) SA Decachlor...	8.694	8.640	201.7E6	48700127	25.228	25.887
Target Compounds						
26) L6 AR-1254-1	5.561	5.534	149.8E6	90186516	266.972	270.708
27) L6 AR-1254-2	5.709	5.682	132.8E6	79565133	266.923	272.772
28) L6 AR-1254-3	6.113	6.083	197.0E6	115.6E6	256.833	264.830
29) L6 AR-1254-4	6.343	6.311	143.1E6	71723033	255.366	262.540
30) L6 AR-1254-5	6.761	6.727	186.0E6	89336234	254.281	262.279



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO072325\
Data File : PO112434.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Jul 2025 17:39
Operator : YP/AJ
Sample : AR1254ICC250
Misc :
ALS Vial : 23 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1254ICC250

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 23 18:10:33 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO072325.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jul 23 17:59:47 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0072325\
 Data File : P0112435.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 23 Jul 2025 17:57
 Operator : YP/AJ
 Sample : AR1254ICC050
 Misc :
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1254ICC050

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 07/24/2025
 Supervised By :mohammad ahmed 07/25/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 23 18:14:12 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0072325.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Jul 23 17:59:47 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.669	3.661	35024549	22021040	4.120	4.264
2) SA Decachlor...	8.694	8.640	31608253	7869748	4.125	4.324
Target Compounds						
26) L6 AR-1254-1	5.560	5.534	25025407	15525110	45.951m	47.285m
27) L6 AR-1254-2	5.710	5.682	21962545	14034521	45.717m	48.955m
28) L6 AR-1254-3	6.114	6.082	32662856	19574397	43.885	45.822m
29) L6 AR-1254-4	6.344	6.311	25352878	13603049	46.121	50.646
30) L6 AR-1254-5	6.762	6.726	30690471	15381815	43.359	46.051

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO072325\
Data File : PO112435.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Jul 2025 17:57
Operator : YP/AJ
Sample : AR1254ICC050
Misc :
ALS Vial : 24 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
AR1254ICC050

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 07/24/2025
Supervised By :mohammad ahmed 07/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 23 18:14:12 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO072325.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jul 23 17:59:47 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_0\Data\P0072325\
 Data File : P0112436.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 23 Jul 2025 18:16
 Operator : YP/AJ
 Sample : AR1262ICC500
 Misc :
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 ECD_0
 ClientSampleId :
 AR1262ICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 24 04:23:57 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\P0072325.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Jul 24 04:22:29 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.670	3.663	431.0E6	259.6E6	50.000	50.000
2) SA Decachlor...	8.695	8.640	390.0E6	91817563	50.000	50.000
Target Compounds						
36) L8 AR-1262-1	6.801	6.766	464.6E6	207.4E6	500.000	500.000
37) L8 AR-1262-2	7.303	7.265	743.5E6	257.8E6	500.000	500.000
38) L8 AR-1262-3	7.586	7.546	295.1E6	93262624	500.000	500.000
39) L8 AR-1262-4	7.650	7.611	497.1E6	144.1E6	500.000	500.000
40) L8 AR-1262-5	8.146	8.103	217.0E6	49751686	500.000	500.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO072325\
Data File : PO112436.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Jul 2025 18:16
Operator : YP/AJ
Sample : AR1262ICC500
Misc :
ALS Vial : 25 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1262ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 24 04:23:57 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO072325.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Jul 24 04:22:29 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0072325\
 Data File : P0112437.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 23 Jul 2025 18:34
 Operator : YP/AJ
 Sample : AR1268ICC1000
 Misc :
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1268ICC1000

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 24 04:26:45 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0072325.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Jul 24 04:26:17 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.670	3.663	891.6E6	532.6E6	96.171	96.108
2) SA Decachlor...	8.695	8.642	1453.9E6	322.4E6	98.386	96.213
Target Compounds						
41) L9 AR-1268-1	7.587	7.547	1805.2E6	514.6E6	975.507	965.897
42) L9 AR-1268-2	7.652	7.613	1523.8E6	421.6E6	977.268	965.543
43) L9 AR-1268-3	7.858	7.817	1298.6E6	318.8E6	974.006	962.314
44) L9 AR-1268-4	8.147	8.103	502.3E6	111.5E6	976.399	971.430
45) L9 AR-1268-5	8.440	8.392	3537.9E6	742.7E6	993.352	978.791



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_0\Data\PO072325\
Data File : PO112437.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Jul 2025 18:34
Operator : YP/AJ
Sample : AR1268ICC1000
Misc :
ALS Vial : 26 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1268ICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 24 04:26:45 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\PO072325.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Jul 24 04:26:17 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_0\Data\P0072325\
 Data File : P0112438.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 23 Jul 2025 18:53
 Operator : YP/AJ
 Sample : AR1268ICC750
 Misc :
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 ECD_0
 ClientSampleId :
 AR1268ICC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 24 04:27:02 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\P0072325.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Jul 24 04:26:17 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.670	3.663	692.2E6	413.4E6	74.664	74.591
2) SA Decachlor...	8.695	8.641	1101.3E6	247.6E6	74.526	73.871
Target Compounds						
41) L9 AR-1268-1	7.586	7.547	1377.7E6	387.5E6	744.484	727.418
42) L9 AR-1268-2	7.652	7.613	1161.1E6	319.5E6	744.679	731.710
43) L9 AR-1268-3	7.857	7.816	987.1E6	243.4E6	740.359	734.748
44) L9 AR-1268-4	8.147	8.103	386.4E6	86779378	751.011	755.940
45) L9 AR-1268-5	8.439	8.391	2676.6E6	567.6E6	751.521	747.952



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO072325\
Data File : PO112438.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Jul 2025 18:53
Operator : YP/AJ
Sample : AR1268ICC750
Misc :
ALS Vial : 27 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1268ICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 24 04:27:02 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO072325.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Jul 24 04:26:17 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0072325\
 Data File : P0112439.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 23 Jul 2025 19:11
 Operator : YP/AJ
 Sample : AR1268ICC500
 Misc :
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1268ICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 24 04:27:17 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0072325.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Jul 24 04:26:17 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.669	3.661	463.6E6	277.1E6	50.000	50.000
2) SA Decachlor...	8.695	8.641	738.9E6	167.6E6	50.000	50.000
Target Compounds						
41) L9 AR-1268-1	7.586	7.546	925.3E6	266.4E6	500.000	500.000
42) L9 AR-1268-2	7.652	7.611	779.6E6	218.3E6	500.000	500.000
43) L9 AR-1268-3	7.857	7.816	666.6E6	165.6E6	500.000	500.000
44) L9 AR-1268-4	8.146	8.103	257.2E6	57398296	500.000	500.000
45) L9 AR-1268-5	8.439	8.391	1780.8E6	379.4E6	500.000	500.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO072325\
Data File : PO112439.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Jul 2025 19:11
Operator : YP/AJ
Sample : AR1268ICC500
Misc :
ALS Vial : 28 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1268ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 24 04:27:17 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO072325.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Jul 24 04:26:17 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0072325\
 Data File : P0112440.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 23 Jul 2025 19:28
 Operator : YP/AJ
 Sample : AR1268ICC250
 Misc :
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1268ICC250

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 24 04:27:32 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0072325.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Jul 24 04:26:17 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.668	3.663	233.9E6	141.8E6	25.226	25.590
2) SA Decachlor...	8.693	8.640	370.0E6	87026360	25.041	25.967
Target Compounds						
41) L9 AR-1268-1	7.584	7.546	463.2E6	137.9E6	250.297	258.785
42) L9 AR-1268-2	7.649	7.611	391.7E6	113.6E6	251.211	260.244
43) L9 AR-1268-3	7.855	7.815	332.3E6	86494420	249.235	261.122
44) L9 AR-1268-4	8.144	8.102	129.0E6	30404180	250.733	264.853
45) L9 AR-1268-5	8.438	8.390	884.1E6	195.1E6	248.250	257.124



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO072325\
Data File : PO112440.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Jul 2025 19:28
Operator : YP/AJ
Sample : AR1268ICC250
Misc :
ALS Vial : 29 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1268ICC250

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 24 04:27:32 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO072325.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Jul 24 04:26:17 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0072325\
 Data File : P0112441.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 23 Jul 2025 19:47
 Operator : YP/AJ
 Sample : AR1268ICC050
 Misc :
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1268ICC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 24 04:27:47 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0072325.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Jul 24 04:26:17 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.670	3.662	32580103	20954895	3.514	3.781
2) SA Decachlor...	8.695	8.641	54814922	13633947	3.709	4.068
Target Compounds						
41) L9 AR-1268-1	7.587	7.546	70515596	23692896	38.106	44.474
42) L9 AR-1268-2	7.653	7.612	58352509	19332274	37.423	44.277
43) L9 AR-1268-3	7.857	7.816	49192646	14292573	36.896	43.148
44) L9 AR-1268-4	8.148	8.103	18069640	4642090	35.123	40.438
45) L9 AR-1268-5	8.441	8.391	129.1E6	30999751	36.236	40.853



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO072325\
Data File : PO112441.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Jul 2025 19:47
Operator : YP/AJ
Sample : AR1268ICC050
Misc :
ALS Vial : 30 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1268ICC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 24 04:27:47 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO072325.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Jul 24 04:26:17 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0072325\
 Data File : P0112442.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 23 Jul 2025 20:05
 Operator : YP/AJ
 Sample : P0072325ICV500
 Misc :
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 ICVPO072325

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 24 05:04:14 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0072325.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Jul 24 04:54:06 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.670	3.662	429.5E6	257.8E6	52.813	51.889
2) SA Decachlor...	8.695	8.640	375.0E6	87671423	51.286	49.613
Target Compounds						
3) L1 AR-1016-1	4.756	4.737	141.9E6	87193424	528.782	498.228
4) L1 AR-1016-2	4.774	4.756	214.8E6	130.5E6	528.864	500.904
5) L1 AR-1016-3	4.832	4.931	137.7E6	68749292	526.111	505.634
6) L1 AR-1016-4	4.951	4.973	109.7E6	56111590	517.883	505.124
7) L1 AR-1016-5	5.208	5.186	116.9E6	72935368	532.523	508.959
31) L7 AR-1260-1	6.245	6.214	222.6E6	126.1E6	506.590	507.535
32) L7 AR-1260-2	6.435	6.402	342.3E6	162.4E6	497.536	492.983
33) L7 AR-1260-3	6.800	6.553	311.1E6	133.0E6	528.992	517.596
34) L7 AR-1260-4	7.059	7.023	226.3E6	98281247	501.824	513.726
35) L7 AR-1260-5	7.303	7.265	621.4E6	214.6E6	501.504	520.509



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0072325\
Data File : P0112442.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Jul 2025 20:05
Operator : YP/AJ
Sample : P0072325ICV500
Misc :
ALS Vial : 31 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
ICVPO072325

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 24 05:04:14 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0072325.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Jul 24 04:54:06 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0072325\
 Data File : P0112443.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 23 Jul 2025 20:42
 Operator : YP/AJ
 Sample : AR1242ICV500
 Misc :
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 ICVPO072325AR1242

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 24 05:04:36 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0072325.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Jul 24 04:54:06 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.670	3.663	460.7E6	275.9E6	56.651	55.526
2) SA Decachlor...	8.694	8.640	401.5E6	93838611	54.911	53.103
Target Compounds						
16) L4 AR-1242-1	4.756	4.738	129.0E6	80174107	521.171	506.169
17) L4 AR-1242-2	4.776	4.756	194.8E6	118.6E6	526.705	507.457
18) L4 AR-1242-3	4.832	4.931	123.7E6	62408526	522.828	514.968
19) L4 AR-1242-4	4.952	5.015	99083869	61266680	510.009	499.161
20) L4 AR-1242-5	5.603	5.535	107.3E6	81609556	501.589	506.982



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0072325\
Data File : P0112443.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Jul 2025 20:42
Operator : YP/AJ
Sample : AR1242ICV500
Misc :
ALS Vial : 32 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
ICVPO072325AR1242

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 24 05:04:36 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0072325.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Jul 24 04:54:06 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0072325\
 Data File : P0112444.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 23 Jul 2025 21:18
 Operator : YP/AJ
 Sample : AR1248ICV500
 Misc :
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 ICVPO072325AR1248

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 24 05:04:58 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0072325.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Jul 24 04:54:06 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.670	3.662	452.2E6	269.7E6	55.615	54.268
2) SA Decachlor...	8.694	8.641	403.2E6	94089624	55.141	53.245
Target Compounds						
21) L5 AR-1248-1	4.755	4.737	96403586	60735474	514.952	499.842
22) L5 AR-1248-2	4.994	4.973	132.8E6	84845389	520.457	501.214
23) L5 AR-1248-3	5.207	5.015	170.0E6	89430727	494.627	503.418
24) L5 AR-1248-4	5.560	5.185	261.6E6	105.8E6	514.800	507.962
25) L5 AR-1248-5	5.602	5.576	177.0E6	109.0E6	506.243	507.656



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO072325\
Data File : PO112444.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Jul 2025 21:18
Operator : YP/AJ
Sample : AR1248ICV500
Misc :
ALS Vial : 33 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
ICVPO072325AR1248

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 24 05:04:58 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO072325.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Jul 24 04:54:06 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0072325\
 Data File : P0112445.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 23 Jul 2025 21:54
 Operator : YP/AJ
 Sample : AR1254ICV500
 Misc :
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 ICVPO072325AR1254

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 24 05:05:20 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0072325.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Jul 24 04:54:06 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.669	3.662	445.4E6	266.6E6	54.778	53.658
2) SA Decachlor...	8.693	8.639	391.6E6	91844886	53.551	51.975
Target Compounds						
26) L6 AR-1254-1	5.561	5.534	276.3E6	163.9E6	503.299	498.849
27) L6 AR-1254-2	5.711	5.683	249.7E6	144.3E6	513.931	498.389
28) L6 AR-1254-3	6.114	6.083	379.5E6	213.7E6	509.918	499.814
29) L6 AR-1254-4	6.342	6.310	247.0E6	123.1E6	449.340	450.950
30) L6 AR-1254-5	6.761	6.726	356.3E6	167.5E6	503.357	501.388



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO072325\
Data File : PO112445.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Jul 2025 21:54
Operator : YP/AJ
Sample : AR1254ICV500
Misc :
ALS Vial : 34 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
ICVPO072325AR1254

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 24 05:05:20 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO072325.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Jul 24 04:54:06 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0072325\
 Data File : P0112446.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 23 Jul 2025 22:31
 Operator : YP/AJ
 Sample : AR1268ICV500
 Misc :
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 ICVPO072325AR1268

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 24 04:54:33 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0072325.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Jul 24 04:52:41 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.669	3.662	460.9E6	275.4E6	53.238	52.473
2) SA Decachlor...	8.694	8.640	731.3E6	166.7E6	52.412	51.834
Target Compounds						
41) L9 AR-1268-1	7.586	7.546	910.3E6	250.7E6	519.844	484.189
42) L9 AR-1268-2	7.651	7.611	767.0E6	207.2E6	520.717	487.400
43) L9 AR-1268-3	7.856	7.815	656.2E6	158.5E6	524.050	493.242
44) L9 AR-1268-4	8.146	8.102	253.2E6	55462942	525.396	498.339
45) L9 AR-1268-5	8.439	8.391	1754.5E6	373.2E6	522.622	509.970



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0072325\
Data File : P0112446.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Jul 2025 22:31
Operator : YP/AJ
Sample : AR1268ICV500
Misc :
ALS Vial : 35 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
ICVPO072325AR1268

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 24 04:54:33 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0072325.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Jul 24 04:52:41 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Calibration Times: 12:05 20:28

LAB FILE ID: RT 1000 = PP074168.D RT 750 = PP074169.D
RT 500 = PP074170.D RT 250 = PP074171.D RT 050 = PP074172.D

[illegible]

Calibration Times: 12:05 20:28

LAB FILE ID: RT 1000 = PP074168.D RT 750 = PP074169.D
RT 500 = PP074170.D RT 250 = PP074171.D RT 050 = PP074172.D

[illegible]



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION FACTOR OF INITIAL CALIBRATION

Lab Name: Alliance
Lab Code: ACE
Instrument ID: ECD_P

Contract: EARTH03
SDG NO.: Q2818

Calibration Date(s): 08/01/2025 08/01/2025
Calibration Times: 12:05 20:28

GC Column: ZB-MR1 ID: 0.32 (mm)

LAB FILE ID:		CF 1000 = <u>PP074168.D</u>		CF 750 = <u>PP074169.D</u>			
CF 500 = <u>PP074170.D</u>		CF 250 = <u>PP074171.D</u>		CF 050 = <u>PP074172.D</u>			
COMPOUND		CF 1000	CF 750	CF 500	CF 250	CF 050	% RSD
Aroclor-1016-1	(1)	35273925	37423669	39112938	43859036	50614460	41256806
Aroclor-1016-2	(2)	51795728	55199127	58347288	64689960	73676540	60741729
Aroclor-1016-3	(3)	33422245	35685564	37869938	42702196	48622620	39660513
Aroclor-1016-4	(4)	27792815	29586895	31487664	35862436	37992560	32544474
Aroclor-1016-5	(5)	27690329	29450228	31441930	35606544	38151200	32468046
Aroclor-1260-1	(1)	47709498	50888675	54142460	59119820	69919860	56356063
Aroclor-1260-2	(2)	56508230	60166887	63913596	71401348	88771760	68152364
Aroclor-1260-3	(3)	45067030	48123128	50650728	57005548	65633220	53295931
Aroclor-1260-4	(4)	54035441	57066597	60264502	66664376	75043960	62614975
Aroclor-1260-5	(5)	97992993	102942959	107738868	117332464	140517260	113304909
Decachlorobiphenyl		852434670	894679120	940440840	1016872040	1151380600	971161454
Tetrachloro-m-xylene		994587660	1038636867	1076051520	1154524400	1327726000	1118305289
Aroclor-1242-1	(1)	31512428	33419607	35918512	39696020	45782500	37265813
Aroclor-1242-2	(2)	47421171	49356251	53063334	57890692	64313200	54408930
Aroclor-1242-3	(3)	30290298	32076312	34724744	38188808	43399140	35735860
Aroclor-1242-4	(4)	25177060	26633319	28691812	30950152	34640880	29218645
Aroclor-1242-5	(5)	29128304	32087727	34406488	38696772	43470800	35558018
Decachlorobiphenyl		880786000	919506067	984418560	1056643120	1205252000	1009321149
Tetrachloro-m-xylene		1043240650	1075688493	1133074840	1202593000	1370780600	1165075517
Aroclor-1254-1	(1)	49559228	49811068	53907260	53993024	55934640	52641044
Aroclor-1254-2	(2)	70079706	72686637	78166692	86577056	90012700	79504558
Aroclor-1254-3	(3)	76495408	79340601	84479860	93420496	102844700	87316213
Aroclor-1254-4	(4)	57834344	59926583	63428804	69599176	68258480	63809477
Aroclor-1254-5	(5)	73335985	76078215	80654566	88040632	87614400	81144760
Decachlorobiphenyl		901613940	929059080	979122040	1057269880	1005303000	974473588
Tetrachloro-m-xylene		1055318110	1070331613	1114290760	1189360720	1187072000	1123274641



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION FACTOR OF INITIAL CALIBRATION

Lab Name: Alliance
Lab Code: ACE
Instrument ID: ECD_P

Contract: EARTH03
SDG NO.: Q2818

Calibration Date(s): 08/01/2025 08/01/2025
Calibration Times: 12:05 20:28

GC Column: ZB-MR2 ID: 0.32 (mm)

LAB FILE ID:		CF 1000 = <u>PP074168.D</u>		CF 750 = <u>PP074169.D</u>			
CF 500 = <u>PP074170.D</u>		CF 250 = <u>PP074171.D</u>		CF 050 = <u>PP074172.D</u>			
COMPOUND		CF 1000	CF 750	CF 500	CF 250	CF 050	% RSD
Aroclor-1016-1	(1)	401748995	398824608	399912360	403735356	390215520	398887368
Aroclor-1016-2	(2)	186854426	192495525	187577416	198284052	173833620	187809008
Aroclor-1016-3	(3)	104048217	106580327	106062448	114925596	98731580	106069634
Aroclor-1016-4	(4)	102945349	106576167	111089380	109624088	115865320	109220061
Aroclor-1016-5	(5)	116097268	118066511	118643196	118883084	113126240	116963260
Aroclor-1260-1	(1)	398303416	397444124	405725820	438106428	380206020	403957162
Aroclor-1260-2	(2)	301304557	308158192	308266150	305734868	340030740	312698901
Aroclor-1260-3	(3)	408378449	417603121	404068616	383320072	357221180	394118288
Aroclor-1260-4	(4)	295359710	299954812	306136602	280781712	272663920	290979351
Aroclor-1260-5	(5)	811772813	805780147	794112086	732491376	642892660	757409816
Decachlorobiphenyl		6091085760	6145498480	6064775080	5991469080	5775302800	6013626240
Tetrachloro-m-xylene		4204747790	4209506627	3997167900	3658659200	3102045200	3834425343
Aroclor-1242-1	(1)	341272725	348537165	356175586	358260688	321457320	345140697
Aroclor-1242-2	(2)	162660327	171629959	165033416	168697644	148866940	163377657
Aroclor-1242-3	(3)	93075141	96957855	92088854	97223352	81897380	92248516
Aroclor-1242-4	(4)	117292839	116784915	126377070	120581608	116071680	119421622
Aroclor-1242-5	(5)	140731652	152438257	148354640	155481924	150793560	149560007
Decachlorobiphenyl		6182103300	6198353267	6343202860	6254049920	5638925600	6123326989
Tetrachloro-m-xylene		4380731720	4319851627	4244718880	3977934440	3138312200	4012309773
Aroclor-1254-1	(1)	384557283	361450625	385081202	377494716	322325960	366181957
Aroclor-1254-2	(2)	284389782	276601852	284135592	290728916	255228600	278216948
Aroclor-1254-3	(3)	520239168	520693099	508814756	526662268	369647720	489211402
Aroclor-1254-4	(4)	378053305	389619320	377098978	394351108	298703960	367565334
Aroclor-1254-5	(5)	425051302	410967521	398226716	393575444	298014540	385167105
Decachlorobiphenyl		6206832720	6179384467	6118770460	6279475360	4842174000	5925327401
Tetrachloro-m-xylene		4358828150	4220486587	4040036280	3680489520	2529500800	3765868267



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Lab Name: Alliance Contract: EARTH03

Lab Code: ACE SDG NO.: Q2818

Instrument ID: ECD_P Date(s) Analyzed: 08/01/2025 08/01/2025

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	AMOUNT (ng)	PEAK	RT	RT WINDOW		CALIBRATION FACTOR
				FROM	TO	
Aroclor-1221	500	1	4.86	4.76	4.96	16763600
		2	4.94	4.84	5.04	12601100
		3	5.02	4.92	5.12	36635200
		4	0.00			0
		5	0.00			0
Aroclor-1232	500	1	5.02	4.92	5.12	29037400
		2	5.55	5.45	5.65	15042700
		3	5.83	5.73	5.93	29001800
		4	5.99	5.89	6.09	15468800
		5	6.08	5.98	6.18	11716000
Aroclor-1248	500	1	5.81	5.71	5.91	28382600
		2	6.08	5.98	6.18	40187600
		3	6.28	6.18	6.38	44640800
		4	6.68	6.58	6.78	51582200
		5	6.72	6.62	6.82	53240200
Aroclor-1262	500	1	8.24	8.14	8.34	75789600
		2	8.56	8.46	8.66	130430000
		3	8.89	8.79	8.99	94070800
		4	8.97	8.87	9.07	72851600
		5	9.65	9.55	9.75	50922600
Aroclor-1268	500	1	8.88	8.78	8.98	158028000
		2	8.98	8.88	9.08	146028000
		3	9.22	9.12	9.32	121402000
		4	9.64	9.54	9.74	59648200
		5	10.08	9.98	10.18	357940000



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Lab Name: Alliance **Contract:** EARTH03
Lab Code: ACE **SDG NO.:** Q2818
Instrument ID: ECD_P **Date(s) Analyzed:** 08/01/2025 08/01/2025
GC Column: ZB-MR2 **ID:** 0.32 (mm)

COMPOUND	AMOUNT (ng)	PEAK	RT	RT WINDOW		CALIBRATION FACTOR
				FROM	TO	
Aroclor-1221	500	1	4.02	3.92	4.12	49085200
		2	4.10	4.00	4.20	35889400
		3	4.18	4.08	4.28	135225000
		4	0.00			0
		5	0.00			0
Aroclor-1232	500	1	4.18	4.08	4.28	101362000
		2	4.90	4.80	5.00	174266000
		3	5.08	4.98	5.18	45816600
		4	5.16	5.06	5.26	58010800
		5	5.34	5.24	5.44	46979400
Aroclor-1248	500	1	4.90	4.80	5.00	223628000
		2	5.12	5.02	5.22	147770000
		3	5.16	5.06	5.26	179546000
		4	5.34	5.24	5.44	167057000
		5	5.73	5.63	5.83	302782000
Aroclor-1262	500	1	6.92	6.82	7.02	563142000
		2	7.18	7.08	7.28	401024000
		3	7.70	7.60	7.80	359470000
		4	7.76	7.66	7.86	666922000
		5	8.26	8.16	8.36	283268000
Aroclor-1268	500	1	7.70	7.60	7.80	1169220000
		2	7.76	7.66	7.86	1148130000
		3	7.97	7.87	8.07	891866000
		4	8.26	8.16	8.36	329470000
		5	8.56	8.46	8.66	2772300000

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP080125\
 Data File : PP074168.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 01 Aug 2025 12:05
 Operator : YP\AJ
 Sample : AR1660ICC1000
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1660ICC1000

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 01 13:05:03 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP080125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Aug 01 13:02:12 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.660	3.805	99458766	420.5E6	96.066	102.531
2)	SA Decachlor...	10.437	8.826	85243467	609.1E6	95.091	100.216
Target Compounds							
3)	L1 AR-1016-1	5.812	4.905	35273925	401.7E6	948.391	1002.291
4)	L1 AR-1016-2	5.833	4.963	51795728	186.9E6	940.518	998.069
5)	L1 AR-1016-3	5.896	5.083	33422245	104.0E6	937.613	990.413
6)	L1 AR-1016-4	5.993	5.124	27792815	102.9E6	937.672	961.950
7)	L1 AR-1016-5	6.285	5.338	27690329	116.1E6	936.556	989.154
31)	L7 AR-1260-1	7.403	6.555	47709498	398.3E6	936.840	990.768
32)	L7 AR-1260-2	7.656	6.711	56508230	301.3E6	938.505	988.580
33)	L7 AR-1260-3	8.013	6.921	45067030	408.4E6	941.665	1005.305
34)	L7 AR-1260-4	8.241	7.180	54035441	295.4E6	945.503	982.083
35)	L7 AR-1260-5	8.567	7.419	97992993	811.8E6	952.628	1010.998



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP080125\
Data File : PP074168.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 01 Aug 2025 12:05
Operator : YP\AJ
Sample : AR1660ICC1000
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660ICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 01 13:05:03 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP080125.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Aug 01 13:02:12 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP080125\
 Data File : PP074169.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 01 Aug 2025 12:22
 Operator : YP\AJ
 Sample : AR1660ICC750
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1660ICC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 01 13:07:47 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP080125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Aug 01 13:02:12 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.662	3.804	77897765	315.7E6	75.160	76.312
2) SA Decachlor...	10.440	8.825	67100934	460.9E6	74.902	75.554
Target Compounds						
3) L1 AR-1016-1	5.814	4.904	28067752	299.1E6	753.089	747.493
4) L1 AR-1016-2	5.836	4.962	41399345	144.4E6	751.158	763.969
5) L1 AR-1016-3	5.898	5.082	26764173	79935245	750.553	757.223
6) L1 AR-1016-4	5.995	5.122	22190171	79932125	749.100	747.936
7) L1 AR-1016-5	6.288	5.337	22087671	88549883	748.037	752.960
31) L7 AR-1260-1	7.405	6.554	38166506	298.1E6	749.634	744.294
32) L7 AR-1260-2	7.658	6.709	45125165	231.1E6	749.634	755.513
33) L7 AR-1260-3	8.016	6.919	36092346	313.2E6	752.756	763.877
34) L7 AR-1260-4	8.244	7.178	42799948	225.0E6	749.270	748.680
35) L7 AR-1260-5	8.569	7.417	77207219	604.3E6	750.374	751.765



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP080125\
Data File : PP074169.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 01 Aug 2025 12:22
Operator : YP\AJ
Sample : AR1660ICC750
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660ICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 01 13:07:47 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP080125.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Aug 01 13:02:12 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP080125\
 Data File : PP074170.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 01 Aug 2025 12:38
 Operator : YP\AJ
 Sample : AR1660ICC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1660ICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 01 13:02:29 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP080125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Aug 01 13:02:12 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.660	3.805	53802576	199.9E6	50.000	50.000
2) SA Decachlor...	10.438	8.826	47022042	303.2E6	50.000	50.000
Target Compounds						
3) L1 AR-1016-1	5.812	4.906	19556469	200.0E6	500.000	500.000
4) L1 AR-1016-2	5.833	4.963	29173644	93788708	500.000	500.000
5) L1 AR-1016-3	5.896	5.083	18934969	53031224	500.000	500.000
6) L1 AR-1016-4	5.993	5.123	15743832	55544690	500.000	500.000
7) L1 AR-1016-5	6.286	5.339	15720965	59321598	500.000	500.000
31) L7 AR-1260-1	7.403	6.555	27071230	202.9E6	500.000	500.000
32) L7 AR-1260-2	7.655	6.709	31956798	154.1E6	500.000	500.000
33) L7 AR-1260-3	8.013	6.919	25325364	202.0E6	500.000	500.000
34) L7 AR-1260-4	8.242	7.179	30132251	153.1E6	500.000	500.000
35) L7 AR-1260-5	8.568	7.418	53869434	397.1E6	500.000	500.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP080125\
Data File : PP074170.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 01 Aug 2025 12:38
Operator : YP\AJ
Sample : AR1660ICC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 01 13:02:29 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP080125.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Aug 01 13:02:12 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP080125\
 Data File : PP074171.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 01 Aug 2025 12:54
 Operator : YP\AJ
 Sample : AR1660ICC250
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1660ICC250

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 08/05/2025
 Supervised By :mohammad ahmed 08/08/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 01 13:10:47 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP080125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Aug 01 13:02:12 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.658	3.807	28863110	91466480	27.077	22.767
2) SA Decachlor...	10.435	8.827	25421801	149.8E6	27.450	24.664
Target Compounds						
3) L1 AR-1016-1	5.810	4.907	10964759	100.9E6	281.744	251.671
4) L1 AR-1016-2	5.832	4.965	16172490	49571013	281.221	259.123
5) L1 AR-1016-3	5.895	5.085	10675549	28731399	285.290	266.268
6) L1 AR-1016-4	5.991	5.126	8965609	27406022	287.521	254.800
7) L1 AR-1016-5	6.282	5.340	8901636	29720771	287.291m	252.036
31) L7 AR-1260-1	7.399	6.557	14779955	109.5E6	277.879m	267.207
32) L7 AR-1260-2	7.652	6.711	17850337	76433717	284.356m	249.893
33) L7 AR-1260-3	8.011	6.922	14251387	95830018	283.827	237.590
34) L7 AR-1260-4	8.240	7.182	16666094	70195428	280.066	237.501
35) L7 AR-1260-5	8.565	7.421	29333116	183.1E6	275.424	232.969

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP080125\
Data File : PP074171.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 01 Aug 2025 12:54
Operator : YP\AJ
Sample : AR1660ICC250
Misc :
ALS Vial : 6 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
AR1660ICC250

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/05/2025
Supervised By :mohammad ahmed 08/08/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 01 13:10:47 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP080125.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Aug 01 13:02:12 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP080125\
 Data File : PP074172.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 01 Aug 2025 13:42
 Operator : YP\AJ
 Sample : AR1660ICC050
 Misc :
 ALS Vial : 50 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1660ICC050

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 08/05/2025
 Supervised By :mohammad ahmed 08/08/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 01 13:58:51 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP080125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Aug 01 13:58:39 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.660	3.804	6638630	15510226	5.935m	4.043m#
2) SA Decachlor...	10.437	8.827	5756903	28876514	5.928	4.802
Target Compounds						
3) L1 AR-1016-1	5.813	4.906	2530723	19510776	61.341	48.913
4) L1 AR-1016-2	5.835	4.966	3683827	8691681	60.647	46.279
5) L1 AR-1016-3	5.897	5.084	2431131	4936579	61.299	46.541
6) L1 AR-1016-4	5.994	5.124	1899628	5793266	58.370	53.042
7) L1 AR-1016-5	6.287	5.341	1907560	5656312	58.752	48.360
31) L7 AR-1260-1	7.404	6.556	3495993	19010301	62.034	47.060
32) L7 AR-1260-2	7.656	6.709	4438588	17001537	64.615m	54.991m
33) L7 AR-1260-3	8.015	6.920	3281661	17861059	61.574	45.319 #
34) L7 AR-1260-4	8.242	7.180	3752198	13633196	59.829m	46.853
35) L7 AR-1260-5	8.568	7.418	7025863	32144633	62.008	42.440 #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP080125\
Data File : PP074172.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 01 Aug 2025 13:42
Operator : YP\AJ
Sample : AR1660ICC050
Misc :
ALS Vial : 50 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
AR1660ICC050

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/05/2025
Supervised By :mohammad ahmed 08/08/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 01 13:58:51 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP080125.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Aug 01 13:58:39 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP080125\
 Data File : PP074173.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 01 Aug 2025 13:58
 Operator : YP\AJ
 Sample : AR1221ICC500
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1221ICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 01 15:14:09 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP080125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Aug 01 15:13:54 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.658	3.803	54821028	195.8E6	50.000	50.000
2) SA Decachlor...	10.435	8.825	46278402	298.9E6	50.000	50.000
Target Compounds						
8) L2 AR-1221-1	4.858	4.015	8381808	24542587	500.000	500.000
9) L2 AR-1221-2	4.944	4.100	6300527	17944660	500.000	500.000
10) L2 AR-1221-3	5.019	4.176	18317605	67612277	500.000	500.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP080125\
Data File : PP074173.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 01 Aug 2025 13:58
Operator : YP\AJ
Sample : AR1221ICC500
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1221ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 01 15:14:09 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP080125.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Aug 01 15:13:54 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP080125\
 Data File : PP074174.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 01 Aug 2025 14:15
 Operator : YP\AJ
 Sample : AR1232ICC500
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1232ICC500

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 08/05/2025
 Supervised By :mohammad ahmed 08/08/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 01 15:17:09 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP080125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Aug 01 15:13:54 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.660	3.804	54684846	201.8E6	50.000	50.000
2) SA Decachlor...	10.437	8.826	47117835	301.0E6	50.000	50.000
Target Compounds						
11) L3 AR-1232-1	5.020	4.177	14518661	50681220	500.000	500.000
12) L3 AR-1232-2	5.546	4.904	7521357	87133099	500.000	500.000
13) L3 AR-1232-3	5.832	5.081	14500854	22908311	500.000	500.000
14) L3 AR-1232-4	5.993	5.164	7734413	29005356	500.000	502.399m
15) L3 AR-1232-5	6.082	5.338	5858006	23489726	500.000	500.000

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP080125\
Data File : PP074174.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 01 Aug 2025 14:15
Operator : YP\AJ
Sample : AR1232ICC500
Misc :
ALS Vial : 9 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
AR1232ICC500

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/05/2025
Supervised By :mohammad ahmed 08/08/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 01 15:17:09 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP080125.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Aug 01 15:13:54 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP080125\
 Data File : PP074175.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 01 Aug 2025 14:31
 Operator : YP\AJ
 Sample : AR1242ICC1000
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1242ICC1000

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 08/05/2025
 Supervised By :mohammad ahmed 08/08/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 01 15:23:53 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP080125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Aug 01 15:20:45 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.659	3.804	104.3E6	438.1E6	95.872	101.577
2)	SA Decachlor...	10.436	8.825	88078600	618.2E6	94.444	98.714
Target Compounds							
16)	L4 AR-1242-1	5.811	4.905	31512428	341.3E6	934.658	978.632
17)	L4 AR-1242-2	5.832	4.962	47421171	162.7E6	943.850	992.758
18)	L4 AR-1242-3	5.896	5.082	30290298	93075141	931.794	1005.327
19)	L4 AR-1242-4	5.992	5.164	25177060	117.3E6	934.754	940.884m
20)	L4 AR-1242-5	6.721	5.686	29128304	140.7E6	916.924	973.631

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP080125\
Data File : PP074175.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 01 Aug 2025 14:31
Operator : YP\AJ
Sample : AR1242ICC1000
Misc :
ALS Vial : 10 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
AR1242ICC1000

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/05/2025
Supervised By :mohammad ahmed 08/08/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 01 15:23:53 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP080125.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Aug 01 15:20:45 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP080125\
 Data File : PP074176.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 01 Aug 2025 14:47
 Operator : YP\AJ
 Sample : AR1242ICC750
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1242ICC750

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 08/05/2025
 Supervised By :mohammad ahmed 08/08/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 01 15:26:53 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP080125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Aug 01 15:20:45 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.662	3.803	80676637	324.0E6	74.425	75.083
2)	SA Decachlor...	10.438	8.824	68962955	464.9E6	74.295	74.485
Target Compounds							
16)	L4 AR-1242-1	5.813	4.904	25064705	261.4E6	745.599	749.732
17)	L4 AR-1242-2	5.835	4.960	37017188	128.7E6	741.131	773.381
18)	L4 AR-1242-3	5.898	5.081	24057234	72718391	743.338	773.266
19)	L4 AR-1242-4	5.995	5.163	19974989	87588686	744.389	713.205m
20)	L4 AR-1242-5	6.724	5.685	24065795	114.3E6	755.025	776.822

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP080125\
Data File : PP074176.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 01 Aug 2025 14:47
Operator : YP\AJ
Sample : AR1242ICC750
Misc :
ALS Vial : 11 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
AR1242ICC750

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/05/2025
Supervised By :mohammad ahmed 08/08/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 01 15:26:53 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP080125.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Aug 01 15:20:45 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP080125\
 Data File : PP074177.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 01 Aug 2025 15:03
 Operator : YP\AJ
 Sample : AR1242ICC500
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1242ICC500

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 08/05/2025
 Supervised By :mohammad ahmed 08/08/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 01 15:21:01 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP080125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Aug 01 15:20:45 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.657	3.805	56653742	212.2E6	50.000	50.000
2)	SA Decachlor...	10.434	8.826	49220928	317.2E6	50.000	50.000
Target Compounds							
16)	L4 AR-1242-1	5.808	4.905	17959256	178.1E6	500.000	500.000
17)	L4 AR-1242-2	5.830	4.963	26531667	82516708	500.000	500.000
18)	L4 AR-1242-3	5.893	5.084	17362372	46044427	500.000	500.000
19)	L4 AR-1242-4	5.990	5.164	14345906	63188535	500.000	505.329m
20)	L4 AR-1242-5	6.719	5.687	17203244	74177320	500.000	500.000

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP080125\
Data File : PP074177.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 01 Aug 2025 15:03
Operator : YP\AJ
Sample : AR1242ICC500
Misc :
ALS Vial : 12 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
AR1242ICC500

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/05/2025
Supervised By :mohammad ahmed 08/08/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 01 15:21:01 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP080125.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Aug 01 15:20:45 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP080125\
 Data File : PP074178.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 01 Aug 2025 15:19
 Operator : YP\AJ
 Sample : AR1242ICC250
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1242ICC250

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 08/05/2025
 Supervised By :mohammad ahmed 08/08/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 01 15:30:20 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP080125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Aug 01 15:20:45 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.662	3.805	30064825	99448361	26.997	23.506
2) SA Decachlor...	10.436	8.824	26416078	156.4E6	27.507	25.039
Target Compounds						
16) L4 AR-1242-1	5.812	4.906	9924005	89565172	282.440	255.127
17) L4 AR-1242-2	5.834	4.964	14472673	42174411	278.680	252.533
18) L4 AR-1242-3	5.897	5.084	9547202	24305838	282.294	256.293
19) L4 AR-1242-4	5.994	5.165	7737538	30145402	277.699	249.527m
20) L4 AR-1242-5	6.721	5.687	9674193	38870481	289.648m	260.436

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP080125\
Data File : PP074178.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 01 Aug 2025 15:19
Operator : YP\AJ
Sample : AR1242ICC250
Misc :
ALS Vial : 13 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
AR1242ICC250

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/05/2025
Supervised By :mohammad ahmed 08/08/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 01 15:30:20 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP080125.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Aug 01 15:20:45 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP080125\
 Data File : PP074179.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 01 Aug 2025 15:36
 Operator : YP\AJ
 Sample : AR1242ICC050
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1242ICC050

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 08/05/2025
 Supervised By :mohammad ahmed 08/08/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 01 15:58:06 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP080125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Aug 01 15:57:36 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.657	3.802	6853903	15691561	5.892m	3.915m#
2)	SA Decachlor...	10.432	8.821	6026260	28194628	5.971	4.604
Target Compounds							
16)	L4 AR-1242-1	5.810	4.903	2289125	16072866	61.427	46.569
17)	L4 AR-1242-2	5.832	4.962	3215660	7443347	59.102	45.559
18)	L4 AR-1242-3	5.895	5.082	2169957	4094869	60.722	44.390 #
19)	L4 AR-1242-4	5.991	5.163	1732044	5803584	59.279	48.597
20)	L4 AR-1242-5	6.720	5.686	2173540	7539678	56.364m	50.412

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP080125\
Data File : PP074179.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 01 Aug 2025 15:36
Operator : YP\AJ
Sample : AR1242ICC050
Misc :
ALS Vial : 14 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
AR1242ICC050

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/05/2025
Supervised By :mohammad ahmed 08/08/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 01 15:58:06 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP080125.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Aug 01 15:57:36 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP080125\
 Data File : PP074182.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 01 Aug 2025 16:57
 Operator : YP\AJ
 Sample : AR1248ICC500
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1248ICC500

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 08/05/2025
 Supervised By :mohammad ahmed 08/08/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 01 23:35:19 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP080125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Aug 01 23:33:57 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.660	3.805	54922987	197.2E6	50.000	50.000
2) SA Decachlor...	10.433	8.824	48401219	304.7E6	50.000	50.000
Target Compounds						
21) L5 AR-1248-1	5.810	4.903	14191296	111.8E6	500.000	500.000
22) L5 AR-1248-2	6.082	5.124	20093773	73885161	500.000	500.000
23) L5 AR-1248-3	6.284	5.164	22320416	89773071	500.000	477.333m
24) L5 AR-1248-4	6.682	5.338	25791080	83528426	500.000	500.000
25) L5 AR-1248-5	6.721	5.727	26620059	151.4E6	500.000	500.000

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP080125\
Data File : PP074182.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 01 Aug 2025 16:57
Operator : YP\AJ
Sample : AR1248ICC500
Misc :
ALS Vial : 17 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
AR1248ICC500

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/05/2025
Supervised By :mohammad ahmed 08/08/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 01 23:35:19 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP080125.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Aug 01 23:33:57 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP080125\
 Data File : PP074185.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 01 Aug 2025 18:02
 Operator : YP\AJ
 Sample : AR1254ICC1000
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1254ICC1000

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 08/05/2025
 Supervised By :mohammad ahmed 08/08/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 02 00:47:02 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP080125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Aug 02 00:46:23 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.660	3.805	105.5E6	435.9E6	94.708	107.891
2) SA Decachlor...	10.433	8.824	90161394	620.7E6	92.084	101.439
Target Compounds						
26) L6 AR-1254-1	6.658	5.689	49559228	384.6E6	913.982m	998.639
27) L6 AR-1254-2	6.875	5.835	70079706	284.4E6	896.542	1000.895
28) L6 AR-1254-3	7.237	6.238	76495408	520.2E6	905.487	1022.453
29) L6 AR-1254-4	7.518	6.464	57834344	378.1E6	911.799	1002.531
30) L6 AR-1254-5	7.934	6.880	73335985	425.1E6	909.260	1067.360m

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP080125\
Data File : PP074185.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 01 Aug 2025 18:02
Operator : YP\AJ
Sample : AR1254ICC1000
Misc :
ALS Vial : 20 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
AR1254ICC1000

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/05/2025
Supervised By :mohammad ahmed 08/08/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 02 00:47:02 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP080125.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Aug 02 00:46:23 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP080125\
 Data File : PP074186.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 01 Aug 2025 18:18
 Operator : YP\AJ
 Sample : AR1254ICC750
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1254ICC750

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 08/05/2025
 Supervised By :mohammad ahmed 08/08/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 02 00:47:20 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP080125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Aug 02 00:46:23 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.659	3.804	80274871	316.5E6	72.041	78.350
2)	SA Decachlor...	10.433	8.823	69679431	463.5E6	71.165	75.743
Target Compounds							
26)	L6 AR-1254-1	6.657	5.688	37358301	271.1E6	688.970m	703.976
27)	L6 AR-1254-2	6.874	5.835	54514978	207.5E6	697.420	730.114
28)	L6 AR-1254-3	7.236	6.237	59505451	390.5E6	704.374	767.509
29)	L6 AR-1254-4	7.517	6.464	44944937	292.2E6	708.589	774.901
30)	L6 AR-1254-5	7.933	6.881	57058661	308.2E6	707.445	773.995

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP080125\
Data File : PP074186.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 01 Aug 2025 18:18
Operator : YP\AJ
Sample : AR1254ICC750
Misc :
ALS Vial : 21 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
AR1254ICC750

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/05/2025
Supervised By :mohammad ahmed 08/08/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 02 00:47:20 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP080125.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Aug 02 00:46:23 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP080125\
 Data File : PP074187.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 01 Aug 2025 18:34
 Operator : YP\AJ
 Sample : AR1254ICC500
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1254ICC500

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 08/05/2025
 Supervised By :mohammad ahmed 08/08/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 02 00:47:38 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP080125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Aug 02 00:46:23 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.658	3.803	55714538	202.0E6	50.000	50.000
2) SA Decachlor...	10.432	8.822	48956102	305.9E6	50.000	50.000
Target Compounds						
26) L6 AR-1254-1	6.657	5.687	26953630	192.5E6	497.085m	500.000
27) L6 AR-1254-2	6.874	5.834	39083346	142.1E6	500.000	500.000
28) L6 AR-1254-3	7.236	6.237	42239930	254.4E6	500.000	500.000
29) L6 AR-1254-4	7.517	6.463	31714402	188.5E6	500.000	500.000
30) L6 AR-1254-5	7.933	6.880	40327283	199.1E6	500.000	500.000

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP080125\
Data File : PP074187.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 01 Aug 2025 18:34
Operator : YP\AJ
Sample : AR1254ICC500
Misc :
ALS Vial : 22 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
AR1254ICC500

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/05/2025
Supervised By :mohammad ahmed 08/08/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 02 00:47:38 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP080125.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Aug 02 00:46:23 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP080125\
 Data File : PP074188.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 01 Aug 2025 18:50
 Operator : YP\AJ
 Sample : AR1254ICC250
 Misc :
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1254ICC250

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 08/05/2025
 Supervised By :mohammad ahmed 08/08/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 02 00:47:59 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP080125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Aug 02 00:46:23 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.658	3.803	29734018	92012238	26.684	22.775
2) SA Decachlor...	10.431	8.821	26431747	157.0E6	26.995	25.657
Target Compounds						
26) L6 AR-1254-1	6.656	5.687	13498256	94373679	248.938m	245.075
27) L6 AR-1254-2	6.873	5.834	21644264	72682229	276.899	255.801
28) L6 AR-1254-3	7.235	6.236	23355124	131.7E6	276.458	258.769
29) L6 AR-1254-4	7.516	6.462	17399794	98587777	274.320	261.437
30) L6 AR-1254-5	7.932	6.879	22010158	98393861	272.894	247.080

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP080125\
Data File : PP074188.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 01 Aug 2025 18:50
Operator : YP\AJ
Sample : AR1254ICC250
Misc :
ALS Vial : 23 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
AR1254ICC250

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/05/2025
Supervised By :mohammad ahmed 08/08/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 02 00:47:59 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP080125.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Aug 02 00:46:23 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP080125\
 Data File : PP074189.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 01 Aug 2025 19:23
 Operator : YP\AJ
 Sample : AR1254ICC050
 Misc :
 ALS Vial : 48 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1254ICC050

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 08/05/2025
 Supervised By :mohammad ahmed 08/08/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 02 01:03:58 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP080125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Aug 02 01:03:01 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.657	3.804	5935360	12647504	5.284	3.358 #
2)	SA Decachlor...	10.430	8.823	5026515	24210870	5.158	4.086
Target Compounds							
26)	L6 AR-1254-1	6.655	5.689	2796732	16116298	51.810m	44.012
27)	L6 AR-1254-2	6.872	5.836	4500635	12761430	56.609	45.869
28)	L6 AR-1254-3	7.234	6.237	5142235	18482386	58.892	37.780 #
29)	L6 AR-1254-4	7.515	6.464	3412924	14935198	53.486	40.633
30)	L6 AR-1254-5	7.931	6.881	4380720	14900727	53.986	37.906 #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP080125\
Data File : PP074189.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 01 Aug 2025 19:23
Operator : YP\AJ
Sample : AR1254ICC050
Misc :
ALS Vial : 48 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
AR1254ICC050

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/05/2025
Supervised By :mohammad ahmed 08/08/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 02 01:03:58 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP080125.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Aug 02 01:03:01 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP080125\
 Data File : PP074190.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 01 Aug 2025 19:39
 Operator : YP\AJ
 Sample : AR1262ICC500
 Misc :
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1262ICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 02 01:10:39 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP080125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Aug 02 01:10:03 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.657	3.802	54354703	199.7E6	50.000	50.000
2) SA Decachlor...	10.430	8.820	48321379	306.6E6	50.000	50.000
Target Compounds						
36) L8 AR-1262-1	8.237	6.917	37894848	281.6E6	500.000	500.000
37) L8 AR-1262-2	8.562	7.176	65215013	200.5E6	500.000	500.000
38) L8 AR-1262-3	8.888	7.698	47035440	179.7E6	500.000	500.000
39) L8 AR-1262-4	8.974	7.761	36425846	333.5E6	500.000	500.000
40) L8 AR-1262-5	9.645	8.255	25461320	141.6E6	500.000	500.000


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP080125\
Data File : PP074190.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 01 Aug 2025 19:39
Operator : YP\AJ
Sample : AR1262ICC500
Misc :
ALS Vial : 25 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1262ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 02 01:10:39 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP080125.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Aug 02 01:10:03 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP080125\
 Data File : PP074193.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 01 Aug 2025 20:28
 Operator : YP\AJ
 Sample : AR1268ICC500
 Misc :
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1268ICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 02 01:15:46 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP080125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Aug 02 01:13:43 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.656	3.801	58418637	219.4E6	50.000	50.000
2) SA Decachlor...	10.429	8.820	86270477	592.0E6	50.000	50.000
Target Compounds						
41) L9 AR-1268-1	8.883	7.698	79013824	584.6E6	500.000	500.000
42) L9 AR-1268-2	8.978	7.762	73014004	574.1E6	500.000	500.000
43) L9 AR-1268-3	9.218	7.970	60700868	445.9E6	500.000	500.000
44) L9 AR-1268-4	9.644	8.255	29824069	164.7E6	500.000	500.000
45) L9 AR-1268-5	10.075	8.556	179.0E6	1386.1E6	500.000	500.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP080125\
Data File : PP074193.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 01 Aug 2025 20:28
Operator : YP\AJ
Sample : AR1268ICC500
Misc :
ALS Vial : 28 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1268ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 02 01:15:46 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP080125.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Aug 02 01:13:43 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP080125\
 Data File : PP074196.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 01 Aug 2025 21:16
 Operator : YP\AJ
 Sample : PP080125ICV500
 Misc :
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 ICVPP080125

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 02 01:35:54 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP080125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Aug 02 01:33:31 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.659	3.806	53809613	192.5E6	48.117	50.193
2) SA Decachlor...	10.433	8.823	46918201	290.2E6	48.311	48.263
Target Compounds						
3) L1 AR-1016-1	5.810	4.906	19711119	191.3E6	477.766	479.651
4) L1 AR-1016-2	5.831	4.963	29290082	90085682	482.207	479.666
5) L1 AR-1016-3	5.894	5.083	19171615	49893167	483.393	470.381
6) L1 AR-1016-4	5.991	5.124	15834175	51335124	486.540	470.016
7) L1 AR-1016-5	6.283	5.338	15788703	57011251	486.284	487.429
31) L7 AR-1260-1	7.400	6.554	27159073	192.5E6	481.919	476.633
32) L7 AR-1260-2	7.653	6.708	31876376	149.4E6	467.722	477.644
33) L7 AR-1260-3	8.010	6.918	25282960	196.9E6	474.388	499.492
34) L7 AR-1260-4	8.239	7.178	29964985	140.7E6	478.559	483.661
35) L7 AR-1260-5	8.563	7.417	53436862	370.6E6	471.620	489.248



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP080125\
Data File : PP074196.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 01 Aug 2025 21:16
Operator : YP\AJ
Sample : PP080125ICV500
Misc :
ALS Vial : 31 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
ICVPP080125

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 02 01:35:54 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP080125.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Aug 02 01:33:31 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP080125\
 Data File : PP074197.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 01 Aug 2025 22:05
 Operator : YP\AJ
 Sample : AR1242ICV500
 Misc :
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 ICVPP080125AR1242

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 08/05/2025
 Supervised By :mohammad ahmed 08/08/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 02 01:36:23 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP080125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Aug 02 01:33:31 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.655	3.804	55143327	198.9E6	49.310	51.866
2) SA Decachlor...	10.427	8.822	47800939	295.6E6	49.220	49.156
Target Compounds						
16) L4 AR-1242-1	5.806	4.904	17499270	167.4E6	469.580	484.954
17) L4 AR-1242-2	5.827	4.962	25894594	77861664	475.925	476.575
18) L4 AR-1242-3	5.890	5.082	17050061	43676713	477.113	473.468
19) L4 AR-1242-4	5.987	5.164	14124012	56465948	483.390	472.829m
20) L4 AR-1242-5	6.716	5.686	16823190	70286035	473.119	469.952

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP080125\
Data File : PP074197.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 01 Aug 2025 22:05
Operator : YP\AJ
Sample : AR1242ICV500
Misc :
ALS Vial : 32 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
ICVPP080125AR1242

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/05/2025
Supervised By :mohammad ahmed 08/08/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 02 01:36:23 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP080125.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Aug 02 01:33:31 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP080125\
 Data File : PP074198.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 01 Aug 2025 22:37
 Operator : YP\AJ
 Sample : AR1248ICV500
 Misc :
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 ICVPP080125AR1248

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 08/05/2025
 Supervised By :mohammad ahmed 08/08/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 02 01:36:51 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP080125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Aug 02 01:33:31 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.655	3.805	53554793	191.1E6	47.889	49.847
2) SA Decachlor...	10.428	8.824	47537298	295.3E6	48.949	49.101
Target Compounds						
21) L5 AR-1248-1	5.806	4.904	13951339	107.6E6	498.004	503.095
22) L5 AR-1248-2	6.077	5.124	19721117	71900356	489.831	491.416
23) L5 AR-1248-3	6.280	5.165	21866877	85869463	490.580	500.813m
24) L5 AR-1248-4	6.677	5.338	24961771	82403417	479.691	495.291
25) L5 AR-1248-5	6.717	5.728	25702187	138.2E6	451.365	474.938

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP080125\
Data File : PP074198.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 01 Aug 2025 22:37
Operator : YP\AJ
Sample : AR1248ICV500
Misc :
ALS Vial : 33 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
ICVPP080125AR1248

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/05/2025
Supervised By :mohammad ahmed 08/08/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 02 01:36:51 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP080125.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Aug 02 01:33:31 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP080125\
 Data File : PP074199.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 01 Aug 2025 23:10
 Operator : YP\AJ
 Sample : AR1254ICV500
 Misc :
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 ICVPP080125AR1254

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 08/08/2025
 Supervised By :mohammad ahmed 08/08/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 04 11:04:59 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP080125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Aug 04 11:01:49 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.656	3.803	53831323	191.6E6	48.137	49.971
2) SA Decachlor...	10.430	8.820	47450930	299.2E6	48.860	49.761
Target Compounds						
26) L6 AR-1254-1	6.654	5.686	27290919	178.5E6	518.434m	487.493
27) L6 AR-1254-2	6.871	5.833	37599140	134.8E6	472.918	484.650
28) L6 AR-1254-3	7.233	6.235	40597005	245.2E6	464.942	501.115
29) L6 AR-1254-4	7.514	6.462	30708926	181.0E6	481.260	492.525
30) L6 AR-1254-5	7.930	6.878	38809795	188.2E6	478.279	488.601

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP080125\
Data File : PP074199.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 01 Aug 2025 23:10
Operator : YP\AJ
Sample : AR1254ICV500
Misc :
ALS Vial : 34 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
ICVPP080125AR1254

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/08/2025
Supervised By :mohammad ahmed 08/08/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 04 11:04:59 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP080125.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Aug 04 11:01:49 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP080125\
 Data File : PP074200.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 01 Aug 2025 23:42
 Operator : YP\AJ
 Sample : AR1268ICV500
 Misc :
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 ICVPP080125AR1268

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 02 01:25:27 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP080125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Aug 02 01:23:25 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.654	3.804	55201921	199.0E6	48.099	50.981
2) SA Decachlor...	10.426	8.822	82544950	560.4E6	48.605	50.452
Target Compounds						
41) L9 AR-1268-1	8.879	7.701	75022320	549.0E6	482.047	503.522
42) L9 AR-1268-2	8.975	7.764	69343438	540.7E6	483.480	509.488
43) L9 AR-1268-3	9.215	7.973	57914901	420.0E6	484.752	504.814
44) L9 AR-1268-4	9.641	8.257	28601597	156.0E6	486.318	506.427
45) L9 AR-1268-5	10.071	8.557	170.0E6	1304.3E6	483.271	511.465



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP080125\
Data File : PP074200.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 01 Aug 2025 23:42
Operator : YP\AJ
Sample : AR1268ICV500
Misc :
ALS Vial : 35 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
ICVPP080125AR1268

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 02 01:25:27 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP080125.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Aug 02 01:23:25 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm





284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: EARTH03

Lab Code: ACE

SDG NO.: Q2818

Continuing Calib Date: 08/12/2025

Initial Calibration Date(s): 07/23/2025

07/23/2025

Continuing Calib Time: 08:54

Initial Calibration Time(s): 11:32

19:47

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
Aroclor-1016-1	(1)	4.75	4.76	4.66	4.86	0.01
Aroclor-1016-2	(2)	4.77	4.77	4.67	4.87	0.00
Aroclor-1016-3	(3)	4.83	4.83	4.73	4.93	0.00
Aroclor-1016-4	(4)	4.95	4.95	4.85	5.05	0.00
Aroclor-1016-5	(5)	5.20	5.21	5.11	5.31	0.01
Aroclor-1260-1	(1)	6.24	6.25	6.15	6.35	0.01
Aroclor-1260-2	(2)	6.43	6.43	6.33	6.53	0.00
Aroclor-1260-3	(3)	6.79	6.80	6.70	6.90	0.01
Aroclor-1260-4	(4)	7.05	7.06	6.96	7.16	0.01
Aroclor-1260-5	(5)	7.30	7.30	7.20	7.40	0.00
Tetrachloro-m-xylene		3.67	3.67	3.57	3.77	0.01
Decachlorobiphenyl		8.69	8.69	8.59	8.79	0.00



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: EARTH03

Lab Code: ACE

SDG NO.: Q2818

Continuing Calib Date: 08/12/2025

Initial Calibration Date(s): 07/23/2025 07/23/2025

Continuing Calib Time: 08:54

Initial Calibration Time(s): 11:32 19:47

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	4.73	4.74	4.64	4.84	0.01
Aroclor-1016-2	(2)	4.75	4.76	4.66	4.86	0.01
Aroclor-1016-3	(3)	4.92	4.93	4.83	5.03	0.01
Aroclor-1016-4	(4)	4.97	4.97	4.87	5.07	0.00
Aroclor-1016-5	(5)	5.18	5.19	5.09	5.29	0.01
Aroclor-1260-1	(1)	6.21	6.21	6.11	6.31	0.00
Aroclor-1260-2	(2)	6.39	6.40	6.30	6.50	0.01
Aroclor-1260-3	(3)	6.55	6.55	6.45	6.65	0.00
Aroclor-1260-4	(4)	7.02	7.02	6.92	7.12	0.01
Aroclor-1260-5	(5)	7.26	7.26	7.16	7.36	0.00
Tetrachloro-m-xylene		3.66	3.66	3.56	3.76	0.00
Decachlorobiphenyl		8.63	8.64	8.54	8.74	0.01



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** EARTH03
Lab Code: ACE **SDG NO.:** Q2818
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/23/2025 07/23/2025

Client Sample No.: CCAL01 **Date Analyzed:** 08/12/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO112850.D **Time Analyzed:** 08:54

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.750	4.655	4.855	651.000	500.000	30.2
Aroclor-1016-2	4.769	4.674	4.874	549.480	500.000	9.9
Aroclor-1016-3	4.826	4.731	4.931	565.780	500.000	13.2
Aroclor-1016-4	4.945	4.851	5.051	571.540	500.000	14.3
Aroclor-1016-5	5.202	5.108	5.308	574.860	500.000	15.0
Aroclor-1260-1	6.238	6.145	6.345	579.220	500.000	15.8
Aroclor-1260-2	6.428	6.333	6.533	547.500	500.000	9.5
Aroclor-1260-3	6.794	6.700	6.900	528.750	500.000	5.8
Aroclor-1260-4	7.053	6.959	7.159	498.850	500.000	-0.2
Aroclor-1260-5	7.296	7.203	7.403	553.130	500.000	10.6
Decachlorobiphenyl	8.686	8.593	8.793	46.550	50.000	-6.9
Tetrachloro-m-xylene	3.665	3.569	3.769	60.080	50.000	20.2



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** EARTH03
Lab Code: ACE **SDG NO.:** Q2818
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/23/2025 07/23/2025

Client Sample No.: CCAL01 **Date Analyzed:** 08/12/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO112850.D **Time Analyzed:** 08:54

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.731	4.638	4.838	532.580	500.000	6.5
Aroclor-1016-2	4.749	4.656	4.856	524.330	500.000	4.9
Aroclor-1016-3	4.924	4.831	5.031	531.470	500.000	6.3
Aroclor-1016-4	4.966	4.873	5.073	529.880	500.000	6.0
Aroclor-1016-5	5.179	5.085	5.285	523.970	500.000	4.8
Aroclor-1260-1	6.206	6.113	6.313	535.240	500.000	7.0
Aroclor-1260-2	6.394	6.302	6.502	538.810	500.000	7.8
Aroclor-1260-3	6.545	6.453	6.653	559.410	500.000	11.9
Aroclor-1260-4	7.015	6.923	7.123	530.130	500.000	6.0
Aroclor-1260-5	7.257	7.164	7.364	549.900	500.000	10.0
Decachlorobiphenyl	8.631	8.540	8.740	52.550	50.000	5.1
Tetrachloro-m-xylene	3.658	3.562	3.762	56.390	50.000	12.8

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0081225\
 Data File : PO112850.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 12 Aug 2025 08:54
 Operator : YP/AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1660CCC500

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 08/13/2025
 Supervised By :mohammad ahmed 08/14/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 12 11:01:49 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0072325.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Jul 24 04:54:06 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.665	3.658	488.6E6	280.2E6	60.081	56.386
2) SA Decachlor...	8.686	8.631	340.4E6	92854815	46.549	52.546
Target Compounds						
3) L1 AR-1016-1	4.750	4.731	174.7E6	93205664	650.999m	532.582
4) L1 AR-1016-2	4.769	4.749	223.2E6	136.6E6	549.476m	524.330
5) L1 AR-1016-3	4.826	4.924	148.1E6	72262042	565.785	531.470m
6) L1 AR-1016-4	4.945	4.966	121.1E6	58861848	571.539	529.882m
7) L1 AR-1016-5	5.202	5.179	126.2E6	75085975	574.857	523.966
31) L7 AR-1260-1	6.238	6.206	254.5E6	133.0E6	579.224	535.236
32) L7 AR-1260-2	6.428	6.394	376.6E6	177.5E6	547.504	538.814
33) L7 AR-1260-3	6.794	6.545	310.9E6	143.7E6	528.746	559.414m
34) L7 AR-1260-4	7.053	7.015	225.0E6	101.4E6	498.852	530.125
35) L7 AR-1260-5	7.296	7.257	685.4E6	226.7E6	553.135	549.898

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO081225\
Data File : PO112850.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Aug 2025 08:54
Operator : YP/AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
AR1660CCC500

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/13/2025
Supervised By :mohammad ahmed 08/14/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 12 11:01:49 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO072325.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Jul 24 04:54:06 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm





284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: EARTH03

Lab Code: ACE

SDG NO.: Q2818

Continuing Calib Date: 08/12/2025

Initial Calibration Date(s): 07/23/2025 07/23/2025

Continuing Calib Time: 15:50

Initial Calibration Time(s): 11:32 19:47

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	4.75	4.76	4.66	4.86	0.01
Aroclor-1016-2	(2)	4.77	4.77	4.67	4.87	0.00
Aroclor-1016-3	(3)	4.83	4.83	4.73	4.93	0.00
Aroclor-1016-4	(4)	4.95	4.95	4.85	5.05	0.00
Aroclor-1016-5	(5)	5.20	5.21	5.11	5.31	0.01
Aroclor-1260-1	(1)	6.24	6.25	6.15	6.35	0.01
Aroclor-1260-2	(2)	6.43	6.43	6.33	6.53	0.00
Aroclor-1260-3	(3)	6.80	6.80	6.70	6.90	0.00
Aroclor-1260-4	(4)	7.06	7.06	6.96	7.16	0.01
Aroclor-1260-5	(5)	7.30	7.30	7.20	7.40	0.00
Tetrachloro-m-xylene		3.67	3.67	3.57	3.77	0.00
Decachlorobiphenyl		8.69	8.69	8.59	8.79	0.00



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: EARTH03

Lab Code: ACE

SDG NO.: Q2818

Continuing Calib Date: 08/12/2025

Initial Calibration Date(s): 07/23/2025

07/23/2025

Continuing Calib Time: 15:50

Initial Calibration Time(s): 11:32

19:47

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	4.73	4.74	4.64	4.84	0.01
Aroclor-1016-2	(2)	4.75	4.76	4.66	4.86	0.01
Aroclor-1016-3	(3)	4.93	4.93	4.83	5.03	0.00
Aroclor-1016-4	(4)	4.97	4.97	4.87	5.07	0.00
Aroclor-1016-5	(5)	5.18	5.19	5.09	5.29	0.01
Aroclor-1260-1	(1)	6.21	6.21	6.11	6.31	0.00
Aroclor-1260-2	(2)	6.40	6.40	6.30	6.50	0.00
Aroclor-1260-3	(3)	6.55	6.55	6.45	6.65	0.00
Aroclor-1260-4	(4)	7.02	7.02	6.92	7.12	0.00
Aroclor-1260-5	(5)	7.26	7.26	7.16	7.36	0.00
Tetrachloro-m-xylene		3.66	3.66	3.56	3.76	0.00
Decachlorobiphenyl		8.63	8.64	8.54	8.74	0.01



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** EARTH03
Lab Code: ACE **SDG NO.:** Q2818
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/23/2025 07/23/2025

Client Sample No.: CCAL02 **Date Analyzed:** 08/12/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO112865.D **Time Analyzed:** 15:50

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.751	4.655	4.855	524.640	500.000	4.9
Aroclor-1016-2	4.769	4.674	4.874	553.430	500.000	10.7
Aroclor-1016-3	4.827	4.731	4.931	551.220	500.000	10.2
Aroclor-1016-4	4.947	4.851	5.051	550.720	500.000	10.1
Aroclor-1016-5	5.203	5.108	5.308	584.150	500.000	16.8
Aroclor-1260-1	6.240	6.145	6.345	550.350	500.000	10.1
Aroclor-1260-2	6.430	6.333	6.533	543.410	500.000	8.7
Aroclor-1260-3	6.795	6.700	6.900	570.950	500.000	14.2
Aroclor-1260-4	7.055	6.959	7.159	542.260	500.000	8.5
Aroclor-1260-5	7.297	7.203	7.403	529.970	500.000	6.0
Decachlorobiphenyl	8.687	8.593	8.793	48.970	50.000	-2.1
Tetrachloro-m-xylene	3.667	3.569	3.769	57.760	50.000	15.5



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** EARTH03
Lab Code: ACE **SDG NO.:** Q2818
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/23/2025 07/23/2025

Client Sample No.: CCAL02 **Date Analyzed:** 08/12/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO112865.D **Time Analyzed:** 15:50

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.734	4.638	4.838	522.080	500.000	4.4
Aroclor-1016-2	4.751	4.656	4.856	517.090	500.000	3.4
Aroclor-1016-3	4.926	4.831	5.031	516.260	500.000	3.3
Aroclor-1016-4	4.969	4.873	5.073	515.840	500.000	3.2
Aroclor-1016-5	5.181	5.085	5.285	527.110	500.000	5.4
Aroclor-1260-1	6.208	6.113	6.313	523.960	500.000	4.8
Aroclor-1260-2	6.396	6.302	6.502	531.850	500.000	6.4
Aroclor-1260-3	6.547	6.453	6.653	550.200	500.000	10.0
Aroclor-1260-4	7.017	6.923	7.123	543.780	500.000	8.8
Aroclor-1260-5	7.260	7.164	7.364	560.820	500.000	12.2
Decachlorobiphenyl	8.633	8.540	8.740	53.180	50.000	6.4
Tetrachloro-m-xylene	3.659	3.562	3.762	55.270	50.000	10.5

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0081225\
 Data File : PO112865.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 12 Aug 2025 15:50
 Operator : YP/AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1660CCC500

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 08/13/2025
 Supervised By :mohammad ahmed 08/14/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 13 01:39:22 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0072325.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Jul 24 04:54:06 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.667	3.659	469.7E6	274.6E6	57.765	55.269
2)	SA Decachlor...	8.687	8.633	358.1E6	93971024	48.973	53.178
Target Compounds							
3)	L1 AR-1016-1	4.751	4.734	140.8E6	91367071	524.643m	522.077
4)	L1 AR-1016-2	4.769	4.751	224.8E6	134.7E6	553.429m	517.092
5)	L1 AR-1016-3	4.827	4.926	144.2E6	70193896	551.218	516.259
6)	L1 AR-1016-4	4.947	4.969	116.7E6	57301939	550.721	515.840
7)	L1 AR-1016-5	5.203	5.181	128.2E6	75537038	584.153	527.114
31)	L7 AR-1260-1	6.240	6.208	241.8E6	130.2E6	550.345	523.958
32)	L7 AR-1260-2	6.430	6.396	373.8E6	175.2E6	543.411	531.849
33)	L7 AR-1260-3	6.795	6.547	335.7E6	141.4E6	570.949	550.198m
34)	L7 AR-1260-4	7.055	7.017	244.6E6	104.0E6	542.265	543.785
35)	L7 AR-1260-5	7.297	7.260	656.7E6	231.2E6	529.972	560.820

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO081225\
Data File : PO112865.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Aug 2025 15:50
Operator : YP/AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
AR1660CCC500

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/13/2025
Supervised By :mohammad ahmed 08/14/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 13 01:39:22 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO072325.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Jul 24 04:54:06 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm





284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: EARTH03

Lab Code: ACE

SDG NO.: Q2818

Continuing Calib Date: 08/12/2025

Initial Calibration Date(s): 08/01/2025 08/01/2025

Continuing Calib Time: 09:51

Initial Calibration Time(s): 12:05 20:28

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	5.81	5.81	5.71	5.91	0.00
Aroclor-1016-2	(2)	5.83	5.83	5.73	5.93	0.00
Aroclor-1016-3	(3)	5.89	5.90	5.80	6.00	0.01
Aroclor-1016-4	(4)	5.99	5.99	5.89	6.09	0.00
Aroclor-1016-5	(5)	6.28	6.29	6.19	6.39	0.01
Aroclor-1260-1	(1)	7.40	7.40	7.30	7.50	0.00
Aroclor-1260-2	(2)	7.65	7.66	7.56	7.76	0.01
Aroclor-1260-3	(3)	8.01	8.01	7.91	8.11	0.00
Aroclor-1260-4	(4)	8.24	8.24	8.14	8.34	0.00
Aroclor-1260-5	(5)	8.56	8.57	8.47	8.67	0.01
Tetrachloro-m-xylene		4.66	4.66	4.56	4.76	0.00
Decachlorobiphenyl		10.43	10.44	10.34	10.54	0.01



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: EARTH03

Lab Code: ACE

SDG NO.: Q2818

Continuing Calib Date: 08/12/2025

Initial Calibration Date(s): 08/01/2025 08/01/2025

Continuing Calib Time: 09:51

Initial Calibration Time(s): 12:05 20:28

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	4.90	4.91	4.81	5.01	0.01
Aroclor-1016-2	(2)	4.96	4.96	4.86	5.06	0.00
Aroclor-1016-3	(3)	5.08	5.08	4.98	5.18	0.00
Aroclor-1016-4	(4)	5.12	5.12	5.02	5.22	0.00
Aroclor-1016-5	(5)	5.34	5.34	5.24	5.44	0.00
Aroclor-1260-1	(1)	6.55	6.56	6.46	6.66	0.01
Aroclor-1260-2	(2)	6.71	6.71	6.61	6.81	0.01
Aroclor-1260-3	(3)	6.92	6.92	6.82	7.02	0.01
Aroclor-1260-4	(4)	7.17	7.18	7.08	7.28	0.01
Aroclor-1260-5	(5)	7.41	7.42	7.32	7.52	0.01
Tetrachloro-m-xylene		3.80	3.81	3.71	3.91	0.01
Decachlorobiphenyl		8.82	8.83	8.73	8.93	0.01



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** EARTH03
Lab Code: ACE **SDG NO.:** Q2818
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/01/2025 08/01/2025

Client Sample No.: CCAL03 **Date Analyzed:** 08/12/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP074302.D **Time Analyzed:** 09:51

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.809	5.712	5.912	541.110	500.000	8.2
Aroclor-1016-2	5.830	5.733	5.933	542.210	500.000	8.4
Aroclor-1016-3	5.894	5.796	5.996	546.140	500.000	9.2
Aroclor-1016-4	5.990	5.893	6.093	549.740	500.000	9.9
Aroclor-1016-5	6.283	6.186	6.386	547.590	500.000	9.5
Aroclor-1260-1	7.400	7.303	7.503	520.400	500.000	4.1
Aroclor-1260-2	7.653	7.555	7.755	502.220	500.000	0.4
Aroclor-1260-3	8.010	7.913	8.113	510.120	500.000	2.0
Aroclor-1260-4	8.239	8.142	8.342	508.850	500.000	1.8
Aroclor-1260-5	8.564	8.468	8.668	488.630	500.000	-2.3
Decachlorobiphenyl	10.432	10.338	10.538	50.860	50.000	1.7
Tetrachloro-m-xylene	4.657	4.560	4.760	56.700	50.000	13.4



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** EARTH03
Lab Code: ACE **SDG NO.:** Q2818
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/01/2025 08/01/2025

Client Sample No.: CCAL03 **Date Analyzed:** 08/12/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP074302.D **Time Analyzed:** 09:51

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.901	4.806	5.006	583.180	500.000	16.6
Aroclor-1016-2	4.958	4.863	5.063	601.820	500.000	20.4
Aroclor-1016-3	5.080	4.983	5.183	578.000	500.000	15.6
Aroclor-1016-4	5.120	5.023	5.223	590.230	500.000	18.0
Aroclor-1016-5	5.335	5.239	5.439	592.640	500.000	18.5
Aroclor-1260-1	6.551	6.455	6.655	550.000	500.000	10.0
Aroclor-1260-2	6.705	6.609	6.809	554.960	500.000	11.0
Aroclor-1260-3	6.915	6.819	7.019	569.790	500.000	14.0
Aroclor-1260-4	7.174	7.079	7.279	581.890	500.000	16.4
Aroclor-1260-5	7.413	7.318	7.518	587.660	500.000	17.5
Decachlorobiphenyl	8.818	8.726	8.926	52.860	50.000	5.7
Tetrachloro-m-xylene	3.802	3.705	3.905	61.630	50.000	23.3

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP081225\
 Data File : PP074302.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 12 Aug 2025 09:51
 Operator : YP\AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1660CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 12 11:29:45 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP080125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Aug 04 11:01:49 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.657	3.802	63412779	236.3E6	56.704	61.629
2) SA Decachlor...	10.432	8.818	49396168	317.9E6	50.863	52.859
Target Compounds						
3) L1 AR-1016-1	5.809	4.901	22324389	232.6E6	541.108	583.175
4) L1 AR-1016-2	5.830	4.958	32934644	113.0E6	542.208	601.822
5) L1 AR-1016-3	5.894	5.080	21660183	61308172	546.140	577.999
6) L1 AR-1016-4	5.990	5.120	17890857	64465367	549.736	590.234
7) L1 AR-1016-5	6.283	5.335	17779139	69316623	547.589	592.636
31) L7 AR-1260-1	7.400	6.551	29327545	222.2E6	520.397	549.999
32) L7 AR-1260-2	7.653	6.705	34227789	173.5E6	502.225	554.959
33) L7 AR-1260-3	8.010	6.915	27187361	224.6E6	510.121	569.795
34) L7 AR-1260-4	8.239	7.174	31861536	169.3E6	508.848	581.890
35) L7 AR-1260-5	8.564	7.413	55364530	445.1E6	488.633	587.662



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP081225\
Data File : PP074302.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Aug 2025 09:51
Operator : YP\AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 12 11:29:45 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP080125.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Aug 04 11:01:49 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm





284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: EARTH03

Lab Code: ACE

SDG NO.: Q2818

Continuing Calib Date: 08/12/2025

Initial Calibration Date(s): 08/01/2025 08/01/2025

Continuing Calib Time: 15:48

Initial Calibration Time(s): 12:05 20:28

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	5.81	5.81	5.71	5.91	0.00
Aroclor-1016-2	(2)	5.83	5.83	5.73	5.93	0.00
Aroclor-1016-3	(3)	5.90	5.90	5.80	6.00	0.01
Aroclor-1016-4	(4)	5.99	5.99	5.89	6.09	0.00
Aroclor-1016-5	(5)	6.28	6.29	6.19	6.39	0.01
Aroclor-1260-1	(1)	7.40	7.40	7.30	7.50	0.00
Aroclor-1260-2	(2)	7.65	7.66	7.56	7.76	0.01
Aroclor-1260-3	(3)	8.01	8.01	7.91	8.11	0.00
Aroclor-1260-4	(4)	8.24	8.24	8.14	8.34	0.00
Aroclor-1260-5	(5)	8.57	8.57	8.47	8.67	0.01
Tetrachloro-m-xylene		4.66	4.66	4.56	4.76	0.00
Decachlorobiphenyl		10.44	10.44	10.34	10.54	0.00



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: EARTH03

Lab Code: ACE

SDG NO.: Q2818

Continuing Calib Date: 08/12/2025

Initial Calibration Date(s): 08/01/2025

08/01/2025

Continuing Calib Time: 15:48

Initial Calibration Time(s): 12:05

20:28

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	4.90	4.91	4.81	5.01	0.01
Aroclor-1016-2	(2)	4.96	4.96	4.86	5.06	0.00
Aroclor-1016-3	(3)	5.08	5.08	4.98	5.18	0.00
Aroclor-1016-4	(4)	5.12	5.12	5.02	5.22	0.00
Aroclor-1016-5	(5)	5.33	5.34	5.24	5.44	0.01
Aroclor-1260-1	(1)	6.55	6.56	6.46	6.66	0.01
Aroclor-1260-2	(2)	6.71	6.71	6.61	6.81	0.00
Aroclor-1260-3	(3)	6.92	6.92	6.82	7.02	0.00
Aroclor-1260-4	(4)	7.18	7.18	7.08	7.28	0.00
Aroclor-1260-5	(5)	7.41	7.42	7.32	7.52	0.01
Tetrachloro-m-xylene		3.80	3.81	3.71	3.91	0.01
Decachlorobiphenyl		8.82	8.83	8.73	8.93	0.01



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** EARTH03
Lab Code: ACE **SDG NO.:** Q2818
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/01/2025 08/01/2025

Client Sample No.: CCAL04 **Date Analyzed:** 08/12/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP074316.D **Time Analyzed:** 15:48

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.811	5.712	5.912	556.880	500.000	11.4
Aroclor-1016-2	5.832	5.733	5.933	550.560	500.000	10.1
Aroclor-1016-3	5.895	5.796	5.996	559.440	500.000	11.9
Aroclor-1016-4	5.992	5.893	6.093	563.510	500.000	12.7
Aroclor-1016-5	6.284	6.186	6.386	554.850	500.000	11.0
Aroclor-1260-1	7.401	7.303	7.503	536.530	500.000	7.3
Aroclor-1260-2	7.654	7.555	7.755	517.500	500.000	3.5
Aroclor-1260-3	8.012	7.913	8.113	522.830	500.000	4.6
Aroclor-1260-4	8.240	8.142	8.342	532.940	500.000	6.6
Aroclor-1260-5	8.565	8.468	8.668	506.790	500.000	1.4
Decachlorobiphenyl	10.436	10.338	10.538	52.310	50.000	4.6
Tetrachloro-m-xylene	4.659	4.560	4.760	57.560	50.000	15.1



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** EARTH03
Lab Code: ACE **SDG NO.:** Q2818
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/01/2025 08/01/2025

Client Sample No.: CCAL04 **Date Analyzed:** 08/12/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP074316.D **Time Analyzed:** 15:48

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.903	4.806	5.006	585.540	500.000	17.1
Aroclor-1016-2	4.958	4.863	5.063	555.580	500.000	11.1
Aroclor-1016-3	5.081	4.983	5.183	591.090	500.000	18.2
Aroclor-1016-4	5.122	5.023	5.223	595.870	500.000	19.2
Aroclor-1016-5	5.334	5.239	5.439	590.590	500.000	18.1
Aroclor-1260-1	6.552	6.455	6.655	580.600	500.000	16.1
Aroclor-1260-2	6.706	6.609	6.809	568.790	500.000	13.8
Aroclor-1260-3	6.917	6.819	7.019	594.280	500.000	18.9
Aroclor-1260-4	7.175	7.079	7.279	583.960	500.000	16.8
Aroclor-1260-5	7.414	7.318	7.518	585.850	500.000	17.2
Decachlorobiphenyl	8.820	8.726	8.926	55.460	50.000	10.9
Tetrachloro-m-xylene	3.802	3.705	3.905	59.630	50.000	19.3

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP081225\
 Data File : PP074316.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 12 Aug 2025 15:48
 Operator : YP\AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1660CCC500

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 08/13/2025
 Supervised By :mohammad ahmed 08/14/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 13 01:54:08 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP080125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Aug 04 11:01:49 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.659	3.802	64369772	228.6E6	57.560	59.625m
2) SA Decachlor...	10.436	8.820	50800485	333.5E6	52.309	55.456
Target Compounds						
3) L1 AR-1016-1	5.811	4.903	22974987	233.6E6	556.877	585.544
4) L1 AR-1016-2	5.832	4.958	33441672	104.3E6	550.555	555.579m
5) L1 AR-1016-3	5.895	5.081	22187737	62696370	559.442	591.087
6) L1 AR-1016-4	5.992	5.122	18339102	65081402	563.509	595.874
7) L1 AR-1016-5	6.284	5.334	18014922	69076982	554.851	590.587m
31) L7 AR-1260-1	7.401	6.552	30236790	234.5E6	536.531	580.596
32) L7 AR-1260-2	7.654	6.706	35268906	177.9E6	517.501	568.793
33) L7 AR-1260-3	8.012	6.917	27864737	234.2E6	522.830	594.283
34) L7 AR-1260-4	8.240	7.175	33370027	169.9E6	532.940	583.961
35) L7 AR-1260-5	8.565	7.414	57421797	443.7E6	506.790	585.848

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP081225\
Data File : PP074316.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Aug 2025 15:48
Operator : YP\AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
AR1660CCC500

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/13/2025
Supervised By :mohammad ahmed 08/14/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 13 01:54:08 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP080125.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Aug 04 11:01:49 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Analytical Sequence

Client: Earth Engineering Inc.

SDG No.: Q2818

Project: Reserve Turgyan Farms

Instrument ID: ECD_O

GC Column: ZB-MR1

ID: 0.32 (mm)

Inst. Calib. Date(s): 07/23/2025

07/23/2025

THE ANALYTICAL SEQUENCE OF PERFORMANCE EVALUATION MIXTURES, BLANKS, SAMPLES,
AND STANDARDS IS GIVEN BELOW:

CLIENT ID	LAB SAMPLE ID	DATE ANALYZED	TIME ANALYZED	DATAFILE	DCB RT #	TCX RT #
IBLK	IBLK	07/23/2025	11:14	PO112413.D	8.69	3.67
AR1660ICC1000	AR1660ICC1000	07/23/2025	11:32	PO112414.D	8.70	3.67
AR1660ICC750	AR1660ICC750	07/23/2025	11:50	PO112415.D	8.69	3.67
AR1660ICC500	AR1660ICC500	07/23/2025	12:08	PO112416.D	8.69	3.67
AR1660ICC250	AR1660ICC250	07/23/2025	12:27	PO112417.D	8.69	3.67
AR1660ICC050	AR1660ICC050	07/23/2025	12:45	PO112418.D	8.69	3.67
AR1221ICC500	AR1221ICC500	07/23/2025	13:03	PO112419.D	8.69	3.67
AR1232ICC500	AR1232ICC500	07/23/2025	13:22	PO112420.D	8.69	3.67
AR1242ICC1000	AR1242ICC1000	07/23/2025	13:40	PO112421.D	8.69	3.67
AR1242ICC750	AR1242ICC750	07/23/2025	13:59	PO112422.D	8.69	3.67
AR1242ICC500	AR1242ICC500	07/23/2025	14:17	PO112423.D	8.69	3.67
AR1242ICC250	AR1242ICC250	07/23/2025	14:36	PO112424.D	8.69	3.67
AR1242ICC050	AR1242ICC050	07/23/2025	14:54	PO112425.D	8.70	3.67
AR1248ICC1000	AR1248ICC1000	07/23/2025	15:13	PO112426.D	8.69	3.67
AR1248ICC750	AR1248ICC750	07/23/2025	15:31	PO112427.D	8.69	3.67
AR1248ICC500	AR1248ICC500	07/23/2025	15:48	PO112428.D	8.69	3.67
AR1248ICC250	AR1248ICC250	07/23/2025	16:07	PO112429.D	8.70	3.67
AR1248ICC050	AR1248ICC050	07/23/2025	16:25	PO112430.D	8.69	3.67
AR1254ICC1000	AR1254ICC1000	07/23/2025	16:44	PO112431.D	8.69	3.67
AR1254ICC750	AR1254ICC750	07/23/2025	17:02	PO112432.D	8.69	3.67
AR1254ICC500	AR1254ICC500	07/23/2025	17:21	PO112433.D	8.70	3.67
AR1254ICC250	AR1254ICC250	07/23/2025	17:39	PO112434.D	8.69	3.67
AR1254ICC050	AR1254ICC050	07/23/2025	17:57	PO112435.D	8.69	3.67
AR1262ICC500	AR1262ICC500	07/23/2025	18:16	PO112436.D	8.70	3.67
AR1268ICC1000	AR1268ICC1000	07/23/2025	18:34	PO112437.D	8.70	3.67
AR1268ICC750	AR1268ICC750	07/23/2025	18:53	PO112438.D	8.70	3.67
AR1268ICC500	AR1268ICC500	07/23/2025	19:11	PO112439.D	8.70	3.67
AR1268ICC250	AR1268ICC250	07/23/2025	19:28	PO112440.D	8.69	3.67
AR1268ICC050	AR1268ICC050	07/23/2025	19:47	PO112441.D	8.70	3.67
AR1660CCC500	AR1660CCC500	08/12/2025	08:54	PO112850.D	8.69	3.67
IBLK	IBLK	08/12/2025	10:05	PO112854.D	8.69	3.66
B-2-5-1	Q2818-01	08/12/2025	12:50	PO112858.D	8.69	3.67
B-3-5-2	Q2818-02	08/12/2025	13:08	PO112859.D	8.69	3.67
AR1660CCC500	AR1660CCC500	08/12/2025	15:50	PO112865.D	8.69	3.67
IBLK	IBLK	08/12/2025	17:00	PO112869.D	8.69	3.67
IBLK	IBLK	08/01/2025	11:49	PP074167.D	10.44	4.66
AR1660ICC1000	AR1660ICC1000	08/01/2025	12:05	PP074168.D	10.44	4.66
AR1660ICC750	AR1660ICC750	08/01/2025	12:22	PP074169.D	10.44	4.66
AR1660ICC500	AR1660ICC500	08/01/2025	12:38	PP074170.D	10.44	4.66
AR1660ICC250	AR1660ICC250	08/01/2025	12:54	PP074171.D	10.44	4.66
AR1660ICC050	AR1660ICC050	08/01/2025	13:42	PP074172.D	10.44	4.66
AR1221ICC500	AR1221ICC500	08/01/2025	13:58	PP074173.D	10.44	4.66

Analytical Sequence

AR1232ICC500	AR1232ICC500	08/01/2025	14:15	PP074174.D	10.44	4.66
AR1242ICC1000	AR1242ICC1000	08/01/2025	14:31	PP074175.D	10.44	4.66
AR1242ICC750	AR1242ICC750	08/01/2025	14:47	PP074176.D	10.44	4.66
AR1242ICC500	AR1242ICC500	08/01/2025	15:03	PP074177.D	10.43	4.66
AR1242ICC250	AR1242ICC250	08/01/2025	15:19	PP074178.D	10.44	4.66
AR1242ICC050	AR1242ICC050	08/01/2025	15:36	PP074179.D	10.43	4.66
AR1248ICC500	AR1248ICC500	08/01/2025	16:57	PP074182.D	10.43	4.66
AR1254ICC1000	AR1254ICC1000	08/01/2025	18:02	PP074185.D	10.43	4.66
AR1254ICC750	AR1254ICC750	08/01/2025	18:18	PP074186.D	10.43	4.66
AR1254ICC500	AR1254ICC500	08/01/2025	18:34	PP074187.D	10.43	4.66
AR1254ICC250	AR1254ICC250	08/01/2025	18:50	PP074188.D	10.43	4.66
AR1254ICC050	AR1254ICC050	08/01/2025	19:23	PP074189.D	10.43	4.66
AR1262ICC500	AR1262ICC500	08/01/2025	19:39	PP074190.D	10.43	4.66
AR1268ICC500	AR1268ICC500	08/01/2025	20:28	PP074193.D	10.43	4.66
AR1660CCC500	AR1660CCC500	08/12/2025	09:51	PP074302.D	10.43	4.66
LBLK	LBLK	08/12/2025	10:43	PP074305.D	10.44	4.66
PB169205BL	PB169205BL	08/12/2025	12:48	PP074308.D	10.44	4.66
PB169205BS	PB169205BS	08/12/2025	13:05	PP074309.D	10.44	4.66
BIN0009-DRIVEWAY-TP-SOUTH-EAST	Q2830-01MS	08/12/2025	13:37	PP074311.D	10.44	4.66
BIN0009-DRIVEWAY-TP-SOUTH-EAST	Q2830-01MSD	08/12/2025	13:53	PP074312.D	10.44	4.66
AR1660CCC500	AR1660CCC500	08/12/2025	15:48	PP074316.D	10.44	4.66
LBLK	LBLK	08/12/2025	16:53	PP074319.D	10.43	4.66

Analytical Sequence

Client: Earth Engineering Inc.

SDG No.: Q2818

Project: Reserve Turgyan Farms

Instrument ID: ECD_O

GC Column: ZB-MR2

ID: 0.32 (mm)

Inst. Calib. Date(s): 07/23/2025

07/23/2025

THE ANALYTICAL SEQUENCE OF PERFORMANCE EVALUATION MIXTURES, BLANKS, SAMPLES,
AND STANDARDS IS GIVEN BELOW:

CLIENT ID	LAB SAMPLE ID	DATE ANALYZED	TIME ANALYZED	DATAFILE	DCB RT #	TCX RT #
IBLK	IBLK	07/23/2025	11:14	PO112413.D	8.64	3.66
AR1660ICC1000	AR1660ICC1000	07/23/2025	11:32	PO112414.D	8.64	3.66
AR1660ICC750	AR1660ICC750	07/23/2025	11:50	PO112415.D	8.64	3.66
AR1660ICC500	AR1660ICC500	07/23/2025	12:08	PO112416.D	8.64	3.66
AR1660ICC250	AR1660ICC250	07/23/2025	12:27	PO112417.D	8.64	3.66
AR1660ICC050	AR1660ICC050	07/23/2025	12:45	PO112418.D	8.64	3.66
AR1221ICC500	AR1221ICC500	07/23/2025	13:03	PO112419.D	8.64	3.66
AR1232ICC500	AR1232ICC500	07/23/2025	13:22	PO112420.D	8.64	3.66
AR1242ICC1000	AR1242ICC1000	07/23/2025	13:40	PO112421.D	8.64	3.66
AR1242ICC750	AR1242ICC750	07/23/2025	13:59	PO112422.D	8.64	3.66
AR1242ICC500	AR1242ICC500	07/23/2025	14:17	PO112423.D	8.64	3.66
AR1242ICC250	AR1242ICC250	07/23/2025	14:36	PO112424.D	8.64	3.66
AR1242ICC050	AR1242ICC050	07/23/2025	14:54	PO112425.D	8.64	3.66
AR1248ICC1000	AR1248ICC1000	07/23/2025	15:13	PO112426.D	8.64	3.66
AR1248ICC750	AR1248ICC750	07/23/2025	15:31	PO112427.D	8.64	3.66
AR1248ICC500	AR1248ICC500	07/23/2025	15:48	PO112428.D	8.64	3.66
AR1248ICC250	AR1248ICC250	07/23/2025	16:07	PO112429.D	8.64	3.66
AR1248ICC050	AR1248ICC050	07/23/2025	16:25	PO112430.D	8.64	3.66
AR1254ICC1000	AR1254ICC1000	07/23/2025	16:44	PO112431.D	8.64	3.66
AR1254ICC750	AR1254ICC750	07/23/2025	17:02	PO112432.D	8.64	3.66
AR1254ICC500	AR1254ICC500	07/23/2025	17:21	PO112433.D	8.64	3.66
AR1254ICC250	AR1254ICC250	07/23/2025	17:39	PO112434.D	8.64	3.66
AR1254ICC050	AR1254ICC050	07/23/2025	17:57	PO112435.D	8.64	3.66
AR1262ICC500	AR1262ICC500	07/23/2025	18:16	PO112436.D	8.64	3.66
AR1268ICC1000	AR1268ICC1000	07/23/2025	18:34	PO112437.D	8.64	3.66
AR1268ICC750	AR1268ICC750	07/23/2025	18:53	PO112438.D	8.64	3.66
AR1268ICC500	AR1268ICC500	07/23/2025	19:11	PO112439.D	8.64	3.66
AR1268ICC250	AR1268ICC250	07/23/2025	19:28	PO112440.D	8.64	3.66
AR1268ICC050	AR1268ICC050	07/23/2025	19:47	PO112441.D	8.64	3.66
AR1660CCC500	AR1660CCC500	08/12/2025	08:54	PO112850.D	8.63	3.66
IBLK	IBLK	08/12/2025	10:05	PO112854.D	8.63	3.66
B-2-5-1	Q2818-01	08/12/2025	12:50	PO112858.D	8.63	3.66
B-3-5-2	Q2818-02	08/12/2025	13:08	PO112859.D	8.63	3.66
AR1660CCC500	AR1660CCC500	08/12/2025	15:50	PO112865.D	8.63	3.66
IBLK	IBLK	08/12/2025	17:00	PO112869.D	8.63	3.66
IBLK	IBLK	08/01/2025	11:49	PP074167.D	8.83	3.80
AR1660ICC1000	AR1660ICC1000	08/01/2025	12:05	PP074168.D	8.83	3.81
AR1660ICC750	AR1660ICC750	08/01/2025	12:22	PP074169.D	8.83	3.80
AR1660ICC500	AR1660ICC500	08/01/2025	12:38	PP074170.D	8.83	3.81
AR1660ICC250	AR1660ICC250	08/01/2025	12:54	PP074171.D	8.83	3.81
AR1660ICC050	AR1660ICC050	08/01/2025	13:42	PP074172.D	8.83	3.80
AR1221ICC500	AR1221ICC500	08/01/2025	13:58	PP074173.D	8.83	3.80

Analytical Sequence

AR1232ICC500	AR1232ICC500	08/01/2025	14:15	PP074174.D	8.83	3.80
AR1242ICC1000	AR1242ICC1000	08/01/2025	14:31	PP074175.D	8.83	3.80
AR1242ICC750	AR1242ICC750	08/01/2025	14:47	PP074176.D	8.82	3.80
AR1242ICC500	AR1242ICC500	08/01/2025	15:03	PP074177.D	8.83	3.81
AR1242ICC250	AR1242ICC250	08/01/2025	15:19	PP074178.D	8.82	3.81
AR1242ICC050	AR1242ICC050	08/01/2025	15:36	PP074179.D	8.82	3.80
AR1248ICC500	AR1248ICC500	08/01/2025	16:57	PP074182.D	8.82	3.81
AR1254ICC1000	AR1254ICC1000	08/01/2025	18:02	PP074185.D	8.82	3.81
AR1254ICC750	AR1254ICC750	08/01/2025	18:18	PP074186.D	8.82	3.80
AR1254ICC500	AR1254ICC500	08/01/2025	18:34	PP074187.D	8.82	3.80
AR1254ICC250	AR1254ICC250	08/01/2025	18:50	PP074188.D	8.82	3.80
AR1254ICC050	AR1254ICC050	08/01/2025	19:23	PP074189.D	8.82	3.80
AR1262ICC500	AR1262ICC500	08/01/2025	19:39	PP074190.D	8.82	3.80
AR1268ICC500	AR1268ICC500	08/01/2025	20:28	PP074193.D	8.82	3.80
AR1660CCC500	AR1660CCC500	08/12/2025	09:51	PP074302.D	8.82	3.80
LBLK	LBLK	08/12/2025	10:43	PP074305.D	8.82	3.81
PB169205BL	PB169205BL	08/12/2025	12:48	PP074308.D	8.82	3.80
PB169205BS	PB169205BS	08/12/2025	13:05	PP074309.D	8.82	3.80
BIN0009-DRIVEWAY-TP-SOUTH-EAST	Q2830-01MS	08/12/2025	13:37	PP074311.D	8.82	3.80
BIN0009-DRIVEWAY-TP-SOUTH-EAST	Q2830-01MSD	08/12/2025	13:53	PP074312.D	8.82	3.80
AR1660CCC500	AR1660CCC500	08/12/2025	15:48	PP074316.D	8.82	3.80
LBLK	LBLK	08/12/2025	16:53	PP074319.D	8.82	3.80



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

IDENTIFICATION SUMMARY
FOR MULTICOMPONENT ANALYTES

SAMPLE NO.

PB169205BS

Lab Name: Alliance

Contract: EARTH03

Lab Code: ACE

SDG NO.: Q2818

Lab Sample ID: PB169205BS

Date(s) Analyzed: 08/12/2025 08/12/2025

Instrument ID (1): ECD_P

Instrument ID (2): ECD_P

GC Column: (1): ZB-MR1 ID: 0.32 (mm)

GC Column: (2): ZB-MR2 ID: 0.32 (mm)

Data file PP074309.D

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	MEAN CONCENTRATION	%RPD
			FROM	TO			
Aroclor-1016	1	5.809	5.759	5.859	183	185	1.09
COLUMN 1	2	5.83	5.78	5.88	187		
	3	5.893	5.843	5.943	187		
	4	5.99	5.94	6.04	188		
	5	6.283	6.233	6.333	180		
	1	4.902	4.852	4.952	191		
COLUMN 2	2	4.959	4.909	5.009	185	183	
	3	5.08	5.03	5.13	176		
	4	5.121	5.071	5.171	172		
	5	5.334	5.284	5.384	191		
Aroclor-1260	1	7.401	7.351	7.451	183	165	8.7
COLUMN 1	2	7.653	7.603	7.703	176		
	3	8.012	7.962	8.062	152		
	4	8.238	8.188	8.288	165		
	5	8.566	8.516	8.616	150		
	1	6.551	6.501	6.601	195	180	
COLUMN 2	2	6.706	6.656	6.756	195		
	3	6.915	6.865	6.965	173		
	4	7.175	7.125	7.225	171		
	5	7.415	7.365	7.465	168		

IDENTIFICATION SUMMARY
FOR MULTICOMPONENT ANALYTES

SAMPLE NO.

BIN0009-DRIVEWAY-TP

Lab Name: Alliance

Contract: EARTH03

Lab Code: ACE

SDG NO.: Q2818

Lab Sample ID: Q2830-01MS

Date(s) Analyzed: 08/12/2025 08/12/2025

Instrument ID (1): ECD_P

Instrument ID (2): ECD_P

GC Column: (1): ZB-MR1 ID: 0.32 (mm)

GC Column: (2): ZB-MR2 ID: 0.32 (mm)

Data file PP074311.D

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	MEAN CONCENTRATION	%RPD
			FROM	TO			
Aroclor-1016	1	5.809	5.759	5.859	200	201	5.8
COLUMN 1	2	5.831	5.781	5.881	201		
	3	5.894	5.844	5.944	200		
	4	5.991	5.941	6.041	209		
	5	6.283	6.233	6.333	195		
	1	4.901	4.851	4.951	213		
COLUMN 2	2	4.958	4.908	5.008	215	213	
	3	5.079	5.029	5.129	219		
	4	5.119	5.069	5.169	206		
	5	5.334	5.284	5.384	212		
Aroclor-1260	1	7.4	7.35	7.45	191	173	4.52
COLUMN 1	2	7.653	7.603	7.703	183		
	3	8.012	7.962	8.062	157		
	4	8.238	8.188	8.288	181		
	5	8.566	8.516	8.616	155		
	1	6.55	6.5	6.6	196		
COLUMN 2	2	6.704	6.654	6.754	193	181	
	3	6.914	6.864	6.964	172		
	4	7.175	7.125	7.225	175		
	5	7.413	7.363	7.463	170		

IDENTIFICATION SUMMARY
FOR MULTICOMPONENT ANALYTES

SAMPLE NO.

BIN0009-DRIVEWAY-TP.

Lab Name: Alliance

Contract: EARTH03

Lab Code: ACE

SDG NO.: Q2818

Lab Sample ID: Q2830-01MSD

Date(s) Analyzed: 08/12/2025 08/12/2025

Instrument ID (1): ECD_P

Instrument ID (2): ECD_P

GC Column: (1): ZB-MR1 ID: 0.32 (mm)

GC Column: (2): ZB-MR2 ID: 0.32 (mm)

Data file PP074312.D

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	MEAN CONCENTRATION	%RPD
			FROM	TO			
Aroclor-1016	1	5.811	5.761	5.861	205	213	2.78
COLUMN 1	2	5.832	5.782	5.882	202		
	3	5.895	5.845	5.945	204		
	4	5.992	5.942	6.042	256		
	5	6.284	6.234	6.334	199		
	1	4.904	4.854	4.954	220		
COLUMN 2	2	4.961	4.911	5.011	222	219	
	3	5.081	5.031	5.131	215		
	4	5.122	5.072	5.172	218		
	5	5.337	5.287	5.387	219		
Aroclor-1260	1	7.402	7.352	7.452	197	177	6.03
COLUMN 1	2	7.654	7.604	7.704	186		
	3	8.012	7.962	8.062	160		
	4	8.239	8.189	8.289	183		
	5	8.565	8.515	8.615	159		
	1	6.552	6.502	6.602	206		
COLUMN 2	2	6.707	6.657	6.757	201	188	
	3	6.916	6.866	6.966	178		
	4	7.176	7.126	7.226	180		
	5	7.415	7.365	7.465	176		



QC SAMPLE DATA

Report of Analysis

Client:	Earth Engineering Inc.	Date Collected:	
Project:	Reserve Turgyan Farms	Date Received:	
Client Sample ID:	PB169205BL	SDG No.:	Q2818
Lab Sample ID:	PB169205BL	Matrix:	SOIL
Analytical Method:	8082A	% Solid:	100
Sample Wt/Vol:	30.03	Units:	g
Soil Aliquot Vol:			uL
Extraction Type:		Test:	PCB
GPC Factor :	1.0	Injection Volume :	
Prep Method :	SW3541B		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PP074308.D	1	08/12/25 08:15	08/12/25 12:48	PB169205

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	3.90	U	3.90	17.0	ug/kg
11104-28-2	Aroclor-1221	4.00	U	4.00	17.0	ug/kg
11141-16-5	Aroclor-1232	3.70	U	3.70	17.0	ug/kg
53469-21-9	Aroclor-1242	4.00	U	4.00	17.0	ug/kg
12672-29-6	Aroclor-1248	5.90	U	5.90	17.0	ug/kg
11097-69-1	Aroclor-1254	3.20	U	3.20	17.0	ug/kg
37324-23-5	Aroclor-1262	5.00	U	5.00	17.0	ug/kg
11100-14-4	Aroclor-1268	3.60	U	3.60	17.0	ug/kg
11096-82-5	Aroclor-1260	3.20	U	3.20	17.0	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	23.1		32 - 144	116%	SPK: 20
2051-24-3	Decachlorobiphenyl	22.2		32 - 175	111%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP081225\
Data File : PP074308.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Aug 2025 12:48
Operator : YP\AJ
Sample : PB169205BL
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
PB169205BL

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 12 13:31:26 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP080125.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Aug 04 11:01:49 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.656	3.798	25834612	77167851	23.102	20.125
2)	SA Decachlor...	10.439	8.817	21524564	120.6E6	22.164	20.052

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP081225\
Data File : PP074308.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Aug 2025 12:48
Operator : YP\AJ
Sample : PB169205BL
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
PB169205BL

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 12 13:31:26 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP080125.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Aug 04 11:01:49 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.656 min
Delta R.T.: -0.004 min
Response: 25834612
Conc: 23.10 ng/ml

Instrument :
ECD_P
ClientSampleId :
PB169205BL




#1 Tetrachloro-m-xylene

R.T.: 3.798 min
Delta R.T.: -0.007 min
Response: 77167851
Conc: 20.13 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.439 min
Delta R.T.: 0.000 min
Response: 21524564
Conc: 22.16 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.817 min
Delta R.T.: -0.009 min
Response: 120583654
Conc: 20.05 ng/ml

Report of Analysis

Client:	Earth Engineering Inc.		Date Collected:	07/23/25	
Project:	Reserve Turgyan Farms		Date Received:	07/23/25	
Client Sample ID:	PIBLK-PO112413.D		SDG No.:	Q2818	
Lab Sample ID:	I.BLK-PO112413.D		Matrix:	WATER	
Analytical Method:	8082A		% Solid:	0	Decanted:
Sample Wt/Vol:	1000	Units: mL	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	PCB	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	5030				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO112413.D	1		07/23/25	PO072325

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
12674-11-2	Aroclor-1016	0.097	U	0.097	0.50	ug/L
11104-28-2	Aroclor-1221	0.13	U	0.13	0.50	ug/L
11141-16-5	Aroclor-1232	0.096	U	0.096	0.50	ug/L
53469-21-9	Aroclor-1242	0.12	U	0.12	0.50	ug/L
12672-29-6	Aroclor-1248	0.071	U	0.071	0.50	ug/L
11097-69-1	Aroclor-1254	0.094	U	0.094	0.50	ug/L
11096-82-5	Aroclor-1260	0.081	U	0.081	0.50	ug/L
37324-23-5	Aroclor-1262	0.14	U	0.14	0.50	ug/L
11100-14-4	Aroclor-1268	0.11	U	0.11	0.50	ug/L
SURROGATES						
877-09-8	Tetrachloro-m-xylene	19.1		60 - 140	95%	SPK: 20
2051-24-3	Decachlorobiphenyl	20.0		60 - 140	100%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0072325\
Data File : PO112413.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Jul 2025 11:14
Operator : YP/AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 24 05:03:52 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0072325.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Jul 24 04:54:06 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.670	3.662	155.1E6	97072114	19.073	19.535
2) SA Decachlor...	8.694	8.640	146.0E6	36122645	19.970	20.442

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_0\Data\P0072325\
Data File : P0112413.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Jul 2025 11:14
Operator : YP/AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 24 05:03:52 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\P0072325.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Jul 24 04:54:06 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.670 min
Delta R.T.: 0.001 min
Response: 155090113
Conc: 19.07 ng/ml

Instrument :
ECD_O
ClientSampleId :
I.BLK




#1 Tetrachloro-m-xylene

R.T.: 3.662 min
Delta R.T.: 0.000 min
Response: 97072114
Conc: 19.53 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.694 min
Delta R.T.: 0.001 min
Response: 146014518
Conc: 19.97 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.640 min
Delta R.T.: 0.000 min
Response: 36122645
Conc: 20.44 ng/ml

Report of Analysis

Client:	Earth Engineering Inc.	Date Collected:	08/12/25
Project:	Reserve Turgyan Farms	Date Received:	08/12/25
Client Sample ID:	PIBLK-PO112854.D	SDG No.:	Q2818
Lab Sample ID:	I.BLK-PO112854.D	Matrix:	WATER
Analytical Method:	8082A	% Solid:	0
Sample Wt/Vol:	1000	Units:	mL
Soil Aliquot Vol:			uL
Extraction Type:		Test:	PCB
GPC Factor :	1.0	PH :	
Prep Method :	5030	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO112854.D	1		08/12/25	PO081225

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
12674-11-2	Aroclor-1016	0.097	U	0.097	0.50	ug/L
11104-28-2	Aroclor-1221	0.13	U	0.13	0.50	ug/L
11141-16-5	Aroclor-1232	0.096	U	0.096	0.50	ug/L
53469-21-9	Aroclor-1242	0.12	U	0.12	0.50	ug/L
12672-29-6	Aroclor-1248	0.071	U	0.071	0.50	ug/L
11097-69-1	Aroclor-1254	0.094	U	0.094	0.50	ug/L
11096-82-5	Aroclor-1260	0.081	U	0.081	0.50	ug/L
37324-23-5	Aroclor-1262	0.14	U	0.14	0.50	ug/L
11100-14-4	Aroclor-1268	0.11	U	0.11	0.50	ug/L
SURROGATES						
877-09-8	Tetrachloro-m-xylene	20.5		60 - 140	102%	SPK: 20
2051-24-3	Decachlorobiphenyl	17.2		60 - 140	86%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0081225\
 Data File : PO112854.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 12 Aug 2025 10:05
 Operator : YP/AJ
 Sample : I.BLK
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 I.BLK

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 12 11:05:34 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0072325.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Jul 24 04:54:06 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.664	3.658	170.0E6	101.6E6	20.911	20.456
2) SA Decachlor...	8.685	8.632	126.0E6	36010722	17.231	20.378

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO081225\
Data File : PO112854.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Aug 2025 10:05
Operator : YP/AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 12 11:05:34 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO072325.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Jul 24 04:54:06 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.664 min
Delta R.T.: -0.005 min
Response: 170038746
Conc: 20.91 ng/ml

Instrument :
ECD_O
ClientSampleId :
I.BLK




#1 Tetrachloro-m-xylene

R.T.: 3.658 min
Delta R.T.: -0.004 min
Response: 101647548
Conc: 20.46 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.685 min
Delta R.T.: -0.008 min
Response: 125986606
Conc: 17.23 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.632 min
Delta R.T.: -0.008 min
Response: 36010722
Conc: 20.38 ng/ml

Report of Analysis

Client:	Earth Engineering Inc.	Date Collected:	08/12/25
Project:	Reserve Turgyan Farms	Date Received:	08/12/25
Client Sample ID:	PIBLK-PO112869.D	SDG No.:	Q2818
Lab Sample ID:	I.BLK-PO112869.D	Matrix:	WATER
Analytical Method:	8082A	% Solid:	0
Sample Wt/Vol:	1000	Units:	mL
Soil Aliquot Vol:			uL
Extraction Type:		Test:	PCB
GPC Factor :	1.0	PH :	
Prep Method :	5030	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO112869.D	1		08/12/25	po081225

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
12674-11-2	Aroclor-1016	0.097	U	0.097	0.50	ug/L
11104-28-2	Aroclor-1221	0.13	U	0.13	0.50	ug/L
11141-16-5	Aroclor-1232	0.096	U	0.096	0.50	ug/L
53469-21-9	Aroclor-1242	0.12	U	0.12	0.50	ug/L
12672-29-6	Aroclor-1248	0.071	U	0.071	0.50	ug/L
11097-69-1	Aroclor-1254	0.094	U	0.094	0.50	ug/L
11096-82-5	Aroclor-1260	0.081	U	0.081	0.50	ug/L
37324-23-5	Aroclor-1262	0.14	U	0.14	0.50	ug/L
11100-14-4	Aroclor-1268	0.11	U	0.11	0.50	ug/L
SURROGATES						
877-09-8	Tetrachloro-m-xylene	20.3		60 - 140	102%	SPK: 20
2051-24-3	Decachlorobiphenyl	19.2		60 - 140	96%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0081225\
Data File : PO112869.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Aug 2025 17:00
Operator : YP/AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 13 01:40:19 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0072325.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Jul 24 04:54:06 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.667	3.660	168.1E6	101.0E6	20.674	20.323
2) SA Decachlor...	8.688	8.633	140.3E6	36246288	19.192	20.512

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO081225\
Data File : PO112869.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Aug 2025 17:00
Operator : YP/AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 13 01:40:19 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO072325.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Jul 24 04:54:06 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.667 min
Delta R.T.: -0.002 min
Response: 168114131
Conc: 20.67 ng/ml

Instrument :
ECD_O
ClientSampleId :
I.BLK




#1 Tetrachloro-m-xylene

R.T.: 3.660 min
Delta R.T.: -0.003 min
Response: 100987501
Conc: 20.32 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.688 min
Delta R.T.: -0.004 min
Response: 140325611
Conc: 19.19 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.633 min
Delta R.T.: -0.007 min
Response: 36246288
Conc: 20.51 ng/ml

Report of Analysis

Client:	Earth Engineering Inc.	Date Collected:	08/01/25
Project:	Reserve Turgyan Farms	Date Received:	08/01/25
Client Sample ID:	PIBLK-PP074167.D	SDG No.:	Q2818
Lab Sample ID:	I.BLK-PP074167.D	Matrix:	WATER
Analytical Method:	8082A	% Solid:	0
Sample Wt/Vol:	1000	Units:	mL
Soil Aliquot Vol:			uL
Extraction Type:		Test:	PCB
GPC Factor :	1.0	PH :	
Prep Method :	5030	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PP074167.D	1		08/01/25	PP080125

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
12674-11-2	Aroclor-1016	0.097	U	0.097	0.50	ug/L
11104-28-2	Aroclor-1221	0.13	U	0.13	0.50	ug/L
11141-16-5	Aroclor-1232	0.096	U	0.096	0.50	ug/L
53469-21-9	Aroclor-1242	0.12	U	0.12	0.50	ug/L
12672-29-6	Aroclor-1248	0.071	U	0.071	0.50	ug/L
11097-69-1	Aroclor-1254	0.094	U	0.094	0.50	ug/L
11096-82-5	Aroclor-1260	0.081	U	0.081	0.50	ug/L
37324-23-5	Aroclor-1262	0.14	U	0.14	0.50	ug/L
11100-14-4	Aroclor-1268	0.11	U	0.11	0.50	ug/L
SURROGATES						
877-09-8	Tetrachloro-m-xylene	17.3		60 - 140	86%	SPK: 20
2051-24-3	Decachlorobiphenyl	18.7		60 - 140	93%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP080125\
Data File : PP074167.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 01 Aug 2025 11:49
Operator : YP\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 02 01:35:26 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP080125.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Aug 02 01:33:31 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.658	3.803	21300659	66146859	19.047	17.251
2)	SA Decachlor...	10.435	8.825	19096192	112.2E6	19.663	18.655

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP080125\
Data File : PP074167.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 01 Aug 2025 11:49
Operator : YP\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1


Instrument :
ECD_P
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 02 01:35:26 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP080125.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Aug 02 01:33:31 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Response_ Signal: PP074167.D\ECD1A.ch




#1 Tetrachloro-m-xylene

R.T.: 4.658 min
Delta R.T.: -0.002 min
Response: 21300659
Conc: 19.05 ng/ml

Instrument :
ECD_P
ClientSampleId :
I.BLK


Response_ Signal: PP074167.D\ECD2B.ch



#1 Tetrachloro-m-xylene

R.T.: 3.803 min
Delta R.T.: -0.002 min
Response: 66146859
Conc: 17.25 ng/ml


Response_ Signal: PP074167.D\ECD1A.ch



#2 Decachlorobiphenyl

R.T.: 10.435 min
Delta R.T.: -0.002 min
Response: 19096192
Conc: 19.66 ng/ml

Response_ Signal: PP074167.D\ECD2B.ch



#2 Decachlorobiphenyl

R.T.: 8.825 min
Delta R.T.: -0.002 min
Response: 112184396
Conc: 18.66 ng/ml

Report of Analysis

Client:	Earth Engineering Inc.	Date Collected:	08/12/25
Project:	Reserve Turgyan Farms	Date Received:	08/12/25
Client Sample ID:	PIBLK-PP074305.D	SDG No.:	Q2818
Lab Sample ID:	I.BLK-PP074305.D	Matrix:	WATER
Analytical Method:	8082A	% Solid:	0
Sample Wt/Vol:	1000	Units:	mL
Soil Aliquot Vol:			uL
Extraction Type:		Test:	PCB
GPC Factor :	1.0	PH :	
Prep Method :	5030	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PP074305.D	1		08/12/25	PP081225

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
12674-11-2	Aroclor-1016	0.097	U	0.097	0.50	ug/L
11104-28-2	Aroclor-1221	0.13	U	0.13	0.50	ug/L
11141-16-5	Aroclor-1232	0.096	U	0.096	0.50	ug/L
53469-21-9	Aroclor-1242	0.12	U	0.12	0.50	ug/L
12672-29-6	Aroclor-1248	0.071	U	0.071	0.50	ug/L
11097-69-1	Aroclor-1254	0.094	U	0.094	0.50	ug/L
11096-82-5	Aroclor-1260	0.081	U	0.081	0.50	ug/L
37324-23-5	Aroclor-1262	0.14	U	0.14	0.50	ug/L
11100-14-4	Aroclor-1268	0.11	U	0.11	0.50	ug/L
SURROGATES						
877-09-8	Tetrachloro-m-xylene	19.7		60 - 140	98%	SPK: 20
2051-24-3	Decachlorobiphenyl	18.5		60 - 140	93%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP081225\
Data File : PP074305.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Aug 2025 10:43
Operator : YP\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 12 11:31:24 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP080125.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Aug 04 11:01:49 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.661	3.805	24597780	75474847	21.996	19.683
2)	SA Decachlor...	10.437	8.822	20447388	111.5E6	21.055	18.538

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP081225\
Data File : PP074305.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Aug 2025 10:43
Operator : YP\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 12 11:31:24 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP080125.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Aug 04 11:01:49 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.661 min
Delta R.T.: 0.000 min
Response: 24597780
Conc: 22.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
I.BLK




#1 Tetrachloro-m-xylene

R.T.: 3.805 min
Delta R.T.: 0.000 min
Response: 75474847
Conc: 19.68 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.437 min
Delta R.T.: 0.000 min
Response: 20447388
Conc: 21.05 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.822 min
Delta R.T.: -0.004 min
Response: 111481904
Conc: 18.54 ng/ml

Report of Analysis

Client:	Earth Engineering Inc.	Date Collected:	08/12/25
Project:	Reserve Turgyan Farms	Date Received:	08/12/25
Client Sample ID:	PIBLK-PP074319.D	SDG No.:	Q2818
Lab Sample ID:	I.BLK-PP074319.D	Matrix:	WATER
Analytical Method:	8082A	% Solid:	0
Sample Wt/Vol:	1000	Units:	mL
Soil Aliquot Vol:			uL
Extraction Type:		Test:	PCB
GPC Factor :	1.0	PH :	
Prep Method :	5030	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PP074319.D	1		08/12/25	pp081225

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
12674-11-2	Aroclor-1016	0.097	U	0.097	0.50	ug/L
11104-28-2	Aroclor-1221	0.13	U	0.13	0.50	ug/L
11141-16-5	Aroclor-1232	0.096	U	0.096	0.50	ug/L
53469-21-9	Aroclor-1242	0.12	U	0.12	0.50	ug/L
12672-29-6	Aroclor-1248	0.071	U	0.071	0.50	ug/L
11097-69-1	Aroclor-1254	0.094	U	0.094	0.50	ug/L
11096-82-5	Aroclor-1260	0.081	U	0.081	0.50	ug/L
37324-23-5	Aroclor-1262	0.14	U	0.14	0.50	ug/L
11100-14-4	Aroclor-1268	0.11	U	0.11	0.50	ug/L
SURROGATES						
877-09-8	Tetrachloro-m-xylene	19.8		60 - 140	99%	SPK: 20
2051-24-3	Decachlorobiphenyl	19.8		60 - 140	99%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP081225\
Data File : PP074319.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Aug 2025 16:53
Operator : YP\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 13 01:54:48 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP080125.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Aug 04 11:01:49 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.657	3.800	25617664	75726237	22.908	19.749
2)	SA Decachlor...	10.433	8.818	21108609	119.3E6	21.735	19.843

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP081225\
Data File : PP074319.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Aug 2025 16:53
Operator : YP\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 13 01:54:48 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP080125.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Aug 04 11:01:49 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.657 min
Delta R.T.: -0.003 min
Response: 25617664
Conc: 22.91 ng/ml

Instrument :
ECD_P
ClientSampleId :
I.BLK




#1 Tetrachloro-m-xylene

R.T.: 3.800 min
Delta R.T.: -0.005 min
Response: 75726237
Conc: 19.75 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.433 min
Delta R.T.: -0.005 min
Response: 21108609
Conc: 21.74 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.818 min
Delta R.T.: -0.008 min
Response: 119328926
Conc: 19.84 ng/ml

Report of Analysis

Client:	Earth Engineering Inc.		Date Collected:		
Project:	Reserve Turgyan Farms		Date Received:		
Client Sample ID:	PB169205BS		SDG No.:	Q2818	
Lab Sample ID:	PB169205BS		Matrix:	SOIL	
Analytical Method:	8082A		% Solid:	100	Decanted:
Sample Wt/Vol:	30.01	Units: g	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	PCB	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	SW3541B				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PP074309.D	1	08/12/25 08:15	08/12/25 13:05	PB169205

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	185		3.90	17.0	ug/kg
11104-28-2	Aroclor-1221	4.00	U	4.00	17.0	ug/kg
11141-16-5	Aroclor-1232	3.70	U	3.70	17.0	ug/kg
53469-21-9	Aroclor-1242	4.00	U	4.00	17.0	ug/kg
12672-29-6	Aroclor-1248	5.90	U	5.90	17.0	ug/kg
11097-69-1	Aroclor-1254	3.20	U	3.20	17.0	ug/kg
37324-23-5	Aroclor-1262	5.00	U	5.00	17.0	ug/kg
11100-14-4	Aroclor-1268	3.60	U	3.60	17.0	ug/kg
11096-82-5	Aroclor-1260	180		3.20	17.0	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	24.3		32 - 144	121%	SPK: 20
2051-24-3	Decachlorobiphenyl	22.5		32 - 175	112%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP081225\
 Data File : PP074309.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 12 Aug 2025 13:05
 Operator : YP\AJ
 Sample : PB169205BS
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 PB169205BS

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 08/13/2025
 Supervised By :mohammad ahmed 08/14/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 12 13:31:58 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP080125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Aug 04 11:01:49 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.657	3.801	27178701	79780335	24.303	20.806
2) SA Decachlor...	10.438	8.821	21836747	120.3E6	22.485	20.000
Target Compounds						
3) L1 AR-1016-1	5.809	4.902	22669527	228.6E6	549.474	573.187
4) L1 AR-1016-2	5.830	4.959	34110297	104.5E6	561.563	556.666
5) L1 AR-1016-3	5.893	5.080	22276842	56001049	561.688	527.965
6) L1 AR-1016-4	5.990	5.121	18317589	56446484	562.848	516.814
7) L1 AR-1016-5	6.283	5.334	17574695	66942868	541.292	572.341
31) L7 AR-1260-1	7.401	6.551	31033148	236.1E6	550.662	584.486
32) L7 AR-1260-2	7.653	6.706	36070118	182.5E6	529.257	583.716
33) L7 AR-1260-3	8.012	6.915	24314378	204.9E6	456.215	519.977
34) L7 AR-1260-4	8.238	7.175	30953569	149.4E6	494.348m	513.385
35) L7 AR-1260-5	8.566	7.415	51128845	382.8E6	451.250	505.423

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP081225\
Data File : PP074309.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Aug 2025 13:05
Operator : YP\AJ
Sample : PB169205BS
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :

ECD_P

ClientSampleId :

PB169205BS

Manual Integrations



APPROVED

Reviewed By :Yogesh Patel 08/13/2025

Supervised By :mohammad ahmed 08/14/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 12 13:31:58 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP080125.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Aug 04 11:01:49 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Report of Analysis

Client:	Earth Engineering Inc.	Date Collected:	08/11/25
Project:	Reserve Turgyan Farms	Date Received:	08/11/25
Client Sample ID:	BIN0009-DRIVEWAY-TP-SOUTH-EASTMS	SDG No.:	Q2818
Lab Sample ID:	Q2830-01MS	Matrix:	SOIL
Analytical Method:	8082A	% Solid:	86.9
Sample Wt/Vol:	30.01	Units:	g
Soil Aliquot Vol:			uL
Extraction Type:		Test:	PCB
GPC Factor :	1.0	PH :	
Prep Method :	SW3541B	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PP074311.D	1	08/12/25 08:15	08/12/25 13:37	PB169205

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	213		4.50	19.6	ug/kg
11104-28-2	Aroclor-1221	4.60	U	4.60	19.6	ug/kg
11141-16-5	Aroclor-1232	4.30	U	4.30	19.6	ug/kg
53469-21-9	Aroclor-1242	4.60	U	4.60	19.6	ug/kg
12672-29-6	Aroclor-1248	6.80	U	6.80	19.6	ug/kg
11097-69-1	Aroclor-1254	3.70	U	3.70	19.6	ug/kg
37324-23-5	Aroclor-1262	5.80	U	5.80	19.6	ug/kg
11100-14-4	Aroclor-1268	4.10	U	4.10	19.6	ug/kg
11096-82-5	Aroclor-1260	181		3.70	19.6	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	23.5		32 - 144	117%	SPK: 20
2051-24-3	Decachlorobiphenyl	19.9		32 - 175	100%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP081225\
 Data File : PP074311.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 12 Aug 2025 13:37
 Operator : YP\AJ
 Sample : Q2830-01MS
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 BIN0009-DRIVEWAY-TP-SOUTH-EASTMS

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 08/13/2025
 Supervised By :mohammad ahmed 08/14/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 12 14:50:12 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP080125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Aug 04 11:01:49 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.655	3.800	26250208	84433109	23.473m	22.020
2) SA Decachlor...	10.437	8.820	19335273	105.2E6	19.909	17.488
Target Compounds						
3) L1 AR-1016-1	5.809	4.901	21480557	221.4E6	520.655	555.093
4) L1 AR-1016-2	5.831	4.958	31794685	105.4E6	523.441	561.178
5) L1 AR-1016-3	5.894	5.079	20638950	60560553	520.390	570.951
6) L1 AR-1016-4	5.991	5.119	17736155	58694992	544.982	537.401
7) L1 AR-1016-5	6.283	5.334	16474757	64653966	507.414	552.772
31) L7 AR-1260-1	7.400	6.550	28140334	207.0E6	499.331	512.404
32) L7 AR-1260-2	7.653	6.704	32575808	157.6E6	477.985	503.902
33) L7 AR-1260-3	8.012	6.914	21771827	176.5E6	408.508	447.876
34) L7 AR-1260-4	8.238	7.175	29533684	132.7E6	471.671m	455.876
35) L7 AR-1260-5	8.566	7.413	45863963	335.1E6	404.784	442.376

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP081225\
Data File : PP074311.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Aug 2025 13:37
Operator : YP\AJ
Sample : Q2830-01MS
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :

ECD_P

ClientSampleId :

BIN0009-DRIVEWAY-TP-SOUTH-EASTMS

Manual Integrations



APPROVED

Reviewed By :Yogesh Patel 08/13/2025

Supervised By :mohammad ahmed 08/14/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 12 14:50:12 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP080125.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Aug 04 11:01:49 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Report of Analysis

Client:	Earth Engineering Inc.	Date Collected:	08/11/25
Project:	Reserve Turgyan Farms	Date Received:	08/11/25
Client Sample ID:	BIN0009-DRIVEWAY-TP-SOUTH-EASTMSD	SDG No.:	Q2818
Lab Sample ID:	Q2830-01MSD	Matrix:	SOIL
Analytical Method:	8082A	% Solid:	86.9
Sample Wt/Vol:	30.03	Units:	g
Soil Aliquot Vol:			uL
Extraction Type:		Test:	PCB
GPC Factor :	1.0	Injection Volume :	
Prep Method :	SW3541B		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PP074312.D	1	08/12/25 08:15	08/12/25 13:53	PB169205

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	219		4.50	19.5	ug/kg
11104-28-2	Aroclor-1221	4.60	U	4.60	19.5	ug/kg
11141-16-5	Aroclor-1232	4.30	U	4.30	19.5	ug/kg
53469-21-9	Aroclor-1242	4.60	U	4.60	19.5	ug/kg
12672-29-6	Aroclor-1248	6.80	U	6.80	19.5	ug/kg
11097-69-1	Aroclor-1254	3.70	U	3.70	19.5	ug/kg
37324-23-5	Aroclor-1262	5.80	U	5.80	19.5	ug/kg
11100-14-4	Aroclor-1268	4.10	U	4.10	19.5	ug/kg
11096-82-5	Aroclor-1260	188		3.70	19.5	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	23.3		32 - 144	116%	SPK: 20
2051-24-3	Decachlorobiphenyl	19.9		32 - 175	99%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP081225\
 Data File : PP074312.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 12 Aug 2025 13:53
 Operator : YP\AJ
 Sample : Q2830-01MSD
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 BIN0009-DRIVEWAY-TP-SOUTH-EASTMSD

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 08/13/2025
 Supervised By :mohammad ahmed 08/14/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 12 14:50:47 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP080125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Aug 04 11:01:49 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.658	3.803	26034862	84782674	23.281m	22.111
2) SA Decachlor...	10.435	8.820	19291425	107.3E6	19.864m	17.849
Target Compounds						
3) L1 AR-1016-1	5.811	4.904	22025300	229.5E6	533.859	575.363
4) L1 AR-1016-2	5.832	4.961	31957321	109.0E6	526.118	580.403
5) L1 AR-1016-3	5.895	5.081	21162376	59483478	533.588	560.796
6) L1 AR-1016-4	5.992	5.122	21722805	62077143	667.481	568.368
7) L1 AR-1016-5	6.284	5.337	16835913	66799585	518.538	571.116
31) L7 AR-1260-1	7.402	6.552	28910383	216.7E6	512.995	536.408
32) L7 AR-1260-2	7.654	6.707	33152517	164.0E6	486.447	524.348
33) L7 AR-1260-3	8.012	6.916	22216794	183.6E6	416.857	465.809
34) L7 AR-1260-4	8.239	7.176	29914044	137.0E6	477.746m	470.828
35) L7 AR-1260-5	8.565	7.415	47018435	347.3E6	414.973	458.490

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP081225\
Data File : PP074312.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Aug 2025 13:53
Operator : YP\AJ
Sample : Q2830-01MSD
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :

ECD_P

ClientSampleId :

BIN0009-DRIVEWAY-TP-SOUTH-EASTMSD

Manual Integrations



APPROVED

Reviewed By :Yogesh Patel 08/13/2025

Supervised By :mohammad ahmed 08/14/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 12 14:50:47 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP080125.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Aug 04 11:01:49 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Manual Integration Report

Sequence:	PO072325	Instrument	ECD_o
-----------	----------	------------	-------

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
AR1660ICC250	PO112417.D	AR-1260-1	yogesh	7/24/2025 7:42:50 AM	mohammad	7/25/2025 3:53:10	Peak Integrated by Software
AR1660ICC250	PO112417.D	AR-1260-1 #2	yogesh	7/24/2025 7:42:50 AM	mohammad	7/25/2025 3:53:10	Peak Integrated by Software
AR1660ICC250	PO112417.D	AR-1260-2	yogesh	7/24/2025 7:42:50 AM	mohammad	7/25/2025 3:53:10	Peak Integrated by Software
AR1660ICC250	PO112417.D	AR-1260-3 #2	yogesh	7/24/2025 7:42:50 AM	mohammad	7/25/2025 3:53:10	Peak Integrated by Software
AR1660ICC050	PO112418.D	AR-1016-5	yogesh	7/24/2025 7:42:52 AM	mohammad	7/25/2025 3:53:10	Peak Integrated by Software
AR1660ICC050	PO112418.D	AR-1016-5 #2	yogesh	7/24/2025 7:42:52 AM	mohammad	7/25/2025 3:53:10	Peak Integrated by Software
AR1660ICC050	PO112418.D	AR-1260-1 #2	yogesh	7/24/2025 7:42:52 AM	mohammad	7/25/2025 3:53:10	Peak Integrated by Software
AR1660ICC050	PO112418.D	AR-1260-3 #2	yogesh	7/24/2025 7:42:52 AM	mohammad	7/25/2025 3:53:10	Peak Integrated by Software
AR1242ICC250	PO112424.D	AR-1242-5	yogesh	7/24/2025 7:42:53 AM	mohammad	7/25/2025 3:53:10	Peak Integrated by Software
AR1242ICC250	PO112424.D	AR-1242-5 #2	yogesh	7/24/2025 7:42:53 AM	mohammad	7/25/2025 3:53:10	Peak Integrated by Software
AR1242ICC050	PO112425.D	AR-1242-5	yogesh	7/24/2025 7:42:55 AM	mohammad	7/25/2025 3:53:10	Peak Integrated by Software
AR1242ICC050	PO112425.D	AR-1242-5 #2	yogesh	7/24/2025 7:42:55 AM	mohammad	7/25/2025 3:53:10	Peak Integrated by Software
AR1242ICC050	PO112425.D	Tetrachloro-m-xylene #2	yogesh	7/24/2025 7:42:55 AM	mohammad	7/25/2025 3:53:10	Peak Integrated by Software

Manual Integration Report

Sequence:

PO072325

Instrument

ECD_o

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
AR1248ICC050	PO112430.D	AR-1248-4	yogesh	7/24/2025 7:42:57 AM	mohammad	7/25/2025 3:53:10	Peak Integrated by Software
AR1248ICC050	PO112430.D	AR-1248-4 #2	yogesh	7/24/2025 7:42:57 AM	mohammad	7/25/2025 3:53:10	Peak Integrated by Software
AR1248ICC050	PO112430.D	AR-1248-5	yogesh	7/24/2025 7:42:57 AM	mohammad	7/25/2025 3:53:10	Peak Integrated by Software
AR1248ICC050	PO112430.D	AR-1248-5 #2	yogesh	7/24/2025 7:42:57 AM	mohammad	7/25/2025 3:53:10	Peak Integrated by Software
AR1254ICC050	PO112435.D	AR-1254-1	yogesh	7/24/2025 7:42:59 AM	mohammad	7/25/2025 3:53:10	Peak Integrated by Software
AR1254ICC050	PO112435.D	AR-1254-1 #2	yogesh	7/24/2025 7:42:59 AM	mohammad	7/25/2025 3:53:10	Peak Integrated by Software
AR1254ICC050	PO112435.D	AR-1254-2	yogesh	7/24/2025 7:42:59 AM	mohammad	7/25/2025 3:53:10	Peak Integrated by Software
AR1254ICC050	PO112435.D	AR-1254-2 #2	yogesh	7/24/2025 7:42:59 AM	mohammad	7/25/2025 3:53:10	Peak Integrated by Software
AR1254ICC050	PO112435.D	AR-1254-3 #2	yogesh	7/24/2025 7:42:59 AM	mohammad	7/25/2025 3:53:10	Peak Integrated by Software

Manual Integration Report

Sequence:

PO081225

Instrument

ECD_o

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
AR1660CCC500	PO112850.D	AR-1016-1	yogesh	8/13/2025 11:49:43 AM	mohammad	8/14/2025 1:21:40	Peak Integrated by Software
AR1660CCC500	PO112850.D	AR-1016-2	yogesh	8/13/2025 11:49:43 AM	mohammad	8/14/2025 1:21:40	Peak Integrated by Software
AR1660CCC500	PO112850.D	AR-1016-3 #2	yogesh	8/13/2025 11:49:43 AM	mohammad	8/14/2025 1:21:40	Peak Integrated by Software
AR1660CCC500	PO112850.D	AR-1016-4 #2	yogesh	8/13/2025 11:49:43 AM	mohammad	8/14/2025 1:21:40	Peak Integrated by Software
AR1660CCC500	PO112850.D	AR-1260-3 #2	yogesh	8/13/2025 11:49:43 AM	mohammad	8/14/2025 1:21:40	Peak Integrated by Software
AR1242CCC500	PO112851.D	AR-1242-1	yogesh	8/13/2025 11:49:45 AM	mohammad	8/14/2025 1:21:40	Peak Integrated by Software
AR1242CCC500	PO112851.D	AR-1242-2	yogesh	8/13/2025 11:49:45 AM	mohammad	8/14/2025 1:21:40	Peak Integrated by Software
AR1242CCC500	PO112851.D	AR-1242-3 #2	yogesh	8/13/2025 11:49:45 AM	mohammad	8/14/2025 1:21:40	Peak Integrated by Software
AR1248CCC500	PO112852.D	AR-1248-1	yogesh	8/13/2025 11:49:47 AM	mohammad	8/14/2025 1:21:40	Peak Integrated by Software
Q2818-01	PO112858.D	Decachlorobiphenyl	yogesh	8/13/2025 7:36:42 AM	mohammad	8/14/2025 1:21:40	Peak Integrated by Software
AR1660CCC500	PO112865.D	AR-1016-1	yogesh	8/13/2025 11:49:48 AM	mohammad	8/14/2025 1:21:40	Peak Integrated by Software
AR1660CCC500	PO112865.D	AR-1016-2	yogesh	8/13/2025 11:49:48 AM	mohammad	8/14/2025 1:21:40	Peak Integrated by Software
AR1660CCC500	PO112865.D	AR-1260-3 #2	yogesh	8/13/2025 11:49:48 AM	mohammad	8/14/2025 1:21:40	Peak Integrated by Software

Manual Integration Report

Sequence:	PO081225	Instrument	ECD_o
-----------	----------	------------	-------

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
AR1242CCC500	PO112866.D	AR-1242-1	yogesh	8/13/2025 11:49:52 AM	mohammad	8/14/2025 1:21:40	Peak Integrated by Software
AR1242CCC500	PO112866.D	AR-1242-2	yogesh	8/13/2025 11:49:52 AM	mohammad	8/14/2025 1:21:40	Peak Integrated by Software
AR1242CCC500	PO112866.D	AR-1242-3 #2	yogesh	8/13/2025 11:49:52 AM	mohammad	8/14/2025 1:21:40	Peak Integrated by Software
AR1660CCC500	PO112880.D	AR-1016-1	yogesh	8/13/2025 11:49:56 AM	mohammad	8/14/2025 1:21:40	Peak Integrated by Software
AR1660CCC500	PO112880.D	AR-1016-2	yogesh	8/13/2025 11:49:56 AM	mohammad	8/14/2025 1:21:40	Peak Integrated by Software
AR1660CCC500	PO112880.D	AR-1260-3 #2	yogesh	8/13/2025 11:49:56 AM	mohammad	8/14/2025 1:21:40	Peak Integrated by Software
AR1242CCC500	PO112881.D	AR-1242-1	yogesh	8/13/2025 11:49:58 AM	mohammad	8/14/2025 1:21:40	Peak Integrated by Software
AR1242CCC500	PO112881.D	AR-1242-2	yogesh	8/13/2025 11:49:58 AM	mohammad	8/14/2025 1:21:40	Peak Integrated by Software
AR1660CCC500	PO112890.D	AR-1016-1	yogesh	8/13/2025 11:50:01 AM	mohammad	8/14/2025 1:21:40	Peak Integrated by Software
AR1660CCC500	PO112890.D	AR-1016-2	yogesh	8/13/2025 11:50:01 AM	mohammad	8/14/2025 1:21:40	Peak Integrated by Software
AR1242CCC500	PO112891.D	AR-1242-1	yogesh	8/13/2025 11:50:03 AM	mohammad	8/14/2025 1:21:40	Peak Integrated by Software
AR1242CCC500	PO112891.D	AR-1242-2	yogesh	8/13/2025 11:50:03 AM	mohammad	8/14/2025 1:21:40	Peak Integrated by Software
AR1242CCC500	PO112891.D	AR-1242-3 #2	yogesh	8/13/2025 11:50:03 AM	mohammad	8/14/2025 1:21:40	Peak Integrated by Software

Manual Integration Report

Sequence:

PO081225

Instrument

ECD_o

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
AR1242CCC500	PO112891.D	AR-1242-4 #2	yogesh	8/13/2025 11:50:03 AM	mohammad	8/14/2025 1:21:40	Peak Integrated by Software
AR1254CCC500	PO112893.D	AR-1254-1	yogesh	8/13/2025 7:36:58 AM	mohammad	8/14/2025 1:21:40	Peak Integrated by Software
AR1254CCC500	PO112893.D	AR-1254-2	yogesh	8/13/2025 7:36:58 AM	mohammad	8/14/2025 1:21:40	Peak Integrated by Software

Manual Integration Report

Sequence:	PP080125	Instrument	ECD_p
-----------	----------	------------	-------

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
AR1660ICC250	PP074171.D	AR-1016-5	yogesh	8/5/2025 7:23:21 AM	mohammad	8/8/2025 7:37:53	Peak Integrated by Software
AR1660ICC250	PP074171.D	AR-1260-1	yogesh	8/5/2025 7:23:21 AM	mohammad	8/8/2025 7:37:53	Peak Integrated by Software
AR1660ICC250	PP074171.D	AR-1260-2	yogesh	8/5/2025 7:23:21 AM	mohammad	8/8/2025 7:37:53	Peak Integrated by Software
AR1660ICC050	PP074172.D	AR-1260-2	yogesh	8/5/2025 7:22:19 AM	mohammad	8/8/2025 7:37:53	Peak Integrated by Software
AR1660ICC050	PP074172.D	AR-1260-2 #2	yogesh	8/5/2025 7:22:19 AM	mohammad	8/8/2025 7:37:53	Peak Integrated by Software
AR1660ICC050	PP074172.D	AR-1260-4	yogesh	8/5/2025 7:22:19 AM	mohammad	8/8/2025 7:37:53	Peak Integrated by Software
AR1660ICC050	PP074172.D	Tetrachloro-m-xylene	yogesh	8/5/2025 7:22:19 AM	mohammad	8/8/2025 7:37:53	Peak Integrated by Software
AR1660ICC050	PP074172.D	Tetrachloro-m-xylene #2	yogesh	8/5/2025 7:22:19 AM	mohammad	8/8/2025 7:37:53	Peak Integrated by Software
AR1232ICC500	PP074174.D	AR-1232-4 #2	yogesh	8/5/2025 7:22:20 AM	mohammad	8/8/2025 7:37:53	Peak Integrated by Software
AR1242ICC1000	PP074175.D	AR-1242-4 #2	yogesh	8/5/2025 7:22:22 AM	mohammad	8/8/2025 7:37:53	Peak Integrated by Software
AR1242ICC750	PP074176.D	AR-1242-4 #2	yogesh	8/5/2025 7:22:24 AM	mohammad	8/8/2025 7:37:53	Peak Integrated by Software
AR1242ICC500	PP074177.D	AR-1242-4 #2	yogesh	8/5/2025 7:22:27 AM	mohammad	8/8/2025 7:37:53	Peak Integrated by Software
AR1242ICC250	PP074178.D	AR-1242-4 #2	yogesh	8/5/2025 7:22:29 AM	mohammad	8/8/2025 7:37:53	Peak Integrated by Software

Manual Integration Report

Sequence:	PP080125	Instrument	ECD_p
-----------	----------	------------	-------

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
AR1242ICC250	PP074178.D	AR-1242-5	yogesh	8/5/2025 7:22:29 AM	mohammad	8/8/2025 7:37:53	Peak Integrated by Software
AR1242ICC050	PP074179.D	AR-1242-5	yogesh	8/5/2025 7:22:32 AM	mohammad	8/8/2025 7:37:53	Peak Integrated by Software
AR1242ICC050	PP074179.D	Tetrachloro-m-xylene	yogesh	8/5/2025 7:22:32 AM	mohammad	8/8/2025 7:37:53	Peak Integrated by Software
AR1242ICC050	PP074179.D	Tetrachloro-m-xylene #2	yogesh	8/5/2025 7:22:32 AM	mohammad	8/8/2025 7:37:53	Peak Integrated by Software
AR1248ICC500	PP074182.D	AR-1248-3 #2	yogesh	8/5/2025 7:22:38 AM	mohammad	8/8/2025 7:37:53	Peak Integrated by Software
AR1254ICC1000	PP074185.D	AR-1254-1	yogesh	8/5/2025 7:22:44 AM	mohammad	8/8/2025 7:37:53	Peak Integrated by Software
AR1254ICC1000	PP074185.D	AR-1254-5 #2	yogesh	8/5/2025 7:22:44 AM	mohammad	8/8/2025 7:37:53	Peak Integrated by Software
AR1254ICC750	PP074186.D	AR-1254-1	yogesh	8/5/2025 7:22:45 AM	mohammad	8/8/2025 7:37:53	Peak Integrated by Software
AR1254ICC500	PP074187.D	AR-1254-1	yogesh	8/5/2025 7:22:47 AM	mohammad	8/8/2025 7:37:53	Peak Integrated by Software
AR1254ICC250	PP074188.D	AR-1254-1	yogesh	8/5/2025 7:22:49 AM	mohammad	8/8/2025 7:37:53	Peak Integrated by Software
AR1254ICC050	PP074189.D	AR-1254-1	yogesh	8/5/2025 7:22:51 AM	mohammad	8/8/2025 7:37:53	Peak Integrated by Software
AR1242ICV500	PP074197.D	AR-1242-4 #2	yogesh	8/5/2025 7:22:52 AM	mohammad	8/8/2025 7:37:53	Peak Integrated by Software
AR1248ICV500	PP074198.D	AR-1248-3 #2	yogesh	8/5/2025 7:22:54 AM	mohammad	8/8/2025 7:37:53	Peak Integrated by Software



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

Manual Integration Report

Sequence:	PP080125	Instrument	ECD_p
-----------	----------	------------	-------

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
AR1254ICV500	PP074199.D	AR-1254-1	yogesh	8/8/2025 7:32:21 AM	mohammad	8/8/2025 7:37:53	Peak Integrated by Software

Manual Integration Report

Sequence:	PP081225	Instrument	ECD_p
-----------	----------	------------	-------

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
AR1242CCC500	PP074303.D	AR-1242-3 #2	yogesh	8/13/2025 7:35:08 AM	mohammad	8/14/2025 1:21:51	Peak Integrated by Software
AR1242CCC500	PP074303.D	AR-1242-4 #2	yogesh	8/13/2025 7:35:08 AM	mohammad	8/14/2025 1:21:51	Peak Integrated by Software
AR1254CCC500	PP074304.D	AR-1254-1	yogesh	8/13/2025 7:35:10 AM	mohammad	8/14/2025 1:21:51	Peak Integrated by Software
AR1254CCC500	PP074304.D	AR-1254-1 #2	yogesh	8/13/2025 7:35:10 AM	mohammad	8/14/2025 1:21:51	Peak Integrated by Software
AR1254CCC500	PP074304.D	AR-1254-3 #2	yogesh	8/13/2025 7:35:10 AM	mohammad	8/14/2025 1:21:51	Peak Integrated by Software
PB169205BS	PP074309.D	AR-1260-4	yogesh	8/13/2025 7:35:14 AM	mohammad	8/14/2025 1:21:51	Peak Integrated by Software
Q2830-01MS	PP074311.D	AR-1260-4	yogesh	8/13/2025 7:35:18 AM	mohammad	8/14/2025 1:21:51	Peak Integrated by Software
Q2830-01MS	PP074311.D	Tetrachloro-m-xylene	yogesh	8/13/2025 7:35:18 AM	mohammad	8/14/2025 1:21:51	Peak Integrated by Software
Q2830-01MSD	PP074312.D	AR-1260-4	yogesh	8/13/2025 7:35:19 AM	mohammad	8/14/2025 1:21:51	Peak Integrated by Software
Q2830-01MSD	PP074312.D	Decachlorobiphenyl	yogesh	8/13/2025 7:35:19 AM	mohammad	8/14/2025 1:21:51	Peak Integrated by Software
Q2830-01MSD	PP074312.D	Tetrachloro-m-xylene	yogesh	8/13/2025 7:35:19 AM	mohammad	8/14/2025 1:21:51	Peak Integrated by Software
AR1660CCC500	PP074316.D	AR-1016-2 #2	yogesh	8/13/2025 7:35:23 AM	mohammad	8/14/2025 1:21:51	Peak Integrated by Software
AR1660CCC500	PP074316.D	AR-1016-5 #2	yogesh	8/13/2025 7:35:23 AM	mohammad	8/14/2025 1:21:51	Peak Integrated by Software

Manual Integration Report

Sequence:	PP081225	Instrument	ECD_p
-----------	----------	------------	-------

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
AR1660CCC500	PP074316.D	Tetrachloro-m-xylene #2	yogesh	8/13/2025 7:35:23 AM	mohammad	8/14/2025 1:21:51	Peak Integrated by Software
AR1242CCC500	PP074317.D	AR-1242-2 #2	yogesh	8/13/2025 7:35:24 AM	mohammad	8/14/2025 1:21:51	Peak Integrated by Software
AR1242CCC500	PP074317.D	AR-1242-3 #2	yogesh	8/13/2025 7:35:24 AM	mohammad	8/14/2025 1:21:51	Peak Integrated by Software
AR1242CCC500	PP074317.D	AR-1242-4 #2	yogesh	8/13/2025 7:35:24 AM	mohammad	8/14/2025 1:21:51	Peak Integrated by Software
AR1254CCC500	PP074318.D	AR-1254-1	yogesh	8/13/2025 7:35:26 AM	mohammad	8/14/2025 1:21:51	Peak Integrated by Software
AR1254CCC500	PP074318.D	AR-1254-1 #2	yogesh	8/13/2025 7:35:26 AM	mohammad	8/14/2025 1:21:51	Peak Integrated by Software
AR1254CCC500	PP074318.D	AR-1254-2 #2	yogesh	8/13/2025 7:35:26 AM	mohammad	8/14/2025 1:21:51	Peak Integrated by Software
AR1254CCC500	PP074318.D	AR-1254-3 #2	yogesh	8/13/2025 7:35:26 AM	mohammad	8/14/2025 1:21:51	Peak Integrated by Software
AR1254CCC500	PP074318.D	AR-1254-4 #2	yogesh	8/13/2025 7:35:26 AM	mohammad	8/14/2025 1:21:51	Peak Integrated by Software
AR1254CCC500	PP074318.D	AR-1254-5 #2	yogesh	8/13/2025 7:35:26 AM	mohammad	8/14/2025 1:21:51	Peak Integrated by Software
AR1242CCC500	PP074330.D	AR-1242-2 #2	yogesh	8/13/2025 7:35:40 AM	mohammad	8/14/2025 1:21:51	Peak Integrated by Software
AR1242CCC500	PP074330.D	AR-1242-3 #2	yogesh	8/13/2025 7:35:40 AM	mohammad	8/14/2025 1:21:51	Peak Integrated by Software
AR1242CCC500	PP074330.D	AR-1242-4 #2	yogesh	8/13/2025 7:35:40 AM	mohammad	8/14/2025 1:21:51	Peak Integrated by Software

Manual Integration Report

Sequence:

PP081225

Instrument

ECD_p

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
AR1254CCC500	PP074331.D	AR-1254-1	yogesh	8/13/2025 7:35:42 AM	mohammad	8/14/2025 1:21:51	Peak Integrated by Software
AR1254CCC500	PP074331.D	AR-1254-2 #2	yogesh	8/13/2025 7:35:42 AM	mohammad	8/14/2025 1:21:51	Peak Integrated by Software
AR1254CCC500	PP074331.D	AR-1254-3 #2	yogesh	8/13/2025 7:35:42 AM	mohammad	8/14/2025 1:21:51	Peak Integrated by Software
AR1254CCC500	PP074331.D	AR-1254-4 #2	yogesh	8/13/2025 7:35:42 AM	mohammad	8/14/2025 1:21:51	Peak Integrated by Software
AR1254CCC500	PP074331.D	AR-1254-5 #2	yogesh	8/13/2025 7:35:42 AM	mohammad	8/14/2025 1:21:51	Peak Integrated by Software

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO072325

Review By	yogesh	Review On	7/23/2025 3:39:31 PM
Supervise By	mohammad	Supervise On	7/25/2025 3:53:10 AM
SubDirectory	PO072325	HP Acquire Method	HP Processing Method PO072325
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344 ,PP24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359 ,PP24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369 PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP2		

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	HEXANE	PO112412.D	23 Jul 2025 10:56	YP/AJ	Ok
2	I.BLK	PO112413.D	23 Jul 2025 11:14	YP/AJ	Ok
3	AR1660ICC1000	PO112414.D	23 Jul 2025 11:32	YP/AJ	Ok
4	AR1660ICC750	PO112415.D	23 Jul 2025 11:50	YP/AJ	Ok
5	AR1660ICC500	PO112416.D	23 Jul 2025 12:08	YP/AJ	Ok
6	AR1660ICC250	PO112417.D	23 Jul 2025 12:27	YP/AJ	Ok,M
7	AR1660ICC050	PO112418.D	23 Jul 2025 12:45	YP/AJ	Ok,M
8	AR1221ICC500	PO112419.D	23 Jul 2025 13:03	YP/AJ	Ok
9	AR1232ICC500	PO112420.D	23 Jul 2025 13:22	YP/AJ	Ok
10	AR1242ICC1000	PO112421.D	23 Jul 2025 13:40	YP/AJ	Ok
11	AR1242ICC750	PO112422.D	23 Jul 2025 13:59	YP/AJ	Ok
12	AR1242ICC500	PO112423.D	23 Jul 2025 14:17	YP/AJ	Ok
13	AR1242ICC250	PO112424.D	23 Jul 2025 14:36	YP/AJ	Ok,M
14	AR1242ICC050	PO112425.D	23 Jul 2025 14:54	YP/AJ	Ok,M
15	AR1248ICC1000	PO112426.D	23 Jul 2025 15:13	YP/AJ	Ok
16	AR1248ICC750	PO112427.D	23 Jul 2025 15:31	YP/AJ	Ok
17	AR1248ICC500	PO112428.D	23 Jul 2025 15:48	YP/AJ	Ok
18	AR1248ICC250	PO112429.D	23 Jul 2025 16:07	YP/AJ	Ok
19	AR1248ICC050	PO112430.D	23 Jul 2025 16:25	YP/AJ	Ok,M
20	AR1254ICC1000	PO112431.D	23 Jul 2025 16:44	YP/AJ	Ok
21	AR1254ICC750	PO112432.D	23 Jul 2025 17:02	YP/AJ	Ok

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO072325

Review By	yogesh	Review On	7/23/2025 3:39:31 PM
Supervise By	mohammad	Supervise On	7/25/2025 3:53:10 AM
SubDirectory	PO072325	HP Acquire Method	HP Processing Method PO072325
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,PP24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369		
CCC			
Internal Standard/PEM			
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP2		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

22	AR1254ICC500	PO112433.D	23 Jul 2025 17:21	YP/AJ	Ok
23	AR1254ICC250	PO112434.D	23 Jul 2025 17:39	YP/AJ	Ok
24	AR1254ICC050	PO112435.D	23 Jul 2025 17:57	YP/AJ	Ok,M
25	AR1262ICC500	PO112436.D	23 Jul 2025 18:16	YP/AJ	Ok
26	AR1268ICC1000	PO112437.D	23 Jul 2025 18:34	YP/AJ	Ok
27	AR1268ICC750	PO112438.D	23 Jul 2025 18:53	YP/AJ	Ok
28	AR1268ICC500	PO112439.D	23 Jul 2025 19:11	YP/AJ	Ok
29	AR1268ICC250	PO112440.D	23 Jul 2025 19:28	YP/AJ	Ok
30	AR1268ICC050	PO112441.D	23 Jul 2025 19:47	YP/AJ	Ok
31	PO072325ICV500	PO112442.D	23 Jul 2025 20:05	YP/AJ	Ok
32	AR1242ICV500	PO112443.D	23 Jul 2025 20:42	YP/AJ	Ok
33	AR1248ICV500	PO112444.D	23 Jul 2025 21:18	YP/AJ	Ok
34	AR1254ICV500	PO112445.D	23 Jul 2025 21:54	YP/AJ	Ok
35	AR1268ICV500	PO112446.D	23 Jul 2025 22:31	YP/AJ	Ok
36	DDT ANALOG	PO112447.D	23 Jul 2025 23:07	YP/AJ	Ok

M : Manual Integration

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO081225

Review By	yogesh	Review On	8/12/2025 10:12:49 AM
Supervise By	mohammad	Supervise On	8/14/2025 1:21:40 AM
SubDirectory	PO081225	HP Acquire Method	HP Processing Method PO072325
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,PP24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369		
CCC	PP24332,PP24347,PP24352,PP24357		
Internal Standard/PEM			
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP2		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	HEXANE	PO112849.D	12 Aug 2025 08:35	YP/AJ	Ok
2	AR1660CCC500	PO112850.D	12 Aug 2025 08:54	YP/AJ	Ok,M
3	AR1242CCC500	PO112851.D	12 Aug 2025 09:12	YP/AJ	Ok,M
4	AR1248CCC500	PO112852.D	12 Aug 2025 09:29	YP/AJ	Ok,M
5	AR1254CCC500	PO112853.D	12 Aug 2025 09:48	YP/AJ	Ok
6	I.BLK	PO112854.D	12 Aug 2025 10:05	YP/AJ	Ok
7	DDT ANALOG	PO112855.D	12 Aug 2025 10:24	YP/AJ	Ok
8	Q2808-01DL	PO112856.D	12 Aug 2025 10:42	YP/AJ	Ok,M
9	Q2807-02	PO112857.D	12 Aug 2025 11:01	YP/AJ	Ok
10	Q2818-01	PO112858.D	12 Aug 2025 12:50	YP/AJ	Ok,M
11	Q2818-02	PO112859.D	12 Aug 2025 13:08	YP/AJ	Ok
12	Q2819-06	PO112860.D	12 Aug 2025 13:26	YP/AJ	Ok
13	Q2819-11	PO112861.D	12 Aug 2025 13:43	YP/AJ	Ok
14	Q2819-14	PO112862.D	12 Aug 2025 14:01	YP/AJ	Ok
15	Q2820-09	PO112863.D	12 Aug 2025 14:20	YP/AJ	ReRun
16	Q2820-10	PO112864.D	12 Aug 2025 14:38	YP/AJ	Ok
17	AR1660CCC500	PO112865.D	12 Aug 2025 15:50	YP/AJ	Ok,M
18	AR1242CCC500	PO112866.D	12 Aug 2025 16:07	YP/AJ	Ok,M
19	AR1248CCC500	PO112867.D	12 Aug 2025 16:25	YP/AJ	Ok
20	AR1254CCC500	PO112868.D	12 Aug 2025 16:42	YP/AJ	Ok
21	I.BLK	PO112869.D	12 Aug 2025 17:00	YP/AJ	Ok

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO081225

Review By	yogesh	Review On	8/12/2025 10:12:49 AM
Supervise By	mohammad	Supervise On	8/14/2025 1:21:40 AM
SubDirectory	PO081225	HP Acquire Method	HP Processing Method PO072325
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,PP24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369		
CCC	PP24332,PP24347,PP24352,PP24357		
Internal Standard/PEM			
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP2		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

22	Q2820-24	PO112870.D	12 Aug 2025 17:19	YP/AJ	Ok
23	Q2823-01	PO112871.D	12 Aug 2025 17:37	YP/AJ	Ok,M
24	Q2826-01	PO112872.D	12 Aug 2025 17:56	YP/AJ	Ok
25	Q2827-01	PO112873.D	12 Aug 2025 18:14	YP/AJ	Dilution
26	Q2827-05	PO112874.D	12 Aug 2025 18:31	YP/AJ	Ok
27	Q2830-03	PO112875.D	12 Aug 2025 18:49	YP/AJ	Ok
28	Q2839-01	PO112876.D	12 Aug 2025 19:06	YP/AJ	Ok,M
29	Q2839-02	PO112877.D	12 Aug 2025 19:24	YP/AJ	Ok,M
30	Q2839-03	PO112878.D	12 Aug 2025 19:42	YP/AJ	Ok
31	Q2839-04	PO112879.D	12 Aug 2025 19:59	YP/AJ	Ok,M
32	AR1660CCC500	PO112880.D	12 Aug 2025 21:09	YP/AJ	Ok,M
33	AR1242CCC500	PO112881.D	12 Aug 2025 22:02	YP/AJ	Ok,M
34	AR1248CCC500	PO112882.D	12 Aug 2025 22:21	YP/AJ	Ok
35	AR1254CCC500	PO112883.D	12 Aug 2025 22:38	YP/AJ	Ok
36	I.BLK	PO112884.D	12 Aug 2025 22:56	YP/AJ	Ok
37	Q2839-05	PO112885.D	12 Aug 2025 23:15	YP/AJ	Ok
38	Q2839-06	PO112886.D	12 Aug 2025 23:33	YP/AJ	Ok
39	Q2839-07	PO112887.D	12 Aug 2025 23:51	YP/AJ	Ok,M
40	Q2839-08	PO112888.D	13 Aug 2025 00:09	YP/AJ	Ok,M
41	Q2839-09	PO112889.D	13 Aug 2025 00:27	YP/AJ	Ok
42	AR1660CCC500	PO112890.D	13 Aug 2025 01:41	YP/AJ	Ok,M
43	AR1242CCC500	PO112891.D	13 Aug 2025 02:34	YP/AJ	Ok,M
44	AR1248CCC500	PO112892.D	13 Aug 2025 02:52	YP/AJ	Ok

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO081225

Review By	yogesh	Review On	8/12/2025 10:12:49 AM
Supervise By	mohammad	Supervise On	8/14/2025 1:21:40 AM
SubDirectory	PO081225	HP Acquire Method	HP Processing Method PO072325
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,PP24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369		
CCC	PP24332,PP24347,PP24352,PP24357		
Internal Standard/PEM			
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP2		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

45	AR1254CCC500	PO112893.D	13 Aug 2025 03:11	YP/AJ	Ok,M
46	I.BLK	PO112894.D	13 Aug 2025 03:29	YP/AJ	Ok

M : Manual Integration

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP080125

Review By	yogesh	Review On	8/1/2025 3:30:36 PM
Supervise By	mohammad	Supervise On	8/8/2025 7:37:53 AM
SubDirectory	PP080125	HP Acquire Method	HP Processing Method PP080125
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,PP24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369		
CCC	PP24332,PP24347,PP24352,PP24357		
Internal Standard/PEM			
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP2		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	Sampleld	Data File Name	Date-Time	Operator	Status
1	HEXANE	PP074166.D	01 Aug 2025 11:33	YP\AJ	Ok
2	I.BLK	PP074167.D	01 Aug 2025 11:49	YP\AJ	Ok
3	AR1660ICC1000	PP074168.D	01 Aug 2025 12:05	YP\AJ	Ok
4	AR1660ICC750	PP074169.D	01 Aug 2025 12:22	YP\AJ	Ok
5	AR1660ICC500	PP074170.D	01 Aug 2025 12:38	YP\AJ	Ok
6	AR1660ICC250	PP074171.D	01 Aug 2025 12:54	YP\AJ	Ok,M
7	AR1660ICC050	PP074172.D	01 Aug 2025 13:42	YP\AJ	Ok,M
8	AR1221ICC500	PP074173.D	01 Aug 2025 13:58	YP\AJ	Ok
9	AR1232ICC500	PP074174.D	01 Aug 2025 14:15	YP\AJ	Ok,M
10	AR1242ICC1000	PP074175.D	01 Aug 2025 14:31	YP\AJ	Ok,M
11	AR1242ICC750	PP074176.D	01 Aug 2025 14:47	YP\AJ	Ok,M
12	AR1242ICC500	PP074177.D	01 Aug 2025 15:03	YP\AJ	Ok,M
13	AR1242ICC250	PP074178.D	01 Aug 2025 15:19	YP\AJ	Ok,M
14	AR1242ICC050	PP074179.D	01 Aug 2025 15:36	YP\AJ	Ok,M
15	AR1248ICC1000	PP074180.D	01 Aug 2025 16:24	YP\AJ	Not Ok
16	AR1248ICC750	PP074181.D	01 Aug 2025 16:41	YP\AJ	Not Ok
17	AR1248ICC500	PP074182.D	01 Aug 2025 16:57	YP\AJ	Ok,M
18	AR1248ICC250	PP074183.D	01 Aug 2025 17:13	YP\AJ	Not Ok
19	AR1248ICC050	PP074184.D	01 Aug 2025 17:45	YP\AJ	Not Ok
20	AR1254ICC1000	PP074185.D	01 Aug 2025 18:02	YP\AJ	Ok,M
21	AR1254ICC750	PP074186.D	01 Aug 2025 18:18	YP\AJ	Ok,M

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP080125

Review By	yogesh	Review On	8/1/2025 3:30:36 PM
Supervise By	mohammad	Supervise On	8/8/2025 7:37:53 AM
SubDirectory	PP080125	HP Acquire Method	HP Processing Method PP080125
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,PP24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369		
CCC	PP24332,PP24347,PP24352,PP24357		
Internal Standard/PEM			
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP2		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

22	AR1254ICC500	PP074187.D	01 Aug 2025 18:34	YP\AJ	Ok,M
23	AR1254ICC250	PP074188.D	01 Aug 2025 18:50	YP\AJ	Ok,M
24	AR1254ICC050	PP074189.D	01 Aug 2025 19:23	YP\AJ	Ok,M
25	AR1262ICC500	PP074190.D	01 Aug 2025 19:39	YP\AJ	Ok
26	AR1268ICC1000	PP074191.D	01 Aug 2025 19:55	YP\AJ	Not Ok
27	AR1268ICC750	PP074192.D	01 Aug 2025 20:11	YP\AJ	Not Ok
28	AR1268ICC500	PP074193.D	01 Aug 2025 20:28	YP\AJ	Ok
29	AR1268ICC250	PP074194.D	01 Aug 2025 20:44	YP\AJ	Not Ok
30	AR1268ICC050	PP074195.D	01 Aug 2025 21:00	YP\AJ	Not Ok
31	PP080125ICV500	PP074196.D	01 Aug 2025 21:16	YP\AJ	Ok
32	AR1242ICV500	PP074197.D	01 Aug 2025 22:05	YP\AJ	Ok,M
33	AR1248ICV500	PP074198.D	01 Aug 2025 22:37	YP\AJ	Not Ok
34	AR1254ICV500	PP074199.D	01 Aug 2025 23:10	YP\AJ	Ok,M
35	AR1268ICV500	PP074200.D	01 Aug 2025 23:42	YP\AJ	Not Ok
36	DDT ANALOG	PP074201.D	02 Aug 2025 00:15	YP\AJ	Ok

M : Manual Integration

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP081225

Review By	yogesh	Review On	8/12/2025 10:13:41 AM
Supervise By	mohammad	Supervise On	8/14/2025 1:21:51 AM
SubDirectory	PP081225	HP Acquire Method	HP Processing Method PP080125
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,PP24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369		
CCC	PP24332,PP24347,PP24352,PP24357		
Internal Standard/PEM			
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP2		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	HEXANE	PP074301.D	12 Aug 2025 09:35	YP\AJ	Ok
2	AR1660CCC500	PP074302.D	12 Aug 2025 09:51	YP\AJ	Ok
3	AR1242CCC500	PP074303.D	12 Aug 2025 10:11	YP\AJ	Ok,M
4	AR1254CCC500	PP074304.D	12 Aug 2025 10:27	YP\AJ	Ok,M
5	I.BLK	PP074305.D	12 Aug 2025 10:43	YP\AJ	Ok
6	DDT ANALOG	PP074306.D	12 Aug 2025 10:59	YP\AJ	Ok
7	PP24799	PP074307.D	12 Aug 2025 11:16	YP\AJ	Ok,M
8	PB169205BL	PP074308.D	12 Aug 2025 12:48	YP\AJ	Ok
9	PB169205BS	PP074309.D	12 Aug 2025 13:05	YP\AJ	Ok,M
10	Q2830-01	PP074310.D	12 Aug 2025 13:21	YP\AJ	Ok,M
11	Q2830-01MS	PP074311.D	12 Aug 2025 13:37	YP\AJ	Ok,M
12	Q2830-01MSD	PP074312.D	12 Aug 2025 13:53	YP\AJ	Ok,M
13	Q2830-05	PP074313.D	12 Aug 2025 14:10	YP\AJ	Ok
14	Q2830-07	PP074314.D	12 Aug 2025 14:26	YP\AJ	Ok,M
15	Q2831-01	PP074315.D	12 Aug 2025 14:42	YP\AJ	Ok
16	AR1660CCC500	PP074316.D	12 Aug 2025 15:48	YP\AJ	Ok,M
17	AR1242CCC500	PP074317.D	12 Aug 2025 16:20	YP\AJ	Ok,M
18	AR1254CCC500	PP074318.D	12 Aug 2025 16:37	YP\AJ	Ok,M
19	I.BLK	PP074319.D	12 Aug 2025 16:53	YP\AJ	Ok
20	PB169220BL	PP074320.D	12 Aug 2025 17:26	YP\AJ	Ok
21	PB169220BS	PP074321.D	12 Aug 2025 17:42	YP\AJ	Ok

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP081225

Review By	yogesh	Review On	8/12/2025 10:13:41 AM
Supervise By	mohammad	Supervise On	8/14/2025 1:21:51 AM
SubDirectory	PP081225	HP Acquire Method	HP Processing Method PP080125
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,PP24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369		
CCC	PP24332,PP24347,PP24352,PP24357		
Internal Standard/PEM			
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP2		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

22	Q2839-10	PP074322.D	12 Aug 2025 17:58	YP\AJ	Ok,M
23	Q2839-11	PP074323.D	12 Aug 2025 18:14	YP\AJ	Ok,M
24	Q2839-12	PP074324.D	12 Aug 2025 18:31	YP\AJ	Ok,M
25	Q2839-13	PP074325.D	12 Aug 2025 18:47	YP\AJ	Ok,M
26	Q2839-14	PP074326.D	12 Aug 2025 19:03	YP\AJ	Ok,M
27	Q2839-15	PP074327.D	12 Aug 2025 19:20	YP\AJ	Ok,M
28	Q2839-16	PP074328.D	12 Aug 2025 19:36	YP\AJ	Ok,M
29	AR1660CCC500	PP074329.D	12 Aug 2025 20:41	YP\AJ	Ok
30	AR1242CCC500	PP074330.D	12 Aug 2025 21:30	YP\AJ	Ok,M
31	AR1254CCC500	PP074331.D	12 Aug 2025 21:46	YP\AJ	Ok,M
32	I.BLK	PP074332.D	12 Aug 2025 22:02	YP\AJ	Ok

M : Manual Integration

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO072325

Review By	yogesh	Review On	7/23/2025 3:39:31 PM
Supervise By	mohammad	Supervise On	7/25/2025 3:53:10 AM
SubDirectory	PO072325	HP Acquire Method	HP Processing Method PO072325

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,PP24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369 PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PO112412.D	23 Jul 2025 10:56		YP/AJ	Ok
2	I.BLK	I.BLK	PO112413.D	23 Jul 2025 11:14		YP/AJ	Ok
3	AR1660ICC1000	AR1660ICC1000	PO112414.D	23 Jul 2025 11:32		YP/AJ	Ok
4	AR1660ICC750	AR1660ICC750	PO112415.D	23 Jul 2025 11:50		YP/AJ	Ok
5	AR1660ICC500	AR1660ICC500	PO112416.D	23 Jul 2025 12:08		YP/AJ	Ok
6	AR1660ICC250	AR1660ICC250	PO112417.D	23 Jul 2025 12:27		YP/AJ	Ok,M
7	AR1660ICC050	AR1660ICC050	PO112418.D	23 Jul 2025 12:45		YP/AJ	Ok,M
8	AR1221ICC500	AR1221ICC500	PO112419.D	23 Jul 2025 13:03		YP/AJ	Ok
9	AR1232ICC500	AR1232ICC500	PO112420.D	23 Jul 2025 13:22		YP/AJ	Ok
10	AR1242ICC1000	AR1242ICC1000	PO112421.D	23 Jul 2025 13:40		YP/AJ	Ok
11	AR1242ICC750	AR1242ICC750	PO112422.D	23 Jul 2025 13:59		YP/AJ	Ok
12	AR1242ICC500	AR1242ICC500	PO112423.D	23 Jul 2025 14:17		YP/AJ	Ok
13	AR1242ICC250	AR1242ICC250	PO112424.D	23 Jul 2025 14:36		YP/AJ	Ok,M
14	AR1242ICC050	AR1242ICC050	PO112425.D	23 Jul 2025 14:54		YP/AJ	Ok,M
15	AR1248ICC1000	AR1248ICC1000	PO112426.D	23 Jul 2025 15:13		YP/AJ	Ok
16	AR1248ICC750	AR1248ICC750	PO112427.D	23 Jul 2025 15:31		YP/AJ	Ok
17	AR1248ICC500	AR1248ICC500	PO112428.D	23 Jul 2025 15:48		YP/AJ	Ok
18	AR1248ICC250	AR1248ICC250	PO112429.D	23 Jul 2025 16:07		YP/AJ	Ok

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO072325

Review By	yogesh	Review On	7/23/2025 3:39:31 PM
Supervise By	mohammad	Supervise On	7/25/2025 3:53:10 AM
SubDirectory	PO072325	HP Acquire Method	HP Processing Method PO072325
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,P P24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP 24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369 PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387		

19	AR1248ICC050	AR1248ICC050	PO112430.D	23 Jul 2025 16:25		YP/AJ	Ok,M
20	AR1254ICC1000	AR1254ICC1000	PO112431.D	23 Jul 2025 16:44		YP/AJ	Ok
21	AR1254ICC750	AR1254ICC750	PO112432.D	23 Jul 2025 17:02		YP/AJ	Ok
22	AR1254ICC500	AR1254ICC500	PO112433.D	23 Jul 2025 17:21		YP/AJ	Ok
23	AR1254ICC250	AR1254ICC250	PO112434.D	23 Jul 2025 17:39		YP/AJ	Ok
24	AR1254ICC050	AR1254ICC050	PO112435.D	23 Jul 2025 17:57		YP/AJ	Ok,M
25	AR1262ICC500	AR1262ICC500	PO112436.D	23 Jul 2025 18:16		YP/AJ	Ok
26	AR1268ICC1000	AR1268ICC1000	PO112437.D	23 Jul 2025 18:34		YP/AJ	Ok
27	AR1268ICC750	AR1268ICC750	PO112438.D	23 Jul 2025 18:53		YP/AJ	Ok
28	AR1268ICC500	AR1268ICC500	PO112439.D	23 Jul 2025 19:11		YP/AJ	Ok
29	AR1268ICC250	AR1268ICC250	PO112440.D	23 Jul 2025 19:28		YP/AJ	Ok
30	AR1268ICC050	AR1268ICC050	PO112441.D	23 Jul 2025 19:47		YP/AJ	Ok
31	PO072325ICV500	ICVPO072325	PO112442.D	23 Jul 2025 20:05		YP/AJ	Ok
32	AR1242ICV500	ICVPO072325AR1242	PO112443.D	23 Jul 2025 20:42		YP/AJ	Ok
33	AR1248ICV500	ICVPO072325AR1248	PO112444.D	23 Jul 2025 21:18		YP/AJ	Ok
34	AR1254ICV500	ICVPO072325AR1254	PO112445.D	23 Jul 2025 21:54		YP/AJ	Ok
35	AR1268ICV500	ICVPO072325AR1268	PO112446.D	23 Jul 2025 22:31		YP/AJ	Ok
36	DDT ANALOG	DDT ANALOG	PO112447.D	23 Jul 2025 23:07		YP/AJ	Ok

M : Manual Integration

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO081225

Review By	yogesh	Review On	8/12/2025 10:12:49 AM
Supervise By	mohammad	Supervise On	8/14/2025 1:21:40 AM
SubDirectory	PO081225	HP Acquire Method	HP Processing Method PO072325

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,PP24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369
CCC	PP24332,PP24347,PP24352,PP24357
Internal Standard/PEM	
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PO112849.D	12 Aug 2025 08:35		YP/AJ	Ok
2	AR1660CCC500	AR1660CCC500	PO112850.D	12 Aug 2025 08:54		YP/AJ	Ok,M
3	AR1242CCC500	AR1242CCC500	PO112851.D	12 Aug 2025 09:12		YP/AJ	Ok,M
4	AR1248CCC500	AR1248CCC500	PO112852.D	12 Aug 2025 09:29		YP/AJ	Ok,M
5	AR1254CCC500	AR1254CCC500	PO112853.D	12 Aug 2025 09:48		YP/AJ	Ok
6	I.BLK	I.BLK	PO112854.D	12 Aug 2025 10:05		YP/AJ	Ok
7	DDT ANALOG	DDT ANALOG	PO112855.D	12 Aug 2025 10:24		YP/AJ	Ok
8	Q2808-01DL	TP-7DL	PO112856.D	12 Aug 2025 10:42	AR1242 + 1254 Hit	YP/AJ	Ok,M
9	Q2807-02	COMP-5	PO112857.D	12 Aug 2025 11:01		YP/AJ	Ok
10	Q2818-01	B-2-5-1	PO112858.D	12 Aug 2025 12:50		YP/AJ	Ok,M
11	Q2818-02	B-3-5-2	PO112859.D	12 Aug 2025 13:08		YP/AJ	Ok
12	Q2819-06	11M-S	PO112860.D	12 Aug 2025 13:26		YP/AJ	Ok
13	Q2819-11	84SB-W	PO112861.D	12 Aug 2025 13:43	DCB low in 1ST column	YP/AJ	Ok
14	Q2819-14	17M-W	PO112862.D	12 Aug 2025 14:01		YP/AJ	Ok
15	Q2820-09	705R-S	PO112863.D	12 Aug 2025 14:20	DCB low both column	YP/AJ	ReRun
16	Q2820-10	SOIL-DUP-2	PO112864.D	12 Aug 2025 14:38		YP/AJ	Ok
17	AR1660CCC500	AR1660CCC500	PO112865.D	12 Aug 2025 15:50		YP/AJ	Ok,M
18	AR1242CCC500	AR1242CCC500	PO112866.D	12 Aug 2025 16:07		YP/AJ	Ok,M

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO081225

Review By	yogesh	Review On	8/12/2025 10:12:49 AM
Supervise By	mohammad	Supervise On	8/14/2025 1:21:40 AM
SubDirectory	PO081225	HP Acquire Method	HP Processing Method PO072325
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,P P24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP 24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369		
CCC	PP24332,PP24347,PP24352,PP24357		
Internal Standard/PEM			
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	AR1248CCC500	AR1248CCC500	PO112867.D	12 Aug 2025 16:25		YP/AJ	Ok
20	AR1254CCC500	AR1254CCC500	PO112868.D	12 Aug 2025 16:42	TCMX high in 1st column	YP/AJ	Ok
21	I.BLK	I.BLK	PO112869.D	12 Aug 2025 17:00		YP/AJ	Ok
22	Q2820-24	22M-S	PO112870.D	12 Aug 2025 17:19		YP/AJ	Ok
23	Q2823-01	SU-04-081125	PO112871.D	12 Aug 2025 17:37		YP/AJ	Ok,M
24	Q2826-01	WC1	PO112872.D	12 Aug 2025 17:56	AR1254 Hit	YP/AJ	Ok
25	Q2827-01	TP-8	PO112873.D	12 Aug 2025 18:14	AR1248 & AR1260 Hit , Need 2x dilution	YP/AJ	Dilution
26	Q2827-05	TP-9	PO112874.D	12 Aug 2025 18:31	AR1260 Hit	YP/AJ	Ok
27	Q2830-03	BIN0009-DRIVEWAY-T	PO112875.D	12 Aug 2025 18:49	DCB high in 2nd column	YP/AJ	Ok
28	Q2839-01	BC274436-1-1	PO112876.D	12 Aug 2025 19:06		YP/AJ	Ok,M
29	Q2839-02	BC274436-1-2	PO112877.D	12 Aug 2025 19:24		YP/AJ	Ok,M
30	Q2839-03	BC151973-1-1	PO112878.D	12 Aug 2025 19:42		YP/AJ	Ok
31	Q2839-04	BC151973-1-2	PO112879.D	12 Aug 2025 19:59	AR1254 Hit	YP/AJ	Ok,M
32	AR1660CCC500	AR1660CCC500	PO112880.D	12 Aug 2025 21:09		YP/AJ	Ok,M
33	AR1242CCC500	AR1242CCC500	PO112881.D	12 Aug 2025 22:02		YP/AJ	Ok,M
34	AR1248CCC500	AR1248CCC500	PO112882.D	12 Aug 2025 22:21		YP/AJ	Ok
35	AR1254CCC500	AR1254CCC500	PO112883.D	12 Aug 2025 22:38	TCMX high in 1st column	YP/AJ	Ok
36	I.BLK	I.BLK	PO112884.D	12 Aug 2025 22:56		YP/AJ	Ok
37	Q2839-05	BC271336-1-1	PO112885.D	12 Aug 2025 23:15		YP/AJ	Ok

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO081225

Review By	yogesh	Review On	8/12/2025 10:12:49 AM
Supervise By	mohammad	Supervise On	8/14/2025 1:21:40 AM
SubDirectory	PO081225	HP Acquire Method	HP Processing Method PO072325
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,P P24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP 24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369		
CCC	PP24332,PP24347,PP24352,PP24357		
Internal Standard/PEM			
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

38	Q2839-06	BC271336-1-2	PO112886.D	12 Aug 2025 23:33		YP/AJ	Ok
39	Q2839-07	BC271242-1-1	PO112887.D	12 Aug 2025 23:51	AR1254 Hit	YP/AJ	Ok,M
40	Q2839-08	BC271242-1-2	PO112888.D	13 Aug 2025 00:09		YP/AJ	Ok,M
41	Q2839-09	BC271242-2-1	PO112889.D	13 Aug 2025 00:27		YP/AJ	Ok
42	AR1660CCC500	AR1660CCC500	PO112890.D	13 Aug 2025 01:41		YP/AJ	Ok,M
43	AR1242CCC500	AR1242CCC500	PO112891.D	13 Aug 2025 02:34		YP/AJ	Ok,M
44	AR1248CCC500	AR1248CCC500	PO112892.D	13 Aug 2025 02:52		YP/AJ	Ok
45	AR1254CCC500	AR1254CCC500	PO112893.D	13 Aug 2025 03:11	TCMX high in 1st column	YP/AJ	Ok,M
46	I.BLK	I.BLK	PO112894.D	13 Aug 2025 03:29		YP/AJ	Ok

M : Manual Integration

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP080125

Review By	yogesh	Review On	8/1/2025 3:30:36 PM
Supervise By	mohammad	Supervise On	8/8/2025 7:37:53 AM
SubDirectory	PP080125	HP Acquire Method	HP Processing Method PP080125

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,PP24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369
CCC	PP24332,PP24347,PP24352,PP24357
Internal Standard/PEM	
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PP074166.D	01 Aug 2025 11:33		YP\AJ	Ok
2	I.BLK	I.BLK	PP074167.D	01 Aug 2025 11:49		YP\AJ	Ok
3	AR1660ICC1000	AR1660ICC1000	PP074168.D	01 Aug 2025 12:05		YP\AJ	Ok
4	AR1660ICC750	AR1660ICC750	PP074169.D	01 Aug 2025 12:22		YP\AJ	Ok
5	AR1660ICC500	AR1660ICC500	PP074170.D	01 Aug 2025 12:38		YP\AJ	Ok
6	AR1660ICC250	AR1660ICC250	PP074171.D	01 Aug 2025 12:54		YP\AJ	Ok,M
7	AR1660ICC050	AR1660ICC050	PP074172.D	01 Aug 2025 13:42		YP\AJ	Ok,M
8	AR1221ICC500	AR1221ICC500	PP074173.D	01 Aug 2025 13:58		YP\AJ	Ok
9	AR1232ICC500	AR1232ICC500	PP074174.D	01 Aug 2025 14:15		YP\AJ	Ok,M
10	AR1242ICC1000	AR1242ICC1000	PP074175.D	01 Aug 2025 14:31		YP\AJ	Ok,M
11	AR1242ICC750	AR1242ICC750	PP074176.D	01 Aug 2025 14:47		YP\AJ	Ok,M
12	AR1242ICC500	AR1242ICC500	PP074177.D	01 Aug 2025 15:03		YP\AJ	Ok,M
13	AR1242ICC250	AR1242ICC250	PP074178.D	01 Aug 2025 15:19		YP\AJ	Ok,M
14	AR1242ICC050	AR1242ICC050	PP074179.D	01 Aug 2025 15:36		YP\AJ	Ok,M
15	AR1248ICC1000	AR1248ICC1000	PP074180.D	01 Aug 2025 16:24	Not Use	YP\AJ	Not Ok
16	AR1248ICC750	AR1248ICC750	PP074181.D	01 Aug 2025 16:41	Not Use	YP\AJ	Not Ok
17	AR1248ICC500	AR1248ICC500	PP074182.D	01 Aug 2025 16:57		YP\AJ	Ok,M
18	AR1248ICC250	AR1248ICC250	PP074183.D	01 Aug 2025 17:13	Not Use	YP\AJ	Not Ok

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP080125

Review By	yogesh	Review On	8/1/2025 3:30:36 PM
Supervise By	mohammad	Supervise On	8/8/2025 7:37:53 AM
SubDirectory	PP080125	HP Acquire Method	HP Processing Method PP080125
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,P P24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP 24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369		
CCC	PP24332,PP24347,PP24352,PP24357		
Internal Standard/PEM			
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	AR1248ICC050	AR1248ICC050	PP074184.D	01 Aug 2025 17:45	Not Use	YPIAJ	Not Ok
20	AR1254ICC1000	AR1254ICC1000	PP074185.D	01 Aug 2025 18:02		YPIAJ	Ok,M
21	AR1254ICC750	AR1254ICC750	PP074186.D	01 Aug 2025 18:18		YPIAJ	Ok,M
22	AR1254ICC500	AR1254ICC500	PP074187.D	01 Aug 2025 18:34		YPIAJ	Ok,M
23	AR1254ICC250	AR1254ICC250	PP074188.D	01 Aug 2025 18:50		YPIAJ	Ok,M
24	AR1254ICC050	AR1254ICC050	PP074189.D	01 Aug 2025 19:23		YPIAJ	Ok,M
25	AR1262ICC500	AR1262ICC500	PP074190.D	01 Aug 2025 19:39		YPIAJ	Ok
26	AR1268ICC1000	AR1268ICC1000	PP074191.D	01 Aug 2025 19:55	Not Use	YPIAJ	Not Ok
27	AR1268ICC750	AR1268ICC750	PP074192.D	01 Aug 2025 20:11	Not Use	YPIAJ	Not Ok
28	AR1268ICC500	AR1268ICC500	PP074193.D	01 Aug 2025 20:28		YPIAJ	Ok
29	AR1268ICC250	AR1268ICC250	PP074194.D	01 Aug 2025 20:44	Not Use	YPIAJ	Not Ok
30	AR1268ICC050	AR1268ICC050	PP074195.D	01 Aug 2025 21:00	Not Use	YPIAJ	Not Ok
31	PP080125ICV500	ICVPP080125	PP074196.D	01 Aug 2025 21:16		YPIAJ	Ok
32	AR1242ICV500	ICVPP080125AR1242	PP074197.D	01 Aug 2025 22:05		YPIAJ	Ok,M
33	AR1248ICV500	ICVPP080125AR1248	PP074198.D	01 Aug 2025 22:37	Not Use	YPIAJ	Not Ok
34	AR1254ICV500	ICVPP080125AR1254	PP074199.D	01 Aug 2025 23:10		YPIAJ	Ok,M
35	AR1268ICV500	ICVPP080125AR1268	PP074200.D	01 Aug 2025 23:42	Not Use	YPIAJ	Not Ok
36	DDT ANALOG	DDT ANALOG	PP074201.D	02 Aug 2025 00:15		YPIAJ	Ok

M : Manual Integration

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP081225

Review By	yogesh	Review On	8/12/2025 10:13:41 AM
Supervise By	mohammad	Supervise On	8/14/2025 1:21:51 AM
SubDirectory	PP081225	HP Acquire Method	HP Processing Method PP080125

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,PP24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369
CCC	PP24332,PP24347,PP24352,PP24357
Internal Standard/PEM	
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PP074301.D	12 Aug 2025 09:35		YPIAJ	Ok
2	AR1660CCC500	AR1660CCC500	PP074302.D	12 Aug 2025 09:51	AR1016-02& TCMX high in sec column	YPIAJ	Ok
3	AR1242CCC500	AR1242CCC500	PP074303.D	12 Aug 2025 10:11	TCMX high in sec column	YPIAJ	Ok,M
4	AR1254CCC500	AR1254CCC500	PP074304.D	12 Aug 2025 10:27	TCMX high in sec column	YPIAJ	Ok,M
5	I.BLK	I.BLK	PP074305.D	12 Aug 2025 10:43		YPIAJ	Ok
6	DDT ANALOG	DDT ANALOG	PP074306.D	12 Aug 2025 10:59		YPIAJ	Ok
7	PP24799	PP24799	PP074307.D	12 Aug 2025 11:16		YPIAJ	Ok,M
8	PB169205BL	PB169205BL	PP074308.D	12 Aug 2025 12:48		YPIAJ	Ok
9	PB169205BS	PB169205BS	PP074309.D	12 Aug 2025 13:05		YPIAJ	Ok,M
10	Q2830-01	BIN0009-DRIVEWAY-T	PP074310.D	12 Aug 2025 13:21		YPIAJ	Ok,M
11	Q2830-01MS	BIN0009-DRIVEWAY-T	PP074311.D	12 Aug 2025 13:37		YPIAJ	Ok,M
12	Q2830-01MSD	BIN0009-DRIVEWAY-T	PP074312.D	12 Aug 2025 13:53		YPIAJ	Ok,M
13	Q2830-05	BIN0009-DRIVEWAY-T	PP074313.D	12 Aug 2025 14:10		YPIAJ	Ok
14	Q2830-07	BIN0009-DRIVEWAY-T	PP074314.D	12 Aug 2025 14:26		YPIAJ	Ok,M
15	Q2831-01	VNJ-238	PP074315.D	12 Aug 2025 14:42		YPIAJ	Ok
16	AR1660CCC500	AR1660CCC500	PP074316.D	12 Aug 2025 15:48		YPIAJ	Ok,M
17	AR1242CCC500	AR1242CCC500	PP074317.D	12 Aug 2025 16:20	TCMX high in 2nd column	YPIAJ	Ok,M

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP081225

Review By	yogesh	Review On	8/12/2025 10:13:41 AM
Supervise By	mohammad	Supervise On	8/14/2025 1:21:51 AM
SubDirectory	PP081225	HP Acquire Method	HP Processing Method PP080125
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,P P24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP 24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369		
CCC	PP24332,PP24347,PP24352,PP24357		
Internal Standard/PEM			
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

18	AR1254CCC500	AR1254CCC500	PP074318.D	12 Aug 2025 16:37		YPIAJ	Ok,M
19	I.BLK	I.BLK	PP074319.D	12 Aug 2025 16:53		YPIAJ	Ok
20	PB169220BL	PB169220BL	PP074320.D	12 Aug 2025 17:26		YPIAJ	Ok
21	PB169220BS	PB169220BS	PP074321.D	12 Aug 2025 17:42		YPIAJ	Ok
22	Q2839-10	BC271242-2-2	PP074322.D	12 Aug 2025 17:58		YPIAJ	Ok,M
23	Q2839-11	BC226751-1-1	PP074323.D	12 Aug 2025 18:14		YPIAJ	Ok,M
24	Q2839-12	BC226751-1-2	PP074324.D	12 Aug 2025 18:31		YPIAJ	Ok,M
25	Q2839-13	BC226751-2-1	PP074325.D	12 Aug 2025 18:47	AR1254 Hit	YPIAJ	Ok,M
26	Q2839-14	BC226751-2-2	PP074326.D	12 Aug 2025 19:03		YPIAJ	Ok,M
27	Q2839-15	JEC773V-1-1	PP074327.D	12 Aug 2025 19:20	TCMX low in 2nd column	YPIAJ	Ok,M
28	Q2839-16	JEC773V-1-2	PP074328.D	12 Aug 2025 19:36		YPIAJ	Ok,M
29	AR1660CCC500	AR1660CCC500	PP074329.D	12 Aug 2025 20:41		YPIAJ	Ok
30	AR1242CCC500	AR1242CCC500	PP074330.D	12 Aug 2025 21:30	TCMX high in 2nd column	YPIAJ	Ok,M
31	AR1254CCC500	AR1254CCC500	PP074331.D	12 Aug 2025 21:46	TCMX high in 2nd column	YPIAJ	Ok,M
32	I.BLK	I.BLK	PP074332.D	12 Aug 2025 22:02		YPIAJ	Ok

M : Manual Integration

PERCENT SOLID

Supervisor: rubina
Analyst: jignesh
Date: 8/12/2025

OVENTEMP IN Celsius(°C): 108
Time IN: 17:10
In Date: 08/11/2025
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 104
Time OUT: 08:22
Out Date: 08/12/2025
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID-OVEN

QC:LB136776

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
Q2817-01	LEAD-PAINT-CHIPS	1	1.00	1.00	2.00	2.00	100.0	LEAD-PAINT-CHIPS
Q2818-01	B-2-5-1	2	1.15	10.59	11.74	10.42	87.5	
Q2818-02	B-3-5-2	3	1.16	10.66	11.82	10.66	89.1	
Q2823-01	SU-04-081125	4	1.13	10.70	11.83	11.11	93.3	
Q2823-02	SU-04-081125-E2	5	1.14	11.10	12.24	11.34	91.9	
Q2826-01	WC1	6	1.18	10.81	11.99	10.7	88.1	
Q2827-01	TP-8	7	1.16	11.41	12.57	7.33	54.1	
Q2827-02	TP-8-EPH	8	1.16	10.21	11.37	7.68	63.9	
Q2827-03	TP-8-VOC	9	1.18	10.87	12.05	9.22	74.0	
Q2827-05	TP-9	10	1.19	10.61	11.8	9.66	79.8	
Q2827-06	TP-9-EPH	11	1.16	10.06	11.22	8.93	77.2	
Q2827-07	TP-9-VOC	12	1.15	10.48	11.63	9.59	80.5	
Q2828-01	POWDER	13	1.19	10.32	11.51	11.45	99.4	
Q2829-01	SILICA	14	1.00	1.00	2.00	2.00	100.0	silica
Q2830-01	BIN0009-DRIVEWAY-TP-SOUTH-EAST	15	1.18	10.57	11.75	10.37	86.9	
Q2830-02	BIN0009-DRIVEWAY-TP-SOUTH-EAST	16	1.18	10.03	11.21	9.83	86.2	
Q2830-03	BIN0009-DRIVEWAY-TP-WEST	17	1.18	10.74	11.92	10.37	85.6	
Q2830-04	BIN0009-DRIVEWAY-TP-WEST	18	1.15	10.59	11.74	10.21	85.6	
Q2830-05	BIN0009-DRIVEWAY-TP-WESTSIDE	19	1.19	10.46	11.65	9.66	81.0	
Q2830-06	BIN0009-DRIVEWAY-TP-WESTSIDE	20	1.15	10.94	12.09	9.36	75.0	
Q2830-07	BIN0009-DRIVEWAY-TP-EASTSIDE	21	1.18	10.48	11.66	9.22	76.7	
Q2830-08	BIN0009-DRIVEWAY-TP-EASTSIDE	22	1.19	10.66	11.85	9.67	79.5	
Q2831-01	VNJ-238	23	1.18	10.59	11.77	11.04	93.1	
Q2831-02	VNJ-238-VOC	24	1.19	10.62	11.81	11.04	92.7	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$

WORKLIST(Hardcopy Internal Chain)

136716

WorkList Name : %1-081125 WorkList ID : 191189 Department : Wet-Chemistry Date : 08-11-2025 08:22:53

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q2817-01	LEAD-PAINT-CHIPS	Solid	Percent Solids	Cool 4 deg C	HOME01	D31	08/11/2025	Chemtech -SO
Q2818-01	B-2-5-1	Solid	Percent Solids	Cool 4 deg C	EARTH03	J21	08/11/2025	Chemtech -SO
Q2818-02	B-3-5-2	Solid	Percent Solids	Cool 4 deg C	EARTH03	J21	08/11/2025	Chemtech -SO
Q2823-01	SU-04-081125	Solid	Percent Solids	Cool 4 deg C	PSEG05	D31	08/11/2025	Chemtech -SO
Q2823-02	SU-04-081125-E2	Solid	Percent Solids	Cool 4 deg C	PSEG05	D31	08/11/2025	Chemtech -SO
Q2826-01	WC1	Solid	Percent Solids	Cool 4 deg C	GENV01	D31	08/11/2025	Chemtech -SO
Q2827-01	TP-8	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	08/11/2025	Chemtech -SO
Q2827-02	TP-8-EPH	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	08/11/2025	Chemtech -SO
Q2827-03	TP-8-VOC	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	08/11/2025	Chemtech -SO
Q2827-05	TP-9	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	08/11/2025	Chemtech -SO
Q2827-06	TP-9-EPH	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	08/11/2025	Chemtech -SO
Q2827-07	TP-9-VOC	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	08/11/2025	Chemtech -SO
Q2828-01	POWDER	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	08/11/2025	Chemtech -SO
Q2829-01	SILICA	Solid	Percent Solids	Cool 4 deg C	VIVA01	D21	08/11/2025	Chemtech -SO
Q2830-01	BIN009-DRIVEWAY-TP-SOUTH	Solid	Percent Solids	Cool 4 deg C	VIVA01	D21	08/11/2025	Chemtech -SO
Q2830-02	BIN009-DRIVEWAY-TP-SOUTH	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	08/11/2025	Chemtech -SO
Q2830-03	BIN009-DRIVEWAY-TP-WEST	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	08/11/2025	Chemtech -SO
Q2830-04	BIN009-DRIVEWAY-TP-WEST	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	08/11/2025	Chemtech -SO
Q2830-05	BIN009-DRIVEWAY-TP-WEST	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	08/11/2025	Chemtech -SO
Q2830-06	BIN009-DRIVEWAY-TP-WEST	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	08/11/2025	Chemtech -SO
Q2830-07	BIN009-DRIVEWAY-TP-EASTS	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	08/11/2025	Chemtech -SO

Date/Time 08-11-25 15:10
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]
 Date/Time 08-11-25 17:20
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]

WORKLIST(Hardcopy Internal Chain)

136776

WorkList Name : %1-081125 WorkList ID : 191189 Department : Wet-Chemistry Date : 08-11-2025 08:22:53

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q2830-08	BIN009-DRIVEWAY-TP-EASTS	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	08/11/2025	Chemtech -SO
Q2831-01	VNJ-238	Solid	Percent Solids	Cool 4 deg C	PSEG03	D21	08/11/2025	Chemtech -SO
Q2831-02	VNJ-238-VOC	Solid	Percent Solids	Cool 4 deg C	PSEG03	D21	08/11/2025	Chemtech -SO

Date/Time 08.11.25 15:10
 Raw Sample Received by: Sh Woc
 Raw Sample Relinquished by: Sh Woc

Date/Time 08.11.25 17:20
 Raw Sample Received by: Sh Woc
 Raw Sample Relinquished by: Sh Woc

SOP ID: M3541-ASE Extraction-15

Clean Up SOP #: Acid Cleanup

Extraction Start Date : 08/12/2025

Matrix : Solid

Extraction Start Time : 08:15

Weigh By: RJ

Extraction By: RJ

Extraction End Date : 08/12/2025

Balance check: RJ

Filter By: RJ

Extraction End Time : 11:30

Balance ID: EX-SC-2

pH Meter ID: N/A

Concentration By: RS

pH Strip Lot#: N/A

Hood ID: 3,7

Supervisor By : RUPESH

Extraction Method: ☐ Separatory Funnel ☐ Continuous Liquid/Liquid ☐ Sonication ☐ Waste Dilution ☒ Soxhlet

Standard Name	MLS USED	Concentration ug/mL	STD REF. # FROM LOG
Spike Sol 1	1.0ML	5000 PPB	PP24650
Surrogate	1.0ML	200 PPB	PP24663
N/A	N/A	N/A	N/A
N/A	N/A	N/A	N/A
N/A	N/A	N/A	N/A

Chemical Used	ML/SAMPLE USED	Lot Number
Hexane/Acetone/1:1	N/A	EP2627
Baked Na2SO4	N/A	EP2632
Sand	N/A	E3951
Hexane	N/A	E3962
H2SO4 1:1	N/A	EP2610
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A

Extraction Conformance/Non-Conformance Comments:

40ML Vial Lot # 03-40BTS721.

KD Bath ID: N/A

Envap ID: NEVAP-02

KD Bath Temperature: N/A

Envap Temperature: 40 °C

Date / Time	Prepped Sample Relinquished By/Location	Received By/Location
8/12/25	RS (Ext Lab)	Y. P. PestipCB
11:35	Preparation Group	Analysis Group

Analytical Method: N/A

Concentration Date: 08/12/2025

Sample ID	Client Sample ID	Test	g/mL	PH	Surr/Spike By:		Final Vol. (mL)	JarID	Comments	Prep Pos
					AddedBy	VerifiedBy				
PB169205BL	ABLK205	PCB	30.03	N/A	ritesh	RUPESH	10			U1-1
PB169205BS	ALCS205	PCB	30.01	N/A	ritesh	RUPESH	10			2
Q2818-01	B-2-5-1	PCB	30.07	N/A	ritesh	RUPESH	10	B		3
Q2818-02	B-3-5-2	PCB	30.03	N/A	ritesh	RUPESH	10	B		4
Q2819-06	11M-S	PCB	30.05	N/A	ritesh	RUPESH	10	E		5
Q2819-11	84SB-W	PCB	30.02	N/A	ritesh	RUPESH	10	E		6
Q2819-14	17M-W	PCB	30.08	N/A	ritesh	RUPESH	10	E		U2-1
Q2820-09	705R-S	PCB	30.06	N/A	ritesh	RUPESH	10	E		2
Q2820-10	SOIL-DUP-2	PCB	30.09	N/A	ritesh	RUPESH	10	E		3
Q2820-24	22M-S	PCB	30.01	N/A	ritesh	RUPESH	10	E		4
Q2823-01	SU-04-081125	PCB	30.04	N/A	ritesh	RUPESH	10	E		5
Q2826-01	WC1	PCB	30.08	N/A	ritesh	RUPESH	10			6
Q2827-01	TP-8	PCB	30.07	N/A	ritesh	RUPESH	10	E		U3-1
Q2827-05	TP-9	PCB	30.02	N/A	ritesh	RUPESH	10	E	Gasoline Smell	2
Q2830-01	BIN0009-DRIVEWAY-TP-S OUTH-EAST	PCB	30.06	N/A	ritesh	RUPESH	10	E		3
Q2830-01MS	BIN0009-DRIVEWAY-TP-S OUTH-EASTMS	PCB	30.01	N/A	ritesh	RUPESH	10	E		4
Q2830-01MS D	BIN0009-DRIVEWAY-TP-S OUTH-EASTMSD	PCB	30.03	N/A	ritesh	RUPESH	10	E		5
Q2830-03	BIN0009-DRIVEWAY-TP- WEST	PCB	30.09	N/A	ritesh	RUPESH	10	E		6
Q2830-05	BIN0009-DRIVEWAY-TP- WESTSIDE	PCB	30.06	N/A	ritesh	RUPESH	10	E		U6-1
Q2830-07	BIN0009-DRIVEWAY-TP-E ASTSIDE	PCB	30.04	N/A	ritesh	RUPESH	10	E		2
Q2831-01	VNJ-238	PCB	30.08	N/A	ritesh	RUPESH	10	B		3

69205
08-15

WORKLIST(Hardcopy Internal Chain)

WorkList Name : Q2818 WorkList ID : 191213 Department : Extraction Date : 08-12-2025 08:10:04

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q2818-01	B-2-5-1	Solid	PCB	Cool 4 deg C	EARTH03	D31	08/11/2025	8082A
Q2818-02	B-3-5-2	Solid	PCB	Cool 4 deg C	EARTH03	D31	08/11/2025	8082A
Q2819-06	11M-S	Solid	PCB	Cool 4 deg C	FIRS02	D31	08/05/2025	8082A
Q2819-11	84SB-W	Solid	PCB	Cool 4 deg C	FIRS02	D31	08/05/2025	8082A
Q2819-14	17M-W	Solid	PCB	Cool 4 deg C	FIRS02	D31	08/05/2025	8082A
Q2820-09	705R-S	Solid	PCB	Cool 4 deg C	FIRS02	D31	08/07/2025	8082A
Q2820-10	SOIL-DUP-2	Solid	PCB	Cool 4 deg C	FIRS02	D31	08/07/2025	8082A
Q2820-24	22M-S	Solid	PCB	Cool 4 deg C	FIRS02	D31	08/08/2025	8082A
Q2823-01	SU-04-081125	Solid	PCB	Cool 4 deg C	PSEG05	D31	08/11/2025	8082A
Q2826-01	WC1	Solid	PCB	Cool 4 deg C	GENV01	D31	08/11/2025	8082A
Q2827-01	TP-8	Solid	PCB	Cool 4 deg C	PSEG03	D31	08/11/2025	8082A
Q2827-05	TP-9	Solid	PCB	Cool 4 deg C	PSEG03	D31	08/11/2025	8082A
Q2830-01	BIN009-DRIVEWAY-TP-SOUTH	Solid	PCB	Cool 4 deg C	PSEG03	D31	08/11/2025	8082A
Q2830-03	BIN009-DRIVEWAY-TP-WEST	Solid	PCB	Cool 4 deg C	PSEG03	D31	08/11/2025	8082A
Q2830-05	BIN009-DRIVEWAY-TP-WEST	Solid	PCB	Cool 4 deg C	PSEG03	D31	08/11/2025	8082A
Q2830-07	BIN009-DRIVEWAY-TP-EASTS	Solid	PCB	Cool 4 deg C	PSEG03	D31	08/11/2025	8082A
Q2831-01	VNJ-238	Solid	PCB	Cool 4 deg C	PSEG03	D21	08/11/2025	8082A

Date/Time 08/12/25 8:10
Raw Sample Received by: RJ (EAT-106)
Raw Sample Relinquished by: CR

Date/Time 08/12/25 8:40
Raw Sample Received by: CR
Raw Sample Relinquished by: RJ (EAT-106)

Prep Standard - Chemical Standard Summary

Order ID : Q2818

Test : PCB

Prepbatch ID : PB169205,

Sequence ID/Qc Batch ID: PO081225,PP081225,

Standard ID :

EP2610,EP2627,EP2632,PP24329,PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,PP24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369,PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387,PP24650,PP24663,

Chemical ID :

E3804,E3875,E3877,E3940,E3944,E3949,E3950,E3951,E3962,M6157,P11522,P12699,P12702,P12931,P12936,P12949,P12950,P12957,P13356,P13373,P13381,P13589,P13591,P13697,P13702,P13786,P13830,P13878,P13883,W3112,W3177,



<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
314	1.1 H2SO4 SOLN	EP2610	05/07/2025	11/07/2025	RUPESHKUMAR SHAH	Extraction_SCALE_2 (EX-SC-2)	None	Riteshkumar Patel 05/07/2025
<u>FROM</u>	1000.00000ml of M6157 + 1000.00000ml of W3112 = Final Quantity: 2000.000 ml							

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
230	1:1ACETONE/HEXANE	EP2627	07/15/2025	01/15/2026	RUPESHKUMAR SHAH	None	None	Riteshkumar Patel 07/15/2025
<u>FROM</u> 4000.00000ml of E3949 + 4000.00000ml of E3950 = Final Quantity: 8000.000 ml								

Extractions STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3923	Baked Sodium Sulfate	EP2632	08/11/2025	01/28/2026	RUPESHKUMAR SHAH	Extraction_SC ALE_2 (EX-SC-2)	None	Riteshkumar Patel 08/11/2025

FROM 4000.00000gram of E3875 = Final Quantity: 4000.000 gram

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
84	Pest/PCB Surrogate Stock 20 PPM	PP24329	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza 04/03/2025

FROM 1.00000ml of P13356 + 9.00000ml of W3177 = Final Quantity: 10.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
202	AR1660 1000/100 ppb working solution 1st source	PP24330	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.10000ml of P13697 + 99.40000ml of W3177 + 0.50000ml of PP24329 = Final Quantity: 100.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
203	AR1660 750 PPB STD	PP24331	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.25000ml of W3177 + 0.75000ml of PP24330 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
204	AR1660 500 PPB STD	PP24332	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.50000ml of W3177 + 0.50000ml of PP24330 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
205	AR1660 250 PPB STD	PP24333	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.75000ml of W3177 + 0.25000ml of PP24330 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
206	AR1660 50 PPB STD	PP24334	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.90000ml of W3177 + 0.10000ml of PP24332 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
213	AR1221 1000 PPB WORKING SOLUTION	PP24335	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.10000ml of P13702 + 99.40000ml of W3177 + 0.50000ml of PP24329 = Final Quantity: 100.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1079	AR1221 750 PPB STD	PP24336	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.25000ml of W3177 + 0.75000ml of PP24335 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
222	AR1221 500 PPB STD	PP24337	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.50000ml of W3177 + 0.50000ml of PP24335 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1080	AR1221 250 PPB STD	PP24338	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.75000ml of W3177 + 0.25000ml of PP24335 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1081	AR1221 50 PPB STD	PP24339	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.90000ml of W3177 + 0.10000ml of PP24337 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
214	AR1232 1000 PPB WORKING SOLUTION	PP24340	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.10000ml of P13878 + 99.40000ml of W3177 + 0.50000ml of PP24329 = Final Quantity: 100.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1063	AR1232 750 PPB STD	PP24341	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.25000ml of W3177 + 0.75000ml of PP24340 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
223	AR1232 500 PPB STD	PP24342	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.50000ml of W3177 + 0.50000ml of PP24340 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1064	AR1232 250 PPB STD	PP24343	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.75000ml of W3177 + 0.25000ml of PP24340 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1065	AR1232 50 PPB STD	PP24344	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.90000ml of W3177 + 0.10000ml of PP24342 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
215	AR1242 1000 PPB WORKING STD	PP24345	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.10000ml of P12931 + 99.40000ml of W3177 + 0.50000ml of PP24329 = Final Quantity: 100.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1067	AR1242 750 PPB STD	PP24346	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.75000ml of W3177 + 0.75000ml of PP24345 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
224	AR1242 500 PPB STD	PP24347	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.50000ml of W3177 + 0.50000ml of PP24345 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1068	AR1242 250 PPB STD	PP24348	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.75000ml of W3177 + 0.25000ml of PP24345 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1069	AR1242 50 PPB STD	PP24349	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.90000ml of W3177 + 0.10000ml of PP24347 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
216	AR1248 1000 PPB WORKING STD	PP24350	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.10000ml of P12936 + 99.40000ml of W3177 + 0.50000ml of PP24329 = Final Quantity: 100.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1075	AR1248 750 PPB STD	PP24351	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.25000ml of W3177 + 0.75000ml of PP24350 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
225	AR1248 500 PPB STD	PP24352	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.50000ml of W3177 + 0.50000ml of PP24350 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1076	AR1248 250 PPB STD	PP24353	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.75000ml of W3177 + 0.25000ml of PP24350 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1077	AR1248 50 PPB STD	PP24354	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.90000ml of W3177 + 0.10000ml of PP24352 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
217	AR1254 1000 PPB WORKING STD	PP24355	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.10000ml of P13830 + 99.40000ml of W3177 + 0.50000ml of PP24329 = Final Quantity: 100.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1071	AR1254 750 PPB STD	PP24356	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.25000ml of W3177 + 0.75000ml of PP24355 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
226	AR1254 500 PPB STD	PP24357	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.50000ml of W3177 + 0.50000ml of PP24355 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1072	AR1254 250 PPB STD	PP24358	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.75000ml of W3177 + 0.25000ml of PP24355 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1073	AR1254 50 PPB STD	PP24359	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.90000ml of W3177 + 0.10000ml of PP24357 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1529	AR1262 1000 PPB Working Solution	PP24360	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.10000ml of P13883 + 99.40000ml of W3177 + 0.50000ml of PP24329 = Final Quantity: 100.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3753	AR1262 750 PPB STD	PP24361	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.25000ml of W3177 + 0.75000ml of PP24360 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1530	AR1262 500 PPB STD	PP24362	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.50000ml of W3177 + 0.50000ml of PP24360 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3754	AR1262 250 PPB STD	PP24363	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.75000ml of W3177 + 0.25000ml of PP24360 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3755	AR1262 50 PPB STD	PP24364	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.90000ml of W3177 + 0.10000ml of PP24362 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1532	AR1268 1000 PPB Working Solution	PP24365	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.10000ml of P13381 + 99.40000ml of W3177 + 0.50000ml of PP24329 = Final Quantity: 100.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3820	AR1268 750 PPB STD	PP24366	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.25000ml of W3177 + 0.75000ml of PP24365 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1533	AR1268 500 PPB STD	PP24367	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.50000ml of W3177 + 0.50000ml of PP24365 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3821	AR1268 250 PPB STD	PP24368	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.75000ml of W3177 + 0.25000ml of PP24365 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3822	AR1268 50 PPB STD	PP24369	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.90000ml of W3177 + 0.10000ml of PP24367 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
404	AR1660 100 PPM Stock Solution 2nd Source	PP24370	03/18/2025	09/18/2025	Yogesh Patel	None	None	Abdul Mirza 04/03/2025

FROM 1.00000ml of P12949 + 9.00000ml of E3804 = Final Quantity: 10.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
405	AR1660 1000/100 PPB ICV STD	PP24371	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza 04/03/2025

FROM 98.50000ml of W3177 + 0.50000ml of PP24329 + 1.00000ml of PP24370 = Final Quantity: 100.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
406	AR1660 500 PPB ICV	PP24372	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.50000ml of W3177 + 0.50000ml of PP24371 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3789	AR1221 1000 PPB WORKING SOL.2ND SOURCE(AGILENT)	PP24373	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 1.00000ml of P13373 + 98.50000ml of W3177 + 0.50000ml of PP24329 = Final Quantity: 100.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1886	AR1221 500 PPB ICV	PP24374	03/18/2025	08/12/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.50000ml of E3877 + 0.50000ml of W3177 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1887	AR1232 1000 PPB Working Sol. 2nd Source	PP24375	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 1.00000ml of P12699 + 98.50000ml of W3177 + 0.50000ml of PP24329 = Final Quantity: 100.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1888	AR1232 500 PPB ICV	PP24376	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.50000ml of W3177 + 0.50000ml of PP24375 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1889	AR1242 1000 PPB Working Sol. 2nd Source	PP24377	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 1.00000ml of P13589 + 98.50000ml of W3177 + 0.50000ml of PP24329 = Final Quantity: 100.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1891	AR1242 500 PPB ICV	PP24378	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.50000ml of W3177 + 0.50000ml of PP24377 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1890	AR1248 1000 PPB Working Sol. 2nd Source	PP24379	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 1.00000ml of P13591 + 98.50000ml of W3177 + 0.50000ml of PP24329 = Final Quantity: 100.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1892	AR1248 500 PPB ICV	PP24380	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.50000ml of W3177 + 0.50000ml of PP24379 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1893	AR1254 1000 PPB Working Sol. 2nd Source	PP24381	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 1.00000ml of P12957 + 98.50000ml of W3177 + 0.50000ml of PP24329 = Final Quantity: 100.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1894	AR1254 500 PPB ICV	PP24382	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.50000ml of W3177 + 0.50000ml of PP24381 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3757	AR1262 1000 PPB Working Solution second source	PP24384	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 1.00000ml of P12702 + 98.50000ml of W3177 + 0.50000ml of PP24329 = Final Quantity: 100.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3758	AR1262 500 PPB STD ICV	PP24385	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.50000ml of W3177 + 0.50000ml of PP24384 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3817	AR1268 1000 ppb Working Soln. 2nd source	PP24386	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 1.00000ml of P11522 + 98.50000ml of W3177 + 0.50000ml of PP24329 = Final Quantity: 100.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3823	AR1268 500 PPB STD ICV	PP24387	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.50000ml of W3177 + 0.50000ml of PP24386 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3857	5000 PPB PCB SPIKE SOLUTION 2ND SOURCE	PP24650	06/16/2025	12/11/2025	Abdul Mirza	None	None	Yogesh Patel
								07/21/2025

FROM 0.50000ml of P12950 + 99.50000ml of E3940 = Final Quantity: 100.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
465	200 PPB Pest/PCB Surrogate Spike	PP24663	06/24/2025	12/24/2025	Abdul Mirza	None	None	Yogesh Patel
								07/21/2025

FROM 1.00000ml of P13786 + 999.00000ml of E3944 = Final Quantity: 1000.000 ml

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	9005-05 / Acetone Ultra (cs/4x4L)	24E0761004	11/05/2025	10/01/2024 / Rajesh	09/25/2024 / Rajesh	E3804

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	PC19631-100 / SODIUM SULFATE, ANHYDROUS, PEST GRADE, 1	417203	01/28/2026	07/28/2025 / RUPESH	01/29/2025 / Rajesh	E3875

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9262-03 / Hexane, Ultra-Resi (cs/4x4L)	243570	08/12/2025	02/12/2025 / Rajesh	02/12/2025 / Rajesh	E3877

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9254-03 / Acetone, Ultra Resi (cs/4x4L)	24H1462005	12/11/2025	06/11/2025 / Rajesh	06/04/2025 / Rajesh	E3940

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9254-03 / Acetone, Ultra Resi (cs/4x4L)	24H1462005	05/24/2027	06/20/2025 / RUPESH	05/14/2025 / RUPESH	E3944

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9254-03 / Acetone, Ultra Resi (cs/4x4L)	24H2762008	04/18/2027	07/08/2025 / RITESHKUMAR	07/03/2025 / RUPESH	E3949

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9262-03 / Hexane, Ultra-Resi (cs/4x4L)	25C0362005	04/30/2026	07/08/2025 / RITESHKUMAR	07/03/2025 / RUPESH	E3950

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-3382-05 / Sand, Purified (cs/4x2.5kg)	25A2756718	12/31/2028	07/09/2025 / RUPESH	04/28/2020 / RUPESH	E3951

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9262-03 / Hexane, Ultra-Resi (cs/4x4L)	25C0362005	04/30/2026	08/05/2025 / RUPESH	07/30/2025 / RUPESH	E3962

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9673-33 / Sulfuric Acid, Instra-Analyzed (cs/6c2.5L)	24i1262013	11/07/2025	05/07/2025 / RUPESH	02/18/2025 / Mohan	M6157

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Agilent Technologies	PP-382-1 / Aroclor 1268	0006587800	09/18/2025	03/18/2025 / yogesh	02/21/2022 / Ankita	P11522

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc	91867 / Aroclor 1232 100 ug/mL	020823	09/18/2025	03/18/2025 / yogesh	08/07/2023 / Ankita	P12699

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc	x9166 / Aroclor 1262 100 ug/mL	060523	09/18/2025	03/18/2025 / yogesh	08/07/2023 / Ankita	P12702

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32009 / PCB Mix, Aroclor 1242, 1000ug/mL, Hexane, 1mL/ampul	a0203672	09/18/2025	03/18/2025 / yogesh	12/07/2023 / Ankita	P12931

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32010 / PCB Mix, Aroclor 1248, 1000ug/mL, Hexane, 1mL/ampul	a0202803	09/18/2025	03/18/2025 / yogesh	12/07/2023 / Ankita	P12936

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	20064 / Aroclor 1016/1260	022023	09/18/2025	03/18/2025 / yogesh	12/20/2023 / Yogesh	P12949

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	20064 / Aroclor 1016/1260	022023	12/16/2025	06/16/2025 / Abdul	12/20/2023 / Yogesh	P12950

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	/ Aroclor 1254	121823	04/03/2025	10/03/2024 / Ankita	12/20/2023 / Yogesh	P12957

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32000 / Pesticide Mix, CLP method, Pesticide Surrogate Mix, 200ug/mL, Acetone, 1mL	A0206810	09/18/2025	03/18/2025 / yogesh	04/22/2024 / Abdul	P13356

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Agilent Technologies	PP-292-1 / Aroclor 1221	0006783205	09/18/2025	03/18/2025 / yogesh	05/02/2024 / Ankita	P13373

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32410 / PCB Stock Solution, Aroclor 1268 Std, 1mL, Hexane	A0207475	09/18/2025	03/18/2025 / yogesh	05/03/2024 / Abdul	P13381

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Agilent Technologies	PP-312-1 / Aroclor 1242	0006665550	09/18/2025	03/18/2025 / yogesh	10/14/2024 / Ankita	P13589

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Agilent Technologies	PP-342-1 / Aroclor 1248	0006726317	09/18/2025	03/18/2025 / yogesh	10/14/2024 / Ankita	P13591

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32039 / PCB Mix, Aroclor 1016/1260, 1000ug/mL, hexane, 1mL/ampul	A0210629	09/18/2025	03/18/2025 / yogesh	10/17/2024 / yogesh	P13697

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32007 / PCB Mix, Aroclor 1221, 1000ug/mL, Hexane, 1mL/ampul	A0215270	09/18/2025	03/18/2025 / yogesh	10/17/2024 / yogesh	P13702

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32000 / Pesticide Mix, CLP method, Pesticide Surrogate Mix, 200ug/mL, Acetone, 1mL	A0214495	12/24/2025	06/24/2025 / Abdul	11/19/2024 / Ankita	P13786

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32011 / PCB Mix, Aroclor 1254, 1000ug/mL, Hexane, 1mL/ampul	A0217391	09/18/2025	03/18/2025 / yogesh	12/09/2024 / Ankita	P13830

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32008 / PCB Mix, Aroclor 1232, 1000ug/mL, Hexane, 1mL/ampul	A0219655	09/18/2025	03/18/2025 / yogesh	01/23/2025 / Ankita	P13878

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32409 / PCB Stock Solution, Aroclor 1262 Std, 1mL, Hexane	A0220950	09/18/2025	03/18/2025 / yogesh	01/23/2025 / Ankita	P13883

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	DIW / DI Water	Daily Lab-Certified	07/03/2029	07/03/2024 / lwona	07/03/2024 / lwona	W3112

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9262-03 / Hexane, Ultra-Resi (cs/4x4L)	24G1962003	08/22/2025	02/03/2025 / jignesh	01/31/2025 / jignesh	W3177

Material No.: 9005-05
Batch No.: 24E0761004
Manufactured Date: 2024-05-02
Retest Date: 2029-05-01
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
Assay ((CH ₃) ₂ CO) (by GC, corrected for water)	≥ 99.5 %	99.8 %
Color (APHA)	≤ 10	< 5
Residue after Evaporation	≤ 5 ppm	< 1 ppm
Titration Acid (μeq/g)	≤ 0.3	0.1
Titration Base (μeq/g)	≤ 0.5	0.1
Water (H ₂ O)	≤ 0.5 %	0.1 %
Solubility in H ₂ O	Passes Test	Passes Test
Chloride (Cl)	≤ 0.2 ppm	< 0.2 ppm
Phosphate (PO ₄)	≤ 0.05 ppm	< 0.05 ppm
Trace Impurities – Aluminum (Al)	≤ 50.0 ppb	< 5.0 ppb
Arsenic and Antimony (as As)	≤ 5.0 ppb	< 5.0 ppb
Trace Impurities – Barium (Ba)	≤ 20.0 ppb	< 1.0 ppb
Trace Impurities – Beryllium (Be)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities – Bismuth (Bi)	≤ 20.0 ppb	< 10.0 ppb
Trace Impurities – Boron (B)	≤ 10.0 ppb	< 5.0 ppb
Trace Impurities – Cadmium (Cd)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities – Calcium (Ca)	≤ 25.0 ppb	3.6 ppb
Trace Impurities – Chromium (Cr)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities – Cobalt (Co)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities – Copper (Cu)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities – Gallium (Ga)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities – Germanium (Ge)	≤ 10.0 ppb	< 10.0 ppb
Trace Impurities – Gold (Au)	≤ 20 ppb	< 5 ppb
Trace Impurities – Iron (Fe)	≤ 20.0 ppb	< 1.0 ppb
Trace Impurities – Lead (Pb)	≤ 10.0 ppb	< 10.0 ppb
Trace Impurities – Lithium (Li)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities – Magnesium (Mg)	≤ 20 ppb	< 1 ppb
Trace Impurities – Manganese (Mn)	≤ 10.0 ppb	< 1.0 ppb

Recd by RP on 9/25/24

E 3804

>>> Continued on page 2 >>>

Acetone
CMOS

 **avantor™**



Material No.: 9005-05
Batch No.: 24E0761004

Test	Specification	Result
Trace Impurities – Molybdenum (Mo)	≤ 10.0 ppb	< 5.0 ppb
Trace Impurities – Nickel (Ni)	≤ 10.0 ppb	< 5.0 ppb
Trace Impurities – Niobium (Nb)	≤ 50.0 ppb	< 1.0 ppb
Trace Impurities – Potassium (K)	≤ 10.0 ppb	< 10.0 ppb
Trace Impurities – Silicon (Si)	≤ 50 ppb	< 10 ppb
Trace Impurities – Silver (Ag)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities – Sodium (Na)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities – Strontium (Sr)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities – Tantalum (Ta)	≤ 50.0 ppb	< 5.0 ppb
Trace Impurities – Thallium (Tl)	≤ 10.0 ppb	< 5.0 ppb
Trace Impurities – Tin (Sn)	≤ 20.0 ppb	< 10.0 ppb
Trace Impurities – Titanium (Ti)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities – Vanadium (V)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities – Zinc (Zn)	≤ 20.0 ppb	7.9 ppb
Trace Impurities – Zirconium (Zr)	≤ 10.0 ppb	< 1.0 ppb
Particle Count – 0.5 µm and greater (Rion KS42AF)	≤ 100 par/ml	8 par/ml
Particle Count – 1.0 µm and greater (Rion KS42AF)	≤ 8 par/ml	2 par/ml

>>> Continued on page 3 >>>

Acetone
CMOS

 **avantor™**



Material No.: 9005-05
Batch No.: 24E0761004

Test	Specification	Result
------	---------------	--------

For Microelectronic Use

Country of Origin: USA
Packaging Site: Paris Mfg Ctr & DC



Michelle Bales
Sr. Manager, Quality Assurance



**PRODUCTOS
QUÍMICOS
MONTERREY, S.A. DE C.V.**

MIRADOR 201, COL. MIRADOR
MONTERREY, N.L. MÉXICO
CP 64070
TEL +52 81 13 52 67 67
www.pqm.com.mx

CERTIFICATE OF ANALYSIS

PRODUCT :	SODIUM SULFATE CRYSTALS ANHYDROUS		
QUALITY :	ACS (CODE RMB3375)	FORMULA :	Na ₂ SO ₄
SPECIFICATION NUMBER:	6399	RELEASE DATE:	MAY/23/2024
LOT NUMBER :	417203		

TEST	SPECIFICATIONS	LOT VALUES
Assay (Na ₂ SO ₄)	Min. 99.0%	99.8 %
pH of a 5% solution at 25°C	5.2 - 9.2	6.2
Insoluble matter	Max. 0.01%	0.001 %
Loss on ignition	Max. 0.5%	0.1 %
Chloride (Cl)	Max. 0.001%	<0.001 %
Nitrogen compounds (as N)	Max. 5 ppm	<5 ppm
Phosphate (PO ₄)	Max. 0.001%	<0.001 %
Heavy metals (as Pb)	Max. 5 ppm	<5 ppm
Iron (Fe)	Max. 0.001%	<0.001 %
Calcium (Ca)	Max. 0.01%	0.001 %
Magnesium (Mg)	Max. 0.005%	0.001 %
Potassium (K)	Max. 0.008%	0.001 %
Extraction-concentration suitability	Passes test	Passes test
Appearance	Passes test	Passes test
Identification	Passes test	Passes test
Solubility and foreign matter	Passes test	Passes test
Retained on US Standard No. 10 sieve	Max. 1%	0.2 %
Retained on US Standard No. 60 sieve	Min. 94%	96.2 %
Through US Standard No. 60 sieve	Max. 5%	3.5 %
Through US Standard No. 100 sieve	Max. 10%	0.1 %

COMMENTS

QC: PhC Irma Belmares

If you need further details, please call our factory or contact our local distributor.

RE-02-01, Ed. 3

E 3875

Certificate of Analysis

1 Reagent Lane
 Fair Lawn, NJ 07410
 201.796.7100 tel
 201.796.1329 fax

Thermo Fisher Scientific's Quality System has been found to conform to Quality Management System
 Standard ISO9001:2015 by SAI Global Certificate Number CERT – 0120633

This is to certify that units of the lot number below were tested and found to comply with the specifications of the grade listed. Certain data have been supplied by third parties. Thermo Fisher Scientific expressly disclaims all warranties, expressed or implied, including the implied warranties of merchantability and fitness for a particular purpose. Products are for research use or further manufacturing. Not for direct administration to humans or animals. It is the responsibility of the final formulator and end user to determine suitability based upon the intended use of the end product. Products are tested to meet the analytical requirements of the noted grade. The following information is the actual analytical results obtained.

Catalog Number	H303	Quality Test / Release Date	11/07/2024
Lot Number	243570		
Description	HEXANES - OPTIMA		
Country of Origin	United States	Suggested Retest Date	Nov/2029
Chemical Origin	Organic - non animal		
BSE/TSE Comment	No animal products are used as starting raw material ingredients, or used in processing, including lubricants, processing aids, or any other material that might migrate to the finished product.		

N/A			
Result Name	Units	Specifications	Test Value
APPEARANCE		REPORT	Clear, colorless liquid
ASSAY (N-HEXANE)	%	>= 60	69
ASSAY (SUM C6 HYDROCARBONS)	%	>= 99.9	>99.9
COLOR	APHA	<= 5	<5
DENSITY AT 25 DEGREES C	GM/ML	Inclusive Between 0.653 - 0.673	0.669
EVAPORATION RESIDUE	ppm	<= 1	<1
FLUORESCENCE BACKGROUND	ppb	<= 1	<1
IDENTIFICATION	PASS/FAIL	= PASS TEST	PASS TEST
OPTICAL ABS AT 195 NM	ABS. UNITS	<= 1	0.74
OPTICAL ABS AT 210 NM	ABS. UNITS	<= 0.25	0.17
OPTICAL ABS AT 220 NM	ABS. UNITS	<= 0.07	0.05
OPTICAL ABS AT 254 NM	ABS. UNITS	<= 0.005	0.001
PESTICIDE RESIDUE ANALYSIS	NG/L	<= 10	<10
REFRACTIVE INDEX @ 25 DEG C		Inclusive Between 1.375 - 1.385	1.379
SUITABILITY FOR GC/MS		= PASS TEST	PASS TEST
SULFUR COMPOUNDS	%	<= 0.005	<0.005
THIOPHENE	PASS/FAIL	= PASS TEST	PASS TEST
WATER (H2O)	%	<= 0.01	<0.01
WATER-SOLUBLE TITRABLE ACID	MEQ/G	<= 0.0003	0.0001

Recd. by RP on 2/12/25

Harout Sahagian **E3877**

Harout Sahagian - Quality Control Manager - Fair Lawn

Note: The data listed is valid for all package sizes of this lot of this product, expressed as an extension of this catalog number listed above.
 If there are any questions with this certificate, please call at (800) 227-6701.

*Based on suggested storage condition.

Acetone

BAKER RESI-ANALYZED® Reagent
For Organic Residue Analysis

avantor™



Material No.: 9254-03

Batch No.: 24H1462005

Manufactured Date: 2024-05-24

Expiration Date: 2027-05-24

Revision No.: 0

Certificate of Analysis

Test	Specification	Result
Assay ((CH ₃) ₂ CO) (by GC, corrected for water)	>= 99.4 %	99.8 %
Color (APHA)	<= 10	5
Residue after Evaporation	<= 1.0 ppm	0.2 ppm
Substances Reducing Permanganate	Passes Test	Passes Test
Titration Acid (µeq/g)	<= 0.3	0.2
Titration Base (µeq/g)	<= 0.6	<0.1
Water (H ₂ O)	<= 0.5 %	0.2 %
FID-Sensitive Impurities (as 2-Octanol) Single Impurity Peak (ng/mL)	<= 5	<1
ECD Sensitive Impurities (as Heptachlor Epoxide) Single Peak (pg/mL)	<= 10	1

For Laboratory, Research, or Manufacturing Use

MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: United States

Packaging Site: Phillipsburg Mfg Ctr & DC

Recd by RP on 6/11/25

E3940

Jamie Croak
Director Quality Operations, Bioscience Production

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700

Avantor Performance Materials LLC

Acetone
BAKER RESI-ANALYZED® Reagent
For Organic Residue Analysis

Avantor™



Material No.: 9254-03

Batch No.: 24H1462005

Manufactured Date: 2024-05-24

Expiration Date: 2027-05-24

Revision No.: 0

Certificate of Analysis

Test	Specification	Result
Assay ((CH ₃) ₂ CO) (by GC, corrected for water)	>= 99.4 %	99.8 %
Color (APHA)	<= 10	5
Residue after Evaporation	<= 1.0 ppm	0.2 ppm
Substances Reducing Permanganate	Passes Test	Passes Test
Titration Acid (μeq/g)	<= 0.3	0.2
Titration Base (μeq/g)	<= 0.6	<0.1
Water (H ₂ O)	<= 0.5 %	0.2 %
FID-Sensitive Impurities (as 2-Octanol) Single Impurity Peak (ng/mL)	<= 5	<1
ECD Sensitive Impurities (as Heptachlor Epoxide) Single Peak (pg/mL)	<= 10	1

For Laboratory, Research, or Manufacturing Use
MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: United States
Packaging Site: Phillipsburg Mfg Ctr & DC

E3944

Jamie Croak
Director Quality Operations, Bioscience Production

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700

Avantor Performance Materials LLC

Acetone

BAKER RESI-ANALYZED® Reagent
For Organic Residue Analysis

Avantor™



Material No.: 9254-03

Batch No.: 24H2762008

Manufactured Date: 2024-04-18

Expiration Date: 2027-04-18

Revision No.: 0

Certificate of Analysis

Test	Specification	Result
Assay ((CH ₃) ₂ CO) (by GC, corrected for water)	>= 99.4 %	100.0 %
Color (APHA)	<= 10	5
Residue after Evaporation	<= 1.0 ppm	0.0 ppm
Substances Reducing Permanganate	Passes Test	Passes Test
Titration Acid (µeq/g)	<= 0.3	0.2
Titration Base (µeq/g)	<= 0.6	<0.1
Water (H ₂ O)	<= 0.5 %	<0.1 %
FID-Sensitive Impurities (as 2-Octanol) Single Impurity Peak (ng/mL)	<= 5	1
ECD Sensitive Impurities (as Heptachlor Epoxide) Single Peak (pg/mL)	<= 10	1

For Laboratory, Research, or Manufacturing Use

MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: United States

Packaging Site: Phillipsburg Mfg Ctr & DC

Reel on 7/2/25

E3949

Jamie Croak
Director Quality Operations, Bioscience Production

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700

Avantor Performance Materials LLC

n-Hexane 95%
ULTRA RESI-ANALYZED
For Organic Residue Analysis



Material No.: 9262-03

Batch No.: 25C0362005

Manufactured Date: 2025-01-29

Expiration Date: 2026-04-30

Revision No.: 0

Certificate of Analysis

Test	Specification	Result
FID-Sensitive Impurities (as 2-Octanol) Single Impurity Peak (ng/mL)	≤ 5	1
ECD Sensitive Impurities (as Heptachlor Epoxide) Single Peak (pg/mL)	≤ 10	6
ECD-Sensitive Impurities (as Ethylene Dibromide) - Single Impurity Peak (ng/mL)	≤ 5	5
Assay (Total Saturated C ₆ Isomers) (by GC, corrected for water)	$\geq 99.5 \%$	100.0 %
Assay (as n-Hexane) (by GC, corrected for water)	$\geq 95 \%$	100 %
Color (APHA)	≤ 10	10
Residue after Evaporation	≤ 1.0 ppm	0.1 ppm
Substances Darkened by H ₂ SO ₄	Passes Test	Passes Test
Water (by KF, coulometric)	$\leq 0.05 \%$	$< 0.01 \%$

For Laboratory, Research, or Manufacturing Use
MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: United States
Packaging Site: Phillipsburg Mfg Ctr & DC

3950

Recd on 7/02/25

Jamie Croak
Director Quality Operations, Bioscience Production



Material	BDH9274-2.5KG
Material Description	BDH SAND STDD OTTAWA W+I 2.5KG
Grade	NOT APPLICABLE
Batch	25A2756718
Reassay Date	12/31/2028
CAS Number	14808-60-7
Molecular Formula	SiO ₂
Molecular Mass	60.09
Date of Manufacture	12/05/2024
Storage	Room Temperature

Characteristics	Specifications	Measured Values
Appearance	Beige granules.	Beige granules.
Moisture	<= 0.1 %	0.1 %
Particle Size 30-40 mesh	>= 80 %	99 %
CUSTOMER PART # BDH9274-2.5KG		

Received on 1/12/25.

E3951

Internal ID #: 793

Signature	Additional Information
We certify that this batch conforms to the specifications listed above. This document has been electronically produced and is valid without a signature. Leona Edwardson, Quality Control Sr. Manager - Solon VWR Chemicals, LLC. 28600 Fountain Parkway, Solon OH 44139 USA	Analysis may have been rounded to significant digits in specification limits Product meets analytical specifications of the grades listed.

n-Hexane 95%
ULTRA RESI-ANALYZED
For Organic Residue Analysis



Material No.: 9262-03
Batch No.: 25C0362005
Manufactured Date: 2025-01-29
Expiration Date: 2026-04-30
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
FID-Sensitive Impurities (as 2-Octanol) Single Impurity Peak (ng/mL)	≤ 5	1
ECD Sensitive Impurities (as HeptachlorEpoxide) Single Peak (pg/mL)	≤ 10	6
ECD-Sensitive Impurities (as EthyleneDibromide) - Single Impurity Peak (ng/mL)	≤ 5	5
Assay (Total Saturated C ₆ Isomers) (by GC, corrected for water)	$\geq 99.5 \%$	100.0 %
Assay (as n-Hexane) (by GC, corrected for water)	$\geq 95 \%$	100 %
Color (APHA)	≤ 10	10
Residue after Evaporation	≤ 1.0 ppm	0.1 ppm
Substances Darkened by H ₂ SO ₄	Passes Test	Passes Test
Water (by KF, coulometric)	$\leq 0.05 \%$	$< 0.01 \%$

For Laboratory, Research, or Manufacturing Use
MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: United States
Packaging Site: Phillipsburg Mfg Ctr & DC

Received on 7/30/25

E3962

Jamie Croak
Director Quality Operations, Bioscience Production

Sulfuric Acid
BAKER INSTRA-ANALYZED® Reagent
For Trace Metal Analysis
Low Selenium

avantor™



MG157
MS

Material No.: 9673-33

Batch No.: 24I1262013

Manufactured Date: 2024-08-07

Retest Date: 2029-08-06

Revision No.: 0

Certificate of Analysis

Test	Specification	Result
ACS – Assay (H ₂ SO ₄)	95.0 – 98.0 %	96.2 %
Appearance	Passes Test	Passes Test
ACS – Color (APHA)	<= 10	5
ACS – Residue after Ignition	<= 3 ppm	<1 ppm
ACS – Substances Reducing Permanganate(as SO ₂)	<= 2 ppm	<2 ppm
Ammonium (NH ₄)	<= 1 ppm	<1 ppm
Chloride (Cl)	<= 0.1 ppm	<0.1 ppm
Nitrate (NO ₃)	<= 0.2 ppm	0.1 ppm
Phosphate (PO ₄)	<= 0.5 ppm	<0.1 ppm
Trace Impurities – Aluminum (Al)	<= 30.0 ppb	<5.0 ppb
Arsenic & Antimony (as As)	<= 4.0 ppb	<2.0 ppb
Trace Impurities – Boron (B)	<= 10.0 ppb	<5.0 ppb
Trace Impurities – Cadmium (Cd)	<= 2.0 ppb	<1.0 ppb
Trace Impurities – Chromium (Cr)	<= 6.0 ppb	<1.0 ppb
Trace Impurities – Cobalt (Co)	<= 0.5 ppb	<0.3 ppb
Trace Impurities – Copper (Cu)	<= 1.0 ppb	<1.0 ppb
Trace Impurities – Gold (Au)	<= 10.0 ppb	<5.0 ppb
Heavy Metals (as Pb)	<= 500.0 ppb	<100.0 ppb
Trace Impurities – Iron (Fe)	<= 50.0 ppb	<1.0 ppb
Trace Impurities – Lead (Pb)	<= 0.5 ppb	<0.5 ppb
Trace Impurities – Magnesium (Mg)	<= 7.0 ppb	<1.0 ppb
Trace Impurities – Manganese (Mn)	<= 1.0 ppb	<1.0 ppb
Trace Impurities – Mercury (Hg)	<= 0.5 ppb	<0.1 ppb
Trace Impurities – Nickel (Ni)	<= 2.0 ppb	<0.3 ppb
Trace Impurities – Potassium (K)	<= 500.0 ppb	<10.0 ppb
Trace Impurities – Selenium (Se)	<= 50.0 ppb	7.2 ppb
Trace Impurities – Silicon (Si)	<= 100.0 ppb	12.8 ppb
Trace Impurities – Silver (Ag)	<= 1.0 ppb	<1.0 ppb

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700

Avantor Performance Materials LLC

Sulfuric Acid

BAKER INSTRA-ANALYZED® Reagent

For Trace Metal Analysis

Low Selenium

 **avantors**™



Material No.: 9673-33

Batch No.: 2411262013

Test	Specification	Result
Trace Impurities – Sodium (Na)	≤ 500.0 ppb	< 5.0 ppb
Trace Impurities – Strontium (Sr)	≤ 5.0 ppb	< 1.0 ppb
Trace Impurities – Tin (Sn)	≤ 5.0 ppb	1.1 ppb
Trace Impurities – Zinc (Zn)	≤ 5.0 ppb	< 1.0 ppb

For Laboratory, Research, or Manufacturing Use

Country of Origin: United States

Packaging Site: Phillipsburg Mfg Ctr & DC



Jamie Croak

Director Quality Operations, Bioscience Production

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700

Avantor Performance Materials LLC



Certificate of Analysis

P11518
↓
P11522
AJ
02/21/22

Product Name: Aroclor 1268 Standard

Product Number: PP-382-1

Lot Issue Date: 09-Feb-2021

Lot Number: 0006587800

Expiration Date: 31-Mar-2029

Description:

This analytical reference material (RM) was manufactured and verified in accordance with an ISO 9001 registered quality system and analyte concentrations were verified by an ISO 17025 accredited laboratory. The concentration and uncertainty value at the 95% confidence level for each analyte, determined gravimetrically, is listed below.

Analyte	CAS#	Analyte Lot	Concentration ± Uncertainty
Aroclor 1268	011100-14-4	RM00937	100.0 ± 0.5 µg/mL

Matrix: isooctane (2,2,4-trimethylpentane)

Storage Conditions: Store at Room Temperature (15° to 30°C).

Traceability:

The balances used for these measurements are calibrated with weights traceable to NIST in compliance with ANSI/NCCL Z540.3, ISO 9001, ISO 17025, and ISO 17034. Calibrated Class A glassware is used for volumetric measurements. Thermometers are calibrated against a NIST traceable thermometer in accordance with NIST Special Publication 1088.

Homogeneity:

This RM was unitized according to an in-house procedure and is guaranteed to be homogeneous. There is no minimum sub-sample size required.

Intended Use:

This RM is intended for the preparation of working reference samples for use in routine laboratory analyses, calibration of instruments, validation of analytical methods, assessments of measurement methods, and continuing calibration verification.

Instructions for Use:

Sample aliquots for analysis should be withdrawn at 20°C to 25°C immediately after opening the container and should be processed without delay for the certified values to be valid within the stated uncertainties.

Hazards:

Refer to the Safety Data Sheet on www.agilent.com for information regarding this RM.


Expiration of Certification:

The certification of this RM is valid until the expiration date specified above, provided the RM is handled and stored in accordance with the instructions given in this certificate. This certification is nullified if the RM is damaged, contaminated, or otherwise modified.

Maintenance of Certification:

If substantive changes are noted that affect the certification before the expiration of this certificate, Agilent will notify the purchaser.

Sample lot approver:


Monica Bourgeois
QMS Representative



ISO 17034 Cert
No. AR-1936

RM was produced in accordance with TUV USA Inc registered ISO 9001 Quality Management System. Cert # 56 100 18560026
Page: 1 of 1

www.agilent.com/quality/
CSD-QA-015.1



ISO 17025 Cert
No. AT-1937



CERTIFIED WEIGHT REPORT

Part Number: **91867**
 Lot Number: **020823**
 Description: **WP 037 - Aroclor 1232**
PCB Technical Mixture
 Expiration Date: **020833**
 Recommended Storage: **Ambient (20 °C)**
 Nominal Concentration ($\mu\text{g/mL}$): **100**
 NIST Test ID#: **6UTB**

Solvent(
Aceton


Weight(s) shown below were combined and diluted to (mL): **100.0**
 5E-05 Balance Uncertainty
 0.057 Flask Uncertainty

Compound	RM#	Lot Number	Nominal Conc ($\mu\text{g/mL}$)	Purity (%)	Uncertainty Purity	Target Weight (g)
1. Aroclor 1232	17	45-6A	100	100	0.5	0.01000

- The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
- Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
- Standards are certified (+/-) 0.5% of the stated value, unless otherwise stated.
- All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
- Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measure Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).

Comments

GC3-M1 Analysis by Melissa Stonier
 Column ID SPB-608 30 meter X 0.53mm X 5 μm film thickness
 Flow rates: Helium (carrier) = 5mL/min, Helium (make-up) = 25mL/min
 Hydrogen (make-up) = 30mL/min, Air (make-up) = 350mL/min
 Oven Profile: Temp 1 = 150°C (Time 1 = 4 min), Temp 2 = 290°C (Time 2 = 13.5 min)
 Rate = 8°C/min, Total run time = 35 min
 Injector temp. = 200°C, FID Temp. = 300°C. FID Signal = Edaq Channel 1
 Standard injection = 1.5 μL , Range=3






110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No.: 32009 **Lot No.:** A0203672

Description: Aroclor® 1242 Standard

Aroclor® 1242 Standard 1,000 µg/mL, Hexane, 1mL/ampul

Container Size: 2 mL **Pkg Amt:** > 1 mL

Expiration Date: January 31, 2030 **Storage:** 25°C nominal

Handling: This product contains PCBs. **Ship:** Ambient

p12928

→

p12932

AJ

12/27/23

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty* (95% C.L.; K=2)
1	Aroclor 1242	53469-21-9	01141	—%	1,004.7 µg/mL	+/- 55.7515

Solvent: Hexane

CAS # 110-54-3

Purity 99%

* Expanded Uncertainty displayed in same units as Grav. Conc.

Quality Confirmation Test

Column:

30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:

helium-constant pressure 20 psi.

Temp. Program:

200°C to 300°C
@ 25°C/min. (hold 10 min.)

Inj. Temp:

250°C

Det. Temp:

300°C

Det. Type:


ECD

Split Vent:

10 ml/min.

Inj. Vol

0.2ul



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Russ Boothamer - Operations Technician I

Date Mixed: 26-Oct-2023 Balance Serial # B442140311

Jennifer Polino - Operations Tech III - APM QC

Date Passed: 06-Nov-2023

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FW 80397




110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No.: 32010 **Lot No.:** A0202803

Description: Aroclor® 1248 Standard

Aroclor® 1248 Standard 1,000µg/mL, Hexane, 1mL/ampul

Container Size: 2 mL **Pkg Amt:** > 1 mL

Expiration Date: January 31, 2030 **Storage:** 25°C nominal

Handling: This product contains PCBs. **Ship:** Ambient

P1219697
P1219697
A1
12104123

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Aroclor 1248	12672-29-6	13897600	---%	1,001.7 µg/mL	+/- 55.5850

Solvent: Hexane

CAS # 110-54-3

Purity 99%

* Expanded Uncertainty displayed in same units as Grav. Conc.

Quality Confirmation Test

Column:

30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:

helium-constant pressure 20 psi.

Temp. Program:

200°C to 300°C
@ 25°C/min. (hold 10 min.)

Inj. Temp:

250°C

Det. Temp:

300°C

Det. Type:


ECD

Split Vent:

10 ml/min.

Inj. Vol

0.2ul



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

[Signature]
Laith Clemente - Operations Technician I

Date Mixed: 03-Oct-2023

Balance Serial # 1128360905

[Signature]
Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 09-Oct-2023

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FW 80397



Certified Reference Material CRM

CERTIFIED WEIGHT REPORT

Part Number:
20064

Lot Number:
022023

Description:
CLP PCB'S - Aroclor Mix
Aroclors 1016 & 1260

Expiration Date:
022033

Recommended Storage:
Ambient (20 °C)

Nominal Concentration (µg/mL):
1000

NIST Test ID#:
6UTB

Solvent(s):
Hexane

Lot#
273615

Formulated By:
Benson Chan

Reviewed By:
Pedro L. Rentas

DATE
022023

DATE
022023

Weight(s) shown below were combined and diluted to (mL.):
200.0

Balance Uncertainty
0.010

Flask Uncertainty
0.010

Part Number:
20064

Lot Number:
022023

Description:
CLP PCB'S - Aroclor Mix
Aroclors 1016 & 1260

Expiration Date:
022033

Recommended Storage:
Ambient (20 °C)

Nominal Concentration (µg/mL):
1000

NIST Test ID#:
6UTB

Solvent(s):
Hexane

Lot#
273615

Formulated By:
Benson Chan

Reviewed By:
Pedro L. Rentas

DATE
022023

DATE
022023

Weight(s) shown below were combined and diluted to (mL.):
200.0


Balance Uncertainty
0.010

Flask Uncertainty
0.010

Compound	RM#	Lot Number	Nominal Conc (µg/mL)	Purity (%)	Purity Uncertainty (%)	Target Weight(g)	Actual Weight(g)	Actual Conc (µg/mL)	Expanded Uncertainty (+/-) (µg/mL)	CAS#	OSHA PEL (TWA)	LD50
1. Aroclor 1016	15	020491JC	1000	100	0.2	0.20004	0.20060	1002.8	4.0	12674-11-2	N/A	N/A
2. Aroclor 1260	21	020491JC	1000	100	0.2	0.20004	0.20081	1003.9	4.0	11096-82-5	0.5mg/m3	orl-rat 1315mg/kg

* The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
* Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
* All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
* Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).

Comments
GC3-M1 Analysis by Melissa Stortier
Column ID SPB-608 30 meter X 0.53mm X5µm film thickness
Flow rates: Helium (carrier) = 5mL/min, Helium (make-up) = 25mL/min
Hydrogen (make-up) = 30mL/min, Air (make-up) = 350mL/min
Oven Profile: Temp 1 = 150°C (Time 1 = 4 min), Temp 2 = 280°C (Time 2 = 13.5 min)
Rate = 8°C/min, Total run time = 35 min
Injector temp. = 200°C, FID Temp. = 300°C, FID Signal = Etdaq Channel 1
Standard Injection = 1.5µL, Range=3





Certified Reference Material CRM

CERTIFIED WEIGHT REPORT

Part Number:
20064

Solvent(s):
Hexane

Lot#
273615

Lot Number:
022023

Formulated By:
Benson Chan

DATE
022023

Description:
CLP PCB'S - Aroclor Mix
Aroclors 1016 & 1260

Reviewed By:
Pedro L. Rentas

DATE
022023

Expiration Date:
Ambient (20 °C)

Expanded
Uncertainty

(+/-) (µg/mL)

Recommended Storage:
Nominal Concentration (µg/mL):
200.0

Actual
Conc (µg/mL)

Actual
Weight(g)

NIST Test ID#:
6UTB

Purity
(%)

Target
Weight(g)

Weight(s) shown below were combined and diluted to (mL):

Actual
Conc (µg/mL)

Actual
Weight(g)

SDS Information
(Solvent Safety Info. On Attached pg.)

OSHA PEL (TWA)


LD50

PR2946 718
↓
12/20/23
P12955

Compound	RM#	Lot Number	Nominal Conc (µg/mL)	Purity (%)	Purity Uncertainty	Target Weight(g)	Actual Weight(g)	Actual Conc (µg/mL)	Expanded Uncertainty (+/-) (µg/mL)	CAS#	OSHA PEL (TWA)	LD50
1. Aroclor 1016	15	020491JC	1000	100	0.2	0.20004	0.20060	1002.8	4.0	12674-11-2	N/A	N/A
2. Aroclor 1260	21	020491JC	1000	100	0.2	0.20004	0.20081	1003.9	4.0	11086-82-5	0.5mg/m3	orl-rat 1315mg/kg

* The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
* Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
* All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
* Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).

Comments
GC3-M1 Analysis by Melissa Stortier
Column ID SPB-608 30 meter X 0.53mm X5µm film thickness
Flow rates: Helium (carrier) = 5mL/min, Helium (make-up) = 25mL/min
Hydrogen (make-up) = 30mL/min, Air (make-up) = 350mL/min
Oven Profile: Temp 1 = 150°C (Time 1 = 4 min), Temp 2 = 280°C (Time 2 = 13.5 min)
Rate = 8°C/min, Total run time = 35 min
Injector temp. = 200°C, FID Temp. = 300°C, FID Signal = Etdaq Channel 1
Standard Injection = 1.5µL, Range=3





110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32000 **Lot No.:** A0206810

Description : Pesticide Surrogate Mix
Pesticide Surrogate Mix 200 µg/mL, Acetone, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : April 30, 2030 **Storage:** 10°C or colder

Handling: Contains PCBs - sonicate prior to use. **Ship:** Ambient

P13348
↓
P13357
10
DAUF
04/25/2024

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	2,4,5,6-Tetrachloro-m-xylene	877-09-8	RP220407	99%	200.3 µg/mL	+/- 11.1143
2	Decachlorobiphenyl (BZ# 209)	2051-24-3	30638	99%	200.6 µg/mL	+/- 11.1298

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Acetone
CAS # 67-64-1
Purity 99%

Tech Tips:

Decachlorobiphenyl has poor solubility in most organic solvents. The maximum concentration that can be prepared in acetone, hexane, or isooctane is 200µg/mL. Temperature will affect the solubility as well. Storing solutions at reduced temperatures will cause decachlorobiphenyl to precipitate.

Products containing decachlorobiphenyl must be sonicated for a minimum of 10 minutes prior to opening the ampul. Because each ultrasonic bath operates at a different energy level, 10 minutes is a guideline only. Longer sonication time will not affect product quality.

These precautions apply to working solutions prepared in your laboratory as well. The amount of compound that precipitates depends on concentration AND temperature. If you store your standards at a temperature lower than 4°C (even dilute solutions), allow extra sonication time.

Quality Confirmation Test

Column:

30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:

helium-constant pressure 20 psi.

Temp. Program:

200°C to 300°C
@ 25°C/min. (hold 10 min.)

Inj. Temp:

250°C

Det. Temp:

300°C

Det. Type:


ECD

Split Vent:


10 ml/min.

Inj. Vol

1µl




This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


Laith Clemente - Operations Technician I

Date Mixed: 22-Jan-2024


Balance Serial # 1128360905


Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 24-Jan-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

P 13348
↓
P 13357
10


SAUF
04/25/2025



ISO 17034

Reference Material Certificate

Product Information Sheet

Product Name: Aroclor 1221 Standard
Product Number: PP-292-1
Storage Conditions: Store at Room Temperature (15° to 30°C).

Lot Number: 0006783205
Lot Issue Date: 20-Feb-2024
Expiration Date: 31-Mar-2032

Component Name	Concentration	Uncertainty	CAS#	Analyte Lot
Aroclor 1221	100.3 ±	0.5 µg/mL	011104-28-2	NT01017

Matrix: isooctane (2,2,4-trimethylpentane)

Description:

This document is prepared in accordance with ISO 17034 and Guide 31. This analytical reference material standard was manufactured and verified in accordance with an ISO 9001 registered quality system and analyte concentrations were verified by an ISO 17025 accredited laboratory. The concentration and uncertainty value at the 95% confidence level for each analyte, determined gravimetrically, is listed above.

Traceability:

The balances used for these measurements are calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z540.3, ISO 9001, ISO 17025, and ISO 17034. Calibrated Class A glassware is used for volumetric measurements. Thermometers are calibrated against a NIST traceable thermometer in accordance with NIST Special Publication 1088.

Homogeneity:

This analytical reference standard was unitized according to an in-house procedure and is guaranteed to be homogeneous. There is no minimum sub-sample size required.

Instructions for Use:

Sample aliquots for analysis should be withdrawn at 20°C to 25°C immediately after opening the container and should be processed without delay for the certified values to be valid within the stated uncertainties.

Safety:

Refer to the Safety Data Sheet on www.agilent.com for information regarding this analytical reference material.

Intended Use:

This analytical reference standard is intended for the preparation of working reference samples for use in routine laboratory analyses, calibration of instruments, validation of analytical methods, assessments of measurement methods, and continuing calibration verification.

Expiration of Certification:

The certification of this analytical reference standard is valid until the expiration date specified above, provided the material is handled and stored in accordance with the instructions given in this certificate. This certification is nullified if the material is damaged, contaminated, or otherwise modified.

P13342
AJ
05106124
P13343

Page: 1 of 2

CSD-QA-015.2

ISO 17025

Cert No. AT-1937


250 Smith Street North Kingstown, Rhode Island 02852 www.agilent.com/quality



Maintenance of Certification:

If substantive changes are noted that affect the certification before the expiration of this certificate, Agilent will notify the purchaser.

Sample lot approver:


Monica Bourgeois
QMS Representative



ISO 17034
Cert No. AR-1936

RM was produced in accordance with the TUV/SUD registered ISO
9001:2015 Quality Management System. Cert# 951215321

Page: 2 of 2

www.agilent.com/quality/
CSD-QA-015.2

ISO 17025
Cert No. AT-1937

250 Smith Street North Kingstown, Rhode Island 02852 www.agilent.com/quality



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32410 **Lot No.:** A0207475

Description : Aroclor® 1268 Standard

Aroclor® 1268 Standard 1,000 µg/mL, 1mL/ampul, Hexane

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : May 31, 2030 **Storage:** 25°C nominal

Handling: This product contains PCBs. **Ship:** Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Aroclor 1268	11100-14-4	10947000	---%	1,000.0 µg/mL	+/- 55.4925

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Hexane
CAS # 110-54-3
Purity 99%

P 13380
↓
P 13381
↓
SAZU
05/6/2024

Quality Confirmation Test

Column:

30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:

helium-constant pressure 20 psi.

Temp. Program:

200°C to 300°C
@ 25°C/min. (hold 10 min.)

Inj. Temp:

250°C

Det. Temp:

300°C

Det. Type:


ECD

Split Vent:

Split ratio 500:1

Inj. Vol

0.2µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Michael Maye
Michael Maye - Operations Tech I

Date Mixed: 06-Feb-2024

Balance Serial # B442140311

Dillon Murphy
Dillon Murphy - Operations Technician I


Date Passed: 09-Feb-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

P13380 } ②
↓
P13381
1

[Signature]
05/6/2024

ISO 17034



Agilent

Trusted Answers

Reference Material Certificate

Product Name: Aroclor 1242 Standard

Lot Number: 0006665550

Product Number: PP-312-1

Lot Issue Date: 08-Feb-2022

Storage Conditions: Store at Room Temperature (15° to 30°C).

Expiration Date: 31-Jan-2027

Component Name	CERTIFIED VALUES		CAS#	Analyte Lot
	Concentration	Expanded Uncertainty		
Aroclor 1242	100.4	± 0.5 µg/mL	053469-21-9	NT01020

Matrix: isooctane (2,2,4-trimethylpentane)

Description:

This document is prepared in accordance with ISO 17034 and Guide 31. This analytical reference material standard was manufactured and verified in accordance with an ISO 9001 registered quality system and analyte concentrations were verified by an ISO 17025 accredited laboratory. The concentration and uncertainty value at the 95% confidence level for each analyte, determined gravimetrically, is listed above.

Traceability:

The balances used for these measurements are calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z540.3, ISO 9001, ISO 17025, and ISO 17034. Calibrated Class A glassware is used for volumetric measurements. Thermometers are calibrated against a NIST traceable thermometer in accordance with NIST Special Publication 1088.

Homogeneity:

This analytical reference standard was unitized according to an in-house procedure and is guaranteed to be homogeneous. There is no minimum sub-sample size required.

Instructions for Use:

Sample aliquots for analysis should be withdrawn at 20°C to 25°C immediately after opening the container and should be processed without delay for the certified values to be valid within the stated uncertainties.

Safety:

Refer to the Safety Data Sheet on www.agilent.com for information regarding this analytical reference material.

Intended Use:

This analytical reference standard is intended for the preparation of working reference samples for use in routine laboratory analyses, calibration of instruments, validation of analytical methods, assessments of measurement methods, and continuing calibration verification.

Expiration of Certification:

The certification of this analytical reference standard is valid until the expiration date specified above, provided the material is handled and stored in accordance with the instructions given in this certificate. This certification is nullified if the material is damaged, contaminated, or otherwise modified.

P13589
↓
P13590AJ
10/14/24

ISO 17034



Maintenance of Certification:

If substantive changes are noted that affect the certification before the expiration of this certificate, Agilent will notify the purchaser.

Sample lot approver:

Monica Bourgeois
QMS Representative



ISO 17034 Cert
No. AR-1936

RM was produced in accordance with the TUV/SUD registered ISO 9001:2015
Quality Management System. Cert# 95T215321

Page: 2 of 2

www.agilent.com/quality/

CSD-QA-015.1



ISO 17025
Cert No. AT-

Reference Material Certificate
Product Information Sheet

Product Name: Aroclor 1248 Standard

Lot Number: 0006726317

Product Number: PP-342-1

Lot Issue Date: 27-Jan-2023

Storage Conditions: Store at Room Temperature (15° to 30°C).

Expiration Date: 28-Feb-2031

Component Name	Concentration	Uncertainty	CAS#	Analyte Lot
Aroclor 1248	100.3 ±	0.5 µg/mL	012672-29-6	NT01582

Matrix: isooctane (2,2,4-trimethylpentane)

Description:

This document is prepared in accordance with ISO 17034 and Guide 31. This analytical reference material (RM) standard was manufactured and verified in accordance with an ISO 9001 registered quality system and analyte concentrations were verified by an ISO 17025 accredited laboratory. The concentration and uncertainty value at the 95% confidence level for each analyte, determined gravimetrically, is listed above. Purity values are taken from approved vendor raw material certificates.

Traceability:

The balances used for these measurements are calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z540.3, ISO 9001, ISO 17025, and ISO 17034. Calibrated Class A glassware is used for volumetric measurements. Thermometers are calibrated against a NIST traceable thermometer in accordance with NIST Special Publication 1088.

Homogeneity:

This analytical reference (RM) standard was unitized according to an in-house procedure and is guaranteed to be homogeneous. There is no minimum sub-sample size required.

Instructions for Use:

Sample aliquots for analysis should be withdrawn at 20°C to 25°C immediately after opening the container and should be processed without delay for the certified values to be valid within the stated uncertainties.

Safety:

Refer to the Safety Data Sheet on www.agilent.com for information regarding this analytical reference material.

Intended Use:

This analytical reference (RM) standard is intended for the preparation of working reference samples for use in routine laboratory analyses, calibration of instruments, validation of analytical methods, assessments of measurement methods, and continuing calibration verification.

Expiration of Certification:

The certification of this analytical reference standard (RM) is valid until the expiration date specified above, provided the material is handled and stored in accordance with the instructions given in this certificate. This certification is nullified if the material is damaged, contaminated, or otherwise modified.

P13591
↓
P13592

AS
10/14/2024



Maintenance of Certification:

If substantive changes are noted that affect the certification before the expiration of this certificate, Agilent will notify the purchaser.

Sample lot approver:

Monica Bourgeois
QMS Representative



ISO 17034 Cert
No. AR-1936

RM was produced in accordance with the TUV/SUD registered ISO
9001:2015 Quality Management System. Cert# 951215321

Page: 2 of 2

www.agilent.com/quality/

CSD-QA-015.1

ISO 17025



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32039 **Lot No.:** A0210629

Description : Aroclor® 1016/1260 Mix
Aroclor® 1016/1260 Mix 1,000 µg/mL, Hexane, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : July 31, 2030 **Storage:** 25°C nominal

Handling: This product contains PCBs. **Ship:** Ambient

P13697
↓
P13701 } Y.P.
10/19/24

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Aroclor 1016	12674-11-2	07	----%	1,005.3 µg/mL	+/- 55.7809
2	Aroclor 1260	11096-82-5	1320657	----%	1,000.0 µg/mL	+/- 55.4850

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Hexane
CAS # 110-54-3
Purity 99%

Quality Confirmation Test

Column:

30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:

helium-constant pressure 20 psi.

Temp. Program:

200°C to 300°C
@ 25°C/min. (hold 10 min.)

Inj. Temp:

250°C

Det. Temp:

300°C

Det. Type:


ECD

Split Vent:


10 ml/min.

Inj. Vol

0.2µl




This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


Laith Clemente - Operations Technician I

Date Mixed: 22-Apr-2024

Balance Serial # B442140311


Dillan Murphy - Operations Technician I

Date Passed: 24-Apr-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32007 **Lot No.:** A0215270

Description : Aroclor® 1221 Standard

Aroclor® 1221 Standard 1,000 µg/mL, Hexane, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : November 30, 2030 **Storage:** 25°C nominal

Handling: This product contains PCBs. **Ship:** Ambient

P13902 } Y.P.
P13703 } 10/17/24

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Aroclor 1221	11104-28-2	14969200	----%	1,005.0 µg/mL	+/- 55.7700

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Hexane
CAS # 110-54-3
Purity 99%

Quality Confirmation Test

Column:

30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:

helium-constant pressure 20 psi.

Temp. Program:

200°C to 300°C
@ 25°C/min. (hold 10 min.)

Inj. Temp:

250°C

Det. Temp:

300°C

Det. Type:


ECD

Split Vent:

10 ml/min.

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Michael Maye
Michael Maye - Operations Tech I

Date Mixed: 16-Aug-2024

Balance Serial # 1128360905

Jennifer Pollino
Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 20-Aug-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.




110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32000 **Lot No.:** A0214495

Description : Pesticide Surrogate Mix

Pesticide Surrogate Mix 200 µg/mL, Acetone, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : October 31, 2030 **Storage:** 10°C or colder

Handling: Contains PCBs - sonicate prior to use. **Ship:** Ambient

P19785
AJ
11/19/24
P19789

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	2,4,5,6-Tetrachloro-m-xylene	877-09-8	RP220407	99%	200.2 µg/mL	+/- 11.1087
2	Decachlorobiphenyl (BZ# 209)	2051-24-3	30679	99%	201.4 µg/mL	+/- 11.1753

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Acetone
CAS # 67-64-1
Purity 99%

Tech Tips:

Decachlorobiphenyl has poor solubility in most organic solvents. The maximum concentration that can be prepared in acetone, hexane, or isooctane is 200µg/mL. Temperature will affect the solubility as well. Storing solutions at reduced temperatures will cause decachlorobiphenyl to precipitate.

Products containing decachlorobiphenyl must be sonicated for a minimum of 10 minutes prior to opening the ampul. Because each ultrasonic bath operates at a different energy level, 10 minutes is a guideline only. Longer sonication time will not affect product quality.

These precautions apply to working solutions prepared in your laboratory as well. The amount of compound that precipitates depends on concentration AND temperature. If you store your standards at a temperature lower than 4°C (even dilute solutions), allow extra sonication time.

Quality Confirmation Test

Column:
30m x 25mm x 2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:
helium-constant pressure 20 psi.

Temp. Program:
200°C to 300°C
@ 25°C/min. (hold 10 min.)


Inj. Temp:
250°C

Det. Temp:
300°C

Det. Type:
ECD

Split Vent:
10 ml/min.

Inj. Vol
1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

A. O. P.
Aaron Enyart - Operations Tech I

Date Mixed: 29-Jul-2024 **Balance Serial #** B345965662

Jennifer Pollino
Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 01-Aug-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32011 **Lot No.:** A0217391

Description : Aroclor® 1254 Standard

Aroclor® 1254 Standard 1,000 µg/mL, Hexane, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : January 31, 2031 **Storage:** 25°C nominal

Handling: This product contains PCBs. **Ship:** Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Aroclor 1254	11097-69-1	124-191-B	----%	1,004.7 µg/mL	+/- 55.7515

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Hexane
CAS # 110-54-3
Purity 99%

P13830
↓
P13832
AJ
12/09/24

Quality Confirmation Test

Column:

30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:

helium-constant pressure 20 psi.

Temp. Program:

200°C to 300°C
@ 25°C/min. (hold 10 min.)

Inj. Temp:

250°C

Det. Temp:

300°C

Det. Type:


ECD

Split Vent:

300 ml/min.

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Michael Maye
Michael Maye - Operations Tech I

Date Mixed: 02-Oct-2024

Balance Serial # C322230531

Jennifer Pollino
Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 07-Oct-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32008

Lot No.: A0219655

Description : Aroclor® 1232 Standard

Aroclor® 1232 Standard 1,000 µg/mL, Hexane, 1mL/ampul

Container Size : 2 mL

Pkg Amt: > 1 mL

Expiration Date : March 31, 2031

Storage: 25°C nominal

Handling:

This product contains PCBs.

Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Aroclor 1232	11141-16-5	15665-01	---%	1,007.0 µg/mL	+/- 55.8810

Solvent: Hexane

CAS # 110-54-3

Purity 99%

* Expanded Uncertainty displayed in same units as Grav. Conc.

P13878



AJ
01/28/25

P13880

Quality Confirmation Test

Column:

30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:

helium-constant pressure 20 psi.

Temp. Program:

200°C to 300°C
@ 25°C/min. (hold 10 min.)

Inj. Temp:

250°C

Det. Temp:

300°C

Det. Type:


ECD

Split Vent:

10 ml/min.

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Michael Maye
Michael Maye - Operations Tech I

Date Mixed: 02-Dec-2024

Balance Serial #

C322230531

Brittany Federhik

Brittany Federhiko - Operations Tech I

Date Passed: 05-Dec-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397




110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No.: 32409 **Lot No.:** A0220950

Description: Aroclor® 1262 Standard

Aroclor® 1262 Standard 1,000 µg/mL, 1mL/ampul, Hexane

Container Size: 2 mL **Pkg Amt:** > 1 mL

Expiration Date: April 30, 2031 **Storage:** 25°C nominal

Handling: This product contains PCBs. **Ship:** Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Aroclor 1262	37324-23-5	10849100	----%	1,002.0 µg/mL	+/- 55.6035

Solvent: Hexane

CAS # 110-54-3
Purity 99%

P13882

↓

AJ
01/28/25

P13883

* Expanded Uncertainty displayed in same units as Grav. Conc.

Quality Confirmation Test

Column:

30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:

helium-constant pressure 20 psi.

Temp. Program:

200°C to 300°C
@ 25°C/min. (hold 10 min.)

Inj. Temp:

250°C

Det. Temp:

300°C

Det. Type:


ECD

Split Vent:

300 ml/min.

Inj. Vol

0.2µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


Tom Suckal - Mix Technician

Date Mixed: 09-Jan-2025 Balance Serial # C322230531



Britiany Federenko - Operations Tech I

Date Passed: 14-Jan-2025

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

n-Hexane 95%
ULTRA RESI-ANALYZED
For Organic Residue Analysis



Material No.: 9262-03
Batch No.: 24G1962003
Manufactured Date: 2024-05-23
Expiration Date: 2025-08-22
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
FID-Sensitive Impurities (as 2-Octanol) Single Impurity Peak (ng/mL)	≤ 5	3
ECD Sensitive Impurities (as Heptachlor Epoxide) Single Peak (pg/mL)	≤ 10	1
ECD-Sensitive Impurities (as Ethylene Dibromide) - Single Impurity Peak (ng/mL)	≤ 5	1
Assay (Total Saturated C ₆ Isomers) (by GC, corrected for water)	$\geq 99.5 \%$	99.7 %
Assay (as n-Hexane) (by GC, corrected for water)	$\geq 95 \%$	98 %
Color (APHA)	≤ 10	5
Residue after Evaporation	≤ 1.0 ppm	0.1 ppm
Substances Darkened by H ₂ SO ₄	Passes Test	Passes Test
Water (by KF, coulometric)	$\leq 0.05 \%$	< 0.01 %

For Laboratory, Research, or Manufacturing Use
MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: USA
Packaging Site: Phillipsburg Mfg Ctr & DC

Jamie Croak
Director Quality Operations, Bioscience Production



SHIPPING DOCUMENTS

CLIENT INFORMATION

REPORT TO BE SENT TO:

COMPANY: EARTH ENGINEERING INCORPORATED
ADDRESS: 403 Commerce Lane
CITY West Berlin STATE: NJ ZIP: 08091
ATTENTION: Frank Dougherty
PHONE: 856-768-1001 FAX:

CLIENT PROJECT INFORMATION

PROJECT NAME: Reserve @ Turgyan Farms
PROJECT NO. 38745.J0 LOCATION: NJ
PROJECT MANAGER: Frank Dougherty
e-mail: Frank.d@earthengineering.com
PHONE: 856-768-1001 FAX:

CLIENT BILLING INFORMATION

BILL TO: SAME PO#:
ADDRESS:
CITY STATE: ZIP:
ATTENTION: PHONE:

ANALYSIS

DATA TURNAROUND INFORMATION

FAX (RUSH) 3 DAYS*
HARDCOPY (DATA PACKAGE): DAYS*
EDD: DAYS*

*TO BE APPROVED BY CHEMTECH

STANDARD HARDCOPY TURNAROUND TIME IS 10 BUSINESS

DATA DELIVERABLE INFORMATION

☒ Level 1 (Results Only) ☐ Level 4 (QC + Full Raw Data)
☐ Level 2 (Results + QC) ☐ NJ Reduced ☐ US EPA CLP
☐ Level 3 (Results + QC) ☐ NYS ASP A ☐ NYS ASP B
+ Raw Data ☐ Other
☐ EDD FORMAT

1. TCL/TAL
2. EPH AF
3.
4.
5.
6.
7.
8.
9.

ALLIANCE SAMPLE ID	PROJECT SAMPLE IDENTIFICATION	SAMPLE MATRIX	SAMPLE TYPE		SAMPLE COLLECTION		# OF BOTTLES	PRESERVATIVES									COMMENTS
			COMP	GRAB	DATE	TIME		1	2	3	4	5	6	7	8	9	
1.	B-2-5-1	Soil	☒	☒	8-11-25	8:50	3	X	X								
2.	B-3-5-2	Soil	☒	☒	8-11-25	9:35	3	X	X								
3.																	
4.																	
5.																	
6.																	
7.																	
8.																	
9.																	
10.																	

SAMPLE CUSTODY MUST BE DOCUMENTED BELOW EACH TIME SAMPLES CHANGE POSSESSION INCLUDING COURIER DELIVERY

RELINQUISHED BY SAMPLER: <u>1. [Signature]</u>	DATE/TIME:	RECEIVED BY: <u>1. [Signature]</u>	1147 8-11-25	Conditions of bottles or coolers at receipt: ☐ COMPLIANT ☐ NON COMPLIANT ☐ COOLER TEMP <u>6.1</u> °C
RELINQUISHED BY SAMPLER: <u>2. [Signature]</u>	DATE/TIME:	RECEIVED BY: <u>2. [Signature]</u>		Comments:
RELINQUISHED BY SAMPLER: <u>3. [Signature]</u>	DATE/TIME:	RECEIVED BY: <u>3. [Signature]</u>		Page <u>1</u> of <u>1</u>

CLIENT: ☐ Hand Delivered ☐ Other

Shipment Complete
☐ YES ☐ NO

Laboratory Certification

Certified By	License No.
CAS EPA CLP Contract	68HERH20D0011
Connecticut	PH-0830
DOD ELAP (ANAB)	L2219
Maine	2024021
Maryland	296
New Hampshire	255424 Rev 1
New Jersey	20012
New York	11376
Pennsylvania	68-00548
Soil Permit	525-24-234-08441
Texas	T104704488

LOGIN REPORT/SAMPLE TRANSFER

Order ID : Q2818	EARTH03	Order Date : 8/11/2025 11:57:00 AM	Project Mgr :
Client Name : Earth Engineering Inc.		Project Name : Leon Avenue	Report Type : NJ Reduced
Client Contact : Frank Dougherty, LSRP		Receive DateTime : 8/11/2025 11:47:00 AM	EDD Type : EXCEL NJCLEANUP
Invoice Name : Earth Engineering Inc.		Purchase Order :	Hard Copy Date :
Invoice Contact : Frank Dougherty, LSRP			Date Signoff :

LAB ID	CLIENT ID	MATRIX	SAMPLE DATE	SAMPLE TIME	TEST	TEST GROUP	METHOD	FAX DATE	DUE DATES
Q2818-01	B-2-5-1	Solid	08/11/2025	08:50					
					VOC-TCLVOA-10	TCL+30/TAL	8260D	3 Bus. Days	
Q2818-02	B-3-5-2	Solid	08/11/2025	09:35					
					VOC-TCLVOA-10	TCL+30/TAL	8260D	3 Bus. Days	

JAN

Stored in VOA Freezer #02
after preservation

Relinquished By:

Date / Time: 8-11-24 1240

Received By:

Date / Time:

[Signature]

8-11-25

12:40 pm

Storage Area : VOA Refridgerator Room