

PESTICIDE CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: EARTH03

Lab Code: ACE

SDG NO.: Q2818

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 07/31/2025 07/31/2025

Client Sample No. (PEM): PEM - PD089686.D Date Analyzed: 07/31/2025

Lab Sample No.(PEM): PEM Time Analyzed: 10:47

PEM COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	9.070	8.970	9.170	21.540	20.000	7.7
Tetrachloro-m-xylene	3.548	3.500	3.600	20.670	20.000	3.4
alpha-BHC	3.997	3.950	4.050	9.820	10.000	-1.8
beta-BHC	4.514	4.460	4.560	10.760	10.000	7.6
gamma-BHC (Lindane)	4.328	4.280	4.380	10.080	10.000	0.8
Endrin	6.572	6.500	6.640	54.730	50.000	9.5
4,4'-DDT	7.020	6.950	7.090	108.560	100.000	8.6
Methoxychlor	7.491	7.420	7.560	258.930	250.000	3.6

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 07/31/2025 07/31/2025

Client Sample No. (PEM): PEM - PD089686.D Date Analyzed: 07/31/2025

Lab Sample No.(PEM): PEM Time Analyzed: 10:47

PEM COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	8.068	7.970	8.170	22.080	20.000	10.4
Tetrachloro-m-xylene	2.878	2.830	2.930	21.990	20.000	10.0
alpha-BHC	3.390	3.340	3.440	11.660	10.000	16.6
beta-BHC	4.022	3.970	4.070	11.840	10.000	18.4
gamma-BHC (Lindane)	3.726	3.680	3.780	11.700	10.000	17.0
Endrin	5.785	5.710	5.860	54.470	50.000	8.9
4,4'-DDT	6.179	6.110	6.250	108.020	100.000	8.0
Methoxychlor	6.750	6.680	6.820	237.190	250.000	-5.1