

Instrument ID: ECD_D

Daily Analysis Runlog For Sequence/QC Batch ID # PD073125

Review By	Abdul	Review On	8/1/2025 8:29:12 AM
Supervise By	mohammad	Supervise On	8/1/2025 9:27:16 AM
SubDirectory	PD073125	HP Acquire Method	HP Processing Method pd073125 8081

STD. NAME	STD REF.#
Tune/Reschk	PP24433,PP24651
Initial Calibration Stds	PP24260,PP24261,PP24262,PP24269,PP24266,PP24267,PP24268,PP24269,PP24270,PP24271,PP24272,PP24273,PP24274,PP24275,PP24277,P P24278,PP24279,PP24280,PP24281,PP24282,PP24283,PP24284
CCC	PP24261,PP24273,PP24279,PP24284
Internal Standard/PEM	
ICV/I.BLK	PP24273,PP24279,PP24284
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	Sampleld	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PD089684.D	31 Jul 2025 09:13		AR\AJ	Ok
2	I.BLK	I.BLK	PD089685.D	31 Jul 2025 10:33		AR\AJ	Ok
3	PEM	PEM	PD089686.D	31 Jul 2025 10:47		AR\AJ	Ok,M
4	RESCHK	RESCHK	PD089687.D	31 Jul 2025 11:34		AR\AJ	Ok
5	PSTDICC100	PSTDICC100	PD089688.D	31 Jul 2025 11:47		AR\AJ	Ok
6	PSTDICC075	PSTDICC075	PD089689.D	31 Jul 2025 12:01		AR\AJ	Ok,M
7	PSTDICC050	PSTDICC050	PD089690.D	31 Jul 2025 12:14		AR\AJ	Ok
8	PSTDICC025	PSTDICC025	PD089691.D	31 Jul 2025 12:28		AR\AJ	Ok
9	PSTDICC005	PSTDICC005	PD089692.D	31 Jul 2025 12:41		AR\AJ	Ok,M
10	PCHLORICC1000	PCHLORICC1000	PD089693.D	31 Jul 2025 12:56		AR\AJ	Ok
11	PCHLORICC750	PCHLORICC750	PD089694.D	31 Jul 2025 13:09		AR\AJ	Ok
12	PCHLORICC500	PCHLORICC500	PD089695.D	31 Jul 2025 13:23		AR\AJ	Ok
13	PCHLORICC250	PCHLORICC250	PD089696.D	31 Jul 2025 13:36		AR\AJ	Ok
14	PCHLORICC050	PCHLORICC050	PD089697.D	31 Jul 2025 13:50		AR\AJ	Ok
15	PTOXICC1000	PTOXICC1000	PD089698.D	31 Jul 2025 14:03		AR\AJ	Ok
16	PTOXICC750	PTOXICC750	PD089699.D	31 Jul 2025 14:17		AR\AJ	Ok
17	PTOXICC500	PTOXICC500	PD089700.D	31 Jul 2025 14:31		AR\AJ	Ok
18	PTOXICC250	PTOXICC250	PD089701.D	31 Jul 2025 14:44		AR\AJ	Ok,M

Instrument ID: ECD_D

Daily Analysis Runlog For Sequence/QC Batch ID # PD073125

Review By	Abdul	Review On	8/1/2025 8:29:12 AM
Supervise By	mohammad	Supervise On	8/1/2025 9:27:16 AM
SubDirectory	PD073125	HP Acquire Method	HP Processing Method pd073125 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP24433,PP24651		
Initial Calibration Stds	PP24260,PP24261,PP24262,PP24269,PP24266,PP24267,PP24268,PP24269,PP24270,PP24271,PP24272,PP24273,PP24274,PP24275,PP24277,P P24278,PP24279,PP24280,PP24281,PP24282,PP24283,PP24284		
CCC	PP24261,PP24273,PP24279,PP24284		
Internal Standard/PEM			
ICV/I.BLK	PP24273,PP24279,PP24284		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	PTOXICC100	PTOXICC100	PD089702.D	31 Jul 2025 14:58		AR\AJ	Ok,M
20	PSTDICV050	ICVPD073125	PD089703.D	31 Jul 2025 15:11		AR\AJ	Ok
21	PCHLORICV500	ICVPD073125CHLOR	PD089704.D	31 Jul 2025 15:25		AR\AJ	Ok
22	PTOXICV500	ICVPD073125TOX	PD089705.D	31 Jul 2025 15:39		AR\AJ	Ok

M : Manual Integration

Instrument ID: ECD_D

Daily Analysis Runlog For Sequence/QC Batch ID # PD081325

Review By	Abdul	Review On	8/21/2025 9:33:07 AM
Supervise By	mohammad	Supervise On	8/22/2025 1:53:41 AM
SubDirectory	PD081325	HP Acquire Method	HP Processing Method pd073125 8081

STD. NAME	STD REF.#
Tune/Reschk	PP24433,PP24651
Initial Calibration Stds	PP24744,PP24746,PP24748,PP24750,PP24751,PP24752,PP24753,PP24755,PP24756,PP24757,PP24758,PP24759,PP24760,PP24761,PP24762,P24763,PP24764
CCC	PP24751,PP24756,PP24761
Internal Standard/PEM	
ICV/I.BLK	PP24754,PP24759,PP24761
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PD089861.D	12 Aug 2025 14:35		AR\AJ	Ok
2	I.BLK	I.BLK	PD089862.D	12 Aug 2025 15:22		AR\AJ	Ok
3	PEM	PEM	PD089863.D	12 Aug 2025 15:35		AR\AJ	Ok,M
4	PSTDCCC050	PSTDCCC050	PD089864.D	12 Aug 2025 16:05		AR\AJ	Ok
5	PB169206BL	PB169206BL	PD089865.D	12 Aug 2025 16:33		AR\AJ	Ok
6	PB169206BS	PB169206BS	PD089866.D	12 Aug 2025 16:47		AR\AJ	Ok
7	Q2818-01	B-2-5-1	PD089867.D	12 Aug 2025 17:00		AR\AJ	Ok,M
8	Q2818-02	B-3-5-2	PD089868.D	12 Aug 2025 17:14		AR\AJ	Ok,M
9	Q2823-01	SU-04-081125	PD089869.D	12 Aug 2025 17:27		AR\AJ	Ok,M
10	Q2830-01	BIN0009-DRIVEWAY-T	PD089870.D	12 Aug 2025 17:41		AR\AJ	Ok,M
11	Q2830-03	BIN0009-DRIVEWAY-T	PD089871.D	12 Aug 2025 17:55		AR\AJ	Ok,M
12	Q2830-05	BIN0009-DRIVEWAY-T	PD089872.D	12 Aug 2025 18:08		AR\AJ	Ok,M
13	Q2830-07	BIN0009-DRIVEWAY-T	PD089873.D	12 Aug 2025 18:22		AR\AJ	Ok,M
14	Q2831-01	VNJ-238	PD089874.D	12 Aug 2025 18:36		AR\AJ	Ok,M
15	I.BLK	I.BLK	PD089875.D	12 Aug 2025 18:50		AR\AJ	Ok
16	PSTDCCC050	PSTDCCC050	PD089876.D	12 Aug 2025 20:12		AR\AJ	Ok,M

M : Manual Integration

Instrument ID: ECD_D

Daily Analysis Runlog For Sequence/QC Batch ID # PD081525

Review By	Abdul	Review On	8/15/2025 9:28:00 AM
Supervise By	mohammad	Supervise On	8/18/2025 1:44:47 AM
SubDirectory	PD081525	HP Acquire Method	HP Processing Method pd073125 8081

STD. NAME	STD REF.#
Tune/Reschk	PP24433,PP24651
Initial Calibration Stds	PP24744,PP24746,PP24748,PP24750,PP24751,PP24752,PP24753,PP24755,PP24756,PP24757,PP24758,PP24759,PP24760,PP24761,PP24762,P24763,PP24764
CCC	PP24751,PP24756,PP24761
Internal Standard/PEM	
ICV/I.BLK	PP24754,PP24759,PP24761
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PD089887.D	14 Aug 2025 11:51		AR\AJ	Ok
2	I.BLK	I.BLK	PD089888.D	14 Aug 2025 12:05		AR\AJ	Ok
3	PEM	PEM	PD089889.D	14 Aug 2025 12:19		AR\AJ	Ok,M
4	PSTDCCC050	PSTDCCC050	PD089890.D	14 Aug 2025 12:52		AR\AJ	Ok,M
5	Q2819-06	11M-S	PD089891.D	14 Aug 2025 13:34	Need 5X	AR\AJ	Dilution
6	Q2819-11	84SB-W	PD089892.D	14 Aug 2025 13:47	DCB low in 1ST column	AR\AJ	Ok,M
7	Q2819-14	17M-W	PD089893.D	14 Aug 2025 14:01	Need dilution	AR\AJ	Dilution
8	Q2819-14MS	17M-WMS	PD089894.D	14 Aug 2025 14:14	Recovery fail	AR\AJ	Ok,M
9	Q2819-14MSD	17M-WMSD	PD089895.D	14 Aug 2025 14:28	Recovery fail	AR\AJ	Ok,M
10	Q2820-09	705R-S	PD089896.D	14 Aug 2025 14:41	DCB low in both column	AR\AJ	ReRun
11	Q2820-10	SOIL-DUP-2	PD089897.D	14 Aug 2025 14:55		AR\AJ	Ok,M
12	Q2820-24	22M-S	PD089898.D	14 Aug 2025 15:09		AR\AJ	Ok,M
13	Q2858-01	VNJ 216	PD089899.D	14 Aug 2025 15:23		AR\AJ	Ok,M
14	PB169243BL	PB169243BL	PD089900.D	14 Aug 2025 15:36		AR\AJ	Ok
15	I.BLK	I.BLK	PD089901.D	14 Aug 2025 15:50		AR\AJ	Ok
16	PSTDCCC050	PSTDCCC050	PD089902.D	14 Aug 2025 17:10		AR\AJ	Ok

M : Manual Integration