

Surrogate Summary

SDG No.: **Q2818**

Client: **Earth Engineering Inc.**

Analytical Method: **8081B**

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Recovery(%)	Qual	Limits(%)	
								Low	High
I.BLK-PD089685.D	PIBLK-PD089685.D	Decachlorobiphen	1	20	20.9	104		57	171
		Tetrachloro-m-xyl	1	20	19.6	98		61	148
		Decachlorobiphen	2	20	20.9	104		57	171
		Tetrachloro-m-xyl	2	20	20.4	102		61	148
I.BLK-PD089862.D	PIBLK-PD089862.D	Decachlorobiphen	1	20	20.3	101		57	171
		Tetrachloro-m-xyl	1	20	20.0	100		61	148
		Decachlorobiphen	2	20	23.5	117		57	171
		Tetrachloro-m-xyl	2	20	25.1	125		61	148
PB169206BL	PB169206BL	Decachlorobiphen	1	20	19.4	97		20	144
		Tetrachloro-m-xyl	1	20	18.8	94		19	148
		Decachlorobiphen	2	20	22.7	113		20	144
		Tetrachloro-m-xyl	2	20	24.0	120		19	148
PB169206BS	PB169206BS	Decachlorobiphen	1	20	19.0	95		20	144
		Tetrachloro-m-xyl	1	20	18.1	91		19	148
		Decachlorobiphen	2	20	22.3	112		20	144
		Tetrachloro-m-xyl	2	20	23.2	116		19	148
Q2818-01	B-2-5-1	Decachlorobiphen	1	20	12.7	64		20	144
		Tetrachloro-m-xyl	1	20	14.2	71		19	148
		Decachlorobiphen	2	20	17.7	89		20	144
		Tetrachloro-m-xyl	2	20	16.8	84		19	148
Q2818-02	B-3-5-2	Decachlorobiphen	1	20	14.5	72		20	144
		Tetrachloro-m-xyl	1	20	18.3	91		19	148
		Decachlorobiphen	2	20	16.6	83		20	144
		Tetrachloro-m-xyl	2	20	19.3	96		19	148
I.BLK-PD089875.D	PIBLK-PD089875.D	Decachlorobiphen	1	20	20.0	100		57	171
		Tetrachloro-m-xyl	1	20	20.0	100		61	148
		Decachlorobiphen	2	20	21.9	110		57	171
		Tetrachloro-m-xyl	2	20	25.0	125		61	148
I.BLK-PD089888.D	PIBLK-PD089888.D	Decachlorobiphen	1	20	19.4	97		57	171
		Tetrachloro-m-xyl	1	20	19.3	96		61	148
		Decachlorobiphen	2	20	23.5	117		57	171
		Tetrachloro-m-xyl	2	20	25.4	127		61	148
Q2819-14MS	17M-WMS	Decachlorobiphen	1	20	12.3	62		20	144
		Tetrachloro-m-xyl	1	20	17.3	86		19	148
		Decachlorobiphen	2	20	17.7	88		20	144
		Tetrachloro-m-xyl	2	20	21.3	107		19	148
Q2819-14MSD	17M-WMSD	Decachlorobiphen	1	20	12.2	61		20	144
		Tetrachloro-m-xyl	1	20	17.1	85		19	148
		Decachlorobiphen	2	20	17.9	89		20	144
		Tetrachloro-m-xyl	2	20	21.6	108		19	148
I.BLK-PD089901.D	PIBLK-PD089901.D	Decachlorobiphen	1	20	19.2	96		57	171

Surrogate Summary

SDG No.: Q2818

Client: Earth Engineering Inc.

Analytical Method: 8081B

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Recovery(%)	Qual	Limits(%)	
								Low	High
I.BLK-PD089901.D	PIBLK-PD089901.D	Tetrachloro-m-xyl	1	20	19.2	96		61	148
		Decachlorobiphen	2	20	23.4	117		57	171
		Tetrachloro-m-xyl	2	20	24.5	123		61	148