

Surrogate Summary

SW-846

SDG No.: Q2818

Client: Earth Engineering Inc.

Analytical Method: 8270E

Lab Sample ID	Client ID	Parameter	Spike (PPM)	Result (PPM)	Recovery (%)	Qual	Limits (%)	
							Low	High
PB169210BL	PB169210BL	2-Fluorophenol	150	124	82		18	112
		Phenol-d6	150	109	73		15	107
		Nitrobenzene-d5	100	86.3	86		18	107
		2-Fluorobiphenyl	100	82.2	82		20	109
		2,4,6-Tribromophenol	150	156	104		10	116
		Terphenyl-d14	100	79.4	79		10	105
PB169210BS	PB169210BS	2-Fluorophenol	150	136	91		18	112
		Phenol-d6	150	138	92		15	107
		Nitrobenzene-d5	100	79.6	80		18	107
		2-Fluorobiphenyl	100	79.7	80		20	109
		2,4,6-Tribromophenol	150	138	92		10	116
		Terphenyl-d14	100	83.1	83		10	105
Q2818-01	B-2-5-1	2-Fluorophenol	150	60.5	40		18	112
		Phenol-d6	150	57.1	38		15	107
		Nitrobenzene-d5	100	42.0	42		18	107
		2-Fluorobiphenyl	100	41.9	42		20	109
		2,4,6-Tribromophenol	150	80.0	53		10	116
		Terphenyl-d14	100	39.6	40		10	105
Q2818-02	B-3-5-2	2-Fluorophenol	150	82.8	55		18	112
		Phenol-d6	150	76.9	51		15	107
		Nitrobenzene-d5	100	56.6	57		18	107
		2-Fluorobiphenyl	100	54.6	55		20	109
		2,4,6-Tribromophenol	150	109	72		10	116
		Terphenyl-d14	100	54.3	54		10	105
Q2830-05MS	BIN0009-DRIVEWAY-TP-WESTSID	2-Fluorophenol	150	95.7	64		18	112
		Phenol-d6	150	95.4	64		15	107
		Nitrobenzene-d5	100	75.4	75		18	107
		2-Fluorobiphenyl	100	68.9	69		20	109
		2,4,6-Tribromophenol	150	121	81		10	116
		Terphenyl-d14	100	58.4	58		10	105
Q2830-05MSD	BIN0009-DRIVEWAY-TP-WESTSID	2-Fluorophenol	150	97.8	65		18	112
		Phenol-d6	150	101	67		15	107
		Nitrobenzene-d5	100	77.3	77		18	107
		2-Fluorobiphenyl	100	65.9	66		20	109
		2,4,6-Tribromophenol	150	142	95		10	116
		Terphenyl-d14	100	77.5	78		10	105