

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q2818

Client ID : SSTDCCC040

Date Analyzed: 08/13/2025

Lab File ID: BF143365.D

Time Analyzed: 10:00

Instrument ID: BNA_F

GC Column: DB-UI ID: 0.18 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	155874	6.934	589243	8.21	314886	9.96
UPPER LIMIT	311748	7.434	1178490	8.71	629772	10.463
LOWER LIMIT	77937	6.434	294622	7.71	157443	9.463
EPA SAMPLE NO.						
01 BIN0009-DRIVEWAY-TP-WESTSIDEMS	137804	6.93	517024	8.21	272568	9.96
02 BIN0009-DRIVEWAY-TP-WESTSIDEMS	182122	6.93	707644	8.21	400799	9.97

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.