

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q2818

Client ID : SSTDCCC040

Date Analyzed: 08/12/2025

Lab File ID: BP025360.D

Time Analyzed: 10:07

Instrument ID: BNA\_P

GC Column: ZB-GR ID: 0.25 (mm)

|                | IS1 (DCB)<br>AREA # | RT #  | IS2 (NPT)<br>AREA # | RT #   | IS3 (ANT)<br>AREA # | RT #   |
|----------------|---------------------|-------|---------------------|--------|---------------------|--------|
| 12 HOUR STD    | 195999              | 7.796 | 806810              | 10.58  | 510213              | 14.43  |
| UPPER LIMIT    | 391998              | 8.296 | 1613620             | 11.078 | 1020430             | 14.931 |
| LOWER LIMIT    | 97999.5             | 7.296 | 403405              | 10.078 | 255107              | 13.931 |
| EPA SAMPLE NO. |                     |       |                     |        |                     |        |
| 01 B-2-5-1     | 233793              | 7.80  | 927693              | 10.57  | 559882              | 14.42  |
| 02 B-3-5-2     | 234324              | 7.80  | 947209              | 10.57  | 581002              | 14.43  |

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

# Column used to flag values outside QC limits with an asterisk.

\* Values outside of QC limits.