

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: Alliance
Lab Code: ACE
Lab File ID: BP025297.D
Instrument ID: BNA_P

Contract: EARTH03
SDG NO.: Q2818
DFTPP Injection Date: 08/05/2025
DFTPP Injection Time: 11:21

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
68	Less than 2.0% of mass 69	0.5 (1.6) 1
69	Mass 69 relative abundance	34.1
70	Less than 2.0% of mass 69	0.2 (0.5) 1
197	Less than 2.0% of mass 198	0.2
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	7
365	Greater than 1% of mass 198	3.3
441	Present, but less than mass 443	13.4
442	Greater than 50% of mass 198	87.3
443	15.0 - 24.0% of mass 442	17.1 (19.6) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDICC2.5	SSTDICC2.5	BP025298.D	08/05/2025	12:03
SSTDICC005	SSTDICC005	BP025299.D	08/05/2025	12:44
SSTDICC010	SSTDICC010	BP025300.D	08/05/2025	13:25
SSTDICC020	SSTDICC020	BP025301.D	08/05/2025	14:06
SSTDICCC040	SSTDICCC040	BP025302.D	08/05/2025	14:47
SSTDICC050	SSTDICC050	BP025303.D	08/05/2025	15:28
SSTDICC060	SSTDICC060	BP025304.D	08/05/2025	16:10
SSTDICC080	SSTDICC080	BP025305.D	08/05/2025	16:51