

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: First Environment, Inc.

SDG No.: Q2819

Contract: FIRS02

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: PLASMA-PURE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV08	Mercury	4.39	4.0	110	90 - 110	CV	08/12/2025	09:56	LB136782

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: First Environment, Inc.

SDG No.: Q2819

Contract: FIRS02

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: PLASMA-PURE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV22	Mercury	5.22	5.0	104	90 - 110	CV	08/12/2025	10:03	LB136782
CCV23	Mercury	5.26	5.0	105	90 - 110	CV	08/12/2025	11:11	LB136782
CCV24	Mercury	5.28	5.0	106	90 - 110	CV	08/12/2025	11:51	LB136782
CCV25	Mercury	5.14	5.0	103	90 - 110	CV	08/12/2025	12:32	LB136782
CCV26	Mercury	4.84	5.0	97	90 - 110	CV	08/12/2025	12:55	LB136782

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: First Environment, Inc.

SDG No.: Q2819

Contract: FIRS02

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Aluminum	7680	8000	96	90 - 110	P	08/13/2025	12:21	LB136801
	Antimony	4000	4000	100	90 - 110	P	08/13/2025	12:21	LB136801
	Arsenic	3870	4000	97	90 - 110	P	08/13/2025	12:21	LB136801
	Barium	7790	8000	97	90 - 110	P	08/13/2025	12:21	LB136801
	Beryllium	189	200	95	90 - 110	P	08/13/2025	12:21	LB136801
	Cadmium	1920	2000	96	90 - 110	P	08/13/2025	12:21	LB136801
	Calcium	19200	20000	96	90 - 110	P	08/13/2025	12:21	LB136801
	Chromium	788	800	98	90 - 110	P	08/13/2025	12:21	LB136801
	Cobalt	1950	2000	98	90 - 110	P	08/13/2025	12:21	LB136801
	Copper	997	1000	100	90 - 110	P	08/13/2025	12:21	LB136801
	Iron	3950	4000	99	90 - 110	P	08/13/2025	12:21	LB136801
	Lead	3770	4000	94	90 - 110	P	08/13/2025	12:21	LB136801
	Magnesium	19100	20000	96	90 - 110	P	08/13/2025	12:21	LB136801
	Manganese	1920	2000	96	90 - 110	P	08/13/2025	12:21	LB136801
	Nickel	1950	2000	97	90 - 110	P	08/13/2025	12:21	LB136801
	Potassium	19400	20000	97	90 - 110	P	08/13/2025	12:21	LB136801
	Selenium	3880	4000	97	90 - 110	P	08/13/2025	12:21	LB136801
	Silver	1030	1000	103	90 - 110	P	08/13/2025	12:21	LB136801
	Sodium	19100	20000	96	90 - 110	P	08/13/2025	12:21	LB136801
	Thallium	3750	4000	94	90 - 110	P	08/13/2025	12:21	LB136801
	Vanadium	1940	2000	97	90 - 110	P	08/13/2025	12:21	LB136801
	Zinc	1950	2000	98	90 - 110	P	08/13/2025	12:21	LB136801

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: First Environment, Inc.

SDG No.: Q2819

Contract: FIRS02

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Aluminum	95.1	100	95	80 - 120	P	08/13/2025	12:28	LB136801
	Antimony	50.1	50.0	100	80 - 120	P	08/13/2025	12:28	LB136801
	Arsenic	21.3	20.0	106	80 - 120	P	08/13/2025	12:28	LB136801
	Barium	94.7	100	95	80 - 120	P	08/13/2025	12:28	LB136801
	Beryllium	6.08	6.0	101	80 - 120	P	08/13/2025	12:28	LB136801
	Cadmium	5.59	6.0	93	80 - 120	P	08/13/2025	12:28	LB136801
	Calcium	2040	2000	102	80 - 120	P	08/13/2025	12:28	LB136801
	Chromium	10.5	10.0	105	80 - 120	P	08/13/2025	12:28	LB136801
	Cobalt	29.4	30.0	98	80 - 120	P	08/13/2025	12:28	LB136801
	Copper	21.3	20.0	107	80 - 120	P	08/13/2025	12:28	LB136801
	Iron	107	100	107	80 - 120	P	08/13/2025	12:28	LB136801
	Lead	11.2	12.0	93	80 - 120	P	08/13/2025	12:28	LB136801
	Magnesium	2120	2000	106	80 - 120	P	08/13/2025	12:28	LB136801
	Manganese	21.0	20.0	105	80 - 120	P	08/13/2025	12:28	LB136801
	Nickel	39.6	40.0	99	80 - 120	P	08/13/2025	12:28	LB136801
	Potassium	1990	2000	100	80 - 120	P	08/13/2025	12:28	LB136801
	Selenium	19.2	20.0	96	80 - 120	P	08/13/2025	12:28	LB136801
	Silver	10.1	10.0	101	80 - 120	P	08/13/2025	12:28	LB136801
	Sodium	1850	2000	92	80 - 120	P	08/13/2025	12:28	LB136801
	Thallium	43.0	40.0	107	80 - 120	P	08/13/2025	12:28	LB136801
	Vanadium	38.9	40.0	97	80 - 120	P	08/13/2025	12:28	LB136801
	Zinc	42.4	40.0	106	80 - 120	P	08/13/2025	12:28	LB136801

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: First Environment, Inc.

SDG No.: Q2819

Contract: FIRS02

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Aluminum	9630	10000	96	90 - 110	P	08/13/2025	12:58	LB136801
	Antimony	4930	5000	99	90 - 110	P	08/13/2025	12:58	LB136801
	Arsenic	4900	5000	98	90 - 110	P	08/13/2025	12:58	LB136801
	Barium	9640	10000	96	90 - 110	P	08/13/2025	12:58	LB136801
	Beryllium	240	250	96	90 - 110	P	08/13/2025	12:58	LB136801
	Cadmium	2430	2500	97	90 - 110	P	08/13/2025	12:58	LB136801
	Calcium	24100	25000	96	90 - 110	P	08/13/2025	12:58	LB136801
	Chromium	1000	1000	100	90 - 110	P	08/13/2025	12:58	LB136801
	Cobalt	2420	2500	97	90 - 110	P	08/13/2025	12:58	LB136801
	Copper	1230	1250	98	90 - 110	P	08/13/2025	12:58	LB136801
	Iron	5030	5000	101	90 - 110	P	08/13/2025	12:58	LB136801
	Lead	4860	5000	97	90 - 110	P	08/13/2025	12:58	LB136801
	Magnesium	24000	25000	96	90 - 110	P	08/13/2025	12:58	LB136801
	Manganese	2390	2500	96	90 - 110	P	08/13/2025	12:58	LB136801
	Nickel	2430	2500	97	90 - 110	P	08/13/2025	12:58	LB136801
	Potassium	25000	25000	100	90 - 110	P	08/13/2025	12:58	LB136801
	Selenium	4930	5000	98	90 - 110	P	08/13/2025	12:58	LB136801
	Silver	1230	1250	99	90 - 110	P	08/13/2025	12:58	LB136801
	Sodium	24600	25000	99	90 - 110	P	08/13/2025	12:58	LB136801
	Thallium	4820	5000	96	90 - 110	P	08/13/2025	12:58	LB136801
	Vanadium	2420	2500	97	90 - 110	P	08/13/2025	12:58	LB136801
	Zinc	2490	2500	99	90 - 110	P	08/13/2025	12:58	LB136801
CCV02	Aluminum	9700	10000	97	90 - 110	P	08/13/2025	13:42	LB136801
	Antimony	4970	5000	99	90 - 110	P	08/13/2025	13:42	LB136801
	Arsenic	4950	5000	99	90 - 110	P	08/13/2025	13:42	LB136801
	Barium	9660	10000	97	90 - 110	P	08/13/2025	13:42	LB136801
	Beryllium	240	250	96	90 - 110	P	08/13/2025	13:42	LB136801
	Cadmium	2430	2500	97	90 - 110	P	08/13/2025	13:42	LB136801
	Calcium	24200	25000	97	90 - 110	P	08/13/2025	13:42	LB136801
	Chromium	987	1000	99	90 - 110	P	08/13/2025	13:42	LB136801
	Cobalt	2430	2500	97	90 - 110	P	08/13/2025	13:42	LB136801
	Copper	1230	1250	99	90 - 110	P	08/13/2025	13:42	LB136801
	Iron	4980	5000	100	90 - 110	P	08/13/2025	13:42	LB136801
	Lead	4850	5000	97	90 - 110	P	08/13/2025	13:42	LB136801

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: First Environment, Inc.

SDG No.: Q2819

Contract: FIRS02

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV02	Magnesium	24100	25000	96	90 - 110	P	08/13/2025	13:42	LB136801
	Manganese	2420	2500	97	90 - 110	P	08/13/2025	13:42	LB136801
	Nickel	2430	2500	97	90 - 110	P	08/13/2025	13:42	LB136801
	Potassium	24600	25000	98	90 - 110	P	08/13/2025	13:42	LB136801
	Selenium	4980	5000	100	90 - 110	P	08/13/2025	13:42	LB136801
	Silver	1240	1250	99	90 - 110	P	08/13/2025	13:42	LB136801
	Sodium	24100	25000	96	90 - 110	P	08/13/2025	13:42	LB136801
	Thallium	4960	5000	99	90 - 110	P	08/13/2025	13:42	LB136801
	Vanadium	2450	2500	98	90 - 110	P	08/13/2025	13:42	LB136801
	Zinc	2460	2500	98	90 - 110	P	08/13/2025	13:42	LB136801
CCV03	Aluminum	9650	10000	96	90 - 110	P	08/13/2025	14:33	LB136801
	Antimony	5020	5000	100	90 - 110	P	08/13/2025	14:33	LB136801
	Arsenic	4990	5000	100	90 - 110	P	08/13/2025	14:33	LB136801
	Barium	9670	10000	97	90 - 110	P	08/13/2025	14:33	LB136801
	Beryllium	238	250	95	90 - 110	P	08/13/2025	14:33	LB136801
	Cadmium	2430	2500	97	90 - 110	P	08/13/2025	14:33	LB136801
	Calcium	24100	25000	96	90 - 110	P	08/13/2025	14:33	LB136801
	Chromium	993	1000	99	90 - 110	P	08/13/2025	14:33	LB136801
	Cobalt	2430	2500	97	90 - 110	P	08/13/2025	14:33	LB136801
	Copper	1240	1250	99	90 - 110	P	08/13/2025	14:33	LB136801
	Iron	5010	5000	100	90 - 110	P	08/13/2025	14:33	LB136801
	Lead	4860	5000	97	90 - 110	P	08/13/2025	14:33	LB136801
	Magnesium	23900	25000	96	90 - 110	P	08/13/2025	14:33	LB136801
	Manganese	2390	2500	96	90 - 110	P	08/13/2025	14:33	LB136801
	Nickel	2440	2500	98	90 - 110	P	08/13/2025	14:33	LB136801
	Potassium	24700	25000	99	90 - 110	P	08/13/2025	14:33	LB136801
	Selenium	5020	5000	100	90 - 110	P	08/13/2025	14:33	LB136801
	Silver	1240	1250	99	90 - 110	P	08/13/2025	14:33	LB136801
	Sodium	23300	25000	93	90 - 110	P	08/13/2025	14:33	LB136801
	Thallium	4820	5000	96	90 - 110	P	08/13/2025	14:33	LB136801
	Vanadium	2450	2500	98	90 - 110	P	08/13/2025	14:33	LB136801
	Zinc	2480	2500	99	90 - 110	P	08/13/2025	14:33	LB136801
CCV04	Aluminum	9760	10000	98	90 - 110	P	08/13/2025	15:23	LB136801
	Antimony	5060	5000	101	90 - 110	P	08/13/2025	15:23	LB136801

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: First Environment, Inc.

SDG No.: Q2819

Contract: FIRS02

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV04	Arsenic	5010	5000	100	90 - 110	P	08/13/2025	15:23	LB136801
	Barium	9660	10000	97	90 - 110	P	08/13/2025	15:23	LB136801
	Beryllium	239	250	96	90 - 110	P	08/13/2025	15:23	LB136801
	Cadmium	2430	2500	97	90 - 110	P	08/13/2025	15:23	LB136801
	Calcium	24300	25000	97	90 - 110	P	08/13/2025	15:23	LB136801
	Chromium	991	1000	99	90 - 110	P	08/13/2025	15:23	LB136801
	Cobalt	2430	2500	97	90 - 110	P	08/13/2025	15:23	LB136801
	Copper	1240	1250	99	90 - 110	P	08/13/2025	15:23	LB136801
	Iron	5000	5000	100	90 - 110	P	08/13/2025	15:23	LB136801
	Lead	4860	5000	97	90 - 110	P	08/13/2025	15:23	LB136801
	Magnesium	24200	25000	97	90 - 110	P	08/13/2025	15:23	LB136801
	Manganese	2390	2500	96	90 - 110	P	08/13/2025	15:23	LB136801
	Nickel	2440	2500	98	90 - 110	P	08/13/2025	15:23	LB136801
	Potassium	24700	25000	99	90 - 110	P	08/13/2025	15:23	LB136801
	Selenium	5070	5000	101	90 - 110	P	08/13/2025	15:23	LB136801
	Silver	1240	1250	99	90 - 110	P	08/13/2025	15:23	LB136801
	Sodium	23700	25000	95	90 - 110	P	08/13/2025	15:23	LB136801
	Thallium	4760	5000	95	90 - 110	P	08/13/2025	15:23	LB136801
	Vanadium	2470	2500	99	90 - 110	P	08/13/2025	15:23	LB136801
	Zinc	2500	2500	100	90 - 110	P	08/13/2025	15:23	LB136801
CCV05	Aluminum	9690	10000	97	90 - 110	P	08/13/2025	16:14	LB136801
	Antimony	5090	5000	102	90 - 110	P	08/13/2025	16:14	LB136801
	Arsenic	5010	5000	100	90 - 110	P	08/13/2025	16:14	LB136801
	Barium	9840	10000	98	90 - 110	P	08/13/2025	16:14	LB136801
	Beryllium	235	250	94	90 - 110	P	08/13/2025	16:14	LB136801
	Cadmium	2400	2500	96	90 - 110	P	08/13/2025	16:14	LB136801
	Calcium	24100	25000	96	90 - 110	P	08/13/2025	16:14	LB136801
	Chromium	991	1000	99	90 - 110	P	08/13/2025	16:14	LB136801
	Cobalt	2420	2500	97	90 - 110	P	08/13/2025	16:14	LB136801
	Copper	1240	1250	100	90 - 110	P	08/13/2025	16:14	LB136801
	Iron	5090	5000	102	90 - 110	P	08/13/2025	16:14	LB136801
	Lead	4810	5000	96	90 - 110	P	08/13/2025	16:14	LB136801
	Magnesium	23800	25000	95	90 - 110	P	08/13/2025	16:14	LB136801
	Manganese	2380	2500	95	90 - 110	P	08/13/2025	16:14	LB136801

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: First Environment, Inc.

SDG No.: Q2819

Contract: FIRS02

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV05	Nickel	2430	2500	97	90 - 110	P	08/13/2025	16:14	LB136801
	Potassium	25000	25000	100	90 - 110	P	08/13/2025	16:14	LB136801
	Selenium	5070	5000	101	90 - 110	P	08/13/2025	16:14	LB136801
	Silver	1250	1250	100	90 - 110	P	08/13/2025	16:14	LB136801
	Sodium	23800	25000	95	90 - 110	P	08/13/2025	16:14	LB136801
	Thallium	4540	5000	91	90 - 110	P	08/13/2025	16:14	LB136801
	Vanadium	2460	2500	98	90 - 110	P	08/13/2025	16:14	LB136801
	Zinc	2510	2500	101	90 - 110	P	08/13/2025	16:14	LB136801
CCV06	Aluminum	9690	10000	97	90 - 110	P	08/13/2025	17:05	LB136801
	Antimony	5110	5000	102	90 - 110	P	08/13/2025	17:05	LB136801
	Arsenic	5040	5000	101	90 - 110	P	08/13/2025	17:05	LB136801
	Barium	9810	10000	98	90 - 110	P	08/13/2025	17:05	LB136801
	Beryllium	238	250	95	90 - 110	P	08/13/2025	17:05	LB136801
	Cadmium	2420	2500	97	90 - 110	P	08/13/2025	17:05	LB136801
	Calcium	24100	25000	97	90 - 110	P	08/13/2025	17:05	LB136801
	Chromium	988	1000	99	90 - 110	P	08/13/2025	17:05	LB136801
	Cobalt	2430	2500	97	90 - 110	P	08/13/2025	17:05	LB136801
	Copper	1250	1250	100	90 - 110	P	08/13/2025	17:05	LB136801
	Iron	5070	5000	102	90 - 110	P	08/13/2025	17:05	LB136801
	Lead	4850	5000	97	90 - 110	P	08/13/2025	17:05	LB136801
	Magnesium	24000	25000	96	90 - 110	P	08/13/2025	17:05	LB136801
	Manganese	2400	2500	96	90 - 110	P	08/13/2025	17:05	LB136801
	Nickel	2440	2500	98	90 - 110	P	08/13/2025	17:05	LB136801
	Potassium	24800	25000	99	90 - 110	P	08/13/2025	17:05	LB136801
	Selenium	5150	5000	103	90 - 110	P	08/13/2025	17:05	LB136801
	Silver	1260	1250	100	90 - 110	P	08/13/2025	17:05	LB136801
	Sodium	23300	25000	93	90 - 110	P	08/13/2025	17:05	LB136801
	Thallium	5380	5000	108	90 - 110	P	08/13/2025	17:05	LB136801
	Vanadium	2460	2500	98	90 - 110	P	08/13/2025	17:05	LB136801
	Zinc	2520	2500	101	90 - 110	P	08/13/2025	17:05	LB136801
CCV07	Aluminum	9840	10000	98	90 - 110	P	08/13/2025	17:46	LB136801
	Antimony	5120	5000	102	90 - 110	P	08/13/2025	17:46	LB136801
	Arsenic	5060	5000	101	90 - 110	P	08/13/2025	17:46	LB136801
	Barium	9850	10000	98	90 - 110	P	08/13/2025	17:46	LB136801

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: First Environment, Inc.

SDG No.: Q2819

Contract: FIRS02

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV07	Beryllium	241	250	96	90 - 110	P	08/13/2025	17:46	LB136801
	Cadmium	2440	2500	98	90 - 110	P	08/13/2025	17:46	LB136801
	Calcium	24600	25000	98	90 - 110	P	08/13/2025	17:46	LB136801
	Chromium	1010	1000	101	90 - 110	P	08/13/2025	17:46	LB136801
	Cobalt	2450	2500	98	90 - 110	P	08/13/2025	17:46	LB136801
	Copper	1250	1250	100	90 - 110	P	08/13/2025	17:46	LB136801
	Iron	5140	5000	103	90 - 110	P	08/13/2025	17:46	LB136801
	Lead	4880	5000	98	90 - 110	P	08/13/2025	17:46	LB136801
	Magnesium	24400	25000	98	90 - 110	P	08/13/2025	17:46	LB136801
	Manganese	2430	2500	97	90 - 110	P	08/13/2025	17:46	LB136801
	Nickel	2460	2500	98	90 - 110	P	08/13/2025	17:46	LB136801
	Potassium	24800	25000	99	90 - 110	P	08/13/2025	17:46	LB136801
	Selenium	5160	5000	103	90 - 110	P	08/13/2025	17:46	LB136801
	Silver	1260	1250	100	90 - 110	P	08/13/2025	17:46	LB136801
	Sodium	23600	25000	95	90 - 110	P	08/13/2025	17:46	LB136801
	Thallium	5360	5000	107	90 - 110	P	08/13/2025	17:46	LB136801
	Vanadium	2490	2500	100	90 - 110	P	08/13/2025	17:46	LB136801
	Zinc	2530	2500	101	90 - 110	P	08/13/2025	17:46	LB136801