

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0072325\
 Data File : P0112416.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 23 Jul 2025 12:08
 Operator : YP/AJ
 Sample : AR1660ICC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1660ICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 23 12:27:27 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0072325.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Jul 23 12:24:33 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.669	3.662	437.7E6	266.5E6	50.000	50.000
2)	SA Decachlor...	8.693	8.640	391.5E6	93890082	50.000	50.000
Target Compounds							
3)	L1 AR-1016-1	4.755	4.738	145.4E6	90128163	500.000	500.000
4)	L1 AR-1016-2	4.774	4.756	216.9E6	136.8E6	500.000	500.000
5)	L1 AR-1016-3	4.831	4.931	136.8E6	71173535	500.000	500.000
6)	L1 AR-1016-4	4.951	4.973	111.7E6	57830575	500.000	500.000
7)	L1 AR-1016-5	5.208	5.185	116.9E6	73968656	500.000	500.000
31)	L7 AR-1260-1	6.245	6.213	238.1E6	130.6E6	500.000	500.000
32)	L7 AR-1260-2	6.433	6.402	361.5E6	167.5E6	500.000	500.000
33)	L7 AR-1260-3	6.800	6.553	305.2E6	134.6E6	500.000	500.000
34)	L7 AR-1260-4	7.059	7.023	236.6E6	98627757	500.000	500.000
35)	L7 AR-1260-5	7.303	7.264	668.6E6	216.5E6	500.000	500.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO072325\
Data File : PO112416.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Jul 2025 12:08
Operator : YP/AJ
Sample : AR1660ICC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1660ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 23 12:27:27 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO072325.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jul 23 12:24:33 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.669 min
Delta R.T.: 0.000 min
Response: 437703621
Conc: 50.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660ICC500




#1 Tetrachloro-m-xylene

R.T.: 3.662 min
Delta R.T.: 0.000 min
Response: 266527729
Conc: 50.00 ng/ml




#2 Decachlorobiphenyl

R.T.: 8.693 min
Delta R.T.: 0.000 min
Response: 391474769
Conc: 50.00 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.640 min
Delta R.T.: 0.000 min
Response: 93890082
Conc: 50.00 ng/ml



#3 AR-1016-1


R.T.: 4.755 min
Delta R.T.: 0.000 min
Response: 145424759
Conc: 500.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660ICC500




#3 AR-1016-1

R.T.: 4.738 min
Delta R.T.: 0.000 min
Response: 90128163
Conc: 500.00 ng/ml




#4 AR-1016-2

R.T.: 4.774 min
Delta R.T.: 0.000 min
Response: 216924568
Conc: 500.00 ng/ml



#4 AR-1016-2


R.T.: 4.756 min
Delta R.T.: 0.000 min
Response: 136766267
Conc: 500.00 ng/ml



#5 AR-1016-3


R.T.: 4.831 min
Delta R.T.: 0.000 min
Response: 136805252
Conc: 500.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660ICC500




#5 AR-1016-3

R.T.: 4.931 min
Delta R.T.: 0.000 min
Response: 71173535
Conc: 500.00 ng/ml




#6 AR-1016-4

R.T.: 4.951 min
Delta R.T.: 0.000 min
Response: 111666354
Conc: 500.00 ng/ml



#6 AR-1016-4


R.T.: 4.973 min
Delta R.T.: 0.000 min
Response: 57830575
Conc: 500.00 ng/ml



#7 AR-1016-5


R.T.: 5.208 min
Delta R.T.: 0.000 min
Response: 116933714
Conc: 500.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660ICC500




#7 AR-1016-5

R.T.: 5.185 min
Delta R.T.: 0.000 min
Response: 73968656
Conc: 500.00 ng/ml




#31 AR-1260-1

R.T.: 6.245 min
Delta R.T.: 0.000 min
Response: 238120143
Conc: 500.00 ng/ml



#31 AR-1260-1


R.T.: 6.213 min
Delta R.T.: 0.000 min
Response: 130567223
Conc: 500.00 ng/ml



#32 AR-1260-2


R.T.: 6.433 min
Delta R.T.: 0.000 min
Response: 361472265
Conc: 500.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660ICC500




#32 AR-1260-2

R.T.: 6.402 min
Delta R.T.: 0.000 min
Response: 167520551
Conc: 500.00 ng/ml




#33 AR-1260-3

R.T.: 6.800 min
Delta R.T.: 0.000 min
Response: 305155942
Conc: 500.00 ng/ml



#33 AR-1260-3


R.T.: 6.553 min
Delta R.T.: 0.000 min
Response: 134648936
Conc: 500.00 ng/ml



#34 AR-1260-4


R.T.: 7.059 min
Delta R.T.: 0.000 min
Response: 236624452
Conc: 500.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660ICC500




#34 AR-1260-4

R.T.: 7.023 min
Delta R.T.: 0.000 min
Response: 98627757
Conc: 500.00 ng/ml



#35 AR-1260-5

R.T.: 7.303 min
Delta R.T.: 0.000 min
Response: 668560746
Conc: 500.00 ng/ml



#35 AR-1260-5

R.T.: 7.264 min
Delta R.T.: 0.000 min
Response: 216537683
Conc: 500.00 ng/ml