

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081325\
 Data File : VY023082.D
 Acq On : 13 Aug 2025 20:00
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Aug 14 03:01:03 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081225S.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 13 02:10:11 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	82	0.00
2 T	Dichlorodifluoromethane	0.408	0.319	21.8	75	0.00
3 P	Chloromethane	0.661	0.681	-3.0	92	0.00
4 C	Vinyl Chloride	0.817	0.849	-3.9#	91	0.00
5 T	Bromomethane	0.610	0.682	-11.8	104	0.00
6 T	Chloroethane	0.543	0.596	-9.8	95	0.00
7 T	Trichlorofluoromethane	0.994	0.946	4.8	83	0.00
8 T	Diethyl Ether	0.263	0.249	5.3	81	0.01
9 T	1,1,2-Trichlorotrifluoroeth	0.490	0.456	6.9	81	0.00
10 T	Methyl Iodide	0.529	0.503	4.9	75	0.00
11 T	Tert butyl alcohol	0.033	0.029	12.1	74	0.00
12 CM	1,1-Dichloroethene	0.482	0.454	5.8#	81	0.00
13 T	Acrolein	0.048	0.034	29.2#	61	0.00
14 T	Allyl chloride	0.709	0.677	4.5	81	0.00
15 T	Acrylonitrile	0.107	0.104	2.8	80	0.00
16 T	Acetone	0.103	0.082	20.4	65	0.01
17 T	Carbon Disulfide	1.571	1.489	5.2	82	0.00
18 T	Methyl Acetate	0.307	0.326	-6.2	83	0.00
19 T	Methyl tert-butyl Ether	1.249	1.220	2.3	81	0.00
20 T	Methylene Chloride	0.620	0.555	10.5	89	0.00
21 T	trans-1,2-Dichloroethene	0.541	0.540	0.2	85	0.00
22 T	Diisopropyl ether	1.541	1.570	-1.9	85	0.00
23 T	Vinyl Acetate	0.852	0.808	5.2	79	0.00
24 P	1,1-Dichloroethane	0.942	0.958	-1.7	86	0.00
25 T	2-Butanone	0.140	0.126	10.0	73	0.00
26 T	2,2-Dichloropropane	0.813	0.711	12.5	74	0.00
27 T	cis-1,2-Dichloroethene	0.626	0.633	-1.1	86	0.00
28 T	Bromochloromethane	0.374	0.394	-5.3	89	0.00
29 T	Tetrahydrofuran	0.087	0.083	4.6	78	0.00
30 C	Chloroform	0.971	0.998	-2.8#	87	0.00
31 T	Cyclohexane	0.888	0.805	9.3	81	0.00
32 T	1,1,1-Trichloroethane	0.862	0.850	1.4	83	0.00
33 S	1,2-Dichloroethane-d4	0.477	0.490	-2.7	87	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	85	0.00
35 S	Dibromofluoromethane	0.299	0.298	0.3	87	0.00
36 T	1,1-Dichloropropene	0.445	0.422	5.2	83	0.00
37 T	Ethyl Acetate	0.187	0.166	11.2	75	0.00
38 T	Carbon Tetrachloride	0.483	0.456	5.6	83	0.00
39 T	Methylcyclohexane	0.586	0.544	7.2	82	0.00
40 TM	Benzene	1.370	1.347	1.7	86	0.00
41 T	Methacrylonitrile	0.108	0.108	0.0	84	0.00
42 TM	1,2-Dichloroethane	0.356	0.351	1.4	86	0.00
43 T	Isopropyl Acetate	0.378	0.348	7.9	80	0.00
44 TM	Trichloroethene	0.354	0.339	4.2	85	0.00
45 C	1,2-Dichloropropane	0.315	0.316	-0.3#	87	0.00
46 T	Dibromomethane	0.181	0.181	0.0	86	0.00
47 T	Bromodichloromethane	0.466	0.462	0.9	86	0.00
48 T	Methyl methacrylate	0.181	0.168	7.2	76	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081325\
 Data File : VY023082.D
 Acq On : 13 Aug 2025 20:00
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Aug 14 03:01:03 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081225S.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 13 02:10:11 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.002	0.002	0.0	81	0.00
50 S	Toluene-d8	1.169	1.161	0.7	87	0.00
51 T	4-Methyl-2-Pentanone	0.195	0.184	5.6	79	0.00
52 CM	Toluene	0.870	0.868	0.2#	86	0.00
53 T	t-1,3-Dichloropropene	0.422	0.400	5.2	81	0.00
54 T	cis-1,3-Dichloropropene	0.499	0.478	4.2	82	0.00
55 T	1,1,2-Trichloroethane	0.242	0.234	3.3	84	0.00
56 T	Ethyl methacrylate	0.310	0.299	3.5	80	0.00
57 T	1,3-Dichloropropane	0.411	0.403	1.9	85	0.00
58 T	2-Chloroethyl Vinyl ether	0.146	0.154	-5.5	85	0.00
59 T	2-Hexanone	0.133	0.122	8.3	76	0.00
60 T	Dibromochloromethane	0.320	0.313	2.2	84	0.00
61 T	1,2-Dibromoethane	0.228	0.222	2.6	84	0.00
62 S	4-Bromofluorobenzene	0.376	0.379	-0.8	87	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	87	0.00
64 T	Tetrachloroethene	0.446	0.442	0.9	88	0.00
65 PM	Chlorobenzene	1.079	1.048	2.9	86	0.00
66 T	1,1,1,2-Tetrachloroethane	0.366	0.356	2.7	86	0.00
67 C	Ethyl Benzene	1.856	1.805	2.7#	84	0.00
68 T	m/p-Xylenes	0.725	0.711	1.9	86	0.00
69 T	o-Xylene	0.679	0.670	1.3	85	0.00
70 T	Styrene	1.110	1.135	-2.3	88	0.00
71 P	Bromoform	0.214	0.196	8.4	80	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	89	0.00
73 T	Isopropylbenzene	3.529	3.316	6.0	85	0.00
74 T	N-amyl acetate	0.748	0.675	9.8	79	0.00
75 P	1,1,2,2-Tetrachloroethane	0.603	0.538	10.8	82	0.00
76 T	1,2,3-Trichloropropane	0.483	0.438	9.3	73	0.00
77 T	Bromobenzene	0.842	0.773	8.2	83	0.00
78 T	n-propylbenzene	4.201	4.000	4.8	85	0.00
79 T	2-Chlorotoluene	2.407	2.273	5.6	85	0.00
80 T	1,3,5-Trimethylbenzene	2.852	2.738	4.0	86	0.00
81 T	trans-1,4-Dichloro-2-butene	0.199	0.165	17.1	75	0.00
82 T	4-Chlorotoluene	2.497	2.371	5.0	86	0.00
83 T	tert-Butylbenzene	2.576	2.398	6.9	83	0.00
84 T	1,2,4-Trimethylbenzene	2.839	2.745	3.3	87	0.00
85 T	sec-Butylbenzene	3.784	3.563	5.8	84	0.00
86 T	p-Isopropyltoluene	3.141	3.017	3.9	85	0.00
87 T	1,3-Dichlorobenzene	1.660	1.561	6.0	87	0.00
88 T	1,4-Dichlorobenzene	1.647	1.530	7.1	85	0.00
89 T	n-Butylbenzene	2.885	2.740	5.0	84	0.00
90 T	Hexachloroethane	0.643	0.613	4.7	87	0.00
91 T	1,2-Dichlorobenzene	1.456	1.362	6.5	86	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.094	0.085	9.6	80	0.00
93 T	1,2,4-Trichlorobenzene	0.861	0.749	13.0	79	0.00
94 T	Hexachlorobutadiene	0.506	0.425	16.0	77	0.00
95 T	Naphthalene	1.468	1.358	7.5	82	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081325\
Data File : VY023082.D
Acq On : 13 Aug 2025 20:00
Operator : SY/MD
Sample : VSTDCCC050
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_Y
LabSampleId :
VSTDCCC050

Quant Time: Aug 14 03:01:03 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081225S.M
Quant Title : SW846 8260
QLast Update : Wed Aug 13 02:10:11 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.742	0.640	13.7	79	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6