

Instrument ID: MSVOA_W

Daily Analysis Runlog For Sequence/QC Batch ID # VW081125

Review By	Mahesh Dadoda	Review On	8/12/2025 1:11:21 PM
Supervise By	Semsettin Yesilyurt	Supervise On	8/12/2025 1:14:36 PM
SubDirectory	VW081125	HP Acquire Method	MSVOA_W HP Processing Method 82w081125s.m

STD. NAME	STD REF.#
Tune/Reschk	VP135074
Initial Calibration Stds	VP135075,VP135076,VP135077,VP135078,VP135079,VP135080
CCC	VP135082
Internal Standard/PEM	VP133934
ICV/I.BLK	VP135081
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VW032051.D	11 Aug 2025 07:53		SY/MD	Ok
2	VSTDICC005	VSTDICC005	VW032052.D	11 Aug 2025 08:25		SY/MD	Ok,M
3	VSTDICC010	VSTDICC010	VW032053.D	11 Aug 2025 09:09		SY/MD	Ok,M
4	VSTDICC020	VSTDICC020	VW032054.D	11 Aug 2025 09:47		SY/MD	Ok,M
5	VSTDICCC050	VSTDICCC050	VW032055.D	11 Aug 2025 10:09		SY/MD	Ok,M
6	VSTDICC100	VSTDICC100	VW032056.D	11 Aug 2025 10:46		SY/MD	Ok,M
7	VSTDICC150	VSTDICC150	VW032057.D	11 Aug 2025 11:08		SY/MD	Ok,M
8	VIBLK	VIBLK	VW032058.D	11 Aug 2025 11:30		SY/MD	Ok
9	VSTDICV050	ICVVW081125	VW032059.D	11 Aug 2025 11:51		SY/MD	Ok,M
10	VW0811SBL01	VW0811SBL01	VW032060.D	11 Aug 2025 12:18		SY/MD	Ok
11	VW0811SBS01	VW0811SBS01	VW032061.D	11 Aug 2025 12:48		SY/MD	Ok,M
12	VW0811SBSD01	VW0811SBSD01	VW032062.D	11 Aug 2025 13:10		SY/MD	Ok,M
13	Q2807-01	COMP-4	VW032063.D	11 Aug 2025 14:00	vial-A	SY/MD	Ok
14	Q2807-02	COMP-5	VW032064.D	11 Aug 2025 14:22	vial-A	SY/MD	Ok
15	Q2807-03	COMP-6	VW032065.D	11 Aug 2025 14:44	vial-A	SY/MD	Ok
16	Q2808-03	TP-7-VOC	VW032066.D	11 Aug 2025 15:06	vial-A	SY/MD	Ok
17	Q2809-01	OR-03-08082025	VW032067.D	11 Aug 2025 15:28	vial-A	SY/MD	Ok
18	Q2823-01	SU-04-081125	VW032068.D	11 Aug 2025 15:50	vial-A	SY/MD	Ok

Instrument ID: MSVOA_W

Daily Analysis Runlog For Sequence/QC Batch ID # VW081125

Review By	Maresh Dadoda	Review On	8/12/2025 1:11:21 PM
Supervise By	Semsettin Yesilyurt	Supervise On	8/12/2025 1:14:36 PM
SubDirectory	VW081125	HP Acquire Method	MSVOA_W HP Processing Method 82w081125s.m

STD. NAME	STD REF.#
Tune/Reschk	VP135074
Initial Calibration Stds	VP135075,VP135076,VP135077,VP135078,VP135079,VP135080
CCC	VP135082
Internal Standard/PEM	VP133934
ICV/I.BLK	VP135081
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

19	Q2818-01	B-2-5-1	VW032069.D	11 Aug 2025 16:12	vial-A	SY/MD	Ok
20	Q2818-02	B-3-5-2	VW032070.D	11 Aug 2025 16:34	vial-A	SY/MD	Ok
21	Q2819-01	22BP-N	VW032071.D	11 Aug 2025 16:56	vial-A	SY/MD	Ok
22	Q2819-02	22BP-E	VW032072.D	11 Aug 2025 17:18	vial-A	SY/MD	Ok
23	Q2819-03	22BP-W	VW032073.D	11 Aug 2025 17:41	vial-A	SY/MD	Ok
24	Q2819-04	22BP-S	VW032074.D	11 Aug 2025 18:03	vial-A	SY/MD	Ok
25	Q2819-05	11M-W	VW032075.D	11 Aug 2025 18:25	vial-A	SY/MD	Ok
26	Q2819-06	11M-S	VW032076.D	11 Aug 2025 18:47	vial-A	SY/MD	Ok
27	VIBLK	VIBLK	VW032077.D	11 Aug 2025 19:09		SY/MD	Ok
28	VSTDCCC050	VSTDCCC050EC	VW032078.D	11 Aug 2025 19:31		SY/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY081225

Review By	Maresh Dadoda	Review On	8/14/2025 7:34:22 PM
Supervise By	Semsettin Yesilyurt	Supervise On	8/14/2025 7:44:13 PM
SubDirectory	VY081225	HP Acquire Method	MSVOA_Y HP Processing Method 82y081225s.m

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP135083 VP135084,VP135085,VP135086,VP135087,VP135088,VP135089
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP133934 VP135090

Sr#	Sampleld	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VY023048.D	12 Aug 2025 08:26		SY/MD	Ok
2	VSTDICC005	VSTDICC005	VY023049.D	12 Aug 2025 14:32	Not used	SY/MD	Not Ok
3	VSTDICC005	VSTDICC005	VY023050.D	12 Aug 2025 14:54		SY/MD	Ok,M
4	VSTDICC010	VSTDICC010	VY023051.D	12 Aug 2025 15:17		SY/MD	Ok,M
5	VSTDICC020	VSTDICC020	VY023052.D	12 Aug 2025 15:40	LR-20	SY/MD	Ok,M
6	VSTDICCC050	VSTDICCC050	VY023053.D	12 Aug 2025 16:02		SY/MD	Ok,M
7	VSTDICC100	VSTDICC100	VY023054.D	12 Aug 2025 16:25		SY/MD	Ok,M
8	VSTDICC150	VSTDICC150	VY023055.D	12 Aug 2025 16:47		SY/MD	Ok,M
9	VIBLK	VIBLK	VY023056.D	12 Aug 2025 17:31		SY/MD	Ok
10	VSTDICV050	ICVVY081225	VY023057.D	12 Aug 2025 17:53		SY/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY081325

Review By	Maresh Dadoda	Review On	8/18/2025 8:46:30 AM
Supervise By	Semsettin Yesilyurt	Supervise On	8/19/2025 4:18:05 AM
SubDirectory	VY081325	HP Acquire Method	MSVOA_Y HP Processing Method 82y081225s.m

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP135111
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP135112,VP135113 VP133934

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VY023058.D	13 Aug 2025 08:21		SY/MD	Ok
2	VSTDCCC050	VSTDCCC050	VY023059.D	13 Aug 2025 09:37		SY/MD	Ok,M
3	VY0813SBL01	VY0813SBL01	VY023060.D	13 Aug 2025 10:39		SY/MD	Ok
4	VY0813SBS01	VY0813SBS01	VY023061.D	13 Aug 2025 11:09		SY/MD	Ok,M
5	VY0813SBSD01	VY0813SBSD01	VY023062.D	13 Aug 2025 11:32		SY/MD	Ok,M
6	Q2831-02	VNJ-238-VOC	VY023063.D	13 Aug 2025 12:34	vial-A	SY/MD	Ok
7	Q2830-01	BIN0009-DRIVEWAY-T	VY023064.D	13 Aug 2025 12:58	vial-A	SY/MD	Ok
8	Q2830-03	BIN0009-DRIVEWAY-T	VY023065.D	13 Aug 2025 13:21	vial-A	SY/MD	Ok
9	Q2830-05	BIN0009-DRIVEWAY-T	VY023066.D	13 Aug 2025 13:45	vial-A Internal Standard fail	SY/MD	ReRun
10	Q2830-07	BIN0009-DRIVEWAY-T	VY023067.D	13 Aug 2025 14:08	vial-A	SY/MD	Ok
11	Q2827-03	TP-8-VOC	VY023068.D	13 Aug 2025 14:32	vial-A	SY/MD	Ok
12	Q2819-07	11M-N	VY023069.D	13 Aug 2025 14:55	vial-A	SY/MD	Ok,M
13	Q2819-08	11M-E	VY023070.D	13 Aug 2025 15:19	vial-A	SY/MD	Ok
14	Q2819-09	84SB-E	VY023071.D	13 Aug 2025 15:43	vial-A	SY/MD	Ok
15	Q2819-10	84SB-S	VY023072.D	13 Aug 2025 16:06	vial-A Surrogate Fail	SY/MD	ReRun
16	Q2819-11	84SB-W	VY023073.D	13 Aug 2025 16:30	vial-A Internal Standard fail;Surrogate Fail	SY/MD	ReRun
17	Q2819-12	17M-S	VY023074.D	13 Aug 2025 16:53	vial-A	SY/MD	Ok

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY081325

Review By	Mahesh Dadoda	Review On	8/18/2025 8:46:30 AM		
Supervise By	Semsettin Yesilyurt	Supervise On	8/19/2025 4:18:05 AM		
SubDirectory	VY081325	HP Acquire Method	MSVOA_Y	HP Processing Method	82y081225s.m
STD. NAME		STD REF.#			
Tune/Reschk Initial Calibration Stds		VP135111			
CCC		VP135112,VP135113			
Internal Standard/PEM		VP133934			
ICV/I.BLK					
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

18	Q2819-13	17M-E	VY023075.D	13 Aug 2025 17:17	vial-A	SY/MD	Ok
19	Q2819-14	17M-W	VY023076.D	13 Aug 2025 17:40	vial-A	SY/MD	Ok
20	Q2819-15	17M-N	VY023077.D	13 Aug 2025 18:04	vial-A Internal Standard fail	SY/MD	ReRun
21	Q2819-16	38M-S	VY023078.D	13 Aug 2025 18:27	vial-A Surrogate Fail	SY/MD	ReRun
22	Q2819-17	38M-N	VY023079.D	13 Aug 2025 18:51	vial-A	SY/MD	Ok
23	Q2819-18	38M-W	VY023080.D	13 Aug 2025 19:14	vial-A	SY/MD	Ok
24	Q2819-19	38M-E	VY023081.D	13 Aug 2025 19:38	vial-A	SY/MD	Ok
25	VSTDCCC050	VSTDCCC050EC	VY023082.D	13 Aug 2025 20:00		SY/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY081425

Review By	Mahesh Dadoda	Review On	8/19/2025 4:23:08 AM				
Supervise By	Semsettin Yesilyurt	Supervise On	8/19/2025 6:52:38 AM				
SubDirectory	VY081425	HP Acquire Method	MSVOA_Y	HP Processing Method	82y081225s.m		
STD. NAME		STD REF.#					
Tune/Reschk Initial Calibration Stds		VP135134					
CCC		VP135135,VP135136					
Internal Standard/PEM		VP133934					
ICV/I.BLK							
Surrogate Standard							
MS/MSD Standard							
LCS Standard							

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VY023083.D	14 Aug 2025 09:48		SY/MD	Ok
2	VSTDCCC050	VSTDCCC050	VY023084.D	14 Aug 2025 10:17		SY/MD	Ok,M
3	VY0814SBL01	VY0814SBL01	VY023085.D	14 Aug 2025 10:50		SY/MD	Ok
4	VY0814SBS01	VY0814SBS01	VY023086.D	14 Aug 2025 11:22		SY/MD	Ok,M
5	VY0814SBSD01	VY0814SBSD01	VY023087.D	14 Aug 2025 11:45		SY/MD	Ok,M
6	Q2830-05	BIN0009-DRIVEWAY-T	VY023088.D	14 Aug 2025 12:26	vial-B	SY/MD	Ok
7	Q2819-10RE	84SB-SRE	VY023089.D	14 Aug 2025 12:50	vial-B Surrogate Fail	SY/MD	Confirms
8	Q2819-11RE	84SB-WRE	VY023090.D	14 Aug 2025 13:13	vial-B Internal Standard fail;Surrogate Fail	SY/MD	Confirms
9	Q2819-15	17M-N	VY023091.D	14 Aug 2025 13:37	vial-B	SY/MD	Ok
10	Q2819-16	38M-S	VY023092.D	14 Aug 2025 14:00	vial-B	SY/MD	Ok
11	Q2819-20	82H-E	VY023093.D	14 Aug 2025 14:24	vial-A Internal Standard fail;Surrogate Fail	SY/MD	ReRun
12	Q2838-03	TP-11-VOC	VY023094.D	14 Aug 2025 14:47	vial-A	SY/MD	Ok
13	Q2838-07	TP-10-VOC	VY023095.D	14 Aug 2025 15:11	vial-A	SY/MD	Ok
14	Q2840-01	0804-SOIL	VY023096.D	14 Aug 2025 15:34	vial-A	SY/MD	Ok
15	Q2858-01	VNJ 216	VY023097.D	14 Aug 2025 15:57	vial-A	SY/MD	Ok
16	Q2857-03	72-12002	VY023098.D	14 Aug 2025 16:21	vial-A	SY/MD	Ok
17	Q2826-01	WC1	VY023099.D	14 Aug 2025 16:44	vial-A	SY/MD	Ok

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY081425

Review By	Maresh Dadoda	Review On	8/19/2025 4:23:08 AM
Supervise By	Semsettin Yesilyurt	Supervise On	8/19/2025 6:52:38 AM
SubDirectory	VY081425	HP Acquire Method	MSVOA_Y HP Processing Method 82y081225s.m
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	VP135134		
CCC	VP135135,VP135136		
Internal Standard/PEM	VP133934		
ICV/I.BLK			
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

18	Q2873-01	DRILL CUTTING 1-3 C	VY023100.D	14 Aug 2025 17:08	vial-A	SY/MD	Ok
19	Q2873-04	DRILL CUTTING 4-6 C	VY023101.D	14 Aug 2025 17:31	vial-A	SY/MD	Ok
20	Q2871-01	DRILL CUTTING 1-3 C	VY023102.D	14 Aug 2025 17:55	vial-A	SY/MD	Ok
21	Q2871-04	DRILL CUTTING 4-6 C	VY023103.D	14 Aug 2025 18:18	vial-A Internal Standard Fail	SY/MD	ReRun
22	Q2871-07	DRILL CUTTING 1-5 C	VY023104.D	14 Aug 2025 18:42	vial-A	SY/MD	Ok
23	Q2872-01	DRILL CUTTING 1-2 C	VY023105.D	14 Aug 2025 19:05	vial-A Internal Standard Fail	SY/MD	ReRun
24	Q2870-01	DRILL CUTTING 1-5 C	VY023106.D	14 Aug 2025 19:29	vial-A Internal Standard Fail	SY/MD	Ok
25	Q2870-04	DRILL CUTTING 6-9 C	VY023107.D	14 Aug 2025 19:52	vial-A	SY/MD	Ok
26	VIBLK	VIBLK	VY023108.D	14 Aug 2025 20:15		SY/MD	Ok
27	VSTDCCC050	VSTDCCC050EC	VY023109.D	14 Aug 2025 20:38		SY/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY081925

Review By	John Carlone	Review On	8/20/2025 9:05:17 AM
Supervise By	Maresh Dadoda	Supervise On	8/20/2025 9:12:04 AM
SubDirectory	VY081925	HP Acquire Method	MSVOA_Y HP Processing Method 82y081225s.m

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP135166
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP135167,VP135168 VP133934

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VY023136.D	19 Aug 2025 07:45		SY/MD	Ok
2	VSTDCCC050	VSTDCCC050	VY023137.D	19 Aug 2025 09:11		SY/MD	Ok,M
3	VY0819SBL01	VY0819SBL01	VY023138.D	19 Aug 2025 09:49		SY/MD	Ok
4	VY0819SBS01	VY0819SBS01	VY023139.D	19 Aug 2025 10:19		SY/MD	Ok,M
5	VY0819SBSD01	VY0819SBSD01	VY023140.D	19 Aug 2025 10:42		SY/MD	Ok,M
6	Q2820-01	82H-S	VY023141.D	19 Aug 2025 11:20	vial-B	SY/MD	Ok
7	Q2820-15	10P-S	VY023142.D	19 Aug 2025 11:44	vial-B	SY/MD	Ok
8	Q2820-23	22M-E	VY023143.D	19 Aug 2025 12:07	vial-B	SY/MD	Ok
9	Q2819-20	82H-E	VY023144.D	19 Aug 2025 12:31	vial-B	SY/MD	Ok
10	Q2870-01	DRILL CUTTING 1-5 C	VY023145.D	19 Aug 2025 12:54	vial-B Not Purge	SY/MD	Not Ok
11	Q2871-04RE	DRILL CUTTING 4-6 C	VY023146.D	19 Aug 2025 13:18	vial-B Internal Standard Fail	SY/MD	Confirms
12	Q2872-01RE	DRILL CUTTING 1-2 C	VY023147.D	19 Aug 2025 13:41	vial-B Internal Standard Fail	SY/MD	Confirms
13	Q2879-01	OU4-TS-GRILLO-TSCF	VY023148.D	19 Aug 2025 14:04	vial-A	SY/MD	Ok
14	Q2879-03	OU4-TS-GRILLO-TSCF	VY023149.D	19 Aug 2025 14:28	vial-A Surrogate fail	SY/MD	ReRun
15	Q2879-05	OU4-TS-GRILLO-TSCF	VY023150.D	19 Aug 2025 14:51	vial-A Surrogate fail	SY/MD	ReRun
16	Q2879-07	OU4-TS-GRILLO-TSCF	VY023151.D	19 Aug 2025 15:15	vial-A	SY/MD	Ok
17	Q2879-09	OU4-TS-GRILLO-TSCF	VY023152.D	19 Aug 2025 15:38	vial-A Surrogate fail	SY/MD	ReRun
18	Q2879-11	OU4-TS-GRILLO-TSCF	VY023153.D	19 Aug 2025 16:02	vial-A Surrogate fail	SY/MD	ReRun

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY081925

Review By	John Carlone	Review On	8/20/2025 9:05:17 AM		
Supervise By	Mahesh Dadoda	Supervise On	8/20/2025 9:12:04 AM		
SubDirectory	VY081925	HP Acquire Method	MSVOA_Y	HP Processing Method	82y081225s.m
STD. NAME		STD REF.#			
Tune/Reschk Initial Calibration Stds		VP135166			
CCC		VP135167,VP135168			
Internal Standard/PEM		VP133934			
ICV/I.BLK					
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

19	Q2879-13	OU4-TS-GRILLO-TSCF	VY023154.D	19 Aug 2025 16:25	vial-A Surrogate fail	SY/MD	ReRun
20	Q2879-15	OU4-TS-GRILLO-TSCF	VY023155.D	19 Aug 2025 16:49	vial-A Surrogate fail	SY/MD	ReRun
21	Q2879-17	OU4-TS-GRILLO-TSCF	VY023156.D	19 Aug 2025 17:12	vial-A Surrogate fail	SY/MD	ReRun
22	Q2897-01	HD-02-08182025	VY023157.D	19 Aug 2025 17:36	vial-A Internal Standard Fail	SY/MD	ReRun
23	Q2889-01	PL-02-081525	VY023158.D	19 Aug 2025 17:59	vial-A	SY/MD	Ok
24	Q2888-01	VNJ-210	VY023159.D	19 Aug 2025 18:22	vial-A	SY/MD	Ok
25	VSTDCCC050	VSTDCCC050EC	VY023160.D	19 Aug 2025 19:08		SY/MD	Ok,M

M : Manual Integration