

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q2819</u>	Analytical Method:	<u>SW8260D</u>
Client:	<u>First Environment, Inc.</u>	Datafile :	<u>VW032061.D</u>

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VW0811SBS01	Dichlorodifluoromethane	20	21.3	ug/Kg	106			29	149	
	Chloromethane	20	20.8	ug/Kg	104			50	136	
	Vinyl chloride	20	21.1	ug/Kg	106			56	135	
	Bromomethane	20	21.5	ug/Kg	108			53	143	
	Chloroethane	20	20.8	ug/Kg	104			59	139	
	Trichlorofluoromethane	20	17.6	ug/Kg	88			62	140	
	1,1,2-Trichlorotrifluoroethane	20	20.1	ug/Kg	101			66	136	
	1,1-Dichloroethene	20	20.4	ug/Kg	102			70	131	
	Acetone	100	120	ug/Kg	120			36	164	
	Carbon disulfide	20	20.2	ug/Kg	101			63	132	
	Methyl tert-butyl Ether	20	21.9	ug/Kg	110			73	125	
	Methyl Acetate	20	21.7	ug/Kg	109			53	144	
	Methylene Chloride	20	21.4	ug/Kg	107			70	128	
	trans-1,2-Dichloroethene	20	20.7	ug/Kg	104			74	125	
	1,1-Dichloroethane	20	21.0	ug/Kg	105			76	125	
	Cyclohexane	20	20.0	ug/Kg	100			67	131	
	2-Butanone	100	110	ug/Kg	110			51	148	
	Carbon Tetrachloride	20	20.0	ug/Kg	100			70	135	
	cis-1,2-Dichloroethene	20	21.2	ug/Kg	106			77	123	
	Bromochloromethane	20	20.5	ug/Kg	103			78	125	
	Chloroform	20	21.1	ug/Kg	106			78	123	
	1,1,1-Trichloroethane	20	20.7	ug/Kg	104			73	130	
	Methylcyclohexane	20	20.0	ug/Kg	100			66	133	
	Benzene	20	21.0	ug/Kg	105			77	121	
	1,2-Dichloroethane	20	20.9	ug/Kg	104			73	128	
	Trichloroethene	20	20.2	ug/Kg	101			77	123	
	1,2-Dichloropropane	20	21.0	ug/Kg	105			76	123	
	Bromodichloromethane	20	20.4	ug/Kg	102			75	127	
	4-Methyl-2-Pentanone	100	110	ug/Kg	110			65	135	
	Toluene	20	21.0	ug/Kg	105			77	121	
	t-1,3-Dichloropropene	20	20.4	ug/Kg	102			71	130	
	cis-1,3-Dichloropropene	20	20.9	ug/Kg	104			74	126	
	1,1,2-Trichloroethane	20	20.8	ug/Kg	104			78	121	
	2-Hexanone	100	110	ug/Kg	110			53	145	
	Dibromochloromethane	20	20.6	ug/Kg	103			74	126	
	1,2-Dibromoethane	20	21.2	ug/Kg	106			78	122	
	Tetrachloroethene	20	21.1	ug/Kg	106			73	128	
	Chlorobenzene	20	20.2	ug/Kg	101			79	120	
	Ethyl Benzene	20	20.1	ug/Kg	101			76	122	
	m/p-Xylenes	40	40.9	ug/Kg	102			77	124	
	o-Xylene	20	21.2	ug/Kg	106			77	123	
	Styrene	20	20.9	ug/Kg	104			76	124	
	Bromoform	20	19.6	ug/Kg	98			67	132	
	Isopropylbenzene	20	19.9	ug/Kg	100			68	134	
	1,1,2,2-Tetrachloroethane	20	20.1	ug/Kg	101			70	124	
	1,3-Dichlorobenzene	20	20.4	ug/Kg	102			77	121	
	1,4-Dichlorobenzene	20	20.7	ug/Kg	104			75	120	
	1,2-Dichlorobenzene	20	20.3	ug/Kg	102			78	121	



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q2819 Analytical Method: SW8260D
Client: First Environment, Inc. Datafile : VW032061.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VW0811SBS01	1,2-Dibromo-3-Chloropropane	20	20.1	ug/Kg	101			61	132	
	1,2,4-Trichlorobenzene	20	19.5	ug/Kg	98			67	129	
	1,2,3-Trichlorobenzene	20	21.4	ug/Kg	107			66	130	