

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: FIRS02
 Lab Code: ACE SDG NO.: Q2819
 Lab File ID: VW032055.D Date Analyzed: 08/11/2025
 Instrument ID: MSVOA_W Time Analyzed: 10:09
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS4 AREA #	RT #				
12 HOUR STD	178995	13.556				
UPPER LIMIT	357990	14.056				
LOWER LIMIT	89497.5	13.056				
EPA SAMPLE NO.						
22BP-N	149992	13.56				
22BP-E	163221	13.56				
22BP-W	147943	13.56				
22BP-S	157540	13.56				
11M-W	158417	13.56				
11M-S	150514	13.56				
VW0811SBL01	169733	13.56				
VW0811SBS01	191594	13.56				
VW0811SBSD01	195548	13.56				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.