

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name:	<u>Alliance</u>	Contract:	<u>FIRS02</u>
Lab Code:	<u>ACE</u>	SDG NO.:	<u>Q2819</u>
Lab File ID:	<u>VY023059.D</u>	Date Analyzed:	<u>08/13/2025</u>
Instrument ID:	<u>MSVOA_Y</u>	Time Analyzed:	<u>09:37</u>
GC Column:	<u>RXI-624</u>	ID:	<u>0.25</u> (mm)
		Heated Purge:	(Y/N) <u>Y</u>

	IS4 AREA #	RT #				
12 HOUR STD	361154	13.346				
UPPER LIMIT	722308	13.846				
LOWER LIMIT	180577	12.846				
EPA SAMPLE NO.						
11M-N	339290	13.34				
11M-E	338962	13.34				
84SB-E	305266	13.34				
84SB-S	298114	13.34				
84SB-W	66820 *	13.34				
17M-S	344922	13.34				
17M-E	487544	13.34				
17M-W	379138	13.34				
38M-N	509432	13.34				
38M-W	453970	13.34				
38M-E	424298	13.34				
VY0813SBL01	384760	13.34				
VY0813SBS01	325106	13.34				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.