

ALLIANCE 284 Sheffield Street, Mountainside New Jersey 07092

NEW JERSEY LAB ID#: 20012; NEW YORK LAB ID#: 11376

METALS CONFORMANCE/NON-CONFORMANCE SUMMARY

ORDER ID: Q2820

MATRIX: Solid

METHOD: 6010D,7471B

	NA	NO	YES
1. Calibration Summary met criteria.			✓
2. ICP Interference Check Sample Results Summary Submitted.			✓
3. Serial Dilution Summary (if applicable) Submitted.		✓	
The Serial Dilution (POWDERL) met criteria for all compounds except for Aluminum, Calcium, Chromium, Copper, Iron, Manganese and Zinc due to sample matrix interference.			
4. Laboratory Control Sample Summary (if applicable) Submitted.			✓
5. Blank Contamination - If yes, list compounds and concentrations in each blank:		✓	
6. Matrix Spike/Matrix Spike Duplicate Recoveries Met Criteria		✓	
If not met, list those compounds and their recoveries which fall outside the acceptable range.			
The Matrix Spike (POWDERMS) analysis met criteria for all compounds except for Antimony, Arsenic, Beryllium, Chromium, Potassium, Selenium, Silver, Thallium and Vanadium due to Chemical Interference during Digestion process. The Matrix Spike Duplicate (POWDERMSD) analysis met criteria for all compounds except for Antimony, Arsenic, Beryllium, Potassium, Selenium, Silver and Vanadium due to Chemical Interference during Digestion process.			
7. Sample Duplicate Analysis Met QC Criteria		✓	
If not met, list those compounds and their recoveries which fall outside the acceptable range.			
The Duplicate (POWDERMSD) analysis met criteria for all compounds except for Iron and Vanadium due to Chemical Interference during Digestion process.			
8. Digestion Holding Time Met			✓
If not met, list number of days exceeded for each sample:			

ALLIANCE 284 Sheffield Street, Mountainside New Jersey 07092

NEW JERSEY LAB ID#: 20012; NEW YORK LAB ID#: 11376

METALS CONFORMANCE/NON-CONFORMANCE SUMMARY (CONTINUED)

	NA	NO	YES
9. Analysis Holding Time Met			✓
If not met, list those compounds and their recoveries which fall outside the acceptable range.			

ADDITIONAL COMMENTS:

The Post Digest Spike (POWDERA) analysis met criteria for all compounds except for Antimony, Arsenic, Beryllium, Chromium, Potassium, Selenium, Silver and Vanadium due to unknown chemical interference of matrix with the addition of spike amount after digestion and before analysis; matrix has suppression effect during addition of spike.

QA REVIEW

Date