

### CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: FIRS02

Lab Code: ACE

SDG NO.: Q2820

Continuing Calib Date: 08/13/2025

Initial Calibration Date(s): 07/23/2025

07/23/2025

Continuing Calib Time: 03:11

Initial Calibration Time(s): 11:32

19:47

GC Column: ZB-MR1 ID: 0.32 (mm)

| COMPOUND             | CCAL<br>RT | AVG<br>RT | RT WINDOW |      | DIFF<br>RT |
|----------------------|------------|-----------|-----------|------|------------|
|                      |            |           | FROM      | TO   |            |
| Aroclor-1254-1 (1)   | 5.56       | 5.56      | 5.46      | 5.66 | 0.00       |
| Aroclor-1254-2 (2)   | 5.71       | 5.71      | 5.61      | 5.81 | 0.01       |
| Aroclor-1254-3 (3)   | 6.11       | 6.11      | 6.01      | 6.21 | 0.00       |
| Aroclor-1254-4 (4)   | 6.34       | 6.34      | 6.24      | 6.44 | 0.00       |
| Aroclor-1254-5 (5)   | 6.76       | 6.76      | 6.66      | 6.86 | 0.00       |
| Tetrachloro-m-xylene | 3.67       | 3.67      | 3.57      | 3.77 | 0.00       |
| Decachlorobiphenyl   | 8.69       | 8.70      | 8.60      | 8.80 | 0.01       |

### CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: FIRS02

Lab Code: ACE

SDG NO.: Q2820

Continuing Calib Date: 08/13/2025

Initial Calibration Date(s): 07/23/2025

07/23/2025

Continuing Calib Time: 03:11

Initial Calibration Time(s): 11:32

19:47

GC Column: ZB-MR2 ID: 0.32 (mm)

| COMPOUND             | CCAL<br>RT | AVG<br>RT | RT WINDOW<br>FROM   TO |      | DIFF<br>RT |
|----------------------|------------|-----------|------------------------|------|------------|
| Aroclor-1254-1 (1)   | 5.53       | 5.54      | 5.44                   | 5.64 | 0.01       |
| Aroclor-1254-2 (2)   | 5.68       | 5.68      | 5.58                   | 5.78 | 0.00       |
| Aroclor-1254-3 (3)   | 6.08       | 6.08      | 5.98                   | 6.18 | 0.00       |
| Aroclor-1254-4 (4)   | 6.31       | 6.31      | 6.21                   | 6.41 | 0.00       |
| Aroclor-1254-5 (5)   | 6.72       | 6.73      | 6.63                   | 6.83 | 0.01       |
| Tetrachloro-m-xylene | 3.66       | 3.66      | 3.56                   | 3.76 | 0.00       |
| Decachlorobiphenyl   | 8.63       | 8.64      | 8.54                   | 8.74 | 0.01       |