

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** FIRS02
Lab Code: ACE **SDG NO.:** Q2820
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/23/2025 07/23/2025

Client Sample No.: CCAL03 **Date Analyzed:** 08/12/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO112880.D **Time Analyzed:** 21:09

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.749	4.655	4.855	602.530	500.000	20.5
Aroclor-1016-2	4.767	4.674	4.874	529.340	500.000	5.9
Aroclor-1016-3	4.825	4.731	4.931	530.990	500.000	6.2
Aroclor-1016-4	4.945	4.851	5.051	538.670	500.000	7.7
Aroclor-1016-5	5.201	5.108	5.308	570.370	500.000	14.1
Aroclor-1260-1	6.238	6.145	6.345	516.390	500.000	3.3
Aroclor-1260-2	6.427	6.333	6.533	497.430	500.000	-0.5
Aroclor-1260-3	6.793	6.700	6.900	519.410	500.000	3.9
Aroclor-1260-4	7.053	6.959	7.159	517.940	500.000	3.6
Aroclor-1260-5	7.296	7.203	7.403	487.190	500.000	-2.6
Decachlorobiphenyl	8.685	8.593	8.793	44.850	50.000	-10.3
Tetrachloro-m-xylene	3.665	3.569	3.769	56.230	50.000	12.5

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** FIRS02
Lab Code: ACE **SDG NO.:** Q2820
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/23/2025 07/23/2025

Client Sample No.: CCAL03 **Date Analyzed:** 08/12/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO112880.D **Time Analyzed:** 21:09

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.733	4.638	4.838	517.630	500.000	3.5
Aroclor-1016-2	4.751	4.656	4.856	517.600	500.000	3.5
Aroclor-1016-3	4.926	4.831	5.031	509.980	500.000	2.0
Aroclor-1016-4	4.969	4.873	5.073	508.980	500.000	1.8
Aroclor-1016-5	5.180	5.085	5.285	524.870	500.000	5.0
Aroclor-1260-1	6.208	6.113	6.313	512.410	500.000	2.5
Aroclor-1260-2	6.396	6.302	6.502	516.240	500.000	3.2
Aroclor-1260-3	6.547	6.453	6.653	523.480	500.000	4.7
Aroclor-1260-4	7.017	6.923	7.123	518.150	500.000	3.6
Aroclor-1260-5	7.259	7.164	7.364	539.020	500.000	7.8
Decachlorobiphenyl	8.633	8.540	8.740	51.820	50.000	3.6
Tetrachloro-m-xylene	3.659	3.562	3.762	54.700	50.000	9.4