

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** FIRS02
Lab Code: ACE **SDG NO.:** Q2820
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/01/2025 08/01/2025

Client Sample No.: CCAL04 **Date Analyzed:** 08/12/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP074302.D **Time Analyzed:** 09:51

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.809	5.712	5.912	541.110	500.000	8.2
Aroclor-1016-2	5.830	5.733	5.933	542.210	500.000	8.4
Aroclor-1016-3	5.894	5.796	5.996	546.140	500.000	9.2
Aroclor-1016-4	5.990	5.893	6.093	549.740	500.000	9.9
Aroclor-1016-5	6.283	6.186	6.386	547.590	500.000	9.5
Aroclor-1260-1	7.400	7.303	7.503	520.400	500.000	4.1
Aroclor-1260-2	7.653	7.555	7.755	502.220	500.000	0.4
Aroclor-1260-3	8.010	7.913	8.113	510.120	500.000	2.0
Aroclor-1260-4	8.239	8.142	8.342	508.850	500.000	1.8
Aroclor-1260-5	8.564	8.468	8.668	488.630	500.000	-2.3
Decachlorobiphenyl	10.432	10.338	10.538	50.860	50.000	1.7
Tetrachloro-m-xylene	4.657	4.560	4.760	56.700	50.000	13.4

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** FIRS02
Lab Code: ACE **SDG NO.:** Q2820
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/01/2025 08/01/2025

Client Sample No.: CCAL04 **Date Analyzed:** 08/12/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP074302.D **Time Analyzed:** 09:51

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.901	4.806	5.006	583.180	500.000	16.6
Aroclor-1016-2	4.958	4.863	5.063	601.820	500.000	20.4
Aroclor-1016-3	5.080	4.983	5.183	578.000	500.000	15.6
Aroclor-1016-4	5.120	5.023	5.223	590.230	500.000	18.0
Aroclor-1016-5	5.335	5.239	5.439	592.640	500.000	18.5
Aroclor-1260-1	6.551	6.455	6.655	550.000	500.000	10.0
Aroclor-1260-2	6.705	6.609	6.809	554.960	500.000	11.0
Aroclor-1260-3	6.915	6.819	7.019	569.790	500.000	14.0
Aroclor-1260-4	7.174	7.079	7.279	581.890	500.000	16.4
Aroclor-1260-5	7.413	7.318	7.518	587.660	500.000	17.5
Decachlorobiphenyl	8.818	8.726	8.926	52.860	50.000	5.7
Tetrachloro-m-xylene	3.802	3.705	3.905	61.630	50.000	23.3