

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** FIRS02
Lab Code: ACE **SDG NO.:** Q2820
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/01/2025 08/01/2025

Client Sample No.: CCAL05 **Date Analyzed:** 08/12/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP074316.D **Time Analyzed:** 15:48

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.811	5.712	5.912	556.880	500.000	11.4
Aroclor-1016-2	5.832	5.733	5.933	550.560	500.000	10.1
Aroclor-1016-3	5.895	5.796	5.996	559.440	500.000	11.9
Aroclor-1016-4	5.992	5.893	6.093	563.510	500.000	12.7
Aroclor-1016-5	6.284	6.186	6.386	554.850	500.000	11.0
Aroclor-1260-1	7.401	7.303	7.503	536.530	500.000	7.3
Aroclor-1260-2	7.654	7.555	7.755	517.500	500.000	3.5
Aroclor-1260-3	8.012	7.913	8.113	522.830	500.000	4.6
Aroclor-1260-4	8.240	8.142	8.342	532.940	500.000	6.6
Aroclor-1260-5	8.565	8.468	8.668	506.790	500.000	1.4
Decachlorobiphenyl	10.436	10.338	10.538	52.310	50.000	4.6
Tetrachloro-m-xylene	4.659	4.560	4.760	57.560	50.000	15.1

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** FIRS02
Lab Code: ACE **SDG NO.:** Q2820
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/01/2025 08/01/2025

Client Sample No.: CCAL05 **Date Analyzed:** 08/12/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP074316.D **Time Analyzed:** 15:48

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.903	4.806	5.006	585.540	500.000	17.1
Aroclor-1016-2	4.958	4.863	5.063	555.580	500.000	11.1
Aroclor-1016-3	5.081	4.983	5.183	591.090	500.000	18.2
Aroclor-1016-4	5.122	5.023	5.223	595.870	500.000	19.2
Aroclor-1016-5	5.334	5.239	5.439	590.590	500.000	18.1
Aroclor-1260-1	6.552	6.455	6.655	580.600	500.000	16.1
Aroclor-1260-2	6.706	6.609	6.809	568.790	500.000	13.8
Aroclor-1260-3	6.917	6.819	7.019	594.280	500.000	18.9
Aroclor-1260-4	7.175	7.079	7.279	583.960	500.000	16.8
Aroclor-1260-5	7.414	7.318	7.518	585.850	500.000	17.2
Decachlorobiphenyl	8.820	8.726	8.926	55.460	50.000	10.9
Tetrachloro-m-xylene	3.802	3.705	3.905	59.630	50.000	19.3