

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>FIRS02</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q2820</u>
Instrument ID:	<u>BNA_P</u>	Calibration Date/Time:	<u>08/19/2025 23:46</u>
Lab File ID:	<u>BP025487.D</u>	Init. Calib. Date(s):	<u>08/14/2025 08/14/2025</u>
EPA Sample No.:	<u>SSTDCCC040</u>	Init. Calib. Time(s):	<u>12:33 17:24</u>
GC Column:	<u>ZB-GR</u>	ID:	<u>0.25</u> (mm)

COMPOUND	RRF	RRF040	MIN RRF	%D	MAX%D
2-Fluorophenol	1.204	1.204		0.0	
Benzaldehyde	1.082	0.876		-19.0	
Phenol-d6	1.544	1.570		1.7	
Phenol	1.620	1.661		2.5	20.0
bis(2-Chloroethyl)ether	1.263	1.268		0.4	
2-Chlorophenol	1.359	1.389		2.2	
2-Methylphenol	1.125	1.152		2.4	
2,2-oxybis(1-Chloropropane)	1.692	1.784		5.4	
Acetophenone	0.502	0.503		0.2	
3+4-Methylphenols	1.560	1.571		0.7	
n-Nitroso-di-n-propylamine	1.023	1.046	0.050	2.2	
Nitrobenzene-d5	0.395	0.401		1.5	
Hexachloroethane	0.563	0.566		0.5	
Nitrobenzene	0.353	0.362		2.5	
Isophorone	0.700	0.705		0.7	
2-Nitrophenol	0.181	0.185		2.2	20.0
2,4-Dimethylphenol	0.312	0.315		1.0	
bis(2-Chloroethoxy)methane	0.423	0.427		0.9	
2,4-Dichlorophenol	0.310	0.313		1.0	20.0
Naphthalene	1.040	1.035		-0.5	
4-Chloroaniline	0.459	0.467		1.7	
Hexachlorobutadiene	0.211	0.207		-1.9	20.0
Caprolactam	0.114	0.117		2.6	
4-Chloro-3-methylphenol	0.355	0.367		3.4	20.0
2-Methylnaphthalene	0.676	0.673		-0.4	
Hexachlorocyclopentadiene	0.349	0.402	0.050	15.2	
2,4,6-Trichlorophenol	0.390	0.400		2.6	20.0
2-Fluorobiphenyl	1.463	1.408		-3.8	
2,4,5-Trichlorophenol	0.427	0.443		3.7	
1,1-Biphenyl	1.453	1.462		0.6	
2-Chloronaphthalene	1.131	1.130		-0.1	
2-Nitroaniline	0.329	0.359		9.1	
Dimethylphthalate	1.465	1.503		2.6	
Acenaphthylene	1.871	1.889		1.0	
2,6-Dinitrotoluene	0.309	0.327		5.8	
3-Nitroaniline	0.347	0.367		5.8	
Acenaphthene	1.098	1.071		-2.5	20.0
2,4-Dinitrophenol	0.165	0.332	0.050	101.2	
4-Nitrophenol	0.285	0.301	0.050	5.6	

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>FIRS02</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q2820</u>
Instrument ID:	<u>BNA_P</u>	Calibration Date/Time:	<u>08/19/2025 23:46</u>
Lab File ID:	<u>BP025487.D</u>	Init. Calib. Date(s):	<u>08/14/2025 08/14/2025</u>
EPA Sample No.:	<u>SSTDCCC040</u>	Init. Calib. Time(s):	<u>12:33 17:24</u>
GC Column:	<u>ZB-GR</u>	ID:	<u>0.25</u> (mm)

COMPOUND	RRF	RRF040	MIN RRF	%D	MAX%D
Dibenzofuran	1.736	1.710		-1.5	
2,4-Dinitrotoluene	0.440	0.475		8.0	
Diethylphthalate	1.488	1.537		3.3	
4-Chlorophenyl-phenylether	0.681	0.664		-2.5	
Fluorene	1.370	1.332		-2.8	
4-Nitroaniline	0.356	0.368		3.4	
4,6-Dinitro-2-methylphenol	0.125	0.178		42.4	
n-Nitrosodiphenylamine	0.610	0.625		2.5	20.0
2,4,6-Tribromophenol	0.269	0.278		3.3	
4-Bromophenyl-phenylether	0.220	0.222		0.9	
Hexachlorobenzene	0.255	0.255		0.0	
Atrazine	0.228	0.232		1.8	
Pentachlorophenol	0.161	0.170		5.6	20.0
Phenanthrene	1.099	1.084		-1.4	
Anthracene	1.122	1.144		2.0	
Carbazole	1.057	1.052		-0.5	
Di-n-butylphthalate	1.300	1.333		2.5	
Fluoranthene	1.311	1.283		-2.1	20.0
Pyrene	1.300	1.285		-1.2	
Terphenyl-d14	1.093	1.061		-2.9	
Butylbenzylphthalate	0.589	0.605		2.7	
3,3-Dichlorobenzidine	0.497	0.485		-2.4	
Benzo (a) anthracene	1.319	1.302		-1.3	
Chrysene	1.235	1.231		-0.3	
Bis(2-ethylhexyl)phthalate	0.867	0.890		2.7	
Di-n-octyl phthalate	1.492	1.526		2.2	20.0
Benzo (b) fluoranthene	1.179	1.172		-0.6	
Benzo (k) fluoranthene	1.180	1.163		-1.4	
Benzo (a) pyrene	1.148	1.141		-0.6	20.0
Indeno (1,2,3-cd) pyrene	1.467	1.434		-2.2	
Dibenzo (a,h) anthracene	1.199	1.180		-1.6	
Benzo (g,h,i) perylene	1.178	1.150		-2.4	
1,2,4,5-Tetrachlorobenzene	0.578	0.572		-1.0	
1,4-Dioxane	0.472	0.461		-2.3	20.0
2,3,4,6-Tetrachlorophenol	0.377	0.388		2.9	

All other compounds must meet a minimum RRF of 0.010.