

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q2820

Client ID : SSTDCCC040

Date Analyzed: 08/19/2025

Lab File ID: BP025470.D

Time Analyzed: 10:44

Instrument ID: BNA_P

GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	152340	7.784	634765	10.56	424501	14.40
UPPER LIMIT	304680	8.284	1269530	11.06	849002	14.895
LOWER LIMIT	76170	7.284	317383	10.06	212251	13.895
EPA SAMPLE NO.						
01 88H-WRE	191159	7.79	749052	10.55	500665	14.41
02 88H-SRE	154647	7.78	609725	10.55	414467	14.40

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.