

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q2820

Client ID : SSTDCCC040

Date Analyzed: 08/19/2025

Lab File ID: BP025487.D

Time Analyzed: 23:46

Instrument ID: BNA_P

GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	159158	7.79	679178	10.56	450864	14.41
UPPER LIMIT	318316	8.29	1358360	11.06	901728	14.913
LOWER LIMIT	79579	7.29	339589	10.06	225432	13.913
EPA SAMPLE NO.						
01 PB169242BL	158125	7.79	604850	10.56	392339	14.40
02 PB169242BS	125615	7.80	505871	10.57	363724	14.40
03 10P-W	145357	7.78	568025	10.55	384903	14.41

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.