

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q2820

Client ID : SSTDCCC040

Date Analyzed: 08/21/2025

Lab File ID: BP025528.D

Time Analyzed: 06:42

Instrument ID: BNA\_P

GC Column: ZB-GR ID: 0.25 (mm)

|                | IS1 (DCB)<br>AREA # | RT # | IS2 (NPT)<br>AREA # | RT #  | IS3 (ANT)<br>AREA # | RT #   |
|----------------|---------------------|------|---------------------|-------|---------------------|--------|
| 12 HOUR STD    | 210336              | 7.79 | 893021              | 10.56 | 582847              | 14.40  |
| UPPER LIMIT    | 420672              | 8.29 | 1786040             | 11.06 | 1165690             | 14.901 |
| LOWER LIMIT    | 105168              | 7.29 | 446511              | 10.06 | 291424              | 13.901 |
| EPA SAMPLE NO. |                     |      |                     |       |                     |        |
| 01 10P-S       | 242296              | 7.78 | 874020              | 10.56 | 479174              | 14.40  |

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

# Column used to flag values outside QC limits with an asterisk.

\* Values outside of QC limits.