

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: Alliance
Lab Code: ACE
Lab File ID: BP025421.D
Instrument ID: BNA_P

Contract: FIRS02
SDG NO.: Q2820
DFTPP Injection Date: 08/15/2025
DFTPP Injection Time: 09:55

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
68	Less than 2.0% of mass 69	0.5 (1.8) 1
69	Mass 69 relative abundance	100
70	Less than 2.0% of mass 69	0.1 (0.5) 1
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.6
365	Greater than 1% of mass 198	4.3
441	Present, but less than mass 443	78.6
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	19.5 (19.5) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC040	SSTDCCC040	BP025422.D	08/15/2025	10:36
518R-SRE	Q2820-07RE	BP025432.D	08/15/2025	18:17
82H-SRE	Q2820-01RE	BP025433.D	08/15/2025	18:58
22M-NRE	Q2820-21RE	BP025434.D	08/15/2025	19:40
22M-WRE	Q2820-22RE	BP025435.D	08/15/2025	20:21
SSTDCCC040EC	SSTDCCC040	BP025436.D	08/15/2025	21:03