

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>FIRS02</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q2820</u>
Instrument ID:	<u>MSVOA_Y</u>	Calibration Date/Time:	<u>08/15/2025 08:39</u>
Lab File ID:	<u>VY023111.D</u>	Init. Calib. Date(s):	<u>08/12/2025 08/12/2025</u>
Heated Purge: (Y/N)	<u>Y</u>	Init. Calib. Time(s):	<u>14:54 16:47</u>
GC Column:	<u>RXI-624</u>	ID:	<u>0.25</u> (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Dichlorodifluoromethane	0.408	0.332		-18.63	20
Chloromethane	0.661	0.651	0.1	-1.51	20
Vinyl Chloride	0.817	0.876		7.22	20
Bromomethane	0.610	0.659		8.03	20
Chloroethane	0.543	0.599		10.31	20
Trichlorofluoromethane	0.994	0.982		-1.21	20
1,1,2-Trichlorotrifluoroethane	0.490	0.484		-1.22	20
1,1-Dichloroethene	0.482	0.471		-2.28	20
Acetone	0.103	0.116		12.62	20
Carbon Disulfide	1.571	1.520		-3.25	20
Methyl tert-butyl Ether	1.249	1.209		-3.2	20
Methyl Acetate	0.307	0.297		-3.26	20
Methylene Chloride	0.620	0.543		-12.42	20
trans-1,2-Dichloroethene	0.541	0.549		1.48	20
1,1-Dichloroethane	0.942	0.968	0.1	2.76	20
Cyclohexane	0.888	0.845		-4.84	20
2-Butanone	0.140	0.143		2.14	20
Carbon Tetrachloride	0.483	0.506		4.76	20
cis-1,2-Dichloroethene	0.626	0.643		2.72	20
Bromochloromethane	0.374	0.395		5.61	20
Chloroform	0.971	1.017		4.74	20
1,1,1-Trichloroethane	0.862	0.888		3.02	20
Methylcyclohexane	0.586	0.583		-0.51	20
Benzene	1.370	1.455		6.2	20
1,2-Dichloroethane	0.356	0.372		4.49	20
Trichloroethene	0.354	0.366		3.39	20
1,2-Dichloropropane	0.315	0.335		6.35	20
Bromodichloromethane	0.466	0.496		6.44	20
4-Methyl-2-Pentanone	0.195	0.194		-0.51	20
Toluene	0.870	0.928		6.67	20
t-1,3-Dichloropropene	0.422	0.434		2.84	20
cis-1,3-Dichloropropene	0.499	0.523		4.81	20
1,1,2-Trichloroethane	0.242	0.252		4.13	20
2-Hexanone	0.133	0.135		1.5	20
Dibromochloromethane	0.320	0.332		3.75	20
1,2-Dibromoethane	0.228	0.230		0.88	20
Tetrachloroethene	0.446	0.469		5.16	20
Chlorobenzene	1.079	1.142	0.3	5.84	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>FIRS02</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q2820</u>
Instrument ID:	<u>MSVOA_Y</u>	Calibration Date/Time:	<u>08/15/2025</u> <u>08:39</u>
Lab File ID:	<u>VY023111.D</u>	Init. Calib. Date(s):	<u>08/12/2025</u> <u>08/12/2025</u>
Heated Purge: (Y/N)	<u>Y</u>	Init. Calib. Time(s):	<u>14:54</u> <u>16:47</u>
GC Column:	<u>RXI-624</u>	ID:	<u>0.25</u> (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Ethyl Benzene	1.856	2.009		8.24	20
m/p-Xylenes	0.725	0.785		8.28	20
o-Xylene	0.679	0.736		8.4	20
Styrene	1.110	1.238		11.53	20
Bromoform	0.214	0.212	0.1	-0.94	20
Isopropylbenzene	3.529	3.734		5.81	20
1,1,2,2-Tetrachloroethane	0.603	0.595	0.3	-1.33	20
1,3-Dichlorobenzene	1.660	1.748		5.3	20
1,4-Dichlorobenzene	1.647	1.696		2.97	20
1,2-Dichlorobenzene	1.456	1.499		2.95	20
1,2-Dibromo-3-Chloropropane	0.094	0.090		-4.26	20
1,2,4-Trichlorobenzene	0.861	0.793		-7.9	20
1,2,3-Trichlorobenzene	0.742	0.679		-8.49	20
1,2-Dichloroethane-d4	0.477	0.480		0.63	20
Dibromofluoromethane	0.299	0.309		3.34	20
Toluene-d8	1.169	1.206		3.16	20
4-Bromofluorobenzene	0.376	0.376		0	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.