



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: Alliance Contract: FIRS02
Lab Code: ACE SDG No.: Q2820
Instrument ID: MSVOA_Y Calibration Date/Time: 08/19/2025 09:11
Lab File ID: VY023137.D Init. Calib. Date(s): 08/12/2025 08/12/2025
Heated Purge: (Y/N) Y Init. Calib. Time(s): 14:54 16:47
GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	Amount	Conc.	%Dev	MAX%D
Methylene Chloride	50.000	47.73	4.545998	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.