

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: Alliance Contract: FIRS02
 Lab Code: ACE SDG No.: Q2820
 Instrument ID: MSVOA_Y Calibration Date/Time: 08/19/2025 19:08
 Lab File ID: VY023160.D Init. Calib. Date(s): 08/12/2025 08/12/2025
 Heated Purge: (Y/N) Y Init. Calib. Time(s): 14:54 16:47
 GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	Amount	Conc.	%Dev	MAX%D
Methylene Chloride	50.000	50.25	-0.4940033	20