

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Contract: FIRS02

Lab Code: ACE

SDG NO.: Q2820

Lab File ID: VW032080.D

Date Analyzed: 08/14/2025

Instrument ID: MSVOA W

Time Analyzed: 09:27

GC Column: RXI-624 ID: 0.25 (mm)

Heated Purge: (Y/N) Y

	IS4 AREA #	RT #				
12 HOUR STD	174813	13.556				
UPPER LIMIT	349626	14.056				
LOWER LIMIT	87406.5	13.056				
EPA SAMPLE NO.						
82H-W	149119	13.56				
82H-N	150411	13.56				
SOIL-DUP-1	147600	13.56				
518R-E	126154	13.56				
518R-N	97060	13.56				
518R-S	122108	13.56				
518R-W	127979	13.56				
705R-S	129964	13.56				
SOIL-DUP-2	103366	13.56				
10PC-W	145639	13.56				
10PC-S	140579	13.56				
10P-W	179419	13.56				
10P-E	140063	13.56				
10P-N	143336	13.56				
88H-E	141036	13.56				
88H-N	149072	13.56				
88H-W	144659	13.56				
88H-S	135214	13.56				
VW0814SBL01	134418	13.56				
VW0814SBS01	173541	13.56				
VW0814SBSD01	177605	13.55				

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: FIRS02
 Lab Code: ACE SDG NO.: Q2820
 Lab File ID: VW032080.D Date Analyzed: 08/14/2025
 Instrument ID: MSVOA_W Time Analyzed: 09:27
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS4 AREA #	RT #				
12 HOUR STD	174813	13.556				
UPPER LIMIT	349626	14.056				
LOWER LIMIT	87406.5	13.056				
EPA SAMPLE NO.						

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.