

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: FIRS02
 Lab Code: ACE SDG NO.: Q2820
 Lab File ID: VY023111.D Date Analyzed: 08/15/2025
 Instrument ID: MSVOA_Y Time Analyzed: 08:39
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS4 AREA #	RT #				
12 HOUR STD	306948	13.347				
UPPER LIMIT	613896	13.847				
LOWER LIMIT	153474	12.847				
EPA SAMPLE NO.						
22M-N	387628	13.34				
22M-W	311657	13.34				
22M-S	376737	13.34				
VY0815SBL01	370974	13.35				
VY0815SBS01	327154	13.34				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.