DATA OF KNOWN QUALITY CONFORMANCE/NON-CONFORMANCE SUMMARY QUESTIONNAIRE

Laboratory Name :		Alliance Technical Group LLC	Client :	G Environmenta	al				
Project Location : NJ			Project Number :						
Laboratory Sample ID(s) : Q2825			Sampling Date(s) :	8/10/2025					
List DKQP Methods Used (e.g., 8260,8270, et Cetra) 8260D,8270-Modified,8270E,SOP									
1	For each analytical method referenced in this laboratory report package, were all specified QA/QC performance criteria followed, including the requirement to explain any criteria falling outside of acceptable guidelines, as specified in the NJDEP Data of Known Quality performance standards?					Yes		No	
1A	Were the method specified handling, preservation, and holding time requirements met?					Yes		No	
1B	EPH Method: Was the EPH method conducted without significant modifications (see Section 11.3 of respective DKQ methods)					Yes		No	✓ N/A
2	Were all samples received by the laboratory in a condition consistent with that described on the associated chain-of-custody document(s)?				V	Yes		No	
3	Were samples received at an appropriate temperature (4±2° C)?					Yes		No	□ N/A
4	Were all QA/Q0 standards ach	C performance criteria specified in the Nieved?	NJDEP DKQP			Yes	V	No	
5	a)Were reporting limits specified or referenced on the chain-of-custody or communicated to the laboratory prior to sample receipt?				\square	Yes		No	
	b)Were these re	eporting limits met?			V	Yes		No	□ N/A
6	results reporte	tical method referenced in this laborato ed for all constituents identified in the m e DKQP documents and/or site-specific	ethod-specific analyte		V	Yes		No	
7	Are project-spe	cific matrix spikes and/or laboratory du	plicates included in th	is data set?	V	Yes		No	

Notes: For all questions to which the response was "No" (with the exception of question #7), additional information should be provided in an attached narrative. If the answer to question #1, #1A, or #1B is "No", the data package does not meet the requirements for "Data of Known Quality."