

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN081325\
 Data File : BN037588.D
 Acq On : 13 Aug 2025 10:28
 Operator : RC/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Aug 13 11:18:48 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN081225.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Aug 13 05:00:58 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	130	0.00
2	1,4-Dioxane	0.383	0.412	-7.6	135	0.00
3	n-Nitrosodimethylamine	0.489	0.454	7.2	121	0.00
4 S	2-Fluorophenol	0.907	0.923	-1.8	135	0.00
5 S	Phenol-d6	1.091	1.135	-4.0	141	0.00
6	bis(2-Chloroethyl)ether	0.983	1.016	-3.4	135	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	133	0.00
8 S	Nitrobenzene-d5	0.282	0.265	6.0	135	0.00
9	Naphthalene	1.065	1.048	1.6	135	0.00
10	Hexachlorobutadiene	0.260	0.246	5.4	125	0.00
11 SURR	2-Methylnaphthalene-d10	0.544	0.508	6.6	133	0.00
12	2-Methylnaphthalene	0.668	0.644	3.6	135	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	134	0.00
14 S	2,4,6-Tribromophenol	0.175	0.166	5.1	142	0.00
15 S	2-Fluorobiphenyl	2.313	2.297	0.7	134	0.00
16	Acenaphthylene	1.793	1.704	5.0	133	0.00
17	Acenaphthene	1.220	1.154	5.4	132	0.00
18	Fluorene	1.596	1.491	6.6	131	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	130	0.00
20	4,6-Dinitro-2-methylphenol	0.053	0.043	18.9	121	0.00
21	4-Bromophenyl-phenylether	0.251	0.242	3.6	134	0.00
22	Hexachlorobenzene	0.356	0.355	0.3	129	0.00
23	Atrazine	0.142	0.139	2.1	139	0.00
24	Pentachlorophenol	0.125	0.122	2.4	139	0.00
25	Phenanthrene	1.216	1.165	4.2	129	0.00
26	Anthracene	1.077	0.982	8.8	128	0.00
27 SURR	Fluoranthene-d10	1.052	0.929	11.7	126	0.00
28	Fluoranthene	1.397	1.275	8.7	127	0.00
29 I	Chrysene-d12	1.000	1.000	0.0	150	0.00
30	Pyrene	1.495	1.288	13.8	132	0.00
31 S	Terphenyl-d14	0.823	0.727	11.7	137	0.00
32	Benzo(a)anthracene	1.336	1.415	-5.9	168#	0.00
33	Chrysene	1.489	1.395	6.3	146	0.00
34	Bis(2-ethylhexyl)phthalate	0.551	0.565	-2.5	162#	0.00
35 I	Perylene-d12	1.000	1.000	0.0	138	0.00
36	Indeno(1,2,3-cd)pyrene	1.686	1.608	4.6	137	0.00
37	Benzo(b)fluoranthene	1.516	1.358	10.4	140	0.00
38	Benzo(k)fluoranthene	1.710	1.544	9.7	129	0.00
39 C	Benzo(a)pyrene	1.257	1.227	2.4	147	0.00
40	Dibenzo(a,h)anthracene	1.308	1.242	5.0	140	0.00
41	Benzo(g,h,i)perylene	1.378	1.279	7.2	140	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4