

Surrogate Summary

SW-846

SDG No.: Q2825

Client: G Environmental

Analytical Method: 8270-Modified

Lab Sample ID	Client ID	Parameter	Spike (PPM)	Result (PPM)	Recovery (%)	Qual	Limits (%)	
							Low	High
PB169222BL	PB169222BL	2-Methylnaphthalene-d10	0.4	0.32	80		30 (20)	150 (139)
		Fluoranthene-d10	0.4	0.32	81		30 (54)	150 (157)
		Nitrobenzene-d5	0.4	0.34	86		30 (27)	130 (154)
		2-Fluorobiphenyl	0.4	0.35	86		30 (30)	130 (155)
		Terphenyl-d14	0.4	0.38	94		30 (54)	130 (175)
PB169222BS	PB169222BS	2-Methylnaphthalene-d10	0.4	0.34	84		30 (20)	150 (139)
		Fluoranthene-d10	0.4	0.30	75		30 (54)	150 (157)
		Nitrobenzene-d5	0.4	0.34	84		30 (27)	130 (154)
		2-Fluorobiphenyl	0.4	0.36	89		30 (30)	130 (155)
		Terphenyl-d14	0.4	0.37	93		30 (54)	130 (175)
PB169222BSD	PB169222BSD	2-Methylnaphthalene-d10	0.4	0.35	88		30 (20)	150 (139)
		Fluoranthene-d10	0.4	0.32	80		30 (54)	150 (157)
		Nitrobenzene-d5	0.4	0.36	91		30 (27)	130 (154)
		2-Fluorobiphenyl	0.4	0.39	97		30 (30)	130 (155)
		Terphenyl-d14	0.4	0.41	101		30 (54)	130 (175)
Q2825-01	TW1	2-Methylnaphthalene-d10	0.4	0.21	53		30 (20)	150 (139)
		Fluoranthene-d10	0.4	0.35	88		30 (54)	150 (157)
		Nitrobenzene-d5	0.4	0.43	107		30 (27)	130 (154)
		2-Fluorobiphenyl	0.4	0.13	31		30 (30)	130 (155)
		Terphenyl-d14	0.4	0.44	109		30 (54)	130 (175)