

## Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

|                 |                        |                           |                      |
|-----------------|------------------------|---------------------------|----------------------|
| <b>SDG No.:</b> | <u>Q2825</u>           | <b>Analytical Method:</b> | <u>8270-Modified</u> |
| <b>Client:</b>  | <u>G Environmental</u> | <b>DataFile:</b>          | <u>BN037597.D</u>    |

| Lab Sample ID | Parameter            | Spike | Result | Unit | Rec | RPD | Qual | RPD  | Limits  |           |         |
|---------------|----------------------|-------|--------|------|-----|-----|------|------|---------|-----------|---------|
|               |                      |       |        |      |     |     |      | Qual | Low     | High      | RPD     |
| PB169222BSD   | Benzo(a)anthracene   | 0.4   | 0.37   | ug/L | 93  | 8   |      |      | 70 (54) | 130 (130) | 20 (20) |
|               | Benzo(b)fluoranthene | 0.4   | 0.38   | ug/L | 95  | 11  |      |      | 70 (65) | 130 (121) | 20 (20) |
|               | Benzo(k)fluoranthene | 0.4   | 0.35   | ug/L | 88  | 9   |      |      | 70 (72) | 130 (119) | 20 (20) |
|               | Benzo(a)pyrene       | 0.4   | 0.39   | ug/L | 98  | 8   |      |      | 70 (68) | 130 (120) | 20 (20) |
|               | Benzo(g,h,i)perylene | 0.4   | 0.41   | ug/L | 103 | 10  |      |      | 70 (76) | 130 (117) | 20 (20) |