

Instrument ID: BNA_P

Daily Analysis Runlog For Sequence/QC Batch ID # BP081425

Review By	Rahul	Review On	8/15/2025 9:35:16 AM
Supervise By	Jagrut	Supervise On	8/15/2025 3:48:12 PM
SubDirectory	BP081425	HP Acquire Method	BNA_P
		HP Processing Method	bp081425

STD. NAME	STD REF.#
Tune/Reschk	SP6757
Initial Calibration Stds	SP6833,SP6834,SP6835,SP6836,SP6837,SP6838,SP6839,SP6840
CCC	SP6836
Internal Standard/PEM	S13170,10ul/1000ul sample
ICV/I.BLK	SP6770
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BP025390.D	14 Aug 2025 10:30		CG/JU	Ok
2	SSTDCCC040	SSTDCCC040	BP025391.D	14 Aug 2025 11:52	A Fresh Calibration is required.	CG/JU	Not Ok
3	SSTDICC2.5	SSTDICC2.5	BP025392.D	14 Aug 2025 12:33		CG/JU	Ok
4	SSTDICC005	SSTDICC005	BP025393.D	14 Aug 2025 13:15		CG/JU	Ok,M
5	SSTDICC010	SSTDICC010	BP025394.D	14 Aug 2025 13:56		CG/JU	Ok,M
6	SSTDICC020	SSTDICC020	BP025395.D	14 Aug 2025 14:38		CG/JU	Ok,M
7	SSTDICCC040	SSTDICCC040	BP025396.D	14 Aug 2025 15:19	Benzidine failed in the Calibration Method.	CG/JU	Ok,M
8	SSTDICC050	SSTDICC050	BP025397.D	14 Aug 2025 16:01		CG/JU	Ok,M
9	SSTDICC060	SSTDICC060	BP025398.D	14 Aug 2025 16:42		CG/JU	Ok
10	SSTDICC080	SSTDICC080	BP025399.D	14 Aug 2025 17:24		CG/JU	Ok,M
11	SSTDICV040	ICVBP081425	BP025400.D	14 Aug 2025 18:05		CG/JU	Ok,M
12	PB169200TB	PB169200TB	BP025401.D	14 Aug 2025 19:28		CG/JU	Ok
13	PB169210BS	PB169210BS	BP025402.D	14 Aug 2025 20:10		CG/JU	Ok,M
14	SSTDCCC040	SSTDCCC040EC	BP025403.D	14 Aug 2025 20:51		CG/JU	Ok,M
15	DFTPP	DFTPP	BP025404.D	14 Aug 2025 22:14		CG/JU	Ok
16	SSTDCCC040	SSTDCCC040	BP025405.D	14 Aug 2025 22:56		CG/JU	Ok,M
17	PB169228BL	PB169228BL	BP025406.D	14 Aug 2025 23:37		CG/JU	Ok

Instrument ID: BNA_P

Daily Analysis Runlog For Sequence/QC Batch ID # BP081425

Review By	Rahul	Review On	8/15/2025 9:35:16 AM		
Supervise By	Jagrut	Supervise On	8/15/2025 3:48:12 PM		
SubDirectory	BP081425	HP Acquire Method	BNA_P	HP Processing Method	bp081425
STD. NAME	STD REF.#				
Tune/Reschk	SP6757				
Initial Calibration Stds	SP6833,SP6834,SP6835,SP6836,SP6837,SP6838,SP6839,SP6840				
CCC	SP6836				
Internal Standard/PEM	S13170,10ul/1000ul sample				
ICV/I.BLK	SP6770				
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

18	PB169228BS	PB169228BS	BP025407.D	15 Aug 2025 00:18		CG/JU	Ok,M
19	Q2820-10	SOIL-DUP-2	BP025408.D	15 Aug 2025 01:00		CG/JU	Ok
20	Q2820-11	10PC-W	BP025409.D	15 Aug 2025 01:41		CG/JU	Ok
21	Q2820-21	22M-N	BP025410.D	15 Aug 2025 02:22	Surrogate Fail	CG/JU	ReRun
22	Q2820-22	22M-W	BP025411.D	15 Aug 2025 03:03	Surrogate Fail	CG/JU	ReRun
23	Q2820-14	10P-E	BP025412.D	15 Aug 2025 03:45		CG/JU	Ok
24	Q2820-16	10P-N	BP025413.D	15 Aug 2025 04:26		CG/JU	Ok
25	Q2820-17	88H-E	BP025414.D	15 Aug 2025 05:07		CG/JU	Ok
26	Q2820-17MS	88H-EMS	BP025415.D	15 Aug 2025 05:48		CG/JU	Ok,M
27	Q2820-17MSD	88H-EMSD	BP025416.D	15 Aug 2025 06:29		CG/JU	Ok,M
28	Q2820-18	88H-N	BP025417.D	15 Aug 2025 07:10		CG/JU	Ok
29	Q2820-19	88H-W	BP025418.D	15 Aug 2025 07:51	Surrogate fail	CG/JU	ReRun
30	Q2820-20	88H-S	BP025419.D	15 Aug 2025 08:32	Surrogate Fail	CG/JU	ReRun
31	SSTDCCC040	SSTDCCC040EC	BP025420.D	15 Aug 2025 09:13		CG/JU	Ok,M

M : Manual Integration

Instrument ID: BNA_P

Daily Analysis Runlog For Sequence/QC Batch ID # BP081525

Review By	Rahul	Review On	8/18/2025 8:39:39 AM
Supervise By	Jagrut	Supervise On	8/18/2025 8:42:55 AM
SubDirectory	BP081525	HP Acquire Method	BNA_P
		HP Processing Method	bp081425

STD. NAME	STD REF.#
Tune/Reschk	SP6757
Initial Calibration Stds	SP6833,SP6834,SP6835,SP6836,SP6837,SP6838,SP6839,SP6840
CCC	SP6836
Internal Standard/PEM	S13170,10ul/1000ul sample
ICV/I.BLK	SP6770
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BP025421.D	15 Aug 2025 09:55		CG/JU	Ok
2	SSTDCCC040	SSTDCCC040	BP025422.D	15 Aug 2025 10:36		CG/JU	Ok,M
3	PB169230BL	PB169230BL	BP025423.D	15 Aug 2025 11:17		CG/JU	Ok
4	PB169230BS	PB169230BS	BP025424.D	15 Aug 2025 11:59		CG/JU	Ok,M
5	PB169230BSD	PB169230BSD	BP025425.D	15 Aug 2025 13:26		CG/JU	Ok
6	Q2814-05RE	TW-82H-ERE	BP025426.D	15 Aug 2025 14:08	Surrogate Fail	CG/JU	Confirms
7	Q2814-20	TW-518R-W	BP025427.D	15 Aug 2025 14:49		CG/JU	Ok
8	Q2814-11	TW-38M-S	BP025428.D	15 Aug 2025 15:31		CG/JU	Ok
9	Q2814-17	TW-518R-S	BP025429.D	15 Aug 2025 16:12		CG/JU	Ok
10	Q2814-18	TW-518R-N	BP025430.D	15 Aug 2025 16:54	Surrogate Fail	CG/JU	ReRun
11	Q2827-05DL	TP-9DL	BP025431.D	15 Aug 2025 17:35		CG/JU	Ok,M
12	Q2820-07RE	518R-SRE	BP025432.D	15 Aug 2025 18:17	Surrogate Fail	CG/JU	Confirms
13	Q2820-01RE	82H-SRE	BP025433.D	15 Aug 2025 18:58	Surrogate Fail	CG/JU	Confirms
14	Q2820-21RE	22M-NRE	BP025434.D	15 Aug 2025 19:40	Surrogate Fail	CG/JU	Confirms
15	Q2820-22RE	22M-WRE	BP025435.D	15 Aug 2025 20:21	Surrogate Fail	CG/JU	Confirms
16	SSTDCCC040	SSTDCCC040EC	BP025436.D	15 Aug 2025 21:03		CG/JU	Ok,M
17	DFTPP	DFTPP	BP025437.D	15 Aug 2025 23:08		CG/JU	Ok
18	SSTDCCC040	SSTDCCC040	BP025438.D	15 Aug 2025 23:49		CG/JU	Ok,M

Instrument ID: BNA_P

Daily Analysis Runlog For Sequence/QC Batch ID # BP081525

Review By	Rahul	Review On	8/18/2025 8:39:39 AM		
Supervise By	Jagrut	Supervise On	8/18/2025 8:42:55 AM		
SubDirectory	BP081525	HP Acquire Method	BNA_P	HP Processing Method	bp081425
STD. NAME		STD REF.#			
Tune/Reschk	SP6757				
Initial Calibration Stds	SP6833,SP6834,SP6835,SP6836,SP6837,SP6838,SP6839,SP6840				
CCC	SP6836				
Internal Standard/PEM	S13170,10ul/1000ul sample				
ICV/I.BLK	SP6770				
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

19	PB169275BL	PB169275BL	BP025439.D	16 Aug 2025 00:31		CG/JU	Ok
20	PB169275BS	PB169275BS	BP025440.D	16 Aug 2025 01:13		CG/JU	Ok,M
21	Q2732-03	WC-A7-01-C	BP025441.D	16 Aug 2025 01:55		CG/JU	Ok
22	Q2836-03	WC-A2-15-C	BP025442.D	16 Aug 2025 02:37		CG/JU	Ok
23	Q2836-07	WC-A2-16-C	BP025443.D	16 Aug 2025 03:19		CG/JU	Ok
24	Q2836-11	WC-A2-17-C	BP025444.D	16 Aug 2025 04:00		CG/JU	Ok
25	Q2836-15	WC-A5-02-C	BP025445.D	16 Aug 2025 04:41		CG/JU	Ok
26	Q2838-04	TP-11	BP025446.D	16 Aug 2025 05:22		CG/JU	Ok
27	Q2838-08	TP-10	BP025447.D	16 Aug 2025 06:04		CG/JU	Ok
28	Q2832-02	TG-S01	BP025448.D	16 Aug 2025 06:45		CG/JU	Ok
29	Q2832-02MS	TG-S01MS	BP025449.D	16 Aug 2025 07:26		CG/JU	Ok,M
30	Q2832-02MSD	TG-S01MSD	BP025450.D	16 Aug 2025 08:07		CG/JU	Ok,M
31	Q2832-04	TG-S02	BP025451.D	16 Aug 2025 08:48		CG/JU	Ok
32	Q2832-06	TG-S03	BP025452.D	16 Aug 2025 09:29		CG/JU	Ok
33	Q2832-08	TG-S04	BP025453.D	16 Aug 2025 10:11		CG/JU	Ok

M : Manual Integration

Instrument ID: BNA_P

Daily Analysis Runlog For Sequence/QC Batch ID # BP081925

Review By	Rahul	Review On	8/20/2025 11:11:45 AM
Supervise By	Jagrut	Supervise On	8/20/2025 12:07:50 PM
SubDirectory	BP081925	HP Acquire Method	BNA_P
		HP Processing Method	bp081425

STD. NAME	STD REF.#
Tune/Reschk	SP6757
Initial Calibration Stds	SP6833,SP6834,SP6835,SP6836,SP6837,SP6838,SP6839,SP6840
CCC	SP6836
Internal Standard/PEM	S13171,10ul/1000ul sample
ICV/I.BLK	SP6770
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BP025469.D	19 Aug 2025 10:03		CG/JU	Ok
2	SSTDCCC040	SSTDCCC040	BP025470.D	19 Aug 2025 10:44		CG/JU	Ok
3	PB169244BL	PB169244BL	BP025471.D	19 Aug 2025 11:25		CG/JU	Ok
4	PB169244BS	PB169244BS	BP025472.D	19 Aug 2025 12:06		CG/JU	Ok,M
5	PB169244BSD	PB169244BSD	BP025473.D	19 Aug 2025 12:47		CG/JU	Ok,M
6	Q2815-18	TW-84SB-W	BP025474.D	19 Aug 2025 13:29		CG/JU	Ok
7	Q2815-19	DUP	BP025475.D	19 Aug 2025 14:10		CG/JU	Ok
8	Q2815-20	TW-11M-W	BP025476.D	19 Aug 2025 14:51		CG/JU	Ok
9	Q2815-21	TW-11M-E	BP025477.D	19 Aug 2025 15:32		CG/JU	Ok
10	Q2815-22	TW-11M-S	BP025478.D	19 Aug 2025 16:13		CG/JU	Ok
11	Q2815-23	TW-11M-N	BP025479.D	19 Aug 2025 16:55		CG/JU	Ok
12	Q2815-26	FB	BP025480.D	19 Aug 2025 17:36		CG/JU	Ok
13	Q2820-19RE	88H-WRE	BP025481.D	19 Aug 2025 18:17	Surrogate fail,	CG/JU	Confirms
14	Q2820-20RE	88H-SRE	BP025482.D	19 Aug 2025 18:58	Surrogate fail.	CG/JU	Confirms
15	Q2815-13	TW-22M-E	BP025483.D	19 Aug 2025 19:39	Surrogates failed.	CG/JU	Ok
16	Q2815-14	TW-22M-N	BP025484.D	19 Aug 2025 20:21		CG/JU	Ok
17	SSTDCCC040	SSTDCCC040EC	BP025485.D	19 Aug 2025 21:43		CG/JU	Ok
18	DFTPP	DFTPP	BP025486.D	19 Aug 2025 23:05		CG/JU	Ok

Instrument ID: BNA_P

Daily Analysis Runlog For Sequence/QC Batch ID # BP081925

Review By	Rahul	Review On	8/20/2025 11:11:45 AM		
Supervise By	Jagrut	Supervise On	8/20/2025 12:07:50 PM		
SubDirectory	BP081925	HP Acquire Method	BNA_P	HP Processing Method	bp081425
STD. NAME		STD REF.#			
Tune/Reschk		SP6757			
Initial Calibration Stds		SP6833,SP6834,SP6835,SP6836,SP6837,SP6838,SP6839,SP6840			
CCC		SP6836			
Internal Standard/PEM		S13171,10ul/1000ul sample			
ICV/I.BLK		SP6770			
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

19	SSTDCCC040	SSTDCCC040	BP025487.D	19 Aug 2025 23:46		CG/JU	Ok,M
20	PB169242BL	PB169242BL	BP025488.D	20 Aug 2025 00:27		CG/JU	Ok
21	PB169242BS	PB169242BS	BP025489.D	20 Aug 2025 01:07		CG/JU	Ok
22	Q2870-04	DRILL CUTTING 6-9 C	BP025490.D	20 Aug 2025 01:48		CG/JU	Ok
23	Q2871-01	DRILL CUTTING 1-3 C	BP025491.D	20 Aug 2025 02:29		CG/JU	Ok
24	Q2870-01	DRILL CUTTING 1-5 C	BP025492.D	20 Aug 2025 03:10		CG/JU	Ok
25	Q2871-04	DRILL CUTTING 4-6 C	BP025493.D	20 Aug 2025 03:51		CG/JU	Ok
26	Q2871-07	DRILL CUTTING 1-5 C	BP025494.D	20 Aug 2025 04:32		CG/JU	Ok
27	Q2873-04	DRILL CUTTING 4-6 C	BP025495.D	20 Aug 2025 05:13		CG/JU	Ok
28	Q2825-01	TW1	BP025496.D	20 Aug 2025 05:54		CG/JU	Ok
29	Q2872-01	DRILL CUTTING 1-2 C	BP025497.D	20 Aug 2025 06:35		CG/JU	Ok
30	Q2815-17	TW-84SB-S	BP025498.D	20 Aug 2025 07:16		CG/JU	Ok
31	Q2815-16	TW-17M-S	BP025499.D	20 Aug 2025 07:57		CG/JU	Ok
32	Q2819-05	11M-W	BP025500.D	20 Aug 2025 08:38		CG/JU	Ok
33	Q2820-13	10P-W	BP025501.D	20 Aug 2025 09:19		CG/JU	Ok
34	SSTDCCC040	SSTDCCC040EC	BP025502.D	20 Aug 2025 10:41		CG/JU	Ok,M

M : Manual Integration