

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q2825

Analytical Method:

8270E

Client: G Environmental

DataFile:

BP025424.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	RPD		Limits	
								Qual	Low	High	RPD
PB169230BS	Benzaldehyde	50	23.9	ug/L	48				20 (10)	160 (162)	
	Phenol	50	40.4	ug/L	81				20 (66)	160 (118)	
	bis(2-Chloroethyl)ether	50	37.1	ug/L	74				70 (62)	130 (103)	
	2-Chlorophenol	50	39.9	ug/L	80				70 (70)	130 (117)	
	2-Methylphenol	50	39.7	ug/L	79				70 (69)	130 (109)	
	2,2-oxybis(1-Chloropropane)	50	37.4	ug/L	75				70 (65)	130 (100)	
	Acetophenone	50	38.4	ug/L	77				70 (60)	130 (104)	
	3+4-Methylphenols	50	38.8	ug/L	78				20 (67)	160 (106)	
	N-Nitroso-di-n-propylamine	50	35.3	ug/L	71				70 (57)	130 (107)	
	Hexachloroethane	50	38.7	ug/L	77				20 (76)	160 (118)	
	Nitrobenzene	50	39.7	ug/L	79				70 (58)	130 (106)	
	Isophorone	50	37.0	ug/L	74				70 (61)	130 (102)	
	2-Nitrophenol	50	39.2	ug/L	78				70 (70)	130 (115)	
	2,4-Dimethylphenol	50	40.8	ug/L	82				70 (42)	130 (142)	
	bis(2-Chloroethoxy)methane	50	37.6	ug/L	75				70 (58)	130 (109)	
	2,4-Dichlorophenol	50	41.5	ug/L	83				70 (66)	130 (115)	
	Naphthalene	50	38.6	ug/L	77				70 (64)	130 (107)	
	4-Chloroaniline	50	25.0	ug/L	50		*		70 (10)	130 (85)	
	Hexachlorobutadiene	50	38.4	ug/L	77				70 (69)	130 (101)	
	Caprolactam	50	38.5	ug/L	77				20 (58)	160 (128)	
	4-Chloro-3-methylphenol	50	40.8	ug/L	82				70 (65)	130 (114)	
	2-Methylnaphthalene	50	38.1	ug/L	76				70 (64)	130 (107)	
	Hexachlorocyclopentadiene	100	86.0	ug/L	86				20 (36)	160 (160)	
	2,4,6-Trichlorophenol	50	41.6	ug/L	83				70 (61)	130 (110)	
	2,4,5-Trichlorophenol	50	42.1	ug/L	84				70 (70)	130 (106)	
	1,1-Biphenyl	50	40.1	ug/L	80				70 (72)	130 (98)	
	2-Chloronaphthalene	50	39.6	ug/L	79				70 (59)	130 (106)	
	2-Nitroaniline	50	42.4	ug/L	85				70 (73)	130 (114)	
	Dimethylphthalate	50	39.4	ug/L	79				70 (64)	130 (103)	
	Acenaphthylene	50	39.5	ug/L	79				70 (79)	130 (103)	
	2,6-Dinitrotoluene	50	41.3	ug/L	83				70 (64)	130 (110)	
	3-Nitroaniline	50	30.4	ug/L	61		*		70 (28)	130 (100)	
	Acenaphthene	50	39.1	ug/L	78				70 (59)	130 (113)	
	2,4-Dinitrophenol	100	91.4	ug/L	91				20 (36)	160 (166)	
	4-Nitrophenol	100	84.2	ug/L	84				20 (45)	160 (147)	
	Dibenzofuran	50	39.0	ug/L	78				70 (65)	130 (106)	
	2,4-Dinitrotoluene	50	41.7	ug/L	83				70 (60)	130 (115)	
	Diethylphthalate	50	39.2	ug/L	78				70 (63)	130 (105)	
	4-Chlorophenyl-phenylether	50	38.0	ug/L	76				70 (61)	130 (104)	
	Fluorene	50	38.7	ug/L	77				70 (64)	130 (107)	
	4-Nitroaniline	50	38.9	ug/L	78				70 (55)	130 (125)	
	4,6-Dinitro-2-methylphenol	50	42.5	ug/L	85				70 (62)	130 (132)	
	N-Nitrosodiphenylamine	50	41.1	ug/L	82				70 (61)	130 (109)	
	4-Bromophenyl-phenylether	50	40.0	ug/L	80				70 (73)	130 (103)	

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: <u>Q2825</u>	Analytical Method: <u>8270E</u>
Client: <u>G Environmental</u>	DataFile: <u>BP025424.D</u>

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	RPD		Limits	
								Qual	Low	High	RPD
PB169230BS	Hexachlorobenzene	50	40.6	ug/L	81				70 (73)	130 (106)	
	Atrazine	50	40.0	ug/L	80				70 (76)	130 (120)	
	Pentachlorophenol	100	87.0	ug/L	87				20 (47)	160 (114)	
	Phenanthrene	50	41.1	ug/L	82				70 (62)	130 (109)	
	Anthracene	50	41.0	ug/L	82				70 (65)	130 (110)	
	Carbazole	50	40.6	ug/L	81				70 (62)	130 (106)	
	Di-n-butylphthalate	50	40.7	ug/L	81				70 (64)	130 (106)	
	Fluoranthene	50	40.6	ug/L	81				70 (64)	130 (110)	
	Pyrene	50	38.2	ug/L	76				70 (71)	130 (103)	
	Butylbenzylphthalate	50	39.3	ug/L	79				70 (61)	130 (105)	
	3,3-Dichlorobenzidine	50	30.3	ug/L	61		*		70 (43)	130 (108)	
	Benzo(a)anthracene	50	41.0	ug/L	82				70 (62)	130 (107)	
	Chrysene	50	41.0	ug/L	82				70 (61)	130 (108)	
	bis(2-Ethylhexyl)phthalate	50	40.4	ug/L	81				70 (59)	130 (110)	
	Di-n-octyl phthalate	50	41.6	ug/L	83				70 (52)	130 (139)	
	Benzo(b)fluoranthene	50	41.4	ug/L	83				70 (77)	130 (113)	
	Benzo(k)fluoranthene	50	40.4	ug/L	81				70 (77)	130 (105)	
	Benzo(a)pyrene	50	41.2	ug/L	82				70 (72)	130 (131)	
	Indeno(1,2,3-cd)pyrene	50	42.2	ug/L	84				70 (72)	130 (105)	
	Dibenz(a,h)anthracene	50	41.9	ug/L	84				70 (78)	130 (115)	
	Benzo(g,h,i)perylene	50	41.8	ug/L	84				70 (75)	130 (118)	
	1,2,4,5-Tetrachlorobenzene	50	39.5	ug/L	79				70 (72)	130 (101)	
	1,4-Dioxane	50	44.4	ug/L	89				20 (38)	160 (125)	
	2,3,4,6-Tetrachlorophenol	50	40.9	ug/L	82				70 (63)	130 (116)	